

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3370

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170192BL	WATER			Batch Number:	PB170192		Prep Date:	10/21/2025	
	Chromium	1.06	<2.5	U	5.00	P	10/21/2025	17:28	LB137613
	Copper	2.30	<5	U	10.0	P	10/21/2025	17:28	LB137613
	Nickel	1.53	<10	U	20.0	P	10/21/2025	17:28	LB137613
	Zinc	8.33	<10	U	20.0	P	10/21/2025	17:28	LB137613