

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3370

Contract: VERI01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	43.7	52.0	84	42	62	10/21/2025	12:59	LB137613
	Copper	17.8	2.0	890	-18	22	10/21/2025	12:59	LB137613
	Nickel	3.80	2.0	190	-38	42	10/21/2025	12:59	LB137613
	Zinc	3.44			-40	40	10/21/2025	12:59	LB137613
ICSAB01	Chromium	536	542	99	460	624	10/21/2025	13:03	LB137613
	Copper	479	511	94	434	588	10/21/2025	13:03	LB137613
	Nickel	973	954	102	810	1100	10/21/2025	13:03	LB137613
	Zinc	1020	952	107	809	1095	10/21/2025	13:03	LB137613