

**Metals**

-9 -

**ICP SERIAL DILUTIONS**

**SAMPLE NO.**

WATER-TREATMENT DISCHARGEL

**Lab Name:** Alliance

**Contract:** VERI01

**Lab Code:** ACE **Lb No.:** lb137613 **Lab Sample ID :** Q3370-01L **SDG No.:** Q3370

**Matrix (soil/water):** Water **Level (low/med):** LOW

**Concentration Units:** ug/L

| Analyte  | Initial Sample<br>Result (I) |   | Serial Dilution<br>Result (S) |   | % Differ-<br>ence |  | Q | M |
|----------|------------------------------|---|-------------------------------|---|-------------------|--|---|---|
|          | C                            |   | C                             |   |                   |  |   |   |
| Chromium | 15.2                         |   | 14.7                          | J | 3                 |  |   | P |
| Copper   | 17.7                         |   | 18.7                          | J | 6                 |  |   | P |
| Nickel   | 16.9                         | J | 17.0                          | J | 0                 |  |   | P |
| Zinc     | 661                          |   | 669                           |   | 1                 |  |   | P |