

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: BN102025

SequenceID : BN102025

|                                                                                                                                                                                                                | NA    | NO          | YES         |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------|-------------|-------------|
| 1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)                                                                                                                               | _____ | _____       | _____✓_____ |
| 2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria<br>(NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)                                                                            | _____ | _____       | _____✓_____ |
| 3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series                                                                                                               | _____ | _____       | _____✓_____ |
| 4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series | _____ | _____       | _____✓_____ |
| 5. GC/MS Calibration Met:                                                                                                                                                                                      | _____ | _____       | _____✓_____ |
| a. Initial calibration Meet Criteria<br>If not met, list those compounds and their recoveries which fall outside the acceptable range.                                                                         | _____ | _____       | _____✓_____ |
| b. Continuous Calibration(CCC) Meet Criteria<br>If not met, list those compounds and their recoveries which fall outside the acceptable range.                                                                 | _____ | _____✓_____ | _____       |
| 6. Blank Contamination - If yes, list compounds and concentrations in each blank:                                                                                                                              | _____ | _____       | _____✓_____ |
| a. B/N Fraction                                                                                                                                                                                                |       |             |             |
| d. Acid Fraction                                                                                                                                                                                               |       |             |             |

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

a. B/N Fraction

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

a. B/N Fraction

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

ADDITIONAL COMMENTS:

Surrogates recoveries in the samples Q3371-03 & Q3371-06 were affected due to matrix interference, as evidenced by the chromatogram. As no additional volume of each sample was available for re-extraction, they could not be re-extracted and as the re-analysis would have given the same results, they were not reanalyzed. This result is final.

dhruv

Analyst

10/20/2025

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: BN102025

SequenceID : BN102025 NA NO YES