

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN092325

Review By	Rahul	Review On	9/24/2025 8:51:58 AM
Supervise By	Jagrut	Supervise On	9/24/2025 9:44:29 AM
SubDirectory	BN092325	HP Acquire Method	BNA_N
		HP Processing Method	bn092325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6842, SP6843, SP6844, SP6845, SP6846, SP6847, SP6848
CCC	SP6846
Internal Standard/PEM	SP6830, 1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037860.D	23 Sep 2025 11:33		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN037861.D	23 Sep 2025 12:13		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN037862.D	23 Sep 2025 12:49		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN037863.D	23 Sep 2025 13:25	Compound #20 kept on QR	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN037864.D	23 Sep 2025 14:02		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN037865.D	23 Sep 2025 14:38		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN037866.D	23 Sep 2025 15:15		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN037867.D	23 Sep 2025 15:51	Compound#14,24 removed from 5.0 ppm	RC/JU	Ok
9	SSTDICV0.4	ICVBN092325	BN037868.D	23 Sep 2025 16:28		RC/JU	Ok
10	PB169743BL	PB169743BL	BN037869.D	23 Sep 2025 17:42	END CCC missing, Analyzed for Contamination check.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN102025

Review By	rahul	Review On	10/21/2025 3:22:53 PM
Supervise By	Jagrut	Supervise On	10/21/2025 3:39:26 PM
SubDirectory	BN102025	HP Acquire Method	BNA_N, 8270_HP Processing Method BN092325

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848
CCC	SP6846
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038100.D	20 Oct 2025 16:41		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038101.D	20 Oct 2025 17:20		RC/JU	Ok
3	PB170167BL	PB170167BL	BN038102.D	20 Oct 2025 17:56		RC/JU	Ok
4	PB170167BS	PB170167BS	BN038103.D	20 Oct 2025 18:33		RC/JU	Ok
5	PB170167BSD	PB170167BSD	BN038104.D	20 Oct 2025 19:09	Double internal standard added	RC/JU	Ok
6	Q3371-02	BP-VPB-184-EB-20251	BN038105.D	20 Oct 2025 19:45		RC/JU	Ok
7	Q3371-03	VPB184-HYD-2025101	BN038106.D	20 Oct 2025 20:21	Surrogate fail.	RC/JU	Ok
8	Q3371-05	BP-VPB-184-GW-920-9	BN038107.D	20 Oct 2025 20:58		RC/JU	Ok
9	Q3371-06	VPB184-HYD-2025101	BN038108.D	20 Oct 2025 21:34	Surrogate fail.	RC/JU	Ok
10	SSTDCCC0.4	SSTDCCC0.4EC	BN038109.D	20 Oct 2025 22:10		RC/JU	Ok

M : Manual Integration