

Surrogate Summary

SW-846

SDG No.: Q3371

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170167BL	PB170167BL	2-Methylnaphthalene-d10	0.4	0.34	85		30	150
		Fluoranthene-d10	0.4	0.33	83		30	150
		Nitrobenzene-d5	0.4	0.36	90		55	111
		2-Fluorobiphenyl	0.4	0.39	98		53	106
		Terphenyl-d14	0.4	0.45	113		58	132
PB170167BS	PB170167BS	2-Methylnaphthalene-d10	0.4	0.40	99		30	150
		Fluoranthene-d10	0.4	0.34	85		30	150
		Nitrobenzene-d5	0.4	0.42	104		55	111
		2-Fluorobiphenyl	0.4	0.44	111	*	53	106
		Terphenyl-d14	0.4	0.51	127		58	132
PB170167BSD	PB170167BSD	2-Methylnaphthalene-d10	0.4	0.36	90		30	150
		Fluoranthene-d10	0.4	0.31	78		30	150
		Nitrobenzene-d5	0.4	0.38	95		55	111
		2-Fluorobiphenyl	0.4	0.42	105		53	106
		Terphenyl-d14	0.4	0.47	117		58	132
Q3371-02	BP-VPB-184-EB-20251016	2-Methylnaphthalene-d10	0.4	0.30	76		30	150
		Fluoranthene-d10	0.4	0.42	105		30	150
		Nitrobenzene-d5	0.4	0.36	91		55	111
		2-Fluorobiphenyl	0.4	0.35	87		53	106
		Terphenyl-d14	0.4	0.55	137	*	58	132
Q3371-03	VPB184-HYD-20251010	2-Methylnaphthalene-d10	0.4	0.016	4	*	30	150
		Fluoranthene-d10	0.4	0.042	10	*	30	150
		Nitrobenzene-d5	0.4	0.34	86		55	111
		2-Fluorobiphenyl	0.4	0.34	84		53	106
		Terphenyl-d14	0.4	0.55	137	*	58	132
Q3371-05	BP-VPB-184-GW-920-922	2-Methylnaphthalene-d10	0.4	0.29	73		30	150
		Fluoranthene-d10	0.4	0.41	101		30	150
		Nitrobenzene-d5	0.4	0.35	88		55	111
		2-Fluorobiphenyl	0.4	0.33	83		53	106
		Terphenyl-d14	0.4	0.55	137	*	58	132
Q3371-06	VPB184-HYD-20251016	2-Methylnaphthalene-d10	0.4	0.014	4	*	30	150
		Fluoranthene-d10	0.4	0.043	11	*	30	150
		Nitrobenzene-d5	0.4	0.31	77		55	111
		2-Fluorobiphenyl	0.4	0.32	79		53	106
		Terphenyl-d14	0.4	0.52	131		58	132