

Surrogate Summary

SDG No.: **Q3371**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3371-01	BP-VPB-184-TB-20251010	1,2-Dichloroethane-d4	50	52.4	105		81	118
		Dibromofluoromethane	50	46.8	94		80	119
		Toluene-d8	50	45.6	91		89	112
		4-Bromofluorobenzene	50	53.0	106		85	114
Q3371-02	BP-VPB-184-EB-20251016	1,2-Dichloroethane-d4	50	54.7	109		81	118
		Dibromofluoromethane	50	49.5	99		80	119
		Toluene-d8	50	47.2	94		89	112
		4-Bromofluorobenzene	50	55.2	110		85	114
Q3371-03	VPB184-HYD-20251010	1,2-Dichloroethane-d4	50	48.4	97		81	118
		Dibromofluoromethane	50	46.4	93		80	119
		Toluene-d8	50	44.9	90		89	112
		4-Bromofluorobenzene	50	50.7	101		85	114
Q3371-04	BP-VPB-184-GW-915-917	1,2-Dichloroethane-d4	50	51.0	102		81	118
		Dibromofluoromethane	50	48.4	97		80	119
		Toluene-d8	50	47.1	94		89	112
		4-Bromofluorobenzene	50	45.0	90		85	114
Q3371-05	BP-VPB-184-GW-920-922	1,2-Dichloroethane-d4	50	55.9	112		81	118
		Dibromofluoromethane	50	53.2	106		80	119
		Toluene-d8	50	51.4	103		89	112
		4-Bromofluorobenzene	50	48.7	97		85	114
Q3371-06	VPB184-HYD-20251016	1,2-Dichloroethane-d4	50	54.8	110		81	118
		Dibromofluoromethane	50	47.2	94		80	119
		Toluene-d8	50	46.8	94		89	112
		4-Bromofluorobenzene	50	55.5	111		85	114
VX1020WBL01	VX1020WBL01	1,2-Dichloroethane-d4	50	56.9	114		81	118
		Dibromofluoromethane	50	53.0	106		80	119
		Toluene-d8	50	51.2	102		89	112
		4-Bromofluorobenzene	50	52.1	104		85	114
VX1020WBS01	VX1020WBS01	1,2-Dichloroethane-d4	50	51.7	103		81	118
		Dibromofluoromethane	50	51.2	102		80	119
		Toluene-d8	50	51.9	104		89	112
		4-Bromofluorobenzene	50	51.1	102		85	114
VX1020WBSD01	VX1020WBSD01	1,2-Dichloroethane-d4	50	57.5	115		81	118
		Dibromofluoromethane	50	53.2	106		80	119
		Toluene-d8	50	51.9	104		89	112
		4-Bromofluorobenzene	50	52.1	104		85	114
VX1021WBL01	VX1021WBL01	1,2-Dichloroethane-d4	50	52.5	105		81	118
		Dibromofluoromethane	50	47.1	94		80	119
		Toluene-d8	50	46.4	93		89	112
		4-Bromofluorobenzene	50	52.6	105		85	114
VX1021WBS01	VX1021WBS01	1,2-Dichloroethane-d4	50	49.2	98		81	118
		Dibromofluoromethane	50	49.3	99		80	119
		Toluene-d8	50	47.8	96		89	112
		4-Bromofluorobenzene	50	49.7	99		85	114