

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG NO.: Q3371
Lab File ID: VX048232.D Date Analyzed: 10/20/2025
Instrument ID: MSVOA_X Time Analyzed: 12:01
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	124491	5.54	201906	6.75	184494	10.04
UPPER LIMIT	248982	6.04	403812	7.25	368988	10.54
LOWER LIMIT	62245.5	5.04	100953	6.25	92247	9.54
EPA SAMPLE NO.						
BP-VPB-184-GW-915-917	107423	5.54	183824	6.75	169870	10.04
BP-VPB-184-GW-920-922	122446	5.54	210478	6.75	193818	10.04
VX1020WBL01	127705	5.53	221220	6.74	203730	10.04
VX1020WBS01	134660	5.54	221833	6.74	200515	10.04
VX1020WBSD01	114781	5.54	195841	6.75	183563	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.