

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG NO.: Q3371
 Lab File ID: VX048232.D Date Analyzed: 10/20/2025
 Instrument ID: MSVOA_X Time Analyzed: 12:01
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	92693	12.01				
UPPER LIMIT	185386	12.51				
LOWER LIMIT	46346.5	11.51				
EPA SAMPLE NO.						
BP-VPB-184-GW-915-917	81048	12.01				
BP-VPB-184-GW-920-922	92863	12.01				
VX1020WBL01	106522	12.01				
VX1020WBS01	101575	12.01				
VX1020WBSD01	92655	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.