

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG No.: Q3373
 Instrument ID: MSVOA_X Calibration Date/Time: 10/22/2025 08:37
 Lab File ID: VX048280.D Init. Calib. Date(s): 10/20/2025 10/20/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 10:59 13:49
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
2-Chloroethyl Vinyl ether	250.000	255.31	-2.124402	20
Hexachlorobutadiene	50.000	54.03	-8.068001	20