284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

DATA PACKAGE

GC SEMI-VOLATILES

PROJECT NAME : AE1-CTY 3.2.Z-PR-DOCK-DES-ODC - 60693795-1719206

AECOM

605 3rd Avenue

29th Floor

New York, NY - 10158

Phone No: 212-973-2900

ORDER ID : Q3375

ATTENTION : Rob Forstner

Laboratory Certification ID # 20012

Table Of Contents for Q3375

1) PCB Data 2

2) Signature Page 4

3) Case Narrative 5

4) Qualifier Page 7

5) Conformance/Non Conformance 8

6) QA Checklist 10

7) Chronicle 11

8) Hit Summary 12

9) QC Data Summary For PCB 13

- 9.1) Deuterated Monitoring Compound Summary 14
- 9.2) MS/MSD Summary 16
- 9.3) LCS/LCSD Summary 18
- 9.4) Method Blank Summary 19

10) Sample Data 20

- 10.1) PR132-S02-074099-20251016 21
- 10.2) PR132-S04-000013-20251016 25
- 10.3) PR132-S05-000057-20251016 29
- 10.4) PR132-S02-010074-20251016 38
- 10.5) PR132-DP02-20251016 42

11) Calibration Data Summary 51

- 11.1) Initial Calibration Data 52
  - 11.1.1) PP092325 52
  - 11.1.2) PQ101025 271
- 11.2) Continued Calibration Data 490
  - 11.2.1) PP075868.D 490
  - 11.2.2) PP075881.D 502
  - 11.2.3) PQ071008.D 514
  - 11.2.4) PQ071023.D 526
- 11.3) Analytical Seq 538

12) Compound Detection Summary 542

13) QC Sample Data 547

- 13.1) Method Blank Data 548
- 13.2) PIBLK Data 552
- 13.3) LCS Data 576
- 13.4) MS Data 585

13.5) MSD Data 594
14) Manual Integration 603
15) Analytical Runlogs 615
16) Percent Solid 629
17) Extraction Logs 634
17.1) PB170162.pdf 634
17.2) PB170162IC.pdf 636
18) Standard Prep Logs 637
19) Shipping Document 729
19.1) Chain Of Custody 730
19.2) Lab Certificate 731
19.3) Internal COC 732

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : Q3375

Project ID : AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

Client : AECOM

Lab Sample Number	Client Sample Number
Q3375-01	PR132-S02-074099-20251016
Q3375-02	PR132-S04-000013-20251016
Q3375-03	PR132-S05-000057-20251016
Q3375-04	PR132-S02-010074-20251016
Q3375-05	PR132-DP02-20251016

I certify that the data package is in compliance with the terms and conditions of the contract,both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : APPROVED
By Nimisha Pandya, QA/QC Supervisor at 3:41 pm, Oct 29, 2025

Date: 10/27/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

AECOM
Project Name: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
Project # N/A
Order ID # Q3375
Test Name: PCB

A. Number of Samples and Date of Receipt:
5 Solid samples were received on 10/16/2025.

B. Parameters
According to the Chain of Custody document, the following analyses were requested: PCB. This data package contains results for PCB.

C. Analytical Techniques:
The analyses were performed on instrument GCECD_Q.The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11.The analyses were performed on instrument GCECD_P. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11.The analysis of PCBs was based on method 8082A and extraction was done based on method 3541.

D. QA/ QC Samples:
The Holding Times were met for all analysis.
The Surrogate recoveries were met for all analysis.
The Retention Times were met for all analysis.
The MS recoveries met the requirements for all compounds.
The MSD recoveries met the requirements for all compounds.
The RPD were met for all analysis.
The Blank Spike met requirements for all compounds.
The Blank analysis did not indicate the presence of lab contamination.
The Initial Calibration met the requirements.

The Continuous Calibration File ID PP075881.D met the requirements except for Decachlorobiphenyl is failing in 1st column AND Aroclor-1016(Peak-02) is failing in 2nd column while another column passing for that therefore no corrective action taken.

E. Additional Comments:
The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature APPROVED
By Nimisha Pandya, QA/QC Supervisor at 3:41 pm, Oct 29, 2025

284 Sheffield Street, Mountainside NJ 07092 Tel.
908-789-8900 Fax: 908-789-8922

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following " Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3375 MATRIX: Solid

METHOD: 8082A/3541

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified.			✓
2.	Standard Summary Submitted.			✓
3.	Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD. The Initial Calibration met the requirements. The Continuous Calibration File ID PP075881.D met the requirements except for Decachlorobiphenyl is failing in 1st column AND Aroclor-1016(Peak-02) is failing in 2nd column while another column passing for that therefore no corrective action taken.			✓
4.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
6.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The MS recoveries met the requirements for all compounds. The MSD recoveries met the requirements for all compounds. The Blank Spike met requirements for all compounds. The RPD were met for all analysis.			✓
7.	Retention Time Shift Meet Criteria (if applicable) Comments:			✓
8.	Extraction Holding Time Met If not met, list number of days exceeded for each sample:			✓
9.	Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.			✓

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

The soil samples results are based on a dry weight basis.

QA REVIEW

REVIEWED

By Sohil Jodhani, QA/QC Director at 2:07 pm, Oct 29, 2025

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3375

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)	✓
Check chain-of-custody for proper relinquish/return of samples	✓
Is the chain of custody signed and complete	✓
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	✓
Collect information for each project id from server. Were all requirements followed	✓
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	✓
Do lab numbers and client Ids on cover page agree with the Chain of Custody	✓
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	✓
Do requested analyses on Chain of Custody agree with the log-in page	✓
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	✓
Were the samples received within hold time	✓
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	✓
ANALYTICAL:	
Was method requirement followed?	✓
Was client requirement followed?	✓
Does the case narrative summarize all QC failure?	✓
All runlogs and manual integration are reviewed for requirements	✓
All manual calculations and /or hand notations verified	✓

QA Review Signature: SOHIL JODHANI **Date:** 10/27/2025

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q3375	**OrderDate:**	10/17/2025 10:32:00 AM
Client:	AECOM	**Project:**	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
Contact:	Rob Forstner	**Location:**	D31,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3375-01	**PR132-S02-074099-20251016**	**SOIL**			**10/16/25**			**10/16/25**
			PCB	8082A		10/20/25	10/20/25	
			Pesticide-TCL	8081B		10/20/25	10/20/25	
Q3375-02	**PR132-S04-000013-20251016**	**SOIL**			**10/16/25**			**10/16/25**
			PCB	8082A		10/20/25	10/20/25	
			Pesticide-TCL	8081B		10/20/25	10/20/25	
Q3375-03	**PR132-S05-000057-20251016**	**SOIL**			**10/16/25**			**10/16/25**
			PCB	8082A		10/20/25	10/20/25	
			Pesticide-TCL	8081B		10/20/25	10/20/25	
Q3375-04	**PR132-S02-010074-20251016**	**SOIL**			**10/16/25**			**10/16/25**
			PCB	8082A		10/20/25	10/20/25	
			Pesticide-TCL	8081B		10/20/25	10/20/25	
Q3375-05	**PR132-DP02-20251016**	**SOIL**			**10/16/25**			**10/16/25**
			PCB	8082A		10/20/25	10/20/25	
			Pesticide-TCL	8081B		10/20/25	10/20/25	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3375 **Order ID:** Q3375

Client: AECOM **Project ID:** AE1-CTY 3.2.Z-PR-DOCK-Des-ODC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : PR132-S05-000057-20251016								
Q3375-03	PR132-S05-000057-2(	SOIL	Aroclor-1242	51.1		6.00	25.4	ug/kg
Q3375-03	PR132-S05-000057-2(	SOIL	Aroclor-1254	21.2	J	4.80	25.4	ug/kg
			Total Concentration:	**72.300**				
Client ID : PR132-DP02-20251016								
Q3375-05	PR132-DP02-2025101	SOIL	Aroclor-1242	61.9		6.00	25.4	ug/kg
Q3375-05	PR132-DP02-2025101	SOIL	Aroclor-1254	25.3	J	4.80	25.4	ug/kg
			Total Concentration:	**87.200**				

QC SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q3375

Client: AECOM

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%) Low	Limits(%) High
I.BLK-PP075246.D	PIBLK-PP075246.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	19.4	97		60	140
		Tetrachloro-m-xyl	2	20	17.9	90		60	140
		Decachlorobiphen	2	20	20.0	100		60	140
I.BLK-PP075872.D	PIBLK-PP075872.D	Tetrachloro-m-xyl	1	20	23.1	115		60	140
		Decachlorobiphen	1	20	20.6	103		60	140
		Tetrachloro-m-xyl	2	20	22.7	113		60	140
		Decachlorobiphen	2	20	22.0	110		60	140
Q3383-01MS	OR-03-101725MS	Tetrachloro-m-xyl	1	20	28.0	140		32	144
		Decachlorobiphen	1	20	22.4	112		32	175
		Tetrachloro-m-xyl	2	20	26.4	132		32	144
		Decachlorobiphen	2	20	22.4	112		32	175
Q3383-01MSD	OR-03-101725MSD	Tetrachloro-m-xyl	1	20	28.4	142		32	144
		Decachlorobiphen	1	20	21.8	109		32	175
		Tetrachloro-m-xyl	2	20	18.9	95		32	144
		Decachlorobiphen	2	20	21.3	107		32	175
I.BLK-PP075885.D	PIBLK-PP075885.D	Tetrachloro-m-xyl	1	20	22.7	114		60	140
		Decachlorobiphen	1	20	20.0	100		60	140
		Tetrachloro-m-xyl	2	20	23.5	117		60	140
		Decachlorobiphen	2	20	20.9	104		60	140
I.BLK-PQ070972.D	PIBLK-PQ070972.D	Tetrachloro-m-xyl	1	20	18.3	91		60	140
		Decachlorobiphen	1	20	19.0	95		60	140
		Tetrachloro-m-xyl	2	20	19.1	96		60	140
		Decachlorobiphen	2	20	19.4	97		60	140
I.BLK-PQ071012.D	PIBLK-PQ071012.D	Tetrachloro-m-xyl	1	20	18.7	94		60	140
		Decachlorobiphen	1	20	22.1	111		60	140
		Tetrachloro-m-xyl	2	20	18.9	95		60	140
		Decachlorobiphen	2	20	22.0	110		60	140
PB170162BL	PB170162BL	Tetrachloro-m-xyl	1	20	21.7	108		32	144
		Decachlorobiphen	1	20	25.1	126		32	175
		Tetrachloro-m-xyl	2	20	21.6	108		32	144
		Decachlorobiphen	2	20	24.8	124		32	175
PB170162BS	PB170162BS	Tetrachloro-m-xyl	1	20	24.0	120		32	144
		Decachlorobiphen	1	20	27.3	137		32	175
		Tetrachloro-m-xyl	2	20	22.5	113		32	144
		Decachlorobiphen	2	20	27.3	137		32	175
Q3375-01	PR132-S02-074099-20251016	Tetrachloro-m-xyl	1	20	20.3	102		32	144
		Decachlorobiphen	1	20	15.7	78		32	175
		Tetrachloro-m-xyl	2	20	21.0	105		32	144
		Decachlorobiphen	2	20	15.1	75		32	175
Q3375-02	PR132-S04-000013-20251016	Tetrachloro-m-xyl	1	20	22.8	114		32	144

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q3375

Client: AECOM

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%) Low	Limits(%) High
Q3375-02	PR132-S04-000013-20251016	Decachlorobiphen	1	20	16.4	82		32	175
		Tetrachloro-m-xyl	2	20	23.8	119		32	144
		Decachlorobiphen	2	20	16.5	83		32	175
Q3375-03	PR132-S05-000057-20251016	Tetrachloro-m-xyl	1	20	18.0	90		32	144
		Decachlorobiphen	1	20	17.0	85		32	175
		Tetrachloro-m-xyl	2	20	18.5	92		32	144
		Decachlorobiphen	2	20	16.8	84		32	175
Q3375-04	PR132-S02-010074-20251016	Tetrachloro-m-xyl	1	20	21.2	106		32	144
		Decachlorobiphen	1	20	18.6	93		32	175
		Tetrachloro-m-xyl	2	20	21.5	108		32	144
		Decachlorobiphen	2	20	18.3	92		32	175
Q3375-05	PR132-DP02-20251016	Tetrachloro-m-xyl	1	20	20.1	100		32	144
		Decachlorobiphen	1	20	18.3	92		32	175
		Tetrachloro-m-xyl	2	20	19.8	99		32	144
		Decachlorobiphen	2	20	17.8	89		32	175
I.BLK-PQ071027.D	PIBLK-PQ071027.D	Tetrachloro-m-xyl	1	20	19.6	98		60	140
		Decachlorobiphen	1	20	22.6	113		60	140
		Tetrachloro-m-xyl	2	20	19.7	98		60	140
		Decachlorobiphen	2	20	22.3	111		60	140

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3375 **Analytical Method:** 8082A

Client: AECOM **DataFile :** PP075879.D

	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Lab Sample ID:	**Q3383-01MS (Column 1)**	**Client Sample ID:**		**OR-03-101725MS**								
	AR1016	180.7	0	189	ug/kg	105				55	146	
	AR1260	180.7	0	184	ug/kg	102				54	119	
Lab Sample ID:	**Q3383-01MS (Column 2)**	**Client Sample ID:**		**OR-03-101725MS**								
	AR1016	180.7	0	260	ug/kg	144				55	146	
	AR1260	180.7	0	173	ug/kg	96				54	119	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3375 **Analytical Method:** 8082A

Client: AECOM **DataFile :** PP075880.D

	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Lab Sample ID:	**Q3383-01MSD (Column 1)**	**Client Sample ID:**		**OR-03-101725MSD**								
	AR1016	180.7	0	187	ug/kg	103		2		55	146	15
	AR1260	180.7	0	177	ug/kg	98		4		54	119	15
Lab Sample ID:	**Q3383-01MSD (Column 2)**	**Client Sample ID:**		**OR-03-101725MSD**								
	AR1016	180.7	0	244	ug/kg	135		6		55	146	15
	AR1260	180.7	0	167	ug/kg	92		4		54	119	15

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3375 **Analytical Method:** 8082A

Client: AECOM **Datafile :** PQ071014.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB170162BS (Column 1)	AR1016	166.5	172	ug/kg	103				71	120	
	AR1260	166.5	165	ug/kg	99				65	130	
PB170162BS (Column 2)	AR1016	166.5	153	ug/kg	92				71	120	
	AR1260	166.5	151	ug/kg	91				65	130	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID: PB170162BL

Lab Name: Alliance | Contract: AECO02

Lab Code: ACE | SDG NO.: Q3375

Lab Sample ID: PB170162BL | Lab File ID: PQ071013.D

Matrix: (soil/water) Solid | Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N | Date Extracted: 10/20/2025

Date Analyzed (1): 10/20/2025 | Date Analyzed (2): 10/20/2025

Time Analyzed (1): 13:20 | Time Analyzed (2): 13:20

Instrument ID (1): ECD_Q | Instrument ID (2): ECD_Q

GC Column (1): ZB-MR1 ID: 0.32 (mm) | GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
OR-03-101725MS	Q3383-01MS	PP075879.D	10/20/2025	10/20/2025
OR-03-101725MSD	Q3383-01MSD	PP075880.D	10/20/2025	10/20/2025
PB170162BS	PB170162BS	PQ071014.D	10/20/2025	10/20/2025
PR132-S02-074099-20251016	Q3375-01	PQ071017.D	10/20/2025	10/20/2025
PR132-S04-000013-20251016	Q3375-02	PQ071018.D	10/20/2025	10/20/2025
PR132-S05-000057-20251016	Q3375-03	PQ071019.D	10/20/2025	10/20/2025
PR132-S02-010074-20251016	Q3375-04	PQ071020.D	10/20/2025	10/20/2025
PR132-DP02-20251016	Q3375-05	PQ071021.D	10/20/2025	10/20/2025

COMMENTS:

SAMPLE DATA

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/16/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/16/25
Client Sample ID:	PR132-S02-074099-20251016	SDG No.:	Q3375
Lab Sample ID:	Q3375-01	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	69.2
Sample Wt/Vol:	30.02 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date 10/20/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	24.6	U	1	5.70	24.6	ug/kg	10/20/25 14:19	PB170162
11104-28-2	Aroclor-1221	24.6	U	1	5.80	24.6	ug/kg	10/20/25 14:19	PB170162
11141-16-5	Aroclor-1232	24.6	U	1	5.40	24.6	ug/kg	10/20/25 14:19	PB170162
53469-21-9	Aroclor-1242	24.6	U	1	5.80	24.6	ug/kg	10/20/25 14:19	PB170162
12672-29-6	Aroclor-1248	24.6	U	1	8.50	24.6	ug/kg	10/20/25 14:19	PB170162
11097-69-1	Aroclor-1254	24.6	U	1	4.60	24.6	ug/kg	10/20/25 14:19	PB170162
37324-23-5	Aroclor-1262	24.6	U	1	7.20	24.6	ug/kg	10/20/25 14:19	PB170162
11100-14-4	Aroclor-1268	24.6	U	1	5.20	24.6	ug/kg	10/20/25 14:19	PB170162
11096-82-5	Aroclor-1260	24.6	U	1	4.70	24.6	ug/kg	10/20/25 14:19	PB170162
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.0			32 - 144	105%	SPK: 20		
2051-24-3	Decachlorobiphenyl	15.7			32 - 175	78%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071017.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 14:19
Operator : YP\AJ
Sample : Q3375-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PR132-S02-074099-20251016

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:25:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.777	88400528	106.5E6	20.321	20.960
2)	SA Decachlor...	8.500	7.533	45528453	76091351	15.678	15.069
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071017.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 14:19
Operator : YP\AJ
Sample : Q3375-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PR132-S02-074099-20251016

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:25:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 88400528
Conc: 20.32 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-S02-074099-20251016



#1 Tetrachloro-m-xylene

R.T.: 2.777 min
Delta R.T.: -0.002 min
Response: 106478035
Conc: 20.96 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.500 min
Delta R.T.: 0.004 min
Response: 45528453
Conc: 15.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 76091351
Conc: 15.07 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/16/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/16/25
Client Sample ID:	PR132-S04-000013-20251016	SDG No.:	Q3375
Lab Sample ID:	Q3375-02	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	74.7
Sample Wt/Vol:	30.05 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date 10/20/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	22.7	U	1	5.30	22.7	ug/kg	10/20/25 14:33	PB170162
11104-28-2	Aroclor-1221	22.7	U	1	5.40	22.7	ug/kg	10/20/25 14:33	PB170162
11141-16-5	Aroclor-1232	22.7	U	1	5.00	22.7	ug/kg	10/20/25 14:33	PB170162
53469-21-9	Aroclor-1242	22.7	U	1	5.40	22.7	ug/kg	10/20/25 14:33	PB170162
12672-29-6	Aroclor-1248	22.7	U	1	7.90	22.7	ug/kg	10/20/25 14:33	PB170162
11097-69-1	Aroclor-1254	22.7	U	1	4.30	22.7	ug/kg	10/20/25 14:33	PB170162
37324-23-5	Aroclor-1262	22.7	U	1	6.70	22.7	ug/kg	10/20/25 14:33	PB170162
11100-14-4	Aroclor-1268	22.7	U	1	4.80	22.7	ug/kg	10/20/25 14:33	PB170162
11096-82-5	Aroclor-1260	22.7	U	1	4.30	22.7	ug/kg	10/20/25 14:33	PB170162
SURROGATES									
877-09-8	Tetrachloro-m-xylene	23.8			32 - 144	119%	SPK: 20		
2051-24-3	Decachlorobiphenyl	16.5			32 - 175	83%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071018.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 14:33
Operator : YP\AJ
Sample : Q3375-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PR132-S04-000013-20251016

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:25:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.399	2.777	99017270	120.7E6	22.761	23.755
2)	SA Decachlor...	8.499	7.533	47733774	83420956	16.437	16.520
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071018.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 14:33
Operator : YP\AJ
Sample : Q3375-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PR132-S04-000013-20251016

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:25:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.399 min
Delta R.T.: 0.000 min
Response: 99017270
Conc: 22.76 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-S04-000013-20251016



#1 Tetrachloro-m-xylene

R.T.: 2.777 min
Delta R.T.: -0.001 min
Response: 120673739
Conc: 23.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.499 min
Delta R.T.: 0.003 min
Response: 47733774
Conc: 16.44 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 83420956
Conc: 16.52 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/16/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/16/25
Client Sample ID:	PR132-S05-000057-20251016	SDG No.:	Q3375
Lab Sample ID:	Q3375-03	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	66.7
Sample Wt/Vol:	30.08 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date 10/20/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	25.4	U	1	5.90	25.4	ug/kg	10/20/25 14:48	PB170162
11104-28-2	Aroclor-1221	25.4	U	1	6.00	25.4	ug/kg	10/20/25 14:48	PB170162
11141-16-5	Aroclor-1232	25.4	U	1	5.60	25.4	ug/kg	10/20/25 14:48	PB170162
53469-21-9	Aroclor-1242	51.1		1	6.00	25.4	ug/kg	10/20/25 14:48	PB170162
12672-29-6	Aroclor-1248	25.4	U	1	8.90	25.4	ug/kg	10/20/25 14:48	PB170162
11097-69-1	Aroclor-1254	21.2	J	1	4.80	25.4	ug/kg	10/20/25 14:48	PB170162
37324-23-5	Aroclor-1262	25.4	U	1	7.50	25.4	ug/kg	10/20/25 14:48	PB170162
11100-14-4	Aroclor-1268	25.4	U	1	5.40	25.4	ug/kg	10/20/25 14:48	PB170162
11096-82-5	Aroclor-1260	25.4	U	1	4.80	25.4	ug/kg	10/20/25 14:48	PB170162
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.5			32 - 144	92%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.0			32 - 175	85%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071019.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 14:48
Operator : YP\AJ
Sample : Q3375-03
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PR132-S05-000057-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:25:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.777	78128584	93906837	17.959	18.486
2)	SA Decachlor...	8.499	7.533	49400540	84757235	17.011	16.785
	Target Compounds						
16)	L4 AR-1242-1	4.482	3.777	8102339	12157215	68.114	74.647
17)	L4 AR-1242-2	4.500	3.792	19066668	20864710	99.976	86.441
18)	L4 AR-1242-3	4.557	3.952	6253036	11227596	52.592	85.173 #
19)	L4 AR-1242-4	4.646	4.038	10203133	15072087	102.487	131.554 #
20)	L4 AR-1242-5	5.353	4.533	10279706	23520425	130.583	135.066
26)	L6 AR-1254-1	5.292	4.532	10870024	21509404	55.381	63.786m
27)	L6 AR-1254-2	5.504	4.678	14022566	13232641	48.273	44.531m
28)	L6 AR-1254-3	5.856	5.059	12859904	15885852	41.861	34.012m
29)	L6 AR-1254-4	6.136	5.286	6463114	10359398	29.439	35.087
30)	L6 AR-1254-5	6.545	5.689	7788540	14652708	33.586	34.858

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071019.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 14:48
Operator : YP\AJ
Sample : Q3375-03
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PR132-S05-000057-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:25:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 78128584
Conc: 17.96 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-S05-000057-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025



#1 Tetrachloro-m-xylene

R.T.: 2.777 min
Delta R.T.: -0.002 min
Response: 93906837
Conc: 18.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.499 min
Delta R.T.: 0.003 min
Response: 49400540
Conc: 17.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 84757235
Conc: 16.79 ng/ml



#16 AR-1242-1

R.T.: 4.482 min
Delta R.T.: 0.002 min
Response: 8102339
Conc: 68.11 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-S05-000057-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025



#16 AR-1242-1

R.T.: 3.777 min
Delta R.T.: -0.001 min
Response: 12157215
Conc: 74.65 ng/ml



#17 AR-1242-2

R.T.: 4.500 min
Delta R.T.: 0.001 min
Response: 19066668
Conc: 99.98 ng/ml



#17 AR-1242-2

R.T.: 3.792 min
Delta R.T.: -0.001 min
Response: 20864710
Conc: 86.44 ng/ml



#18 AR-1242-3

R.T.: 4.557 min
Delta R.T.: 0.001 min
Response: 6253036
Conc: 52.59 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-S05-000057-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025



#18 AR-1242-3

R.T.: 3.952 min
Delta R.T.: -0.002 min
Response: 11227596
Conc: 85.17 ng/ml



#19 AR-1242-4

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 10203133
Conc: 102.49 ng/ml



#19 AR-1242-4

R.T.: 4.038 min
Delta R.T.: -0.002 min
Response: 15072087
Conc: 131.55 ng/ml



#20 AR-1242-5

R.T.: 5.353 min
Delta R.T.: 0.003 min
Response: 10279706
Conc: 130.58 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-S05-000057-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025



#20 AR-1242-5

R.T.: 4.533 min
Delta R.T.: -0.002 min
Response: 23520425
Conc: 135.07 ng/ml



#26 AR-1254-1

R.T.: 5.292 min
Delta R.T.: 0.000 min
Response: 10870024
Conc: 55.38 ng/ml



#26 AR-1254-1

R.T.: 4.532 min
Delta R.T.: 0.000 min
Response: 21509404
Conc: 63.79 ng/ml m



#27 AR-1254-2

R.T.: 5.504 min
Delta R.T.: -0.004 min
Response: 14022566
Conc: 48.27 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-S05-000057-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025



#27 AR-1254-2

R.T.: 4.678 min
Delta R.T.: -0.002 min
Response: 13232641
Conc: 44.53 ng/ml m



#28 AR-1254-3

R.T.: 5.856 min
Delta R.T.: 0.001 min
Response: 12859904
Conc: 41.86 ng/ml



#28 AR-1254-3

R.T.: 5.059 min
Delta R.T.: 0.000 min
Response: 15885852
Conc: 34.01 ng/ml m



#29 AR-1254-4

R.T.: 6.136 min
Delta R.T.: 0.000 min
Response: 6463114
Conc: 29.44 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-S05-000057-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025



#29 AR-1254-4

R.T.: 5.286 min
Delta R.T.: 0.000 min
Response: 10359398
Conc: 35.09 ng/ml



#30 AR-1254-5

R.T.: 6.545 min
Delta R.T.: 0.001 min
Response: 7788540
Conc: 33.59 ng/ml



#30 AR-1254-5

R.T.: 5.689 min
Delta R.T.: 0.000 min
Response: 14652708
Conc: 34.86 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/16/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/16/25
Client Sample ID:	PR132-S02-010074-20251016	SDG No.:	Q3375
Lab Sample ID:	Q3375-04	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	55.3
Sample Wt/Vol:	30.04 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date 10/20/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	30.7	U	1	7.10	30.7	ug/kg	10/20/25 15:02	PB170162
11104-28-2	Aroclor-1221	30.7	U	1	7.30	30.7	ug/kg	10/20/25 15:02	PB170162
11141-16-5	Aroclor-1232	30.7	U	1	6.70	30.7	ug/kg	10/20/25 15:02	PB170162
53469-21-9	Aroclor-1242	30.7	U	1	7.20	30.7	ug/kg	10/20/25 15:02	PB170162
12672-29-6	Aroclor-1248	30.7	U	1	10.7	30.7	ug/kg	10/20/25 15:02	PB170162
11097-69-1	Aroclor-1254	30.7	U	1	5.80	30.7	ug/kg	10/20/25 15:02	PB170162
37324-23-5	Aroclor-1262	30.7	U	1	9.10	30.7	ug/kg	10/20/25 15:02	PB170162
11100-14-4	Aroclor-1268	30.7	U	1	6.50	30.7	ug/kg	10/20/25 15:02	PB170162
11096-82-5	Aroclor-1260	30.7	U	1	5.80	30.7	ug/kg	10/20/25 15:02	PB170162
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.5			32 - 144	108%	SPK: 20		
2051-24-3	Decachlorobiphenyl	18.6			32 - 175	93%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071020.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 15:02
Operator : YP\AJ
Sample : Q3375-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PR132-S02-010074-20251016

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:25:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.777	92147594	109.3E6	21.182	21.509
2)	SA Decachlor...	8.499	7.534	53870378	92559882	18.551	18.330
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071020.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 15:02
Operator : YP\AJ
Sample : Q3375-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PR132-S02-010074-20251016

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:25:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 92147594
Conc: 21.18 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-S02-010074-20251016



#1 Tetrachloro-m-xylene

R.T.: 2.777 min
Delta R.T.: -0.001 min
Response: 109265144
Conc: 21.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.499 min
Delta R.T.: 0.003 min
Response: 53870378
Conc: 18.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.001 min
Response: 92559882
Conc: 18.33 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/16/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/16/25
Client Sample ID:	PR132-DP02-20251016	SDG No.:	Q3375
Lab Sample ID:	Q3375-05	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	66.7
Sample Wt/Vol:	30.07 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date 10/20/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	25.4	U	1	5.90	25.4	ug/kg	10/20/25 15:47	PB170162
11104-28-2	Aroclor-1221	25.4	U	1	6.00	25.4	ug/kg	10/20/25 15:47	PB170162
11141-16-5	Aroclor-1232	25.4	U	1	5.60	25.4	ug/kg	10/20/25 15:47	PB170162
53469-21-9	Aroclor-1242	61.9		1	6.00	25.4	ug/kg	10/20/25 15:47	PB170162
12672-29-6	Aroclor-1248	25.4	U	1	8.90	25.4	ug/kg	10/20/25 15:47	PB170162
11097-69-1	Aroclor-1254	25.3	J	1	4.80	25.4	ug/kg	10/20/25 15:47	PB170162
37324-23-5	Aroclor-1262	25.4	U	1	7.50	25.4	ug/kg	10/20/25 15:47	PB170162
11100-14-4	Aroclor-1268	25.4	U	1	5.40	25.4	ug/kg	10/20/25 15:47	PB170162
11096-82-5	Aroclor-1260	25.4	U	1	4.80	25.4	ug/kg	10/20/25 15:47	PB170162
SURROGATES									
877-09-8	Tetrachloro-m-xylene	20.1			32 - 144	100%	SPK: 20		
2051-24-3	Decachlorobiphenyl	18.3			32 - 175	92%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071021.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 15:47
Operator : YP\AJ
Sample : Q3375-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PR132-DP02-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 08:57:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.415	2.775	87262451	100.6E6	20.059	19.797
2)	SA Decachlor...	8.522	7.538	53157442	89733691	18.305	17.771
	Target Compounds						
16)	L4 AR-1242-1	4.497	3.775	10927776	13705917	91.867m	84.156m
17)	L4 AR-1242-2	4.516	3.791	24278575	21617914	127.304m	89.561m#
18)	L4 AR-1242-3	4.574	3.951	9787731	13497861	82.322	102.396m
19)	L4 AR-1242-4	4.662	4.036	15312978	17487872	153.813	152.640m
20)	L4 AR-1242-5	5.370	4.532	13050205	30209197	165.776m	173.477
26)	L6 AR-1254-1	5.310	4.531	11967246	27079262	60.971	80.303m#
27)	L6 AR-1254-2	5.521	4.677	16915100	15783873	58.231	53.117m
28)	L6 AR-1254-3	5.876	5.060	14046162	20407003	45.722m	43.692
29)	L6 AR-1254-4	6.153	5.285	5815901	11499412	26.491	38.948 #
30)	L6 AR-1254-5	6.563	5.690	8118922	15870553	35.011m	37.756

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071021.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 15:47
Operator : YP\AJ
Sample : Q3375-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PR132-DP02-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 08:57:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.415 min
Delta R.T.: 0.015 min
Response: 87262451
Conc: 20.06 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-DP02-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/25/2025



#1 Tetrachloro-m-xylene

R.T.: 2.775 min
Delta R.T.: -0.003 min
Response: 100568847
Conc: 19.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.522 min
Delta R.T.: 0.022 min
Response: 53157442
Conc: 18.31 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.538 min
Delta R.T.: 0.004 min
Response: 89733691
Conc: 17.77 ng/ml



#16 AR-1242-1

R.T.: 4.497 min
Delta R.T.: 0.017 min
Response: 10927776
Conc: 91.87 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-DP02-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/25/2025



#16 AR-1242-1

R.T.: 3.775 min
Delta R.T.: -0.003 min
Response: 13705917
Conc: 84.16 ng/ml m



#17 AR-1242-2

R.T.: 4.516 min
Delta R.T.: 0.018 min
Response: 24278575
Conc: 127.30 ng/ml m



#17 AR-1242-2

R.T.: 3.791 min
Delta R.T.: -0.003 min
Response: 21617914
Conc: 89.56 ng/ml m



#18 AR-1242-3

R.T.: 4.574 min
Delta R.T.: 0.018 min
Response: 9787731
Conc: 82.32 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-DP02-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/25/2025



#18 AR-1242-3

R.T.: 3.951 min
Delta R.T.: -0.003 min
Response: 13497861
Conc: 102.40 ng/ml m



#19 AR-1242-4

R.T.: 4.662 min
Delta R.T.: 0.016 min
Response: 15312978
Conc: 153.81 ng/ml



#19 AR-1242-4

R.T.: 4.036 min
Delta R.T.: -0.003 min
Response: 17487872
Conc: 152.64 ng/ml m



#20 AR-1242-5

R.T.: 5.370 min
Delta R.T.: 0.020 min
Response: 13050205
Conc: 165.78 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-DP02-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/25/2025



#20 AR-1242-5

R.T.: 4.532 min
Delta R.T.: -0.002 min
Response: 30209197
Conc: 173.48 ng/ml



#26 AR-1254-1

R.T.: 5.310 min
Delta R.T.: 0.017 min
Response: 11967246
Conc: 60.97 ng/ml



#26 AR-1254-1

R.T.: 4.531 min
Delta R.T.: -0.001 min
Response: 27079262
Conc: 80.30 ng/ml m



#27 AR-1254-2

R.T.: 5.521 min
Delta R.T.: 0.014 min
Response: 16915100
Conc: 58.23 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-DP02-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/25/2025



#27 AR-1254-2

R.T.: 4.677 min
Delta R.T.: -0.003 min
Response: 15783873
Conc: 53.12 ng/ml m



#28 AR-1254-3

R.T.: 5.876 min
Delta R.T.: 0.021 min
Response: 14046162
Conc: 45.72 ng/ml m



#28 AR-1254-3

R.T.: 5.060 min
Delta R.T.: 0.000 min
Response: 20407003
Conc: 43.69 ng/ml



#29 AR-1254-4

R.T.: 6.153 min
Delta R.T.: 0.018 min
Response: 5815901
Conc: 26.49 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PR132-DP02-20251016

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/25/2025



#29 AR-1254-4

R.T.: 5.285 min
Delta R.T.: 0.000 min
Response: 11499412
Conc: 38.95 ng/ml



#30 AR-1254-5

R.T.: 6.563 min
Delta R.T.: 0.020 min
Response: 8118922
Conc: 35.01 ng/ml m



#30 AR-1254-5

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 15870553
Conc: 37.76 ng/ml

CALIBRATION SUMMARY

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
Instrument ID: ECD_P **Calibration Date(s):** 09/23/2025 09/23/2025
Calibration Times: 08:45 17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP075247.D RT 750 = PP075248.D
RT 500 = PP075249.D RT 250 = PP075250.D RT 050 = PP075251.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	5.81	5.81	5.81	5.81	5.82	5.81	5.71	5.91
Aroclor-1016-2	(2)	5.83	5.83	5.83	5.83	5.84	5.83	5.73	5.93
Aroclor-1016-3	(3)	5.89	5.90	5.89	5.89	5.90	5.89	5.79	5.99
Aroclor-1016-4	(4)	5.99	5.99	5.99	5.99	6.00	5.99	5.89	6.09
Aroclor-1016-5	(5)	6.28	6.29	6.28	6.28	6.29	6.28	6.18	6.38
Aroclor-1260-1	(1)	7.40	7.40	7.40	7.40	7.41	7.40	7.30	7.50
Aroclor-1260-2	(2)	7.65	7.66	7.65	7.65	7.66	7.65	7.55	7.75
Aroclor-1260-3	(3)	8.01	8.01	8.01	8.01	8.02	8.01	7.91	8.11
Aroclor-1260-4	(4)	8.24	8.24	8.24	8.24	8.25	8.24	8.14	8.34
Aroclor-1260-5	(5)	8.56	8.57	8.56	8.56	8.57	8.57	8.47	8.67
Decachlorobiphenyl		10.43	10.44	10.43	10.43	10.45	10.44	10.34	10.54
Tetrachloro-m-xylene		4.65	4.66	4.66	4.66	4.66	4.66	4.56	4.76
Aroclor-1242-1	(1)	5.81	5.81	5.81	5.81	5.81	5.81	5.71	5.91
Aroclor-1242-2	(2)	5.83	5.83	5.83	5.83	5.83	5.83	5.73	5.93
Aroclor-1242-3	(3)	5.90	5.89	5.90	5.90	5.90	5.90	5.80	6.00
Aroclor-1242-4	(4)	5.99	5.99	5.99	5.99	5.99	5.99	5.89	6.09
Aroclor-1242-5	(5)	6.72	6.72	6.72	6.72	6.72	6.72	6.62	6.82
Decachlorobiphenyl		10.44	10.44	10.44	10.44	10.44	10.44	10.34	10.54
Tetrachloro-m-xylene		4.66	4.66	4.66	4.66	4.66	4.66	4.56	4.76
Aroclor-1248-1	(1)	5.81	5.81	5.81	5.81	5.81	5.81	5.71	5.91
Aroclor-1248-2	(2)	6.08	6.08	6.08	6.08	6.08	6.08	5.98	6.18
Aroclor-1248-3	(3)	6.28	6.28	6.29	6.28	6.28	6.28	6.18	6.38
Aroclor-1248-4	(4)	6.68	6.68	6.68	6.68	6.68	6.68	6.58	6.78
Aroclor-1248-5	(5)	6.72	6.72	6.72	6.72	6.72	6.72	6.62	6.82
Decachlorobiphenyl		10.43	10.44	10.44	10.44	10.44	10.44	10.34	10.54
Tetrachloro-m-xylene		4.66	4.66	4.66	4.66	4.66	4.66	4.56	4.76
Aroclor-1254-1	(1)	6.66	6.66	6.66	6.66	6.66	6.66	6.56	6.76
Aroclor-1254-2	(2)	6.88	6.87	6.88	6.88	6.87	6.87	6.77	6.97
Aroclor-1254-3	(3)	7.24	7.24	7.24	7.24	7.23	7.24	7.14	7.34
Aroclor-1254-4	(4)	7.52	7.52	7.52	7.52	7.52	7.52	7.42	7.62
Aroclor-1254-5	(5)	7.94	7.93	7.94	7.94	7.93	7.93	7.83	8.03
Decachlorobiphenyl		10.44	10.44	10.44	10.44	10.43	10.44	10.34	10.54
Tetrachloro-m-xylene		4.66	4.66	4.66	4.66	4.66	4.66	4.56	4.76
Aroclor-1268-1	(1)	8.89	8.89	8.89	8.89	8.89	8.89	8.79	8.99
Aroclor-1268-2	(2)	8.99	8.99	8.99	8.98	8.98	8.99	8.89	9.09
Aroclor-1268-3	(3)	9.23	9.23	9.23	9.23	9.22	9.23	9.13	9.33
Aroclor-1268-4	(4)	9.66	9.66	9.66	9.65	9.65	9.65	9.55	9.75
Aroclor-1268-5	(5)	10.09	10.09	10.09	10.09	10.08	10.09	9.99	10.19

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.44	10.44	10.44	10.44	10.44	10.44	10.34	10.54
Tetrachloro-m-xylene	4.66	4.66	4.66	4.66	4.66	4.66	4.56	4.76

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Instrument ID: ECD_P **Calibration Date(s):** 09/23/2025 09/23/2025

Calibration Times: 08:45 17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP075247.D RT 750 = PP075248.D

RT 500 = PP075249.D RT 250 = PP075250.D RT 050 = PP075251.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	4.90	4.89	4.89	4.89	4.90	4.90	4.80	5.00
Aroclor-1016-2	(2)	4.95	4.95	4.95	4.95	4.95	4.95	4.85	5.05
Aroclor-1016-3	(3)	5.07	5.07	5.07	5.07	5.07	5.07	4.97	5.17
Aroclor-1016-4	(4)	5.11	5.11	5.11	5.11	5.12	5.11	5.01	5.21
Aroclor-1016-5	(5)	5.33	5.33	5.33	5.33	5.33	5.33	5.23	5.43
Aroclor-1260-1	(1)	6.36	6.36	6.36	6.35	6.36	6.36	6.26	6.46
Aroclor-1260-2	(2)	6.54	6.54	6.54	6.54	6.55	6.54	6.44	6.64
Aroclor-1260-3	(3)	6.70	6.70	6.70	6.70	6.70	6.70	6.60	6.80
Aroclor-1260-4	(4)	7.17	7.17	7.17	7.16	7.17	7.17	7.07	7.27
Aroclor-1260-5	(5)	7.41	7.40	7.40	7.40	7.41	7.41	7.31	7.51
Decachlorobiphenyl		8.81	8.80	8.80	8.80	8.81	8.81	8.71	8.91
Tetrachloro-m-xylene		3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90
Aroclor-1242-1	(1)	4.90	4.90	4.90	4.89	4.89	4.90	4.80	5.00
Aroclor-1242-2	(2)	4.95	4.96	4.95	4.95	4.95	4.95	4.85	5.05
Aroclor-1242-3	(3)	5.07	5.07	5.07	5.07	5.08	5.07	4.97	5.17
Aroclor-1242-4	(4)	5.16	5.16	5.16	5.15	5.16	5.16	5.06	5.26
Aroclor-1242-5	(5)	5.68	5.68	5.68	5.68	5.68	5.68	5.58	5.78
Decachlorobiphenyl		8.81	8.81	8.81	8.81	8.80	8.81	8.71	8.91
Tetrachloro-m-xylene		3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90
Aroclor-1248-1	(1)	4.89	4.90	4.89	4.90	4.90	4.90	4.80	5.00
Aroclor-1248-2	(2)	5.11	5.12	5.11	5.12	5.11	5.11	5.01	5.21
Aroclor-1248-3	(3)	5.15	5.16	5.15	5.16	5.15	5.16	5.06	5.26
Aroclor-1248-4	(4)	5.33	5.33	5.33	5.33	5.33	5.33	5.23	5.43
Aroclor-1248-5	(5)	5.72	5.72	5.71	5.72	5.73	5.72	5.62	5.82
Decachlorobiphenyl		8.80	8.81	8.80	8.81	8.80	8.81	8.71	8.91
Tetrachloro-m-xylene		3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90
Aroclor-1254-1	(1)	5.68	5.68	5.68	5.68	5.68	5.68	5.58	5.78
Aroclor-1254-2	(2)	5.83	5.83	5.83	5.83	5.83	5.83	5.73	5.93
Aroclor-1254-3	(3)	6.23	6.23	6.23	6.23	6.23	6.23	6.13	6.33
Aroclor-1254-4	(4)	6.45	6.45	6.46	6.45	6.45	6.45	6.35	6.55
Aroclor-1254-5	(5)	6.87	6.87	6.87	6.87	6.87	6.87	6.77	6.97
Decachlorobiphenyl		8.81	8.81	8.81	8.81	8.81	8.81	8.71	8.91
Tetrachloro-m-xylene		3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90
Aroclor-1268-1	(1)	7.69	7.69	7.69	7.69	7.69	7.69	7.59	7.79
Aroclor-1268-2	(2)	7.75	7.75	7.75	7.75	7.75	7.75	7.65	7.85
Aroclor-1268-3	(3)	7.96	7.96	7.96	7.96	7.96	7.96	7.86	8.06
Aroclor-1268-4	(4)	8.25	8.25	8.25	8.25	8.25	8.25	8.15	8.35
Aroclor-1268-5	(5)	8.55	8.55	8.55	8.54	8.54	8.54	8.44	8.64

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.81	8.81	8.81	8.80	8.81	8.81	8.71	8.91
Tetrachloro-m-xylene	3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: AECO02
SDG NO.: Q3375
Calibration Date(s): 09/23/2025 09/23/2025
Calibration Times: 08:45 17:14

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID: CF 1000 = PP075247.D CF 750 = PP075248.D
CF 500 = PP075249.D CF 250 = PP075250.D CF 050 = PP075251.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	42347817	45007967	46870170	50153844	56144100	48104780	11
Aroclor-1016-2	(2)	64402738	67720917	69660408	74265036	82593880	71728596	10
Aroclor-1016-3	(3)	39431965	41872236	43807668	47351972	60190040	46530776	18
Aroclor-1016-4	(4)	32893250	34833364	35556514	38336304	36563880	35636662	6
Aroclor-1016-5	(5)	31586228	33455985	34268374	37039376	32256580	33721309	6
Aroclor-1260-1	(1)	52823399	56388379	60551008	64105520	66729180	60119497	9
Aroclor-1260-2	(2)	63303692	67485644	69648556	75704048	87848060	72798000	13
Aroclor-1260-3	(3)	48177125	50964003	52841238	56234364	62508960	54145138	10
Aroclor-1260-4	(4)	52102929	55158468	56921418	59245152	59257840	56537161	5
Aroclor-1260-5	(5)	102126116	107524053	110369268	115809508	113212520	109808293	5
Decachlorobiphenyl		920246130	962712813	986576680	1032362440	1085842800	997548173	6
Tetrachloro-m-xylene		1241637260	1308534467	1312330580	1382545720	1297359400	1308481485	4
Aroclor-1242-1	(1)	36886529	39218005	41223848	44446876	46523260	41659704	9
Aroclor-1242-2	(2)	55040068	57977600	60249464	64089044	60830740	59637383	6
Aroclor-1242-3	(3)	34279044	36622237	38652038	41896284	50681480	40426217	16
Aroclor-1242-4	(4)	28461551	29972721	31322910	33558448	33854640	31434054	7
Aroclor-1242-5	(5)	31568865	33442463	35656256	39523980	35586800	35155673	8
Decachlorobiphenyl		926627030	968517160	1003394580	1047152200	937997000	976737594	5
Tetrachloro-m-xylene		1236238500	1295575613	1320312660	1382228640	1251655600	1297202203	4
Aroclor-1248-1	(1)	29172565	31397577	32249970	33755340	31223340	31559758	5
Aroclor-1248-2	(2)	39359802	41783284	43459792	45079104	47158200	43368036	7
Aroclor-1248-3	(3)	44196144	46704360	48478236	50173996	49240700	47758687	5
Aroclor-1248-4	(4)	53562132	56623921	59266248	61341900	64765060	59111852	7
Aroclor-1248-5	(5)	51867384	54812452	57671502	61547912	69597980	59099446	12
Decachlorobiphenyl		936214370	980173387	1009425800	1018487400	982414600	985343111	3
Tetrachloro-m-xylene		1235222620	1285172653	1318552840	1328682760	1313272800	1296180735	3
Aroclor-1254-1	(1)	61475917	62734304	65986682	62998532	67924500	64223987	4
Aroclor-1254-2	(2)	77552440	82328021	86739650	88146820	94902000	85933786	8
Aroclor-1254-3	(3)	82845377	87949261	92224268	94074964	105017660	92422306	9
Aroclor-1254-4	(4)	64880438	68693089	72666206	72469616	78036900	71349250	7
Aroclor-1254-5	(5)	77310574	81941501	85874638	85948608	77961320	81807328	5
Decachlorobiphenyl		940196120	995283773	1040085460	1002732880	956220600	986903767	4
Tetrachloro-m-xylene		1243265480	1304021133	1335815700	1326930840	1339353000	1309877231	3
Aroclor-1268-1	(1)	152883178	159692704	169160672	171922300	156133040	161958379	5

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	141133811	146979519	155825750	156513748	141348060	148360178	5
Aroclor-1268-3	(3)	118887581	123478560	131058116	130049472	116969640	124088674	5
Aroclor-1268-4	(4)	55874289	58464284	61978114	60989252	60829860	59627160	4
Aroclor-1268-5	(5)	373677572	388367939	407123198	405183456	366620440	388194521	5
Decachlorobiphenyl		1657406580	1735675800	1838831580	1851125440	1610932200	1738794320	6
Tetrachloro-m-xylene		1279137450	1334222493	1390576680	1375184440	1232599000	1322344013	5

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Instrument ID: ECD_P **Calibration Date(s):** 09/23/2025 09/23/2025

Calibration Times: 08:45 17:14

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID: **CF 1000 =** PP075247.D **CF 750 =** PP075248.D

CF 500 = PP075249.D **CF 250 =** PP075250.D **CF 050 =** PP075251.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	550669557	575041595	570968608	557638892	550748680	561013466	2
Aroclor-1016-2	(2)	253692272	266987116	261503738	252594292	250233300	257002144	3
Aroclor-1016-3	(3)	139883237	146298759	147016140	140568364	202575320	155268364	17
Aroclor-1016-4	(4)	139175266	149989423	152767546	146969012	145938660	146967981	3
Aroclor-1016-5	(5)	157135257	166127832	167159970	171673236	154814600	163382179	4
Aroclor-1260-1	(1)	389924896	398807288	400126940	426602072	414215600	405935359	4
Aroclor-1260-2	(2)	565343775	583502056	582903442	580470456	566929720	575829890	2
Aroclor-1260-3	(3)	422351862	433041807	427033404	426203564	431794540	428085035	1
Aroclor-1260-4	(4)	428778820	432299260	434882920	423190072	436968900	431223994	1
Aroclor-1260-5	(5)	1056009939	1104491933	1106785704	1094138908	898515660	1051988429	8
Decachlorobiphenyl		6122260390	6260409867	6205942400	6354732320	5219165200	6032502035	8
Tetrachloro-m-xylene		6636388100	6400989987	6258078260	6278037480	4930931000	6100884965	11
Aroclor-1242-1	(1)	472236904	485846165	486239260	487321948	403955360	467119927	8
Aroclor-1242-2	(2)	217510086	216447451	216513930	219383960	174933800	208957845	9
Aroclor-1242-3	(3)	123250580	122980740	121647250	125144732	116794760	121963612	3
Aroclor-1242-4	(4)	159270704	162580923	163635768	175639772	134330060	159091445	10
Aroclor-1242-5	(5)	192075351	196719551	194226854	201700156	188695940	194683570	3
Decachlorobiphenyl		6210283890	6223405693	6336701840	6098511680	5039253200	5981631261	9
Tetrachloro-m-xylene		6371057840	6261042827	6076223420	6032393800	4834060000	5914955577	10
Aroclor-1248-1	(1)	313001525	314287516	316712298	313219832	245110400	300466314	10
Aroclor-1248-2	(2)	194621063	204942157	200519274	206042640	206270120	202479051	2
Aroclor-1248-3	(3)	238833016	244338553	263622092	258775616	243161480	249746151	4
Aroclor-1248-4	(4)	233390044	236006324	231247510	230022068	233051700	232743529	1
Aroclor-1248-5	(5)	417516536	433645501	453462226	437040196	301192600	408571412	15
Decachlorobiphenyl		6280619880	6334319187	6392002300	6030990520	5377561400	6083098657	7
Tetrachloro-m-xylene		6342135760	6351132560	6173502940	5553544360	4399164000	5763895924	14
Aroclor-1254-1	(1)	503004292	535011767	531987056	521951800	402057780	498802539	11
Aroclor-1254-2	(2)	378219544	403583375	399300676	392234176	422583680	399184290	4
Aroclor-1254-3	(3)	715208754	722755991	718569236	662347256	504737840	664723815	14
Aroclor-1254-4	(4)	505627774	538597859	544931308	492275784	386803020	493647149	13
Aroclor-1254-5	(5)	551845651	631515513	645579268	578311224	446211920	570692715	14
Decachlorobiphenyl		6237706050	6279124973	6507474640	6157241280	5410407600	6118390909	7
Tetrachloro-m-xylene		6240840860	6469002347	6413333980	5502278160	4576318000	5840354669	14
Aroclor-1268-1	(1)	1460168699	1513699808	1559348234	1460709316	1219243940	1442633999	9

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	1471677824	1500245056	1520481630	1382554016	1099105140	1394812733	12
Aroclor-1268-3	(3)	1142468076	1153335345	1195510260	1093929288	869367940	1090922182	12
Aroclor-1268-4	(4)	470517486	472924412	482908940	460922856	362513880	449957515	11
Aroclor-1268-5	(5)	3043336629	3184766664	3278330982	3285110680	2399968480	3038302687	12
Decachlorobiphenyl		12094690940	12187875973	12233092380	12124440600	9791304400	11686280859	9
Tetrachloro-m-xylene		6567329620	6628773733	6640722000	6057317560	4103278800	5999484343	18

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Instrument ID: ECD_P **Date(s) Analyzed:** 09/23/2025 09/23/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	4.86	4.76	4.96	18490700
		2	4.95	4.85	5.05	13697900
		3	5.02	4.92	5.12	41397800
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	5.02	4.92	5.12	32115600
		2	5.54	5.44	5.64	16774900
		3	5.83	5.73	5.93	33440800
		4	5.99	5.89	6.09	17299100
		5	6.08	5.98	6.18	12004400
Aroclor-1262	500	1	8.24	8.14	8.34	75955200
		2	8.57	8.47	8.67	135361000
		3	8.90	8.80	9.00	93451000
		4	8.98	8.88	9.08	72237200
		5	9.66	9.56	9.76	52135400

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Instrument ID: ECD_P **Date(s) Analyzed:** 09/23/2025 09/23/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	4.01	3.91	4.11	76477600
		2	4.09	3.99	4.19	52263200
		3	4.17	4.07	4.27	190406000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.17	4.07	4.27	146566000
		2	4.89	4.79	4.99	252198000
		3	5.07	4.97	5.17	63015000
		4	5.15	5.05	5.25	103164000
		5	5.33	5.23	5.43	64046600
Aroclor-1262	500	1	6.91	6.81	7.01	791116000
		2	7.17	7.07	7.27	552368000
		3	7.69	7.59	7.79	499974000
		4	7.75	7.65	7.85	908084000
		5	8.25	8.15	8.35	429906000

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075247.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 08:45
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:10:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.654	3.798	124.2E6	663.6E6	97.232	102.934
2)	SA Decachlor...	10.434	8.805	92024613	612.2E6	96.521	99.321
	Target Compounds						
3)	L1 AR-1016-1	5.806	4.895	42347817	550.7E6	949.311	981.902
4)	L1 AR-1016-2	5.828	4.954	64402738	253.7E6	960.782	984.838
5)	L1 AR-1016-3	5.890	5.073	39431965	139.9E6	947.432	975.138
6)	L1 AR-1016-4	5.988	5.114	32893250	139.2E6	961.092	953.442
7)	L1 AR-1016-5	6.280	5.328	31586228	157.1E6	959.272	969.088
31)	L7 AR-1260-1	7.397	6.356	52823399	389.9E6	931.840	987.087
32)	L7 AR-1260-2	7.650	6.543	63303692	565.3E6	952.277	984.707
33)	L7 AR-1260-3	8.009	6.696	48177125	422.4E6	953.829	994.488
34)	L7 AR-1260-4	8.236	7.166	52102929	428.8E6	955.804	992.932
35)	L7 AR-1260-5	8.562	7.406	102.1E6	1056.0E6	961.208	976.523

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075247.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 08:45
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:10:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.654 min
Delta R.T.: 0.000 min
Response: 124163726
Conc: 97.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000



#1 Tetrachloro-m-xylene

R.T.: 3.798 min
Delta R.T.: 0.000 min
Response: 663638810
Conc: 102.93 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.434 min
Delta R.T.: 0.000 min
Response: 92024613
Conc: 96.52 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.805 min
Delta R.T.: 0.000 min
Response: 612226039
Conc: 99.32 ng/ml



#3 AR-1016-1

R.T.: 5.806 min
Delta R.T.: 0.000 min
Response: 42347817
Conc: 949.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000



#3 AR-1016-1

R.T.: 4.895 min
Delta R.T.: 0.000 min
Response: 550669557
Conc: 981.90 ng/ml



#4 AR-1016-2

R.T.: 5.828 min
Delta R.T.: 0.000 min
Response: 64402738
Conc: 960.78 ng/ml



#4 AR-1016-2

R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 253692272
Conc: 984.84 ng/ml



#5 AR-1016-3

R.T.: 5.890 min
Delta R.T.: 0.000 min
Response: 39431965
Conc: 947.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000



#5 AR-1016-3

R.T.: 5.073 min
Delta R.T.: 0.000 min
Response: 139883237
Conc: 975.14 ng/ml



#6 AR-1016-4

R.T.: 5.988 min
Delta R.T.: 0.000 min
Response: 32893250
Conc: 961.09 ng/ml



#6 AR-1016-4

R.T.: 5.114 min
Delta R.T.: 0.000 min
Response: 139175266
Conc: 953.44 ng/ml



#7 AR-1016-5

R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 31586228
Conc: 959.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000



#7 AR-1016-5

R.T.: 5.328 min
Delta R.T.: 0.000 min
Response: 157135257
Conc: 969.09 ng/ml



#31 AR-1260-1

R.T.: 7.397 min
Delta R.T.: 0.000 min
Response: 52823399
Conc: 931.84 ng/ml



#31 AR-1260-1

R.T.: 6.356 min
Delta R.T.: 0.000 min
Response: 389924896
Conc: 987.09 ng/ml



#32 AR-1260-2

R.T.: 7.650 min
Delta R.T.: 0.000 min
Response: 63303692
Conc: 952.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000



#32 AR-1260-2

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 565343775
Conc: 984.71 ng/ml



#33 AR-1260-3

R.T.: 8.009 min
Delta R.T.: 0.000 min
Response: 48177125
Conc: 953.83 ng/ml



#33 AR-1260-3

R.T.: 6.696 min
Delta R.T.: 0.000 min
Response: 422351862
Conc: 994.49 ng/ml



#34 AR-1260-4

R.T.: 8.236 min
Delta R.T.: 0.000 min
Response: 52102929
Conc: 955.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000



#34 AR-1260-4

R.T.: 7.166 min
Delta R.T.: 0.000 min
Response: 428778820
Conc: 992.93 ng/ml



#35 AR-1260-5

R.T.: 8.562 min
Delta R.T.: 0.000 min
Response: 102126116
Conc: 961.21 ng/ml



#35 AR-1260-5

R.T.: 7.406 min
Delta R.T.: 0.000 min
Response: 1056009939
Conc: 976.52 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075248.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:01
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:21:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.659	3.796	98140085	480.1E6	76.225	74.641
2)	SA Decachlor...	10.438	8.803	72203461	469.5E6	75.486	75.777
	Target Compounds						
3)	L1 AR-1016-1	5.811	4.894	33755975	431.3E6	754.459	762.574
4)	L1 AR-1016-2	5.832	4.952	50790688	200.2E6	755.124	768.006
5)	L1 AR-1016-3	5.895	5.072	31404177	109.7E6	753.026	759.865
6)	L1 AR-1016-4	5.992	5.111	26125023	112.5E6	758.837	763.638
7)	L1 AR-1016-5	6.285	5.326	25091989	124.6E6	757.985	762.174
31)	L7 AR-1260-1	7.402	6.355	42291284	299.1E6	747.360	754.771
32)	L7 AR-1260-2	7.655	6.541	50614233	437.6E6	757.555	758.123
33)	L7 AR-1260-3	8.013	6.695	38223002	324.8E6	754.489	759.766
34)	L7 AR-1260-4	8.241	7.165	41368851	324.2E6	755.905	750.542
35)	L7 AR-1260-5	8.567	7.404	80643040	828.4E6	755.983	760.602

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075248.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:01
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:21:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 98140085
Conc: 76.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750



#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 480074249
Conc: 74.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.438 min
Delta R.T.: 0.000 min
Response: 72203461
Conc: 75.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 469530740
Conc: 75.78 ng/ml



#3 AR-1016-1

R.T.: 5.811 min
Delta R.T.: 0.000 min
Response: 33755975
Conc: 754.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750



#3 AR-1016-1

R.T.: 4.894 min
Delta R.T.: 0.000 min
Response: 431281196
Conc: 762.57 ng/ml



#4 AR-1016-2

R.T.: 5.832 min
Delta R.T.: 0.000 min
Response: 50790688
Conc: 755.12 ng/ml



#4 AR-1016-2

R.T.: 4.952 min
Delta R.T.: 0.000 min
Response: 200240337
Conc: 768.01 ng/ml



#5 AR-1016-3

R.T.: 5.895 min
Delta R.T.: 0.000 min
Response: 31404177
Conc: 753.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750



#5 AR-1016-3

R.T.: 5.072 min
Delta R.T.: 0.000 min
Response: 109724069
Conc: 759.87 ng/ml



#6 AR-1016-4

R.T.: 5.992 min
Delta R.T.: 0.000 min
Response: 26125023
Conc: 758.84 ng/ml



#6 AR-1016-4

R.T.: 5.111 min
Delta R.T.: 0.000 min
Response: 112492067
Conc: 763.64 ng/ml



#7 AR-1016-5

R.T.: 6.285 min
Delta R.T.: 0.000 min
Response: 25091989
Conc: 757.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750



#7 AR-1016-5

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 124595874
Conc: 762.17 ng/ml



#31 AR-1260-1

R.T.: 7.402 min
Delta R.T.: 0.000 min
Response: 42291284
Conc: 747.36 ng/ml



#31 AR-1260-1

R.T.: 6.355 min
Delta R.T.: 0.000 min
Response: 299105466
Conc: 754.77 ng/ml



#32 AR-1260-2

R.T.: 7.655 min
Delta R.T.: 0.000 min
Response: 50614233
Conc: 757.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750



#32 AR-1260-2

R.T.: 6.541 min
Delta R.T.: 0.000 min
Response: 437626542
Conc: 758.12 ng/ml



#33 AR-1260-3

R.T.: 8.013 min
Delta R.T.: 0.000 min
Response: 38223002
Conc: 754.49 ng/ml



#33 AR-1260-3

R.T.: 6.695 min
Delta R.T.: 0.000 min
Response: 324781355
Conc: 759.77 ng/ml



#34 AR-1260-4

R.T.: 8.241 min
Delta R.T.: 0.000 min
Response: 41368851
Conc: 755.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750



#34 AR-1260-4

R.T.: 7.165 min
Delta R.T.: 0.000 min
Response: 324224445
Conc: 750.54 ng/ml



#35 AR-1260-5

R.T.: 8.567 min
Delta R.T.: 0.000 min
Response: 80643040
Conc: 755.98 ng/ml



#35 AR-1260-5

R.T.: 7.404 min
Delta R.T.: 0.000 min
Response: 828368950
Conc: 760.60 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075249.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:18
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:07:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.655	3.796	65616529	312.9E6	50.000	50.000
2)	SA Decachlor...	10.434	8.803	49328834	310.3E6	50.000	50.000
	Target Compounds						
3)	L1 AR-1016-1	5.808	4.893	23435085	285.5E6	500.000	500.000
4)	L1 AR-1016-2	5.829	4.951	34830204	130.8E6	500.000	500.000
5)	L1 AR-1016-3	5.892	5.071	21903834	73508070	500.000	500.000
6)	L1 AR-1016-4	5.989	5.111	17778257	76383773	500.000	500.000
7)	L1 AR-1016-5	6.281	5.325	17134187	83579985	500.000	500.000
31)	L7 AR-1260-1	7.399	6.355	30275504	200.1E6	500.000	500.000
32)	L7 AR-1260-2	7.651	6.541	34824278	291.5E6	500.000	500.000
33)	L7 AR-1260-3	8.009	6.695	26420619	213.5E6	500.000	500.000
34)	L7 AR-1260-4	8.237	7.165	28460709	217.4E6	500.000	500.000
35)	L7 AR-1260-5	8.564	7.404	55184634	553.4E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075249.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:18
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:07:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 65616529
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500



#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 312903913
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.434 min
Delta R.T.: 0.000 min
Response: 49328834
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 310297120
Conc: 50.00 ng/ml



#3 AR-1016-1

R.T.: 5.808 min
Delta R.T.: 0.000 min
Response: 23435085
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500



#3 AR-1016-1

R.T.: 4.893 min
Delta R.T.: 0.000 min
Response: 285484304
Conc: 500.00 ng/ml



#4 AR-1016-2

R.T.: 5.829 min
Delta R.T.: 0.000 min
Response: 34830204
Conc: 500.00 ng/ml



#4 AR-1016-2

R.T.: 4.951 min
Delta R.T.: 0.000 min
Response: 130751869
Conc: 500.00 ng/ml



#5 AR-1016-3

R.T.: 5.892 min
Delta R.T.: 0.000 min
Response: 21903834
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500



#5 AR-1016-3

R.T.: 5.071 min
Delta R.T.: 0.000 min
Response: 73508070
Conc: 500.00 ng/ml



#6 AR-1016-4

R.T.: 5.989 min
Delta R.T.: 0.000 min
Response: 17778257
Conc: 500.00 ng/ml



#6 AR-1016-4

R.T.: 5.111 min
Delta R.T.: 0.000 min
Response: 76383773
Conc: 500.00 ng/ml



#7 AR-1016-5

R.T.: 6.281 min
Delta R.T.: 0.000 min
Response: 17134187
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500



#7 AR-1016-5

R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 83579985
Conc: 500.00 ng/ml



#31 AR-1260-1

R.T.: 7.399 min
Delta R.T.: 0.000 min
Response: 30275504
Conc: 500.00 ng/ml



#31 AR-1260-1

R.T.: 6.355 min
Delta R.T.: 0.000 min
Response: 200063470
Conc: 500.00 ng/ml



#32 AR-1260-2

R.T.: 7.651 min
Delta R.T.: 0.000 min
Response: 34824278
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500



#32 AR-1260-2

R.T.: 6.541 min
Delta R.T.: 0.000 min
Response: 291451721
Conc: 500.00 ng/ml



#33 AR-1260-3

R.T.: 8.009 min
Delta R.T.: 0.000 min
Response: 26420619
Conc: 500.00 ng/ml



#33 AR-1260-3

R.T.: 6.695 min
Delta R.T.: 0.000 min
Response: 213516702
Conc: 500.00 ng/ml



#34 AR-1260-4

R.T.: 8.237 min
Delta R.T.: 0.000 min
Response: 28460709
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500



#34 AR-1260-4

R.T.: 7.165 min
Delta R.T.: 0.000 min
Response: 217441460
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.564 min
Delta R.T.: 0.000 min
Response: 55184634
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.404 min
Delta R.T.: 0.000 min
Response: 553392852
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075250.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:34
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:25:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.655	3.796	34563643	157.0E6	26.359	24.549
2)	SA Decachlor...	10.434	8.803	25809061	158.9E6	26.458	25.477
	Target Compounds						
3)	L1 AR-1016-1	5.807	4.894	12538461	139.4E6	272.014	247.365
4)	L1 AR-1016-2	5.828	4.952	18566259	63148573	269.028	244.105
5)	L1 AR-1016-3	5.891	5.072	11837993	35142091	274.562	244.992
6)	L1 AR-1016-4	5.988	5.113	9584076	36742253	270.699	249.565
7)	L1 AR-1016-5	6.280	5.326	9259844	42918309	271.649	259.287
31)	L7 AR-1260-1	7.398	6.354	16026380	106.7E6	274.109	264.074
32)	L7 AR-1260-2	7.651	6.542	18926012	145.1E6	274.149	251.045
33)	L7 AR-1260-3	8.009	6.695	14058591	106.6E6	270.076	249.442
34)	L7 AR-1260-4	8.236	7.164	14811288	105.8E6	254.816m	246.162
35)	L7 AR-1260-5	8.563	7.404	28952377	273.5E6	265.722	250.867

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075250.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:34
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:25:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 34563643
Conc: 26.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 156950937
Conc: 24.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.434 min
Delta R.T.: 0.000 min
Response: 25809061
Conc: 26.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 158868308
Conc: 25.48 ng/ml



#3 AR-1016-1

R.T.: 5.807 min
Delta R.T.: 0.000 min
Response: 12538461
Conc: 272.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#3 AR-1016-1

R.T.: 4.894 min
Delta R.T.: 0.000 min
Response: 139409723
Conc: 247.36 ng/ml



#4 AR-1016-2

R.T.: 5.828 min
Delta R.T.: 0.000 min
Response: 18566259
Conc: 269.03 ng/ml



#4 AR-1016-2

R.T.: 4.952 min
Delta R.T.: 0.000 min
Response: 63148573
Conc: 244.10 ng/ml



#5 AR-1016-3

R.T.: 5.891 min
Delta R.T.: 0.000 min
Response: 11837993
Conc: 274.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#5 AR-1016-3

R.T.: 5.072 min
Delta R.T.: 0.000 min
Response: 35142091
Conc: 244.99 ng/ml



#6 AR-1016-4

R.T.: 5.988 min
Delta R.T.: 0.000 min
Response: 9584076
Conc: 270.70 ng/ml



#6 AR-1016-4

R.T.: 5.113 min
Delta R.T.: 0.000 min
Response: 36742253
Conc: 249.56 ng/ml



#7 AR-1016-5

R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 9259844
Conc: 271.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#7 AR-1016-5

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 42918309
Conc: 259.29 ng/ml



#31 AR-1260-1

R.T.: 7.398 min
Delta R.T.: 0.000 min
Response: 16026380
Conc: 274.11 ng/ml



#31 AR-1260-1

R.T.: 6.354 min
Delta R.T.: 0.000 min
Response: 106650518
Conc: 264.07 ng/ml



#32 AR-1260-2

R.T.: 7.651 min
Delta R.T.: 0.000 min
Response: 18926012
Conc: 274.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#32 AR-1260-2

R.T.: 6.542 min
Delta R.T.: 0.000 min
Response: 145117614
Conc: 251.04 ng/ml



#33 AR-1260-3

R.T.: 8.009 min
Delta R.T.: 0.000 min
Response: 14058591
Conc: 270.08 ng/ml



#33 AR-1260-3

R.T.: 6.695 min
Delta R.T.: 0.000 min
Response: 106550891
Conc: 249.44 ng/ml



#34 AR-1260-4

R.T.: 8.236 min
Delta R.T.: -0.001 min
Response: 14811288
Conc: 254.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#34 AR-1260-4

R.T.: 7.164 min
Delta R.T.: 0.000 min
Response: 105797518
Conc: 246.16 ng/ml



#35 AR-1260-5

R.T.: 8.563 min
Delta R.T.: 0.000 min
Response: 28952377
Conc: 265.72 ng/ml



#35 AR-1260-5

R.T.: 7.404 min
Delta R.T.: 0.000 min
Response: 273534727
Conc: 250.87 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075251.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:11
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:30:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.662	3.796	6486797	24654655	4.958	4.041
2)	SA Decachlor...	10.447	8.810	5429214	26095826	5.432m	4.326
	Target Compounds						
3)	L1 AR-1016-1	5.815	4.897	2807205	27537434	58.356	49.085
4)	L1 AR-1016-2	5.837	4.954	4129694	12511665	57.574	48.683
5)	L1 AR-1016-3	5.898	5.071	3009502	10128766	65.402m	65.234
6)	L1 AR-1016-4	5.997	5.115	1828194	7296933	51.301	49.650
7)	L1 AR-1016-5	6.288	5.331	1612829	7740730	48.062m	47.378
31)	L7 AR-1260-1	7.406	6.357	3336459	20710780	56.486m	51.020
32)	L7 AR-1260-2	7.660	6.545	4392403	28346486	62.449	49.227
33)	L7 AR-1260-3	8.018	6.695	3125448	21589727	57.724	50.432m
34)	L7 AR-1260-4	8.246	7.168	2962892	21848445	51.832m	52.185
35)	L7 AR-1260-5	8.573	7.405	5660626	44925783	51.550	43.361m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075251.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:11
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:30:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.662 min
Delta R.T.: 0.000 min
Response: 6486797
Conc: 4.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 24654655
Conc: 4.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.447 min
Delta R.T.: 0.000 min
Response: 5429214
Conc: 5.43 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.810 min
Delta R.T.: 0.000 min
Response: 26095826
Conc: 4.33 ng/ml



#3 AR-1016-1

R.T.: 5.815 min
Delta R.T.: 0.000 min
Response: 2807205
Conc: 58.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#3 AR-1016-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 27537434
Conc: 49.09 ng/ml



#4 AR-1016-2

R.T.: 5.837 min
Delta R.T.: 0.000 min
Response: 4129694
Conc: 57.57 ng/ml



#4 AR-1016-2

R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 12511665
Conc: 48.68 ng/ml



#5 AR-1016-3

R.T.: 5.898 min
Delta R.T.: 0.000 min
Response: 3009502
Conc: 65.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#5 AR-1016-3

R.T.: 5.071 min
Delta R.T.: 0.000 min
Response: 10128766
Conc: 65.23 ng/ml



#6 AR-1016-4

R.T.: 5.997 min
Delta R.T.: 0.000 min
Response: 1828194
Conc: 51.30 ng/ml



#6 AR-1016-4

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 7296933
Conc: 49.65 ng/ml



#7 AR-1016-5

R.T.: 6.288 min
Delta R.T.: 0.000 min
Response: 1612829
Conc: 48.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#7 AR-1016-5

R.T.: 5.331 min
Delta R.T.: 0.000 min
Response: 7740730
Conc: 47.38 ng/ml



#31 AR-1260-1

R.T.: 7.406 min
Delta R.T.: 0.000 min
Response: 3336459
Conc: 56.49 ng/ml m



#31 AR-1260-1

R.T.: 6.357 min
Delta R.T.: 0.000 min
Response: 20710780
Conc: 51.02 ng/ml



#32 AR-1260-2

R.T.: 7.660 min
Delta R.T.: 0.001 min
Response: 4392403
Conc: 62.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#32 AR-1260-2

R.T.: 6.545 min
Delta R.T.: 0.000 min
Response: 28346486
Conc: 49.23 ng/ml



#33 AR-1260-3

R.T.: 8.018 min
Delta R.T.: 0.000 min
Response: 3125448
Conc: 57.72 ng/ml



#33 AR-1260-3

R.T.: 6.695 min
Delta R.T.: 0.000 min
Response: 21589727
Conc: 50.43 ng/ml m



#34 AR-1260-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 2962892
Conc: 51.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#34 AR-1260-4

R.T.: 7.168 min
Delta R.T.: 0.001 min
Response: 21848445
Conc: 52.18 ng/ml



#35 AR-1260-5

R.T.: 8.573 min
Delta R.T.: 0.000 min
Response: 5660626
Conc: 51.55 ng/ml



#35 AR-1260-5

R.T.: 7.405 min
Delta R.T.: 0.000 min
Response: 44925783
Conc: 43.36 ng/ml m

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075252.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:27
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:52:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.660	3.797	66333410	311.2E6	50.000	50.000
2)	SA Decachlor...	10.443	8.805	49168629	311.8E6	50.000	50.000
	Target Compounds						
8)	L2 AR-1221-1	4.860	4.008	9245327	38238805	500.000	500.000
9)	L2 AR-1221-2	4.946	4.093	6848929	26131566	500.000	500.000
10)	L2 AR-1221-3	5.021	4.168	20698858	95202751	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075252.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:27
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:52:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 66333410
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500



#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 311163134
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.443 min
Delta R.T.: 0.000 min
Response: 49168629
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.805 min
Delta R.T.: 0.000 min
Response: 311812597
Conc: 50.00 ng/ml



#8 AR-1221-1

R.T.: 4.860 min
Delta R.T.: 0.000 min
Response: 9245327
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500



#8 AR-1221-1

R.T.: 4.008 min
Delta R.T.: 0.000 min
Response: 38238805
Conc: 500.00 ng/ml



#9 AR-1221-2

R.T.: 4.946 min
Delta R.T.: 0.000 min
Response: 6848929
Conc: 500.00 ng/ml



#9 AR-1221-2

R.T.: 4.093 min
Delta R.T.: 0.000 min
Response: 26131566
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 5.021 min
Delta R.T.: 0.000 min
Response: 20698858
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.168 min
Delta R.T.: 0.000 min
Response: 95202751
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075253.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:43
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:16:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:16:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.657	3.797	66756778	323.1E6	50.000	50.000
2)	SA Decachlor...	10.440	8.805	50078510	307.1E6	50.000	50.000
	Target Compounds						
11)	L3 AR-1232-1	5.018	4.168	16057787	73283142	500.000	500.000
12)	L3 AR-1232-2	5.544	4.894	8387436	126.1E6	500.000	500.000
13)	L3 AR-1232-3	5.831	5.071	16720373	31507455	500.000	500.000
14)	L3 AR-1232-4	5.990	5.154	8649563	51581797	500.000	500.000
15)	L3 AR-1232-5	6.080	5.326	6002207	32023272	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075253.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:43
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:16:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:16:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.657 min
Delta R.T.: 0.000 min
Response: 66756778
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500



#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 323108376
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.440 min
Delta R.T.: 0.000 min
Response: 50078510
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.805 min
Delta R.T.: 0.000 min
Response: 307122792
Conc: 50.00 ng/ml



#11 AR-1232-1

R.T.: 5.018 min
Delta R.T.: 0.000 min
Response: 16057787
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500



#11 AR-1232-1

R.T.: 4.168 min
Delta R.T.: 0.000 min
Response: 73283142
Conc: 500.00 ng/ml



#12 AR-1232-2

R.T.: 5.544 min
Delta R.T.: 0.000 min
Response: 8387436
Conc: 500.00 ng/ml



#12 AR-1232-2

R.T.: 4.894 min
Delta R.T.: 0.000 min
Response: 126098814
Conc: 500.00 ng/ml



#13 AR-1232-3

R.T.: 5.831 min
Delta R.T.: 0.000 min
Response: 16720373
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500



#13 AR-1232-3

R.T.: 5.071 min
Delta R.T.: 0.000 min
Response: 31507455
Conc: 500.00 ng/ml



#14 AR-1232-4

R.T.: 5.990 min
Delta R.T.: 0.000 min
Response: 8649563
Conc: 500.00 ng/ml



#14 AR-1232-4

R.T.: 5.154 min
Delta R.T.: 0.000 min
Response: 51581797
Conc: 500.00 ng/ml



#15 AR-1232-5

R.T.: 6.080 min
Delta R.T.: 0.000 min
Response: 6002207
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500



#15 AR-1232-5

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 32023272
Conc: 500.00 ng/ml

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075254.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:00
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:48:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:43:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.660	3.799	123.6E6	637.1E6	96.711	102.369
2)	SA Decachlor...	10.443	8.807	92662703	621.0E6	96.022	98.992
	Target Compounds						
16)	L4 AR-1242-1	5.812	4.896	36886529	472.2E6	944.472	985.391
17)	L4 AR-1242-2	5.834	4.954	55040068	217.5E6	954.815	1002.295
18)	L4 AR-1242-3	5.896	5.073	34279044	123.3E6	940.039	1006.547
19)	L4 AR-1242-4	5.993	5.155	28461551	159.3E6	952.139	958.832m
20)	L4 AR-1242-5	6.723	5.677	31568865	192.1E6	939.198	994.431

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075254.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:00
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:48:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:43:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 123623850
Conc: 96.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 637105784
Conc: 102.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.443 min
Delta R.T.: 0.000 min
Response: 92662703
Conc: 96.02 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.807 min
Delta R.T.: 0.000 min
Response: 621028389
Conc: 98.99 ng/ml



#16 AR-1242-1

R.T.: 5.812 min
Delta R.T.: 0.000 min
Response: 36886529
Conc: 944.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#16 AR-1242-1

R.T.: 4.896 min
Delta R.T.: 0.000 min
Response: 472236904
Conc: 985.39 ng/ml



#17 AR-1242-2

R.T.: 5.834 min
Delta R.T.: 0.000 min
Response: 55040068
Conc: 954.81 ng/ml



#17 AR-1242-2

R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 217510086
Conc: 1002.30 ng/ml



#18 AR-1242-3

R.T.: 5.896 min
Delta R.T.: 0.000 min
Response: 34279044
Conc: 940.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#18 AR-1242-3

R.T.: 5.073 min
Delta R.T.: 0.000 min
Response: 123250580
Conc: 1006.55 ng/ml



#19 AR-1242-4

R.T.: 5.993 min
Delta R.T.: 0.000 min
Response: 28461551
Conc: 952.14 ng/ml



#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m



#20 AR-1242-5

R.T.: 6.723 min
Delta R.T.: 0.000 min
Response: 31568865
Conc: 939.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#20 AR-1242-5

R.T.: 5.677 min
Delta R.T.: 0.000 min
Response: 192075351
Conc: 994.43 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075255.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:16
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:51:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:43:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.656	3.800	97168171	469.6E6	75.674	75.300
2)	SA Decachlor...	10.439	8.807	72638787	466.8E6	75.181	74.600
	Target Compounds						
16)	L4 AR-1242-1	5.808	4.897	29413504	364.4E6	752.082	756.863
17)	L4 AR-1242-2	5.830	4.955	43483200	162.3E6	752.881	748.698
18)	L4 AR-1242-3	5.892	5.073	27466678	92235555	752.145	752.168
19)	L4 AR-1242-4	5.989	5.156	22479541	121.9E6	751.345	763.610m
20)	L4 AR-1242-5	6.719	5.678	25081847	147.5E6	747.465	759.181

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075255.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:16
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:51:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:43:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.656 min
Delta R.T.: 0.000 min
Response: 97168171
Conc: 75.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.800 min
Delta R.T.: 0.000 min
Response: 469578212
Conc: 75.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.439 min
Delta R.T.: 0.000 min
Response: 72638787
Conc: 75.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.807 min
Delta R.T.: 0.000 min
Response: 466755427
Conc: 74.60 ng/ml



#16 AR-1242-1

R.T.: 5.808 min
Delta R.T.: 0.000 min
Response: 29413504
Conc: 752.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#16 AR-1242-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 364384624
Conc: 756.86 ng/ml



#17 AR-1242-2

R.T.: 5.830 min
Delta R.T.: 0.000 min
Response: 43483200
Conc: 752.88 ng/ml



#17 AR-1242-2

R.T.: 4.955 min
Delta R.T.: 0.000 min
Response: 162335588
Conc: 748.70 ng/ml



#18 AR-1242-3

R.T.: 5.892 min
Delta R.T.: 0.000 min
Response: 27466678
Conc: 752.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#18 AR-1242-3

R.T.: 5.073 min
Delta R.T.: 0.000 min
Response: 92235555
Conc: 752.17 ng/ml



#19 AR-1242-4

R.T.: 5.989 min
Delta R.T.: 0.000 min
Response: 22479541
Conc: 751.35 ng/ml



#19 AR-1242-4

R.T.: 5.156 min
Delta R.T.: 0.000 min
Response: 121935692
Conc: 763.61 ng/ml m



#20 AR-1242-5

R.T.: 6.719 min
Delta R.T.: 0.000 min
Response: 25081847
Conc: 747.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#20 AR-1242-5

R.T.: 5.678 min
Delta R.T.: 0.000 min
Response: 147539663
Conc: 759.18 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075256.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:32
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:45:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:43:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.659	3.799	66015633	303.8E6	50.000	50.000
2)	SA Decachlor...	10.443	8.808	50169729	316.8E6	50.000	50.000
	Target Compounds						
16)	L4 AR-1242-1	5.811	4.895	20611924	243.1E6	500.000	500.000
17)	L4 AR-1242-2	5.832	4.954	30124732	108.3E6	500.000	500.000
18)	L4 AR-1242-3	5.895	5.073	19326019	60823625	500.000	500.000
19)	L4 AR-1242-4	5.992	5.155	15661455	81817884	500.000	481.178m
20)	L4 AR-1242-5	6.722	5.676	17828128	97113427	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075256.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:32
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:45:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:43:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 66015633
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 303811171
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.443 min
Delta R.T.: 0.000 min
Response: 50169729
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.808 min
Delta R.T.: 0.000 min
Response: 316835092
Conc: 50.00 ng/ml



#16 AR-1242-1

R.T.: 5.811 min
Delta R.T.: 0.000 min
Response: 20611924
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#16 AR-1242-1

R.T.: 4.895 min
Delta R.T.: 0.000 min
Response: 243119630
Conc: 500.00 ng/ml



#17 AR-1242-2

R.T.: 5.832 min
Delta R.T.: 0.000 min
Response: 30124732
Conc: 500.00 ng/ml



#17 AR-1242-2

R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 108256965
Conc: 500.00 ng/ml



#18 AR-1242-3

R.T.: 5.895 min
Delta R.T.: 0.000 min
Response: 19326019
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#18 AR-1242-3

R.T.: 5.073 min
Delta R.T.: 0.000 min
Response: 60823625
Conc: 500.00 ng/ml



#19 AR-1242-4

R.T.: 5.992 min
Delta R.T.: 0.000 min
Response: 15661455
Conc: 500.00 ng/ml



#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 81817884
Conc: 481.18 ng/ml m



#20 AR-1242-5

R.T.: 6.722 min
Delta R.T.: 0.000 min
Response: 17828128
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#20 AR-1242-5

R.T.: 5.676 min
Delta R.T.: 0.000 min
Response: 97113427
Conc: 500.00 ng/ml

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075257.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:48
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:55:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:54:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.661	3.797	34555716	150.8E6	26.407	24.382
2)	SA Decachlor...	10.444	8.805	26178805	152.5E6	26.539	24.523
	Target Compounds						
16)	L4 AR-1242-1	5.813	4.894	11111719	121.8E6	274.745	252.283
17)	L4 AR-1242-2	5.834	4.952	16022261	54845990	270.012	252.207
18)	L4 AR-1242-3	5.897	5.072	10474071	31286183	276.635	253.831
19)	L4 AR-1242-4	5.994	5.153	8389612	43909943	272.135	267.938m
20)	L4 AR-1242-5	6.724	5.675	9880995	50425039	281.928	257.034

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075257.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:48
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:55:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:54:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.661 min
Delta R.T.: 0.000 min
Response: 34555716
Conc: 26.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 150809845
Conc: 24.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.444 min
Delta R.T.: 0.000 min
Response: 26178805
Conc: 26.54 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.805 min
Delta R.T.: 0.000 min
Response: 152462792
Conc: 24.52 ng/ml



#16 AR-1242-1

R.T.: 5.813 min
Delta R.T.: 0.000 min
Response: 11111719
Conc: 274.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#16 AR-1242-1

R.T.: 4.894 min
Delta R.T.: 0.000 min
Response: 121830487
Conc: 252.28 ng/ml



#17 AR-1242-2

R.T.: 5.834 min
Delta R.T.: 0.000 min
Response: 16022261
Conc: 270.01 ng/ml



#17 AR-1242-2

R.T.: 4.952 min
Delta R.T.: 0.000 min
Response: 54845990
Conc: 252.21 ng/ml



#18 AR-1242-3

R.T.: 5.897 min
Delta R.T.: 0.000 min
Response: 10474071
Conc: 276.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#18 AR-1242-3

R.T.: 5.072 min
Delta R.T.: 0.000 min
Response: 31286183
Conc: 253.83 ng/ml



#19 AR-1242-4

R.T.: 5.994 min
Delta R.T.: 0.000 min
Response: 8389612
Conc: 272.13 ng/ml



#19 AR-1242-4

R.T.: 5.153 min
Delta R.T.: 0.000 min
Response: 43909943
Conc: 267.94 ng/ml m



#20 AR-1242-5

R.T.: 6.724 min
Delta R.T.: 0.000 min
Response: 9880995
Conc: 281.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#20 AR-1242-5

R.T.: 5.675 min
Delta R.T.: 0.000 min
Response: 50425039
Conc: 257.03 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075258.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:05
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:00:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:59:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.657	3.796	6258278	24170300	4.827m	4.086
2)	SA Decachlor...	10.437	8.803	4689985	25196266	4.802	4.212
	Target Compounds						
16)	L4 AR-1242-1	5.810	4.893	2326163	20197768	55.837	43.239
17)	L4 AR-1242-2	5.831	4.952	3041537	8746690	51.001	41.859
18)	L4 AR-1242-3	5.895	5.076	2534074	5839738	62.684	47.881
19)	L4 AR-1242-4	5.991	5.155	1692732	6716503	53.850	42.218
20)	L4 AR-1242-5	6.719	5.676	1779340	9434797	50.613m	48.462

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075258.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:05
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:00:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:59:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.657 min
Delta R.T.: 0.000 min
Response: 6258278
Conc: 4.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 24170300
Conc: 4.09 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 4689985
Conc: 4.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 25196266
Conc: 4.21 ng/ml



#16 AR-1242-1

R.T.: 5.810 min
Delta R.T.: 0.000 min
Response: 2326163
Conc: 55.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#16 AR-1242-1

R.T.: 4.893 min
Delta R.T.: 0.000 min
Response: 20197768
Conc: 43.24 ng/ml



#17 AR-1242-2

R.T.: 5.831 min
Delta R.T.: 0.000 min
Response: 3041537
Conc: 51.00 ng/ml



#17 AR-1242-2

R.T.: 4.952 min
Delta R.T.: 0.000 min
Response: 8746690
Conc: 41.86 ng/ml



#18 AR-1242-3

R.T.: 5.895 min
Delta R.T.: 0.000 min
Response: 2534074
Conc: 62.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#18 AR-1242-3

R.T.: 5.076 min
Delta R.T.: 0.000 min
Response: 5839738
Conc: 47.88 ng/ml



#19 AR-1242-4

R.T.: 5.991 min
Delta R.T.: 0.000 min
Response: 1692732
Conc: 53.85 ng/ml



#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 6716503
Conc: 42.22 ng/ml



#20 AR-1242-5

R.T.: 6.719 min
Delta R.T.: 0.000 min
Response: 1779340
Conc: 50.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#20 AR-1242-5

R.T.: 5.676 min
Delta R.T.: 0.000 min
Response: 9434797
Conc: 48.46 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075259.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:37
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:08:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.656	3.797	123.5E6	634.2E6	96.737	101.347
2)	SA Decachlor...	10.430	8.803	93621437	628.1E6	96.237	99.121
	Target Compounds						
21)	L5 AR-1248-1	5.807	4.893	29172565	313.0E6	949.898	994.107
22)	L5 AR-1248-2	6.078	5.113	39359802	194.6E6	950.495	985.073
23)	L5 AR-1248-3	6.281	5.153	44196144	238.8E6	953.794	934.027m
24)	L5 AR-1248-4	6.678	5.326	53562132	233.4E6	949.444	1004.611
25)	L5 AR-1248-5	6.718	5.716	51867384	417.5E6	947.013	958.730

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075259.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:37
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:08:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.656 min
Delta R.T.: 0.000 min
Response: 123522262
Conc: 96.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 634213576
Conc: 101.35 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.430 min
Delta R.T.: 0.000 min
Response: 93621437
Conc: 96.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 628061988
Conc: 99.12 ng/ml



#21 AR-1248-1

R.T.: 5.807 min
Delta R.T.: 0.000 min
Response: 29172565
Conc: 949.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#21 AR-1248-1

R.T.: 4.893 min
Delta R.T.: 0.000 min
Response: 313001525
Conc: 994.11 ng/ml



#22 AR-1248-2

R.T.: 6.078 min
Delta R.T.: 0.000 min
Response: 39359802
Conc: 950.49 ng/ml



#22 AR-1248-2

R.T.: 5.113 min
Delta R.T.: 0.000 min
Response: 194621063
Conc: 985.07 ng/ml



#23 AR-1248-3

R.T.: 6.281 min
Delta R.T.: 0.000 min
Response: 44196144
Conc: 953.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#23 AR-1248-3

R.T.: 5.153 min
Delta R.T.: 0.000 min
Response: 238833016
Conc: 934.03 ng/ml m



#24 AR-1248-4

R.T.: 6.678 min
Delta R.T.: 0.000 min
Response: 53562132
Conc: 949.44 ng/ml



#24 AR-1248-4

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 233390044
Conc: 1004.61 ng/ml



#25 AR-1248-5

R.T.: 6.718 min
Delta R.T.: 0.000 min
Response: 51867384
Conc: 947.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#25 AR-1248-5

R.T.: 5.716 min
Delta R.T.: 0.000 min
Response: 417516536
Conc: 958.73 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075260.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:53
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:11:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.658	3.801	96387949	476.3E6	75.324	75.742
2)	SA Decachlor...	10.437	8.808	73513004	475.1E6	75.377	74.984
	Target Compounds						
21)	L5 AR-1248-1	5.810	4.897	23548183	235.7E6	761.091	749.095
22)	L5 AR-1248-2	6.081	5.117	31337463	153.7E6	754.496	768.427
23)	L5 AR-1248-3	6.284	5.157	35028270	183.3E6	753.951	722.769m
24)	L5 AR-1248-4	6.681	5.331	42467941	177.0E6	751.857	757.895
25)	L5 AR-1248-5	6.721	5.719	41109339	325.2E6	750.393	747.880

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075260.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:53
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:11:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.658 min
Delta R.T.: 0.000 min
Response: 96387949
Conc: 75.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.801 min
Delta R.T.: 0.000 min
Response: 476334942
Conc: 75.74 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 73513004
Conc: 75.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.808 min
Delta R.T.: 0.000 min
Response: 475073939
Conc: 74.98 ng/ml



#21 AR-1248-1

R.T.: 5.810 min
Delta R.T.: 0.000 min
Response: 23548183
Conc: 761.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#21 AR-1248-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 235715637
Conc: 749.10 ng/ml



#22 AR-1248-2

R.T.: 6.081 min
Delta R.T.: 0.000 min
Response: 31337463
Conc: 754.50 ng/ml



#22 AR-1248-2

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 153706618
Conc: 768.43 ng/ml



#23 AR-1248-3

R.T.: 6.284 min
Delta R.T.: 0.000 min
Response: 35028270
Conc: 753.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#23 AR-1248-3

R.T.: 5.157 min
Delta R.T.: 0.000 min
Response: 183253915
Conc: 722.77 ng/ml m



#24 AR-1248-4

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 42467941
Conc: 751.86 ng/ml



#24 AR-1248-4

R.T.: 5.331 min
Delta R.T.: 0.000 min
Response: 177004743
Conc: 757.89 ng/ml



#25 AR-1248-5

R.T.: 6.721 min
Delta R.T.: 0.000 min
Response: 41109339
Conc: 750.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#25 AR-1248-5

R.T.: 5.719 min
Delta R.T.: 0.000 min
Response: 325234126
Conc: 747.88 ng/ml

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075261.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 13:10
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:04:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.660	3.796	65927642	308.7E6	50.000	50.000
2)	SA Decachlor...	10.439	8.803	50471290	319.6E6	50.000	50.000
	Target Compounds						
21)	L5 AR-1248-1	5.811	4.892	16124985	158.4E6	500.000	500.000
22)	L5 AR-1248-2	6.083	5.112	21729896	100.3E6	500.000	500.000
23)	L5 AR-1248-3	6.285	5.152	24239118	131.8E6	500.000	515.550m
24)	L5 AR-1248-4	6.683	5.325	29633124	115.6E6	500.000	500.000
25)	L5 AR-1248-5	6.723	5.714	28835751	226.7E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075261.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 13:10
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:04:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 65927642
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 308675147
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.439 min
Delta R.T.: 0.000 min
Response: 50471290
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 319600115
Conc: 50.00 ng/ml



#21 AR-1248-1

R.T.: 5.811 min
Delta R.T.: 0.000 min
Response: 16124985
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#21 AR-1248-1

R.T.: 4.892 min
Delta R.T.: 0.000 min
Response: 158356149
Conc: 500.00 ng/ml



#22 AR-1248-2

R.T.: 6.083 min
Delta R.T.: 0.000 min
Response: 21729896
Conc: 500.00 ng/ml



#22 AR-1248-2

R.T.: 5.112 min
Delta R.T.: 0.000 min
Response: 100259637
Conc: 500.00 ng/ml



#23 AR-1248-3

R.T.: 6.285 min
Delta R.T.: 0.000 min
Response: 24239118
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#23 AR-1248-3

R.T.: 5.152 min
Delta R.T.: 0.000 min
Response: 131811046
Conc: 515.55 ng/ml m



#24 AR-1248-4

R.T.: 6.683 min
Delta R.T.: 0.000 min
Response: 29633124
Conc: 500.00 ng/ml



#24 AR-1248-4

R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 115623755
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 6.723 min
Delta R.T.: 0.000 min
Response: 28835751
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#25 AR-1248-5

R.T.: 5.714 min
Delta R.T.: 0.000 min
Response: 226731113
Conc: 500.00 ng/ml

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075262.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 13:26
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:14:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.657	3.801	33217069	138.8E6	25.712	22.741
2)	SA Decachlor...	10.437	8.806	25462185	150.8E6	25.822	24.087
	Target Compounds						
21)	L5 AR-1248-1	5.809	4.895	8438835	78304958	266.682	249.137
22)	L5 AR-1248-2	6.080	5.116	11269776	51510660	265.668	255.596
23)	L5 AR-1248-3	6.282	5.157	12543499	64693904	264.697	262.977m
24)	L5 AR-1248-4	6.681	5.329	15335475	57505517	265.786	247.159
25)	L5 AR-1248-5	6.720	5.719	15386978	109.3E6	272.457	250.932

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075262.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 13:26
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:14:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.657 min
Delta R.T.: 0.000 min
Response: 33217069
Conc: 25.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.801 min
Delta R.T.: 0.000 min
Response: 138838609
Conc: 22.74 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 25462185
Conc: 25.82 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 150774763
Conc: 24.09 ng/ml



#21 AR-1248-1

R.T.: 5.809 min
Delta R.T.: 0.000 min
Response: 8438835
Conc: 266.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#21 AR-1248-1

R.T.: 4.895 min
Delta R.T.: 0.000 min
Response: 78304958
Conc: 249.14 ng/ml



#22 AR-1248-2

R.T.: 6.080 min
Delta R.T.: 0.000 min
Response: 11269776
Conc: 265.67 ng/ml



#22 AR-1248-2

R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 51510660
Conc: 255.60 ng/ml



#23 AR-1248-3

R.T.: 6.282 min
Delta R.T.: 0.000 min
Response: 12543499
Conc: 264.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#23 AR-1248-3

R.T.: 5.157 min
Delta R.T.: 0.000 min
Response: 64693904
Conc: 262.98 ng/ml m



#24 AR-1248-4

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 15335475
Conc: 265.79 ng/ml



#24 AR-1248-4

R.T.: 5.329 min
Delta R.T.: 0.000 min
Response: 57505517
Conc: 247.16 ng/ml



#25 AR-1248-5

R.T.: 6.720 min
Delta R.T.: 0.000 min
Response: 15386978
Conc: 272.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#25 AR-1248-5

R.T.: 5.719 min
Delta R.T.: 0.000 min
Response: 109260049
Conc: 250.93 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075263.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 13:42
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:19:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.655	3.797	6566364	21995820	5.074m	3.816
2)	SA Decachlor...	10.437	8.804	4912073	26887807	4.985	4.420
	Target Compounds						
21)	L5 AR-1248-1	5.809	4.897	1561167	12255520	49.467	40.788
22)	L5 AR-1248-2	6.080	5.114	2357910	10313506	54.370	50.936
23)	L5 AR-1248-3	6.282	5.154	2462035	12158074	51.552	45.736m
24)	L5 AR-1248-4	6.681	5.326	3238253	11652585	54.782	50.066
25)	L5 AR-1248-5	6.719	5.728	3479899	15059630	59.114m	36.859 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075263.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 13:42
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:19:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 6566364
Conc: 5.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 21995820
Conc: 3.82 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 4912073
Conc: 4.99 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.804 min
Delta R.T.: 0.000 min
Response: 26887807
Conc: 4.42 ng/ml



#21 AR-1248-1

R.T.: 5.809 min
Delta R.T.: 0.000 min
Response: 1561167
Conc: 49.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#21 AR-1248-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 12255520
Conc: 40.79 ng/ml



#22 AR-1248-2

R.T.: 6.080 min
Delta R.T.: 0.000 min
Response: 2357910
Conc: 54.37 ng/ml



#22 AR-1248-2

R.T.: 5.114 min
Delta R.T.: 0.000 min
Response: 10313506
Conc: 50.94 ng/ml



#23 AR-1248-3

R.T.: 6.282 min
Delta R.T.: 0.000 min
Response: 2462035
Conc: 51.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#23 AR-1248-3

R.T.: 5.154 min
Delta R.T.: -0.001 min
Response: 12158074
Conc: 45.74 ng/ml m



#24 AR-1248-4

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 3238253
Conc: 54.78 ng/ml



#24 AR-1248-4

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 11652585
Conc: 50.07 ng/ml



#25 AR-1248-5

R.T.: 6.719 min
Delta R.T.: 0.000 min
Response: 3479899
Conc: 59.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#25 AR-1248-5

R.T.: 5.728 min
Delta R.T.: 0.000 min
Response: 15059630
Conc: 36.86 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075264.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 14:15
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:27:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:25:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.660	3.799	124.3E6	624.1E6	96.412	98.637
2)	SA Decachlor...	10.440	8.805	94019612	623.8E6	94.956	97.883
	Target Compounds						
26)	L6 AR-1254-1	6.659	5.678	61475917	503.0E6	968.019m	971.997
27)	L6 AR-1254-2	6.876	5.825	77552440	378.2E6	944.080	972.887
28)	L6 AR-1254-3	7.238	6.227	82845377	715.2E6	946.428	997.656
29)	L6 AR-1254-4	7.520	6.453	64880438	505.6E6	943.395	962.588
30)	L6 AR-1254-5	7.936	6.870	77310574	551.8E6	947.519	921.721

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075264.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 14:15
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:27:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:25:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 124326548
Conc: 96.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 624084086
Conc: 98.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.440 min
Delta R.T.: 0.000 min
Response: 94019612
Conc: 94.96 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.805 min
Delta R.T.: 0.000 min
Response: 623770605
Conc: 97.88 ng/ml



#26 AR-1254-1

R.T.: 6.659 min
Delta R.T.: 0.000 min
Response: 61475917
Conc: 968.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#26 AR-1254-1

R.T.: 5.678 min
Delta R.T.: 0.000 min
Response: 503004292
Conc: 972.00 ng/ml



#27 AR-1254-2

R.T.: 6.876 min
Delta R.T.: 0.000 min
Response: 77552440
Conc: 944.08 ng/ml



#27 AR-1254-2

R.T.: 5.825 min
Delta R.T.: 0.000 min
Response: 378219544
Conc: 972.89 ng/ml



#28 AR-1254-3

R.T.: 7.238 min
Delta R.T.: 0.000 min
Response: 82845377
Conc: 946.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#28 AR-1254-3

R.T.: 6.227 min
Delta R.T.: 0.000 min
Response: 715208754
Conc: 997.66 ng/ml



#29 AR-1254-4

R.T.: 7.520 min
Delta R.T.: 0.000 min
Response: 64880438
Conc: 943.40 ng/ml



#29 AR-1254-4

R.T.: 6.453 min
Delta R.T.: 0.000 min
Response: 505627774
Conc: 962.59 ng/ml



#30 AR-1254-5

R.T.: 7.936 min
Delta R.T.: 0.000 min
Response: 77310574
Conc: 947.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#30 AR-1254-5

R.T.: 6.870 min
Delta R.T.: 0.000 min
Response: 551845651
Conc: 921.72 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075265.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 14:31
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:30:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:25:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.656	3.799	97801585	485.2E6	75.559	76.113
2)	SA Decachlor...	10.437	8.806	74646283	470.9E6	75.259	74.263
	Target Compounds						
26)	L6 AR-1254-1	6.656	5.678	47050728	401.3E6	729.090m	766.735
27)	L6 AR-1254-2	6.873	5.825	61746016	302.7E6	751.107	768.826
28)	L6 AR-1254-3	7.235	6.227	65961946	542.1E6	752.364	754.081
29)	L6 AR-1254-4	7.517	6.454	51519817	403.9E6	749.416	762.571
30)	L6 AR-1254-5	7.933	6.871	61456126	473.6E6	752.135	776.903

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075265.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 14:31
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:30:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:25:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.656 min
Delta R.T.: 0.000 min
Response: 97801585
Conc: 75.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 485175176
Conc: 76.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 74646283
Conc: 75.26 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 470934373
Conc: 74.26 ng/ml



#26 AR-1254-1

R.T.: 6.656 min
Delta R.T.: 0.000 min
Response: 47050728
Conc: 729.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#26 AR-1254-1

R.T.: 5.678 min
Delta R.T.: 0.000 min
Response: 401258825
Conc: 766.74 ng/ml



#27 AR-1254-2

R.T.: 6.873 min
Delta R.T.: 0.000 min
Response: 61746016
Conc: 751.11 ng/ml



#27 AR-1254-2

R.T.: 5.825 min
Delta R.T.: 0.000 min
Response: 302687531
Conc: 768.83 ng/ml



#28 AR-1254-3

R.T.: 7.235 min
Delta R.T.: 0.000 min
Response: 65961946
Conc: 752.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#28 AR-1254-3

R.T.: 6.227 min
Delta R.T.: 0.000 min
Response: 542066993
Conc: 754.08 ng/ml



#29 AR-1254-4

R.T.: 7.517 min
Delta R.T.: 0.000 min
Response: 51519817
Conc: 749.42 ng/ml



#29 AR-1254-4

R.T.: 6.454 min
Delta R.T.: 0.000 min
Response: 403948394
Conc: 762.57 ng/ml



#30 AR-1254-5

R.T.: 7.933 min
Delta R.T.: 0.000 min
Response: 61456126
Conc: 752.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#30 AR-1254-5

R.T.: 6.871 min
Delta R.T.: 0.000 min
Response: 473636635
Conc: 776.90 ng/ml

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075266.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 14:47
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:14:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:13:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.660	3.799	66790785	320.7E6	50.000	50.000
2)	SA Decachlor...	10.441	8.806	52004273	325.4E6	50.000	50.000
	Target Compounds						
26)	L6 AR-1254-1	6.659	5.679	32993341	266.0E6	519.163m	500.000
27)	L6 AR-1254-2	6.876	5.826	43369825	199.7E6	500.000	500.000
28)	L6 AR-1254-3	7.238	6.228	46112134	359.3E6	500.000	500.000
29)	L6 AR-1254-4	7.520	6.455	36333103	272.5E6	500.000	500.000
30)	L6 AR-1254-5	7.936	6.871	42937319	322.8E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075266.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 14:47
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:14:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:13:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 66790785
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 320666699
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.441 min
Delta R.T.: 0.000 min
Response: 52004273
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 325373732
Conc: 50.00 ng/ml



#26 AR-1254-1

R.T.: 6.659 min
Delta R.T.: 0.000 min
Response: 32993341
Conc: 519.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#26 AR-1254-1

R.T.: 5.679 min
Delta R.T.: 0.000 min
Response: 265993528
Conc: 500.00 ng/ml



#27 AR-1254-2

R.T.: 6.876 min
Delta R.T.: 0.000 min
Response: 43369825
Conc: 500.00 ng/ml



#27 AR-1254-2

R.T.: 5.826 min
Delta R.T.: 0.000 min
Response: 199650338
Conc: 500.00 ng/ml



#28 AR-1254-3

R.T.: 7.238 min
Delta R.T.: 0.000 min
Response: 46112134
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#28 AR-1254-3

R.T.: 6.228 min
Delta R.T.: 0.000 min
Response: 359284618
Conc: 500.00 ng/ml



#29 AR-1254-4

R.T.: 7.520 min
Delta R.T.: 0.000 min
Response: 36333103
Conc: 500.00 ng/ml



#29 AR-1254-4

R.T.: 6.455 min
Delta R.T.: 0.000 min
Response: 272465654
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.936 min
Delta R.T.: 0.000 min
Response: 42937319
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#30 AR-1254-5

R.T.: 6.871 min
Delta R.T.: 0.000 min
Response: 322789634
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 15:04
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:34:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:25:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.659	3.798	33173271	137.6E6	25.469	22.344
2)	SA Decachlor...	10.440	8.806	25068322	153.9E6	25.205	24.451
	Target Compounds						
26)	L6 AR-1254-1	6.658	5.678	15749633	130.5E6	242.688m	249.504
27)	L6 AR-1254-2	6.875	5.825	22036705	98058544	263.308	249.301
28)	L6 AR-1254-3	7.238	6.227	23518741	165.6E6	263.446	234.968
29)	L6 AR-1254-4	7.519	6.453	18117404	123.1E6	260.019	236.508
30)	L6 AR-1254-5	7.935	6.870	21487152	144.6E6	259.604	240.237

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 15:04
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:34:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:25:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 33173271
Conc: 25.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.798 min
Delta R.T.: 0.000 min
Response: 137556954
Conc: 22.34 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.440 min
Delta R.T.: 0.000 min
Response: 25068322
Conc: 25.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 153931032
Conc: 24.45 ng/ml



#26 AR-1254-1

R.T.: 6.658 min
Delta R.T.: 0.000 min
Response: 15749633
Conc: 242.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#26 AR-1254-1

R.T.: 5.678 min
Delta R.T.: 0.000 min
Response: 130487950
Conc: 249.50 ng/ml



#27 AR-1254-2

R.T.: 6.875 min
Delta R.T.: 0.000 min
Response: 22036705
Conc: 263.31 ng/ml



#27 AR-1254-2

R.T.: 5.825 min
Delta R.T.: 0.000 min
Response: 98058544
Conc: 249.30 ng/ml



#28 AR-1254-3

R.T.: 7.238 min
Delta R.T.: 0.000 min
Response: 23518741
Conc: 263.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#28 AR-1254-3

R.T.: 6.227 min
Delta R.T.: 0.000 min
Response: 165586814
Conc: 234.97 ng/ml



#29 AR-1254-4

R.T.: 7.519 min
Delta R.T.: 0.000 min
Response: 18117404
Conc: 260.02 ng/ml



#29 AR-1254-4

R.T.: 6.453 min
Delta R.T.: 0.000 min
Response: 123068946
Conc: 236.51 ng/ml



#30 AR-1254-5

R.T.: 7.935 min
Delta R.T.: 0.000 min
Response: 21487152
Conc: 259.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#30 AR-1254-5

R.T.: 6.870 min
Delta R.T.: 0.000 min
Response: 144577806
Conc: 240.24 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075268.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 15:20
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:51:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:51:36 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.655	3.799	6696765	22881590	5.113	3.918
2)	SA Decachlor...	10.433	8.807	4781103	27052038	4.845	4.421
	Target Compounds						
26)	L6 AR-1254-1	6.655	5.680	3396225	20102889	52.688m	40.302
27)	L6 AR-1254-2	6.872	5.827	4745100	21129184	55.218	52.931
28)	L6 AR-1254-3	7.234	6.228	5250883	25236892	56.814	37.966 #
29)	L6 AR-1254-4	7.516	6.453	3901845	19340151	54.687	39.178 #
30)	L6 AR-1254-5	7.932	6.871	3898066	22310596	47.649	39.094

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075268.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 15:20
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:51:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:51:36 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 6696765
Conc: 5.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 22881590
Conc: 3.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.433 min
Delta R.T.: 0.000 min
Response: 4781103
Conc: 4.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.807 min
Delta R.T.: 0.000 min
Response: 27052038
Conc: 4.42 ng/ml



#26 AR-1254-1

R.T.: 6.655 min
Delta R.T.: 0.000 min
Response: 3396225
Conc: 52.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#26 AR-1254-1

R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 20102889
Conc: 40.30 ng/ml



#27 AR-1254-2

R.T.: 6.872 min
Delta R.T.: 0.000 min
Response: 4745100
Conc: 55.22 ng/ml



#27 AR-1254-2

R.T.: 5.827 min
Delta R.T.: 0.000 min
Response: 21129184
Conc: 52.93 ng/ml



#28 AR-1254-3

R.T.: 7.234 min
Delta R.T.: 0.000 min
Response: 5250883
Conc: 56.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#28 AR-1254-3

R.T.: 6.228 min
Delta R.T.: 0.000 min
Response: 25236892
Conc: 37.97 ng/ml



#29 AR-1254-4

R.T.: 7.516 min
Delta R.T.: 0.000 min
Response: 3901845
Conc: 54.69 ng/ml



#29 AR-1254-4

R.T.: 6.453 min
Delta R.T.: 0.000 min
Response: 19340151
Conc: 39.18 ng/ml



#30 AR-1254-5

R.T.: 7.932 min
Delta R.T.: 0.000 min
Response: 3898066
Conc: 47.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#30 AR-1254-5

R.T.: 6.871 min
Delta R.T.: 0.000 min
Response: 22310596
Conc: 39.09 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075269.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 15:52
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 16:11:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:09:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.659	3.800	64525262	296.7E6	50.000	50.000
2)	SA Decachlor...	10.443	8.807	49716283	321.5E6	50.000	50.000
	Target Compounds						
36)	L8 AR-1262-1	8.243	6.909	37977598	395.6E6	500.000	500.000
37)	L8 AR-1262-2	8.569	7.168	67680488	276.2E6	500.000	500.000
38)	L8 AR-1262-3	8.895	7.690	46725485	250.0E6	500.000	500.000
39)	L8 AR-1262-4	8.984	7.753	36118617	454.0E6	500.000	500.000
40)	L8 AR-1262-5	9.656	8.247	26067736	215.0E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075269.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 15:52
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 16:11:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:09:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 64525262
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500



#1 Tetrachloro-m-xylene

R.T.: 3.800 min
Delta R.T.: 0.000 min
Response: 296718295
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.443 min
Delta R.T.: 0.000 min
Response: 49716283
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.807 min
Delta R.T.: 0.000 min
Response: 321471739
Conc: 50.00 ng/ml



#36 AR-1262-1

R.T.: 8.243 min
Delta R.T.: 0.000 min
Response: 37977598
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500



#36 AR-1262-1

R.T.: 6.909 min
Delta R.T.: 0.000 min
Response: 395558471
Conc: 500.00 ng/ml



#37 AR-1262-2

R.T.: 8.569 min
Delta R.T.: 0.000 min
Response: 67680488
Conc: 500.00 ng/ml



#37 AR-1262-2

R.T.: 7.168 min
Delta R.T.: 0.000 min
Response: 276184209
Conc: 500.00 ng/ml



#38 AR-1262-3

R.T.: 8.895 min
Delta R.T.: 0.000 min
Response: 46725485
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500



#38 AR-1262-3

R.T.: 7.690 min
Delta R.T.: 0.000 min
Response: 249986782
Conc: 500.00 ng/ml



#39 AR-1262-4

R.T.: 8.984 min
Delta R.T.: 0.000 min
Response: 36118617
Conc: 500.00 ng/ml



#39 AR-1262-4

R.T.: 7.753 min
Delta R.T.: 0.000 min
Response: 454041768
Conc: 500.00 ng/ml



#40 AR-1262-5

R.T.: 9.656 min
Delta R.T.: 0.000 min
Response: 26067736
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500



#40 AR-1262-5

R.T.: 8.247 min
Delta R.T.: 0.000 min
Response: 214953059
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075270.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:09
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:03:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:59:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.659	3.799	127.9E6	656.7E6	95.826	99.444
2)	SA Decachlor...	10.442	8.807	165.7E6	1209.5E6	94.811	99.431
	Target Compounds						
41)	L9 AR-1268-1	8.890	7.690	152.9E6	1460.2E6	949.456	967.154
42)	L9 AR-1268-2	8.986	7.754	141.1E6	1471.7E6	950.525	983.689
43)	L9 AR-1268-3	9.226	7.961	118.9E6	1142.5E6	951.307	977.313
44)	L9 AR-1268-4	9.655	8.247	55874289	470.5E6	948.208	987.003
45)	L9 AR-1268-5	10.086	8.545	373.7E6	3043.3E6	957.165	962.827

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075270.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:09
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:03:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:59:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 127913745
Conc: 95.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 656732962
Conc: 99.44 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.442 min
Delta R.T.: 0.000 min
Response: 165740658
Conc: 94.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.807 min
Delta R.T.: 0.000 min
Response: 1209469094
Conc: 99.43 ng/ml



#41 AR-1268-1

R.T.: 8.890 min
Delta R.T.: 0.000 min
Response: 152883178
Conc: 949.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000



#41 AR-1268-1

R.T.: 7.690 min
Delta R.T.: 0.000 min
Response: 1460168699
Conc: 967.15 ng/ml



#42 AR-1268-2

R.T.: 8.986 min
Delta R.T.: 0.000 min
Response: 141133811
Conc: 950.53 ng/ml



#42 AR-1268-2

R.T.: 7.754 min
Delta R.T.: 0.000 min
Response: 1471677824
Conc: 983.69 ng/ml



#43 AR-1268-3

R.T.: 9.226 min
Delta R.T.: 0.000 min
Response: 118887581
Conc: 951.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000



#43 AR-1268-3

R.T.: 7.961 min
Delta R.T.: 0.000 min
Response: 1142468076
Conc: 977.31 ng/ml



#44 AR-1268-4

R.T.: 9.655 min
Delta R.T.: 0.000 min
Response: 55874289
Conc: 948.21 ng/ml



#44 AR-1268-4

R.T.: 8.247 min
Delta R.T.: 0.000 min
Response: 470517486
Conc: 987.00 ng/ml



#45 AR-1268-5

R.T.: 10.086 min
Delta R.T.: 0.000 min
Response: 373677572
Conc: 957.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000



#45 AR-1268-5

R.T.: 8.545 min
Delta R.T.: 0.000 min
Response: 3043336629
Conc: 962.83 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075271.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:25
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:05:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:59:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.660	3.799	100.1E6	497.2E6	74.976	75.187
2)	SA Decachlor...	10.442	8.807	130.2E6	914.1E6	74.643	75.099
	Target Compounds						
41)	L9 AR-1268-1	8.892	7.690	119.8E6	1135.3E6	745.861	751.304
42)	L9 AR-1268-2	8.988	7.754	110.2E6	1125.2E6	744.931	751.391
43)	L9 AR-1268-3	9.228	7.962	92608920	865.0E6	743.998	743.275
44)	L9 AR-1268-4	9.656	8.248	43848213	354.7E6	746.070	746.016
45)	L9 AR-1268-5	10.088	8.546	291.3E6	2388.6E6	747.392	753.776

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075271.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:25
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:05:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:59:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 100066687
Conc: 74.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 497158030
Conc: 75.19 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.442 min
Delta R.T.: 0.000 min
Response: 130175685
Conc: 74.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.807 min
Delta R.T.: 0.000 min
Response: 914090698
Conc: 75.10 ng/ml



#41 AR-1268-1

R.T.: 8.892 min
Delta R.T.: 0.000 min
Response: 119769528
Conc: 745.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750



#41 AR-1268-1

R.T.: 7.690 min
Delta R.T.: 0.000 min
Response: 1135274856
Conc: 751.30 ng/ml



#42 AR-1268-2

R.T.: 8.988 min
Delta R.T.: 0.000 min
Response: 110234639
Conc: 744.93 ng/ml



#42 AR-1268-2

R.T.: 7.754 min
Delta R.T.: 0.000 min
Response: 1125183792
Conc: 751.39 ng/ml



#43 AR-1268-3

R.T.: 9.228 min
Delta R.T.: 0.000 min
Response: 92608920
Conc: 744.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750



#43 AR-1268-3

R.T.: 7.962 min
Delta R.T.: 0.000 min
Response: 865001509
Conc: 743.27 ng/ml



#44 AR-1268-4

R.T.: 9.656 min
Delta R.T.: 0.000 min
Response: 43848213
Conc: 746.07 ng/ml



#44 AR-1268-4

R.T.: 8.248 min
Delta R.T.: 0.000 min
Response: 354693309
Conc: 746.02 ng/ml



#45 AR-1268-5

R.T.: 10.088 min
Delta R.T.: 0.000 min
Response: 291275954
Conc: 747.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750



#45 AR-1268-5

R.T.: 8.546 min
Delta R.T.: 0.000 min
Response: 2388574998
Conc: 753.78 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:41
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:00:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:59:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.660	3.799	69528834	332.0E6	50.000	50.000
2)	SA Decachlor...	10.443	8.806	91941579	611.7E6	50.000	50.000
	Target Compounds						
41)	L9 AR-1268-1	8.891	7.689	84580336	779.7E6	500.000	500.000
42)	L9 AR-1268-2	8.987	7.753	77912875	760.2E6	500.000	500.000
43)	L9 AR-1268-3	9.228	7.960	65529058	597.8E6	500.000	500.000
44)	L9 AR-1268-4	9.655	8.246	30989057	241.5E6	500.000	500.000
45)	L9 AR-1268-5	10.087	8.545	203.6E6	1639.2E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:41
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:00:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:59:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 69528834
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 332036100
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.443 min
Delta R.T.: 0.000 min
Response: 91941579
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 611654619
Conc: 50.00 ng/ml



#41 AR-1268-1

R.T.: 8.891 min
Delta R.T.: 0.000 min
Response: 84580336
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500



#41 AR-1268-1

R.T.: 7.689 min
Delta R.T.: 0.000 min
Response: 779674117
Conc: 500.00 ng/ml



#42 AR-1268-2

R.T.: 8.987 min
Delta R.T.: 0.000 min
Response: 77912875
Conc: 500.00 ng/ml



#42 AR-1268-2

R.T.: 7.753 min
Delta R.T.: 0.000 min
Response: 760240815
Conc: 500.00 ng/ml



#43 AR-1268-3

R.T.: 9.228 min
Delta R.T.: 0.000 min
Response: 65529058
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500



#43 AR-1268-3

R.T.: 7.960 min
Delta R.T.: 0.000 min
Response: 597755130
Conc: 500.00 ng/ml



#44 AR-1268-4

R.T.: 9.655 min
Delta R.T.: 0.000 min
Response: 30989057
Conc: 500.00 ng/ml



#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 241454470
Conc: 500.00 ng/ml



#45 AR-1268-5

R.T.: 10.087 min
Delta R.T.: 0.000 min
Response: 203561599
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500



#45 AR-1268-5

R.T.: 8.545 min
Delta R.T.: 0.000 min
Response: 1639165491
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075273.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:58
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:12:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:12:18 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.658	3.796	34379611	151.4E6	25.565	23.393
2)	SA Decachlor...	10.439	8.804	46278136	303.1E6	26.135	24.927
	Target Compounds						
41)	L9 AR-1268-1	8.889	7.687	42980575	365.2E6	263.015	243.698
42)	L9 AR-1268-2	8.984	7.751	39128437	345.6E6	260.660	235.330
43)	L9 AR-1268-3	9.225	7.958	32512368	273.5E6	258.304	238.576
44)	L9 AR-1268-4	9.653	8.245	15247313	115.2E6	257.007	244.227
45)	L9 AR-1268-5	10.085	8.542	101.3E6	821.3E6	257.365	256.819

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075273.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:58
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:12:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:12:18 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.658 min
Delta R.T.: 0.000 min
Response: 34379611
Conc: 25.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250



#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 151432939
Conc: 23.39 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.439 min
Delta R.T.: 0.000 min
Response: 46278136
Conc: 26.13 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.804 min
Delta R.T.: 0.000 min
Response: 303111015
Conc: 24.93 ng/ml



#41 AR-1268-1

R.T.: 8.889 min
Delta R.T.: 0.000 min
Response: 42980575
Conc: 263.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250



#41 AR-1268-1

R.T.: 7.687 min
Delta R.T.: 0.000 min
Response: 365177329
Conc: 243.70 ng/ml



#42 AR-1268-2

R.T.: 8.984 min
Delta R.T.: 0.000 min
Response: 39128437
Conc: 260.66 ng/ml



#42 AR-1268-2

R.T.: 7.751 min
Delta R.T.: 0.000 min
Response: 345638504
Conc: 235.33 ng/ml



#43 AR-1268-3

R.T.: 9.225 min
Delta R.T.: 0.000 min
Response: 32512368
Conc: 258.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250



#43 AR-1268-3

R.T.: 7.958 min
Delta R.T.: 0.000 min
Response: 273482322
Conc: 238.58 ng/ml



#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: 0.000 min
Response: 15247313
Conc: 257.01 ng/ml



#44 AR-1268-4

R.T.: 8.245 min
Delta R.T.: 0.000 min
Response: 115230714
Conc: 244.23 ng/ml



#45 AR-1268-5

R.T.: 10.085 min
Delta R.T.: 0.000 min
Response: 101295864
Conc: 257.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.542 min
Delta R.T.: 0.000 min
Response: 821277670
Conc: 256.82 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075274.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 17:14
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:35:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:35:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.656	3.796	6162995	20516394	4.661	3.375m#
2)	SA Decachlor...	10.438	8.806	8054661	48956522	4.632	4.189
	Target Compounds						
41)	L9 AR-1268-1	8.887	7.689	7806652	60962197	48.257m	42.258
42)	L9 AR-1268-2	8.984	7.753	7067403	54955257	47.637	39.400
43)	L9 AR-1268-3	9.224	7.960	5848482	43468397	47.131	39.653m
44)	L9 AR-1268-4	9.653	8.247	3041493	18125694	51.009	40.754m
45)	L9 AR-1268-5	10.084	8.544	18331022	120.0E6	47.221	39.458m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075274.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 17:14
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:35:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:35:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.656 min
Delta R.T.: 0.000 min
Response: 6162995
Conc: 4.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: -0.001 min
Response: 20516394
Conc: 3.38 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.438 min
Delta R.T.: 0.000 min
Response: 8054661
Conc: 4.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 48956522
Conc: 4.19 ng/ml



#41 AR-1268-1

R.T.: 8.887 min
Delta R.T.: 0.000 min
Response: 7806652
Conc: 48.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#41 AR-1268-1

R.T.: 7.689 min
Delta R.T.: 0.000 min
Response: 60962197
Conc: 42.26 ng/ml



#42 AR-1268-2

R.T.: 8.984 min
Delta R.T.: 0.000 min
Response: 7067403
Conc: 47.64 ng/ml



#42 AR-1268-2

R.T.: 7.753 min
Delta R.T.: 0.000 min
Response: 54955257
Conc: 39.40 ng/ml



#43 AR-1268-3

R.T.: 9.224 min
Delta R.T.: 0.000 min
Response: 5848482
Conc: 47.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#43 AR-1268-3

R.T.: 7.960 min
Delta R.T.: 0.000 min
Response: 43468397
Conc: 39.65 ng/ml m



#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: 0.000 min
Response: 3041493
Conc: 51.01 ng/ml



#44 AR-1268-4

R.T.: 8.247 min
Delta R.T.: 0.000 min
Response: 18125694
Conc: 40.75 ng/ml m



#45 AR-1268-5

R.T.: 10.084 min
Delta R.T.: 0.000 min
Response: 18331022
Conc: 47.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#45 AR-1268-5

R.T.: 8.544 min
Delta R.T.: 0.000 min
Response: 119998424
Conc: 39.46 ng/ml m

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075275.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 17:46
Operator : YP\AJ
Sample : PP092325ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092325

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 18:28:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.663	3.799	70190259	330.4E6	53.643	54.154
2)	SA Decachlor...	10.445	8.804	53037838	358.3E6	53.168	59.392m
	Target Compounds						
3)	L1 AR-1016-1	5.815	4.896	25480532	304.2E6	529.688	542.185
4)	L1 AR-1016-2	5.836	4.953	36891066	135.1E6	514.315	525.487
5)	L1 AR-1016-3	5.899	5.073	23781479	76410296	511.091	492.118
6)	L1 AR-1016-4	5.996	5.114	19445011	76331861	545.646	519.377
7)	L1 AR-1016-5	6.289	5.328	18500980	85017958	548.644	520.362
31)	L7 AR-1260-1	7.406	6.357	33314276	213.7E6	554.134	526.554
32)	L7 AR-1260-2	7.659	6.543	39216271	298.7E6	538.700	518.643
33)	L7 AR-1260-3	8.017	6.697	31022586	223.0E6	572.952	520.892
34)	L7 AR-1260-4	8.245	7.166	33773955	216.2E6	597.376	501.440
35)	L7 AR-1260-5	8.571	7.405	61446307	573.0E6	559.578	544.665

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075275.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 17:46
Operator : YP\AJ
Sample : PP092325ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092325

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 18:28:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.663 min
Delta R.T.: 0.008 min
Response: 70190259
Conc: 53.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.003 min
Response: 330389764
Conc: 54.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.445 min
Delta R.T.: 0.010 min
Response: 53037838
Conc: 53.17 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.804 min
Delta R.T.: 0.000 min
Response: 358284034
Conc: 59.39 ng/ml m



#3 AR-1016-1

R.T.: 5.815 min
Delta R.T.: 0.007 min
Response: 25480532
Conc: 529.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#3 AR-1016-1

R.T.: 4.896 min
Delta R.T.: 0.003 min
Response: 304173227
Conc: 542.19 ng/ml



#4 AR-1016-2

R.T.: 5.836 min
Delta R.T.: 0.008 min
Response: 36891066
Conc: 514.31 ng/ml



#4 AR-1016-2

R.T.: 4.953 min
Delta R.T.: 0.002 min
Response: 135051180
Conc: 525.49 ng/ml



#5 AR-1016-3

R.T.: 5.899 min
Delta R.T.: 0.007 min
Response: 23781479
Conc: 511.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#5 AR-1016-3

R.T.: 5.073 min
Delta R.T.: 0.002 min
Response: 76410296
Conc: 492.12 ng/ml



#6 AR-1016-4

R.T.: 5.996 min
Delta R.T.: 0.007 min
Response: 19445011
Conc: 545.65 ng/ml



#6 AR-1016-4

R.T.: 5.114 min
Delta R.T.: 0.003 min
Response: 76331861
Conc: 519.38 ng/ml



#7 AR-1016-5

R.T.: 6.289 min
Delta R.T.: 0.007 min
Response: 18500980
Conc: 548.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#7 AR-1016-5

R.T.: 5.328 min
Delta R.T.: 0.003 min
Response: 85017958
Conc: 520.36 ng/ml



#31 AR-1260-1

R.T.: 7.406 min
Delta R.T.: 0.007 min
Response: 33314276
Conc: 554.13 ng/ml



#31 AR-1260-1

R.T.: 6.357 min
Delta R.T.: 0.002 min
Response: 213746874
Conc: 526.55 ng/ml



#32 AR-1260-2

R.T.: 7.659 min
Delta R.T.: 0.007 min
Response: 39216271
Conc: 538.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#32 AR-1260-2

R.T.: 6.543 min
Delta R.T.: 0.002 min
Response: 298650131
Conc: 518.64 ng/ml



#33 AR-1260-3

R.T.: 8.017 min
Delta R.T.: 0.007 min
Response: 31022586
Conc: 572.95 ng/ml



#33 AR-1260-3

R.T.: 6.697 min
Delta R.T.: 0.002 min
Response: 222985953
Conc: 520.89 ng/ml



#34 AR-1260-4

R.T.: 8.245 min
Delta R.T.: 0.008 min
Response: 33773955
Conc: 597.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#34 AR-1260-4

R.T.: 7.166 min
Delta R.T.: 0.000 min
Response: 216232762
Conc: 501.44 ng/ml



#35 AR-1260-5

R.T.: 8.571 min
Delta R.T.: 0.008 min
Response: 61446307
Conc: 559.58 ng/ml



#35 AR-1260-5

R.T.: 7.405 min
Delta R.T.: 0.002 min
Response: 572980909
Conc: 544.66 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 18:35
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 04:35:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.660	3.799	69296262	322.8E6	52.959	52.905
2)	SA Decachlor...	10.440	8.803	52620761	359.1E6	52.750	59.525m
	Target Compounds						
16)	L4 AR-1242-1	5.812	4.896	21512826	253.0E6	516.394	541.693
17)	L4 AR-1242-2	5.834	4.954	31427273	115.0E6	526.973	550.426
18)	L4 AR-1242-3	5.897	5.074	20323849	64997022	502.739	532.921
19)	L4 AR-1242-4	5.993	5.155	16403438	83515061	521.837	524.950m
20)	L4 AR-1242-5	6.723	5.676	18943732	103.0E6	538.853	528.983

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 18:35
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 04:35:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.005 min
Response: 69296262
Conc: 52.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.003 min
Response: 322766613
Conc: 52.90 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.440 min
Delta R.T.: 0.006 min
Response: 52620761
Conc: 52.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 359083335
Conc: 59.52 ng/ml m



#16 AR-1242-1

R.T.: 5.812 min
Delta R.T.: 0.000 min
Response: 21512826
Conc: 516.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#16 AR-1242-1

R.T.: 4.896 min
Delta R.T.: 0.000 min
Response: 253035792
Conc: 541.69 ng/ml



#17 AR-1242-2

R.T.: 5.834 min
Delta R.T.: 0.000 min
Response: 31427273
Conc: 526.97 ng/ml



#17 AR-1242-2

R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 115015732
Conc: 550.43 ng/ml



#18 AR-1242-3

R.T.: 5.897 min
Delta R.T.: 0.000 min
Response: 20323849
Conc: 502.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#18 AR-1242-3

R.T.: 5.074 min
Delta R.T.: 0.000 min
Response: 64997022
Conc: 532.92 ng/ml



#19 AR-1242-4

R.T.: 5.993 min
Delta R.T.: 0.000 min
Response: 16403438
Conc: 521.84 ng/ml



#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 83515061
Conc: 524.95 ng/ml m



#20 AR-1242-5

R.T.: 6.723 min
Delta R.T.: 0.000 min
Response: 18943732
Conc: 538.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#20 AR-1242-5

R.T.: 5.676 min
Delta R.T.: 0.000 min
Response: 102984369
Conc: 528.98 ng/ml

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075277.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 19:08
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 04:30:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.660	3.798	67690061	321.9E6	51.732	52.757
2)	SA Decachlor...	10.441	8.804	52715950	361.4E6	52.846	59.916m
	Target Compounds						
21)	L5 AR-1248-1	5.811	4.894	16390127	167.0E6	519.336	555.702
22)	L5 AR-1248-2	6.082	5.114	22503142	105.0E6	518.888	518.581
23)	L5 AR-1248-3	6.285	5.154	25155093	135.3E6	526.712	541.550m
24)	L5 AR-1248-4	6.683	5.327	30695144	119.4E6	519.272	513.196
25)	L5 AR-1248-5	6.723	5.718	29807919	227.8E6	504.369	557.543

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075277.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 19:08
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 04:30:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.005 min
Response: 67690061
Conc: 51.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.798 min
Delta R.T.: 0.002 min
Response: 321866586
Conc: 52.76 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.441 min
Delta R.T.: 0.006 min
Response: 52715950
Conc: 52.85 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.804 min
Delta R.T.: 0.000 min
Response: 361445303
Conc: 59.92 ng/ml m



#21 AR-1248-1

R.T.: 5.811 min
Delta R.T.: 0.005 min
Response: 16390127
Conc: 519.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#21 AR-1248-1

R.T.: 4.894 min
Delta R.T.: 0.002 min
Response: 166969869
Conc: 555.70 ng/ml



#22 AR-1248-2

R.T.: 6.082 min
Delta R.T.: 0.004 min
Response: 22503142
Conc: 518.89 ng/ml



#22 AR-1248-2

R.T.: 5.114 min
Delta R.T.: 0.001 min
Response: 105001722
Conc: 518.58 ng/ml



#23 AR-1248-3

R.T.: 6.285 min
Delta R.T.: 0.004 min
Response: 25155093
Conc: 526.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#23 AR-1248-3

R.T.: 5.154 min
Delta R.T.: 0.000 min
Response: 135250085
Conc: 541.55 ng/ml m



#24 AR-1248-4

R.T.: 6.683 min
Delta R.T.: 0.005 min
Response: 30695144
Conc: 519.27 ng/ml



#24 AR-1248-4

R.T.: 5.327 min
Delta R.T.: 0.001 min
Response: 119442983
Conc: 513.20 ng/ml



#25 AR-1248-5

R.T.: 6.723 min
Delta R.T.: 0.005 min
Response: 29807919
Conc: 504.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#25 AR-1248-5

R.T.: 5.718 min
Delta R.T.: 0.002 min
Response: 227795946
Conc: 557.54 ng/ml

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 19:40
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1254

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 05:12:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 05:07:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.660	3.799	68095844	321.3E6	52.042	52.664
2)	SA Decachlor...	10.441	8.804	51811352	358.5E6	51.939	59.429m
	Target Compounds						
26)	L6 AR-1254-1	6.659	5.678	28467743	276.6E6	443.257m	554.578 #
27)	L6 AR-1254-2	6.876	5.826	43938129	209.1E6	511.302	523.737
28)	L6 AR-1254-3	7.238	6.227	46807513	363.2E6	506.453	546.317
29)	L6 AR-1254-4	7.520	6.454	36595686	272.7E6	512.909	552.441
30)	L6 AR-1254-5	7.936	6.870	43699326	290.3E6	534.174	508.619

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 19:40
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1254

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 05:12:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 05:07:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.005 min
Response: 68095844
Conc: 52.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1254

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.003 min
Response: 321296251
Conc: 52.66 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.441 min
Delta R.T.: 0.006 min
Response: 51811352
Conc: 51.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.804 min
Delta R.T.: 0.000 min
Response: 358504884
Conc: 59.43 ng/ml m



#26 AR-1254-1

R.T.: 6.659 min
Delta R.T.: 0.000 min
Response: 28467743
Conc: 443.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1254

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#26 AR-1254-1

R.T.: 5.678 min
Delta R.T.: -0.001 min
Response: 276624835
Conc: 554.58 ng/ml



#27 AR-1254-2

R.T.: 6.876 min
Delta R.T.: 0.000 min
Response: 43938129
Conc: 511.30 ng/ml



#27 AR-1254-2

R.T.: 5.826 min
Delta R.T.: 0.000 min
Response: 209067406
Conc: 523.74 ng/ml



#28 AR-1254-3

R.T.: 7.238 min
Delta R.T.: 0.000 min
Response: 46807513
Conc: 506.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1254

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#28 AR-1254-3

R.T.: 6.227 min
Delta R.T.: -0.001 min
Response: 363150092
Conc: 546.32 ng/ml



#29 AR-1254-4

R.T.: 7.520 min
Delta R.T.: 0.000 min
Response: 36595686
Conc: 512.91 ng/ml



#29 AR-1254-4

R.T.: 6.454 min
Delta R.T.: 0.000 min
Response: 272711111
Conc: 552.44 ng/ml



#30 AR-1254-5

R.T.: 7.936 min
Delta R.T.: 0.000 min
Response: 43699326
Conc: 534.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1254

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#30 AR-1254-5

R.T.: 6.870 min
Delta R.T.: 0.000 min
Response: 290265404
Conc: 508.62 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075279.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 19:57
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1268

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 04:49:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 04:47:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.659	3.799	71353406	332.9E6	53.960	55.485
2)	SA Decachlor...	10.440	8.804	92871605	696.3E6	53.411	59.581m
	Target Compounds						
41)	L9 AR-1268-1	8.890	7.689	89479292	799.5E6	552.483	554.201
42)	L9 AR-1268-2	8.985	7.753	81657272	782.2E6	550.399	560.767
43)	L9 AR-1268-3	9.226	7.960	67147736	596.0E6	541.127	546.351
44)	L9 AR-1268-4	9.653	8.246	31706205	236.0E6	531.741	524.416
45)	L9 AR-1268-5	10.085	8.544	206.8E6	1813.0E6	532.693	596.714

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075279.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 19:57
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1268

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 04:49:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 04:47:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 71353406
Conc: 53.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1268

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 332881910
Conc: 55.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.440 min
Delta R.T.: -0.003 min
Response: 92871605
Conc: 53.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.804 min
Delta R.T.: -0.003 min
Response: 696275481
Conc: 59.58 ng/ml m



#41 AR-1268-1

R.T.: 8.890 min
Delta R.T.: -0.001 min
Response: 89479292
Conc: 552.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1268

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#41 AR-1268-1

R.T.: 7.689 min
Delta R.T.: 0.000 min
Response: 799509513
Conc: 554.20 ng/ml



#42 AR-1268-2

R.T.: 8.985 min
Delta R.T.: -0.002 min
Response: 81657272
Conc: 550.40 ng/ml



#42 AR-1268-2

R.T.: 7.753 min
Delta R.T.: 0.000 min
Response: 782164395
Conc: 560.77 ng/ml



#43 AR-1268-3

R.T.: 9.226 min
Delta R.T.: -0.002 min
Response: 67147736
Conc: 541.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1268

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#43 AR-1268-3

R.T.: 7.960 min
Delta R.T.: 0.000 min
Response: 596026528
Conc: 546.35 ng/ml



#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml



#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 235964707
Conc: 524.42 ng/ml



#45 AR-1268-5

R.T.: 10.085 min
Delta R.T.: -0.002 min
Response: 206788671
Conc: 532.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1268

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#45 AR-1268-5

R.T.: 8.544 min
Delta R.T.: 0.000 min
Response: 1812996598
Conc: 596.71 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
Instrument ID: ECD_Q **Calibration Date(s):** 10/10/2025 10/10/2025
Calibration Times: 08:57 16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PQ070973.D RT 750 = PQ070974.D
RT 500 = PQ070975.D RT 250 = PQ070976.D RT 050 = PQ070977.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	4.48	4.48	4.48	4.48	4.48	4.48	4.38	4.58
Aroclor-1016-2	(2)	4.50	4.50	4.50	4.50	4.50	4.50	4.40	4.60
Aroclor-1016-3	(3)	4.56	4.56	4.56	4.56	4.56	4.56	4.46	4.66
Aroclor-1016-4	(4)	4.65	4.65	4.65	4.65	4.65	4.65	4.55	4.75
Aroclor-1016-5	(5)	4.93	4.93	4.93	4.93	4.93	4.93	4.83	5.03
Aroclor-1260-1	(1)	6.02	6.02	6.02	6.02	6.02	6.02	5.92	6.12
Aroclor-1260-2	(2)	6.28	6.28	6.28	6.28	6.27	6.28	6.18	6.38
Aroclor-1260-3	(3)	6.63	6.63	6.63	6.63	6.63	6.63	6.53	6.73
Aroclor-1260-4	(4)	6.84	6.84	6.84	6.84	6.84	6.84	6.74	6.94
Aroclor-1260-5	(5)	7.15	7.15	7.15	7.15	7.15	7.15	7.05	7.25
Decachlorobiphenyl		8.50	8.50	8.50	8.50	8.50	8.50	8.40	8.60
Tetrachloro-m-xylene		3.40	3.40	3.40	3.40	3.40	3.40	3.30	3.50
Aroclor-1242-1	(1)	4.48	4.48	4.48	4.48	4.48	4.48	4.38	4.58
Aroclor-1242-2	(2)	4.50	4.50	4.50	4.50	4.50	4.50	4.40	4.60
Aroclor-1242-3	(3)	4.56	4.56	4.56	4.56	4.56	4.56	4.46	4.66
Aroclor-1242-4	(4)	4.65	4.65	4.65	4.65	4.65	4.65	4.55	4.75
Aroclor-1242-5	(5)	5.35	5.35	5.35	5.35	5.32	5.34	5.24	5.44
Decachlorobiphenyl		8.50	8.50	8.50	8.50	8.50	8.50	8.40	8.60
Tetrachloro-m-xylene		3.40	3.40	3.40	3.40	3.40	3.40	3.30	3.50
Aroclor-1248-1	(1)	4.48	4.48	4.48	4.48	4.48	4.48	4.38	4.58
Aroclor-1248-2	(2)	4.74	4.74	4.74	4.74	4.74	4.74	4.64	4.84
Aroclor-1248-3	(3)	4.93	4.93	4.93	4.93	4.93	4.93	4.83	5.03
Aroclor-1248-4	(4)	5.32	5.32	5.32	5.32	5.32	5.32	5.22	5.42
Aroclor-1248-5	(5)	5.35	5.35	5.35	5.35	5.35	5.35	5.25	5.45
Decachlorobiphenyl		8.50	8.50	8.50	8.50	8.50	8.50	8.40	8.60
Tetrachloro-m-xylene		3.40	3.40	3.40	3.40	3.40	3.40	3.30	3.50
Aroclor-1254-1	(1)	5.29	5.29	5.29	5.29	5.29	5.29	5.19	5.39
Aroclor-1254-2	(2)	5.51	5.51	5.51	5.51	5.51	5.51	5.41	5.61
Aroclor-1254-3	(3)	5.86	5.85	5.86	5.86	5.86	5.86	5.76	5.96
Aroclor-1254-4	(4)	6.14	6.13	6.14	6.14	6.14	6.14	6.04	6.24
Aroclor-1254-5	(5)	6.54	6.54	6.54	6.54	6.54	6.54	6.44	6.64
Decachlorobiphenyl		8.50	8.50	8.50	8.50	8.50	8.50	8.40	8.60
Tetrachloro-m-xylene		3.40	3.40	3.40	3.40	3.40	3.40	3.30	3.50
Aroclor-1268-1	(1)	7.42	7.42	7.42	7.42	7.42	7.42	7.32	7.52
Aroclor-1268-2	(2)	7.49	7.50	7.50	7.50	7.50	7.50	7.40	7.60
Aroclor-1268-3	(3)	7.68	7.68	7.68	7.68	7.68	7.68	7.58	7.78
Aroclor-1268-4	(4)	7.99	7.99	7.99	7.99	7.99	7.99	7.89	8.09
Aroclor-1268-5	(5)	8.27	8.27	8.27	8.27	8.27	8.27	8.17	8.37

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.50	8.50	8.50	8.50	8.50	8.50	8.40	8.60
Tetrachloro-m-xylene	3.40	3.40	3.40	3.40	3.40	3.40	3.30	3.50

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
Instrument ID: ECD_Q **Calibration Date(s):** 10/10/2025 10/10/2025
Calibration Times: 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PQ070973.D RT 750 = PQ070974.D
RT 500 = PQ070975.D RT 250 = PQ070976.D RT 050 = PQ070977.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	3.78	3.78	3.78	3.78	3.78	3.78	3.68	3.88
Aroclor-1016-2	(2)	3.79	3.79	3.79	3.79	3.79	3.79	3.69	3.89
Aroclor-1016-3	(3)	3.95	3.95	3.95	3.95	3.95	3.95	3.85	4.05
Aroclor-1016-4	(4)	4.00	4.00	4.00	4.00	4.00	4.00	3.90	4.10
Aroclor-1016-5	(5)	4.20	4.20	4.20	4.20	4.20	4.20	4.10	4.30
Aroclor-1260-1	(1)	5.19	5.19	5.19	5.19	5.19	5.19	5.09	5.29
Aroclor-1260-2	(2)	5.38	5.38	5.38	5.38	5.38	5.38	5.28	5.48
Aroclor-1260-3	(3)	5.52	5.52	5.52	5.52	5.52	5.52	5.42	5.62
Aroclor-1260-4	(4)	5.98	5.98	5.98	5.98	5.98	5.98	5.88	6.08
Aroclor-1260-5	(5)	6.22	6.22	6.22	6.22	6.22	6.22	6.12	6.32
Decachlorobiphenyl		7.53	7.53	7.53	7.53	7.53	7.53	7.43	7.63
Tetrachloro-m-xylene		2.78	2.78	2.78	2.78	2.78	2.78	2.68	2.88
Aroclor-1242-1	(1)	3.78	3.78	3.78	3.78	3.78	3.78	3.68	3.88
Aroclor-1242-2	(2)	3.79	3.79	3.79	3.79	3.79	3.79	3.69	3.89
Aroclor-1242-3	(3)	3.95	3.95	3.95	3.95	3.95	3.95	3.85	4.05
Aroclor-1242-4	(4)	4.04	4.04	4.04	4.04	4.04	4.04	3.94	4.14
Aroclor-1242-5	(5)	4.53	4.53	4.53	4.53	4.53	4.53	4.43	4.63
Decachlorobiphenyl		7.53	7.53	7.53	7.53	7.53	7.53	7.43	7.63
Tetrachloro-m-xylene		2.78	2.78	2.78	2.78	2.78	2.78	2.68	2.88
Aroclor-1248-1	(1)	3.78	3.78	3.78	3.78	3.78	3.78	3.68	3.88
Aroclor-1248-2	(2)	4.00	4.00	4.00	4.00	4.00	4.00	3.90	4.10
Aroclor-1248-3	(3)	4.04	4.04	4.04	4.04	4.04	4.04	3.94	4.14
Aroclor-1248-4	(4)	4.20	4.20	4.20	4.20	4.20	4.20	4.10	4.30
Aroclor-1248-5	(5)	4.57	4.57	4.57	4.57	4.57	4.57	4.47	4.67
Decachlorobiphenyl		7.53	7.53	7.53	7.53	7.53	7.53	7.43	7.63
Tetrachloro-m-xylene		2.78	2.78	2.78	2.78	2.78	2.78	2.68	2.88
Aroclor-1254-1	(1)	4.53	4.53	4.53	4.53	4.53	4.53	4.43	4.63
Aroclor-1254-2	(2)	4.68	4.68	4.68	4.68	4.68	4.68	4.58	4.78
Aroclor-1254-3	(3)	5.06	5.06	5.06	5.06	5.06	5.06	4.96	5.16
Aroclor-1254-4	(4)	5.29	5.29	5.29	5.29	5.29	5.29	5.19	5.39
Aroclor-1254-5	(5)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Decachlorobiphenyl		7.53	7.53	7.53	7.53	7.53	7.53	7.43	7.63
Tetrachloro-m-xylene		2.78	2.78	2.78	2.78	2.78	2.78	2.68	2.88
Aroclor-1268-1	(1)	6.50	6.50	6.50	6.50	6.50	6.50	6.40	6.60
Aroclor-1268-2	(2)	6.56	6.56	6.56	6.56	6.56	6.56	6.46	6.66
Aroclor-1268-3	(3)	6.76	6.76	6.76	6.76	6.76	6.76	6.66	6.86
Aroclor-1268-4	(4)	7.05	7.05	7.05	7.05	7.05	7.05	6.95	7.15
Aroclor-1268-5	(5)	7.32	7.32	7.32	7.32	7.32	7.32	7.22	7.42

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	7.53	7.53	7.53	7.53	7.53	7.53	7.43	7.63
Tetrachloro-m-xylene	2.78	2.78	2.78	2.78	2.78	2.78	2.68	2.88

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_Q
Contract: AECO02
SDG NO.: Q3375
Calibration Date(s): 10/10/2025 10/10/2025
Calibration Times: 08:57 16:43

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID: CF 1000 = PQ070973.D CF 750 = PQ070974.D
CF 500 = PQ070975.D CF 250 = PQ070976.D CF 050 = PQ070977.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	152632258	161381460	169566884	175492100	161647220	164143984	5
Aroclor-1016-2	(2)	235274296	256838307	264605432	271518004	253235720	256294352	5
Aroclor-1016-3	(3)	145675938	156782704	164416996	169596776	159119400	159118363	6
Aroclor-1016-4	(4)	120162951	129378400	134848252	138303760	126953020	129929277	5
Aroclor-1016-5	(5)	117231883	128545740	133336348	136306768	128593380	128802824	6
Aroclor-1260-1	(1)	208928914	234195365	241877518	247859876	239493660	234471067	6
Aroclor-1260-2	(2)	243695794	275266251	287710272	296720404	279652200	276608984	7
Aroclor-1260-3	(3)	178017013	196424459	204653606	209590628	207855920	199308325	6
Aroclor-1260-4	(4)	191556315	214263107	223690164	228814336	216092480	214883280	7
Aroclor-1260-5	(5)	400986941	436898331	447507618	451939420	431553500	433777162	5
Decachlorobiphenyl		2603997920	2888430733	3017037660	3097201720	2913248400	2903983287	6
Tetrachloro-m-xylene		4048991050	4420325093	4447782420	4497118040	4337178200	4350278961	4
Aroclor-1242-1	(1)	117929900	120501580	121917368	124177888	110233240	118951995	5
Aroclor-1242-2	(2)	179238730	188762288	194471730	197719720	193371880	190712870	4
Aroclor-1242-3	(3)	111298009	117265485	120728326	123754708	121434700	118896246	4
Aroclor-1242-4	(4)	91702396	97376287	98944784	101196952	108558560	99555796	6
Aroclor-1242-5	(5)	63322787	67097277	82718820	102665812	77804800	78721899	20
Decachlorobiphenyl		3082568640	3243406933	3352805760	3359769360	3286790800	3265068299	3
Tetrachloro-m-xylene		4441893010	4577290613	4623065280	4605855840	4702516200	4590124189	2
Aroclor-1248-1	(1)	89951061	92802913	99278242	97399904	67891380	89464700	14
Aroclor-1248-2	(2)	126849870	131764987	138900528	142257364	142462060	136446962	5
Aroclor-1248-3	(3)	152980867	157544657	166146850	167481096	162764480	161383590	4
Aroclor-1248-4	(4)	167077290	173126060	182827440	186713196	157721500	173493097	7
Aroclor-1248-5	(5)	163667698	168136284	132211082	121389384	159185140	148917918	14
Decachlorobiphenyl		3177348440	3248518907	3402766540	3434790520	3353916400	3323468161	3
Tetrachloro-m-xylene		4546508980	4599154480	4739413180	4685828360	4518221600	4617825320	2
Aroclor-1254-1	(1)	186317660	198212097	200308250	222421948	174131700	196278331	9
Aroclor-1254-2	(2)	272465681	301191412	300558232	318683640	259516480	290483089	8
Aroclor-1254-3	(3)	294493285	320137152	318530722	335107472	267769020	307207530	9
Aroclor-1254-4	(4)	212636724	232647637	229798130	237628972	184991000	219540493	10
Aroclor-1254-5	(5)	229095618	241104957	242576814	249415576	197285560	231895705	9
Decachlorobiphenyl		3135993520	3237057933	3300114540	3432442200	2724290800	3165979799	9
Tetrachloro-m-xylene		4449551480	4584042107	4604497820	4625575480	3701892800	4393111937	9
Aroclor-1268-1	(1)	555593291	572438855	583365466	598015136	603747960	582632142	3

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	510325576	524057337	534075228	546100468	539884140	530888550	3
Aroclor-1268-3	(3)	416538407	427532203	436646098	443325192	442459640	433300308	3
Aroclor-1268-4	(4)	179373252	185531188	189624920	193669728	190155160	187670850	3
Aroclor-1268-5	(5)	1283167757	1308010297	1321380362	1326514116	1318957680	1311606042	1
Decachlorobiphenyl		5757514290	5913646120	6037365900	6192444040	6223042400	6024802550	3
Tetrachloro-m-xylene		4734874420	4871269707	4916225920	4964521920	4932484000	4883875193	2

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_Q
Contract: AECO02
SDG NO.: Q3375
Calibration Date(s): 10/10/2025 10/10/2025
Calibration Times: 08:57 16:43

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID: CF 1000 = PQ070973.D CF 750 = PQ070974.D
CF 500 = PQ070975.D CF 250 = PQ070976.D CF 050 = PQ070977.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	198094739	217218485	228189888	237783008	238723760	224001976	8
Aroclor-1016-2	(2)	306974778	336928535	345611618	359075648	352024660	340123048	6
Aroclor-1016-3	(3)	164197912	180535471	187035142	195263012	182262860	181858879	6
Aroclor-1016-4	(4)	139678855	154088441	161162796	170890484	161242760	157412667	7
Aroclor-1016-5	(5)	171310856	187991799	196963118	205633060	197310680	191841903	7
Aroclor-1260-1	(1)	322104344	352779140	365742242	382158164	379245720	360405922	7
Aroclor-1260-2	(2)	392921251	432078265	450339936	466790948	457838380	439993756	7
Aroclor-1260-3	(3)	374441100	409997701	423595070	436050024	430913160	414999411	6
Aroclor-1260-4	(4)	324169048	356082777	368301826	380117112	370143240	359762801	6
Aroclor-1260-5	(5)	773923407	822481391	873089514	880740576	851175200	840282018	5
Decachlorobiphenyl		4511855410	4959881947	5165429800	5359918520	5250729800	5049563095	7
Tetrachloro-m-xylene		4790824740	5183285133	5313707940	5302586120	4809696600	5080020107	5
Aroclor-1242-1	(1)	149120136	160366853	160219200	165386980	179219740	162862582	7
Aroclor-1242-2	(2)	228406508	237737935	244055020	248772624	247903320	241375081	4
Aroclor-1242-3	(3)	123267302	131780304	132089372	135429308	136535900	131820437	4
Aroclor-1242-4	(4)	107929318	113409847	120761922	128344628	102401880	114569519	9
Aroclor-1242-5	(5)	175313049	196019013	175387368	167275268	156704980	174139936	8
Decachlorobiphenyl		5194557400	5483323560	5616478880	5698756520	6104877800	5619598832	6
Tetrachloro-m-xylene		5511103540	5712989200	5725543240	5446452960	5436162600	5566450308	3
Aroclor-1248-1	(1)	118147108	120641916	127443132	132200888	134031720	126492953	6
Aroclor-1248-2	(2)	172958749	177705395	189325858	196101732	206725040	188563355	7
Aroclor-1248-3	(3)	167593672	172934769	178216680	182062792	197467520	179655087	6
Aroclor-1248-4	(4)	207432918	214191484	230649658	232116576	237264740	224331075	6
Aroclor-1248-5	(5)	211078475	216720032	243430792	250826296	242157620	232842643	8
Decachlorobiphenyl		5357284330	5460780160	5730809660	5844375240	6130897000	5704829278	5
Tetrachloro-m-xylene		5499268270	5571294053	5722936200	5664492400	5408993600	5573396905	2
Aroclor-1254-1	(1)	326099704	341297420	345505280	377987212	295169840	337211891	9
Aroclor-1254-2	(2)	283781194	302336180	304849560	326229636	268575420	297154398	7
Aroclor-1254-3	(3)	455841488	477545104	484627186	501619228	415693040	467065209	7
Aroclor-1254-4	(4)	291449906	303039177	308578712	317073776	256115840	295251482	8
Aroclor-1254-5	(5)	417470468	427079459	439495106	452352324	365350180	420349507	8
Decachlorobiphenyl		5360763820	5577066440	5716894620	5972893640	5256420800	5576807864	5
Tetrachloro-m-xylene		5311856950	5451428933	5484489360	5522984680	4238156400	5201783265	10
Aroclor-1268-1	(1)	931455367	961712979	973652708	974423320	1030986200	974446115	4

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	858872879	891158693	900594916	906701084	936258500	898717214	3
Aroclor-1268-3	(3)	746016914	769745548	779615614	791820108	821422560	781724149	4
Aroclor-1268-4	(4)	330162118	339473573	345636662	356393996	350955500	344524370	3
Aroclor-1268-5	(5)	2374681378	2422263161	2432472106	2427996872	2444443520	2420371407	1
Decachlorobiphenyl		10401085910	10736618840	10941521580	11269357960	11896046400	11048926138	5
Tetrachloro-m-xylene		5581659480	5729850200	5733063640	5828463760	5669436400	5708494696	2

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Instrument ID: ECD_Q **Date(s) Analyzed:** 10/10/2025 10/10/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	3.59	3.49	3.69	59190800
		2	3.67	3.57	3.77	43146800
		3	3.74	3.64	3.84	133548000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	3.74	3.64	3.84	108948000
		2	4.23	4.13	4.33	57502800
		3	4.50	4.40	4.60	107099000
		4	4.65	4.55	4.75	54014200
		5	4.79	4.69	4.89	43713200
Aroclor-1262	500	1	6.84	6.74	6.94	274018000
		2	7.15	7.05	7.25	498476000
		3	7.42	7.32	7.52	328660000
		4	7.49	7.39	7.59	252340000
		5	7.99	7.89	8.09	164350000

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Instrument ID: ECD_Q **Date(s) Analyzed:** 10/10/2025 10/10/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	2.98	2.88	3.08	73819200
		2	3.05	2.95	3.15	60003200
		3	3.12	3.02	3.22	171184000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	3.12	3.02	3.22	139342000
		2	3.79	3.69	3.89	138900000
		3	3.95	3.85	4.05	73450200
		4	4.04	3.94	4.14	63034400
		5	4.20	4.10	4.30	71607200
Aroclor-1262	500	1	5.73	5.63	5.83	452486000
		2	5.98	5.88	6.08	427702000
		3	6.50	6.40	6.60	332708000
		4	6.56	6.46	6.66	616820000
		5	7.05	6.95	7.15	302986000

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 08:57
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:54:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.778	404.9E6	479.1E6	95.307	94.825
2)	SA Decachlor...	8.496	7.534	260.4E6	451.2E6	92.652	93.246
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.778	152.6E6	198.1E6	947.441	929.401
4)	L1 AR-1016-2	4.499	3.793	235.3E6	307.0E6	941.324	940.794
5)	L1 AR-1016-3	4.556	3.954	145.7E6	164.2E6	939.563	934.980
6)	L1 AR-1016-4	4.646	4.002	120.2E6	139.7E6	942.413	928.587
7)	L1 AR-1016-5	4.931	4.197	117.2E6	171.3E6	935.728	930.345
31)	L7 AR-1260-1	6.020	5.190	208.9E6	322.1E6	926.912	936.559
32)	L7 AR-1260-2	6.275	5.379	243.7E6	392.9E6	917.174	931.909
33)	L7 AR-1260-3	6.626	5.520	178.0E6	374.4E6	930.393	938.406
34)	L7 AR-1260-4	6.840	5.980	191.6E6	324.2E6	922.615	936.268
35)	L7 AR-1260-5	7.146	6.224	401.0E6	773.9E6	945.173	939.790

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 08:57
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:54:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.401 min
Delta R.T.: 0.000 min
Response: 404899105
Conc: 95.31 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC1000



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 479082474
Conc: 94.83 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 260399792
Conc: 92.65 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 451185541
Conc: 93.25 ng/ml



#3 AR-1016-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 152632258
Conc: 947.44 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC1000



#3 AR-1016-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 198094739
Conc: 929.40 ng/ml



#4 AR-1016-2

R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 235274296
Conc: 941.32 ng/ml



#4 AR-1016-2

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 306974778
Conc: 940.79 ng/ml



#5 AR-1016-3

R.T.: 4.556 min
Delta R.T.: 0.000 min
Response: 145675938
Conc: 939.56 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC1000



#5 AR-1016-3

R.T.: 3.954 min
Delta R.T.: 0.000 min
Response: 164197912
Conc: 934.98 ng/ml



#6 AR-1016-4

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 120162951
Conc: 942.41 ng/ml



#6 AR-1016-4

R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 139678855
Conc: 928.59 ng/ml



#7 AR-1016-5

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 117231883
Conc: 935.73 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC1000



#7 AR-1016-5

R.T.: 4.197 min
Delta R.T.: 0.000 min
Response: 171310856
Conc: 930.34 ng/ml



#31 AR-1260-1

R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 208928914
Conc: 926.91 ng/ml



#31 AR-1260-1

R.T.: 5.190 min
Delta R.T.: 0.000 min
Response: 322104344
Conc: 936.56 ng/ml



#32 AR-1260-2

R.T.: 6.275 min
Delta R.T.: 0.000 min
Response: 243695794
Conc: 917.17 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC1000



#32 AR-1260-2

R.T.: 5.379 min
Delta R.T.: 0.000 min
Response: 392921251
Conc: 931.91 ng/ml



#33 AR-1260-3

R.T.: 6.626 min
Delta R.T.: 0.000 min
Response: 178017013
Conc: 930.39 ng/ml



#33 AR-1260-3

R.T.: 5.520 min
Delta R.T.: 0.000 min
Response: 374441100
Conc: 938.41 ng/ml



#34 AR-1260-4

R.T.: 6.840 min
Delta R.T.: 0.000 min
Response: 191556315
Conc: 922.62 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC1000



#34 AR-1260-4

R.T.: 5.980 min
Delta R.T.: 0.000 min
Response: 324169048
Conc: 936.27 ng/ml



#35 AR-1260-5

R.T.: 7.146 min
Delta R.T.: 0.000 min
Response: 400986941
Conc: 945.17 ng/ml



#35 AR-1260-5

R.T.: 6.224 min
Delta R.T.: 0.000 min
Response: 773923407
Conc: 939.79 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070974.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:11
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:57:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	331.5E6	388.7E6	76.997	76.286
2)	SA Decachlor...	8.496	7.534	216.6E6	372.0E6	76.373	76.242
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.778	121.0E6	162.9E6	750.874	759.502
4)	L1 AR-1016-2	4.499	3.794	192.6E6	252.7E6	763.674	766.122
5)	L1 AR-1016-3	4.556	3.954	117.6E6	135.4E6	755.578	763.875
6)	L1 AR-1016-4	4.646	4.002	97033800	115.6E6	757.308	762.093
7)	L1 AR-1016-5	4.930	4.197	96409305	141.0E6	762.905	760.395
31)	L7 AR-1260-1	6.020	5.189	175.6E6	264.6E6	769.253	762.765
32)	L7 AR-1260-2	6.276	5.378	206.4E6	324.1E6	767.783	762.288
33)	L7 AR-1260-3	6.627	5.519	147.3E6	307.5E6	763.182	763.633
34)	L7 AR-1260-4	6.840	5.980	160.7E6	267.1E6	765.822	764.087
35)	L7 AR-1260-5	7.146	6.224	327.7E6	616.9E6	764.763	749.377

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070974.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:11
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:57:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 331524382
Conc: 77.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC750



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 388746385
Conc: 76.29 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 216632305
Conc: 76.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 371991146
Conc: 76.24 ng/ml



#3 AR-1016-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 121036095
Conc: 750.87 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC750



#3 AR-1016-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 162913864
Conc: 759.50 ng/ml



#4 AR-1016-2

R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 192628730
Conc: 763.67 ng/ml



#4 AR-1016-2

R.T.: 3.794 min
Delta R.T.: 0.000 min
Response: 252696401
Conc: 766.12 ng/ml



#5 AR-1016-3

R.T.: 4.556 min
Delta R.T.: 0.000 min
Response: 117587028
Conc: 755.58 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC750



#5 AR-1016-3

R.T.: 3.954 min
Delta R.T.: 0.000 min
Response: 135401603
Conc: 763.88 ng/ml



#6 AR-1016-4

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 97033800
Conc: 757.31 ng/ml



#6 AR-1016-4

R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 115566331
Conc: 762.09 ng/ml



#7 AR-1016-5

R.T.: 4.930 min
Delta R.T.: 0.000 min
Response: 96409305
Conc: 762.90 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC750



#7 AR-1016-5

R.T.: 4.197 min
Delta R.T.: 0.000 min
Response: 140993849
Conc: 760.39 ng/ml



#31 AR-1260-1

R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 175646524
Conc: 769.25 ng/ml



#31 AR-1260-1

R.T.: 5.189 min
Delta R.T.: 0.000 min
Response: 264584355
Conc: 762.77 ng/ml



#32 AR-1260-2

R.T.: 6.276 min
Delta R.T.: 0.000 min
Response: 206449688
Conc: 767.78 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC750



#32 AR-1260-2

R.T.: 5.378 min
Delta R.T.: 0.000 min
Response: 324058699
Conc: 762.29 ng/ml



#33 AR-1260-3

R.T.: 6.627 min
Delta R.T.: 0.000 min
Response: 147318344
Conc: 763.18 ng/ml



#33 AR-1260-3

R.T.: 5.519 min
Delta R.T.: 0.000 min
Response: 307498276
Conc: 763.63 ng/ml



#34 AR-1260-4

R.T.: 6.840 min
Delta R.T.: 0.000 min
Response: 160697330
Conc: 765.82 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC750



#34 AR-1260-4

R.T.: 5.980 min
Delta R.T.: 0.000 min
Response: 267062083
Conc: 764.09 ng/ml



#35 AR-1260-5

R.T.: 7.146 min
Delta R.T.: 0.000 min
Response: 327673748
Conc: 764.76 ng/ml



#35 AR-1260-5

R.T.: 6.224 min
Delta R.T.: 0.000 min
Response: 616861043
Conc: 749.38 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:26
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:49:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	222.4E6	265.7E6	50.000	50.000
2)	SA Decachlor...	8.496	7.533	150.9E6	258.3E6	50.000	50.000
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.778	84783442	114.1E6	500.000	500.000
4)	L1 AR-1016-2	4.499	3.793	132.3E6	172.8E6	500.000	500.000
5)	L1 AR-1016-3	4.556	3.953	82208498	93517571	500.000	500.000
6)	L1 AR-1016-4	4.646	4.002	67424126	80581398	500.000	500.000
7)	L1 AR-1016-5	4.931	4.197	66668174	98481559	500.000	500.000
31)	L7 AR-1260-1	6.019	5.189	120.9E6	182.9E6	500.000	500.000
32)	L7 AR-1260-2	6.275	5.378	143.9E6	225.2E6	500.000	500.000
33)	L7 AR-1260-3	6.626	5.519	102.3E6	211.8E6	500.000	500.000
34)	L7 AR-1260-4	6.840	5.979	111.8E6	184.2E6	500.000	500.000
35)	L7 AR-1260-5	7.145	6.224	223.8E6	436.5E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:26
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:49:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 222389121
Conc: 50.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC500



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 265685397
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 150851883
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 258271490
Conc: 50.00 ng/ml



#3 AR-1016-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 84783442
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC500



#3 AR-1016-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 114094944
Conc: 500.00 ng/ml



#4 AR-1016-2

R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 132302716
Conc: 500.00 ng/ml



#4 AR-1016-2

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 172805809
Conc: 500.00 ng/ml



#5 AR-1016-3

R.T.: 4.556 min
Delta R.T.: 0.000 min
Response: 82208498
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC500



#5 AR-1016-3

R.T.: 3.953 min
Delta R.T.: 0.000 min
Response: 93517571
Conc: 500.00 ng/ml



#6 AR-1016-4

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 67424126
Conc: 500.00 ng/ml



#6 AR-1016-4

R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 80581398
Conc: 500.00 ng/ml



#7 AR-1016-5

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 66668174
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC500



#7 AR-1016-5

R.T.: 4.197 min
Delta R.T.: 0.000 min
Response: 98481559
Conc: 500.00 ng/ml



#31 AR-1260-1

R.T.: 6.019 min
Delta R.T.: 0.000 min
Response: 120938759
Conc: 500.00 ng/ml



#31 AR-1260-1

R.T.: 5.189 min
Delta R.T.: 0.000 min
Response: 182871121
Conc: 500.00 ng/ml



#32 AR-1260-2

R.T.: 6.275 min
Delta R.T.: 0.000 min
Response: 143855136
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC500



#32 AR-1260-2

R.T.: 5.378 min
Delta R.T.: 0.000 min
Response: 225169968
Conc: 500.00 ng/ml



#33 AR-1260-3

R.T.: 6.626 min
Delta R.T.: 0.000 min
Response: 102326803
Conc: 500.00 ng/ml



#33 AR-1260-3

R.T.: 5.519 min
Delta R.T.: 0.000 min
Response: 211797535
Conc: 500.00 ng/ml



#34 AR-1260-4

R.T.: 6.840 min
Delta R.T.: 0.000 min
Response: 111845082
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC500



#34 AR-1260-4

R.T.: 5.979 min
Delta R.T.: 0.000 min
Response: 184150913
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 223753809
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 6.224 min
Delta R.T.: 0.000 min
Response: 436544757
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:40
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:01:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	112.4E6	132.6E6	25.824	25.753
2)	SA Decachlor...	8.495	7.534	77430043	134.0E6	26.685	26.803
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.778	43873025	59445752	266.271	269.814
4)	L1 AR-1016-2	4.499	3.794	67879501	89768912	264.062	266.260
5)	L1 AR-1016-3	4.557	3.954	42399194	48815753	266.464	268.576
6)	L1 AR-1016-4	4.646	4.003	34575940	42722621	264.598	273.066
7)	L1 AR-1016-5	4.931	4.197	34076692	51408265	264.457	269.895
31)	L7 AR-1260-1	6.019	5.190	61964969	95539541	265.698	268.599
32)	L7 AR-1260-2	6.275	5.379	74180101	116.7E6	268.916	267.943
33)	L7 AR-1260-3	6.626	5.520	52397657	109.0E6	265.747	265.224
34)	L7 AR-1260-4	6.840	5.980	57203584	95029278	266.583	266.063
35)	L7 AR-1260-5	7.146	6.224	113.0E6	220.2E6	260.134	262.889

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:40
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:01:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 112427951
Conc: 25.82 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC250



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 132564653
Conc: 25.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.495 min
Delta R.T.: 0.000 min
Response: 77430043
Conc: 26.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 133997963
Conc: 26.80 ng/ml



#3 AR-1016-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 43873025
Conc: 266.27 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC250



#3 AR-1016-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 59445752
Conc: 269.81 ng/ml



#4 AR-1016-2

R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 67879501
Conc: 264.06 ng/ml



#4 AR-1016-2

R.T.: 3.794 min
Delta R.T.: 0.000 min
Response: 89768912
Conc: 266.26 ng/ml



#5 AR-1016-3

R.T.: 4.557 min
Delta R.T.: 0.000 min
Response: 42399194
Conc: 266.46 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC250



#5 AR-1016-3

R.T.: 3.954 min
Delta R.T.: 0.000 min
Response: 48815753
Conc: 268.58 ng/ml



#6 AR-1016-4

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 34575940
Conc: 264.60 ng/ml



#6 AR-1016-4

R.T.: 4.003 min
Delta R.T.: 0.000 min
Response: 42722621
Conc: 273.07 ng/ml



#7 AR-1016-5

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 34076692
Conc: 264.46 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC250



#7 AR-1016-5

R.T.: 4.197 min
Delta R.T.: 0.000 min
Response: 51408265
Conc: 269.90 ng/ml



#31 AR-1260-1

R.T.: 6.019 min
Delta R.T.: 0.000 min
Response: 61964969
Conc: 265.70 ng/ml



#31 AR-1260-1

R.T.: 5.190 min
Delta R.T.: 0.000 min
Response: 95539541
Conc: 268.60 ng/ml



#32 AR-1260-2

R.T.: 6.275 min
Delta R.T.: 0.000 min
Response: 74180101
Conc: 268.92 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC250



#32 AR-1260-2

R.T.: 5.379 min
Delta R.T.: 0.000 min
Response: 116697737
Conc: 267.94 ng/ml



#33 AR-1260-3

R.T.: 6.626 min
Delta R.T.: 0.000 min
Response: 52397657
Conc: 265.75 ng/ml



#33 AR-1260-3

R.T.: 5.520 min
Delta R.T.: 0.000 min
Response: 109012506
Conc: 265.22 ng/ml



#34 AR-1260-4

R.T.: 6.840 min
Delta R.T.: 0.000 min
Response: 57203584
Conc: 266.58 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC250



#34 AR-1260-4

R.T.: 5.980 min
Delta R.T.: 0.000 min
Response: 95029278
Conc: 266.06 ng/ml



#35 AR-1260-5

R.T.: 7.146 min
Delta R.T.: 0.000 min
Response: 112984855
Conc: 260.13 ng/ml



#35 AR-1260-5

R.T.: 6.224 min
Delta R.T.: 0.000 min
Response: 220185144
Conc: 262.89 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070977.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:55
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:05:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	21685891	24048483	4.985	4.734
2)	SA Decachlor...	8.495	7.533	14566242	26253649	5.016	5.199
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.778	8082361	11936188	49.239	53.286
4)	L1 AR-1016-2	4.498	3.793	12661786	17601233	49.403	51.750
5)	L1 AR-1016-3	4.557	3.953	7955970	9113143	50.000	50.111
6)	L1 AR-1016-4	4.645	4.002	6347651	8062138	48.855	51.217
7)	L1 AR-1016-5	4.930	4.197	6429669	9865534	49.919	51.425
31)	L7 AR-1260-1	6.019	5.190	11974683	18962286	51.162m	52.614
32)	L7 AR-1260-2	6.274	5.378	13982610	22891919	50.322m	52.028
33)	L7 AR-1260-3	6.626	5.519	10392796	21545658	52.144	51.917
34)	L7 AR-1260-4	6.840	5.980	10804624	18507162	50.281	51.443
35)	L7 AR-1260-5	7.145	6.224	21577675	42558760	49.744	50.648

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070977.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:55
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:05:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 21685891
Conc: 4.98 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 24048483
Conc: 4.73 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.495 min
Delta R.T.: 0.000 min
Response: 14566242
Conc: 5.02 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 26253649
Conc: 5.20 ng/ml



#3 AR-1016-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 8082361
Conc: 49.24 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#3 AR-1016-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 11936188
Conc: 53.29 ng/ml



#4 AR-1016-2

R.T.: 4.498 min
Delta R.T.: 0.000 min
Response: 12661786
Conc: 49.40 ng/ml



#4 AR-1016-2

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 17601233
Conc: 51.75 ng/ml



#5 AR-1016-3

R.T.: 4.557 min
Delta R.T.: 0.000 min
Response: 7955970
Conc: 50.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#5 AR-1016-3

R.T.: 3.953 min
Delta R.T.: 0.000 min
Response: 9113143
Conc: 50.11 ng/ml



#6 AR-1016-4

R.T.: 4.645 min
Delta R.T.: 0.000 min
Response: 6347651
Conc: 48.85 ng/ml



#6 AR-1016-4

R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 8062138
Conc: 51.22 ng/ml



#7 AR-1016-5

R.T.: 4.930 min
Delta R.T.: 0.000 min
Response: 6429669
Conc: 49.92 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#7 AR-1016-5

R.T.: 4.197 min
Delta R.T.: 0.000 min
Response: 9865534
Conc: 51.43 ng/ml



#31 AR-1260-1

R.T.: 6.019 min
Delta R.T.: 0.000 min
Response: 11974683
Conc: 51.16 ng/ml m



#31 AR-1260-1

R.T.: 5.190 min
Delta R.T.: 0.000 min
Response: 18962286
Conc: 52.61 ng/ml



#32 AR-1260-2

R.T.: 6.274 min
Delta R.T.: 0.000 min
Response: 13982610
Conc: 50.32 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#32 AR-1260-2

R.T.: 5.378 min
Delta R.T.: 0.000 min
Response: 22891919
Conc: 52.03 ng/ml



#33 AR-1260-3

R.T.: 6.626 min
Delta R.T.: 0.000 min
Response: 10392796
Conc: 52.14 ng/ml



#33 AR-1260-3

R.T.: 5.519 min
Delta R.T.: 0.000 min
Response: 21545658
Conc: 51.92 ng/ml



#34 AR-1260-4

R.T.: 6.840 min
Delta R.T.: 0.000 min
Response: 10804624
Conc: 50.28 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#34 AR-1260-4

R.T.: 5.980 min
Delta R.T.: 0.000 min
Response: 18507162
Conc: 51.44 ng/ml



#35 AR-1260-5

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 21577675
Conc: 49.74 ng/ml



#35 AR-1260-5

R.T.: 6.224 min
Delta R.T.: 0.000 min
Response: 42558760
Conc: 50.65 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070978.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:09
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:55:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 11:55:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	220.4E6	255.3E6	50.000	50.000
2)	SA Decachlor...	8.496	7.534	149.9E6	262.1E6	50.000	50.000
	Target Compounds						
8)	L2 AR-1221-1	3.594	2.977	29595379	36909578	500.000	500.000
9)	L2 AR-1221-2	3.672	3.053	21573372	30001568	500.000	500.000
10)	L2 AR-1221-3	3.739	3.122	66773962	85591891	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070978.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:09
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:55:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 11:55:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 220369967
Conc: 50.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC500



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 255300945
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 149874496
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 262139791
Conc: 50.00 ng/ml



#8 AR-1221-1

R.T.: 3.594 min
Delta R.T.: 0.000 min
Response: 29595379
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC500



#8 AR-1221-1

R.T.: 2.977 min
Delta R.T.: 0.000 min
Response: 36909578
Conc: 500.00 ng/ml



#9 AR-1221-2

R.T.: 3.672 min
Delta R.T.: 0.000 min
Response: 21573372
Conc: 500.00 ng/ml



#9 AR-1221-2

R.T.: 3.053 min
Delta R.T.: 0.000 min
Response: 30001568
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 3.739 min
Delta R.T.: 0.000 min
Response: 66773962
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 3.122 min
Delta R.T.: 0.000 min
Response: 85591891
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:24
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:00:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 11:55:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	234.5E6	276.3E6	50.000	50.000
2)	SA Decachlor...	8.497	7.534	163.6E6	280.6E6	50.000	50.000
	Target Compounds						
11)	L3 AR-1232-1	3.739	3.121	54474047	69671226	500.000	500.000
12)	L3 AR-1232-2	4.226	3.793	28751418	69449831	500.000	500.000
13)	L3 AR-1232-3	4.499	3.953	53549480	36725094	500.000	500.000
14)	L3 AR-1232-4	4.646	4.039	27007118	31517168	500.000	500.000
15)	L3 AR-1232-5	4.785	4.197	21856633	35803624	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:24
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:00:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 11:55:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 234547177
Conc: 50.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1232ICC500



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 276348863
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 163612084
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 280551433
Conc: 50.00 ng/ml



#11 AR-1232-1

R.T.: 3.739 min
Delta R.T.: 0.000 min
Response: 54474047
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1232ICC500



#11 AR-1232-1

R.T.: 3.121 min
Delta R.T.: 0.000 min
Response: 69671226
Conc: 500.00 ng/ml



#12 AR-1232-2

R.T.: 4.226 min
Delta R.T.: 0.000 min
Response: 28751418
Conc: 500.00 ng/ml



#12 AR-1232-2

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 69449831
Conc: 500.00 ng/ml



#13 AR-1232-3

R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 53549480
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1232ICC500



#13 AR-1232-3

R.T.: 3.953 min
Delta R.T.: 0.000 min
Response: 36725094
Conc: 500.00 ng/ml



#14 AR-1232-4

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 27007118
Conc: 500.00 ng/ml



#14 AR-1232-4

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 31517168
Conc: 500.00 ng/ml



#15 AR-1232-5

R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 21856633
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1232ICC500



#15 AR-1232-5

R.T.: 4.197 min
Delta R.T.: 0.000 min
Response: 35803624
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:38
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:08:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:05:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	444.2E6	551.1E6	98.001	98.092
2)	SA Decachlor...	8.495	7.534	308.3E6	519.5E6	95.801	96.097
	Target Compounds						
16)	L4 AR-1242-1	4.480	3.778	117.9E6	149.1E6	983.375	964.120
17)	L4 AR-1242-2	4.499	3.793	179.2E6	228.4E6	959.238	966.879
18)	L4 AR-1242-3	4.556	3.953	111.3E6	123.3E6	959.357	965.452
19)	L4 AR-1242-4	4.646	4.039	91702396	107.9E6	962.012	943.887
20)	L4 AR-1242-5	5.349	4.534	63322787	175.3E6	867.188	999.788

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:38
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:08:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:05:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 444189301
Conc: 98.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC1000



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 551110354
Conc: 98.09 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.495 min
Delta R.T.: 0.000 min
Response: 308256864
Conc: 95.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 519455740
Conc: 96.10 ng/ml



#16 AR-1242-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 117929900
Conc: 983.37 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC1000



#16 AR-1242-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 149120136
Conc: 964.12 ng/ml



#17 AR-1242-2

R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 179238730
Conc: 959.24 ng/ml



#17 AR-1242-2

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 228406508
Conc: 966.88 ng/ml



#18 AR-1242-3

R.T.: 4.556 min
Delta R.T.: 0.000 min
Response: 111298009
Conc: 959.36 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC1000



#18 AR-1242-3

R.T.: 3.953 min
Delta R.T.: 0.000 min
Response: 123267302
Conc: 965.45 ng/ml



#19 AR-1242-4

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 91702396
Conc: 962.01 ng/ml



#19 AR-1242-4

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 107929318
Conc: 943.89 ng/ml



#20 AR-1242-5

R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 63322787
Conc: 867.19 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC1000



#20 AR-1242-5

R.T.: 4.534 min
Delta R.T.: 0.000 min
Response: 175313049
Conc: 999.79 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070981.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:53
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:12:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	343.3E6	428.5E6	75.493	75.838
2)	SA Decachlor...	8.496	7.534	243.3E6	411.2E6	75.399	75.716
	Target Compounds						
16)	L4 AR-1242-1	4.480	3.778	90376185	120.3E6	752.406	768.194
17)	L4 AR-1242-2	4.498	3.794	141.6E6	178.3E6	755.086	753.183
18)	L4 AR-1242-3	4.556	3.954	87949114	98835228	755.378	765.893
19)	L4 AR-1242-4	4.646	4.039	73032215	85057385	760.690	745.897
20)	L4 AR-1242-5	5.348	4.534	50322958	147.0E6	708.312	806.708

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070981.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:53
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:12:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 343296796
Conc: 75.49 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 428474190
Conc: 75.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 243255520
Conc: 75.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 411249267
Conc: 75.72 ng/ml



#16 AR-1242-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 90376185
Conc: 752.41 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750



#16 AR-1242-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 120275140
Conc: 768.19 ng/ml



#17 AR-1242-2

R.T.: 4.498 min
Delta R.T.: 0.000 min
Response: 141571716
Conc: 755.09 ng/ml



#17 AR-1242-2

R.T.: 3.794 min
Delta R.T.: 0.000 min
Response: 178303451
Conc: 753.18 ng/ml



#18 AR-1242-3

R.T.: 4.556 min
Delta R.T.: 0.000 min
Response: 87949114
Conc: 755.38 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750



#18 AR-1242-3

R.T.: 3.954 min
Delta R.T.: 0.000 min
Response: 98835228
Conc: 765.89 ng/ml



#19 AR-1242-4

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 73032215
Conc: 760.69 ng/ml



#19 AR-1242-4

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 85057385
Conc: 745.90 ng/ml



#20 AR-1242-5

R.T.: 5.348 min
Delta R.T.: 0.000 min
Response: 50322958
Conc: 708.31 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 4.534 min
Delta R.T.: 0.000 min
Response: 147014260
Conc: 806.71 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070982.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:08
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:05:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:05:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	231.2E6	286.3E6	50.000	50.000
2)	SA Decachlor...	8.496	7.533	167.6E6	280.8E6	50.000	50.000
	Target Compounds						
16)	L4 AR-1242-1	4.480	3.778	60958684	80109600	500.000	500.000
17)	L4 AR-1242-2	4.499	3.793	97235865	122.0E6	500.000	500.000
18)	L4 AR-1242-3	4.556	3.954	60364163	66044686	500.000	500.000
19)	L4 AR-1242-4	4.646	4.040	49472392	60380961	500.000	500.000
20)	L4 AR-1242-5	5.351	4.534	41359410	87693684	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070982.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:08
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:05:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:05:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 231153264
Conc: 50.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC500



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 286277162
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 167640288
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 280823944
Conc: 50.00 ng/ml



#16 AR-1242-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 60958684
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC500



#16 AR-1242-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 80109600
Conc: 500.00 ng/ml



#17 AR-1242-2

R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 97235865
Conc: 500.00 ng/ml



#17 AR-1242-2

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 122027510
Conc: 500.00 ng/ml



#18 AR-1242-3

R.T.: 4.556 min
Delta R.T.: 0.000 min
Response: 60364163
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC500



#18 AR-1242-3

R.T.: 3.954 min
Delta R.T.: 0.000 min
Response: 66044686
Conc: 500.00 ng/ml



#19 AR-1242-4

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 49472392
Conc: 500.00 ng/ml



#19 AR-1242-4

R.T.: 4.040 min
Delta R.T.: 0.000 min
Response: 60380961
Conc: 500.00 ng/ml



#20 AR-1242-5

R.T.: 5.351 min
Delta R.T.: 0.000 min
Response: 41359410
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC500



#20 AR-1242-5

R.T.: 4.534 min
Delta R.T.: 0.000 min
Response: 87693684
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070983.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:22
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:15:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.778	115.1E6	136.2E6	25.240	24.319
2)	SA Decachlor...	8.497	7.534	83994234	142.5E6	25.768	25.912
	Target Compounds						
16)	L4 AR-1242-1	4.481	3.778	31044472	41346745	256.287	260.414
17)	L4 AR-1242-2	4.499	3.794	49429930	62193156	260.092	259.416
18)	L4 AR-1242-3	4.557	3.954	30938677	33857327	261.612	259.162
19)	L4 AR-1242-4	4.647	4.040	25299238	32086157	259.999	272.815
20)	L4 AR-1242-5	5.352	4.534	25666453	41818817	325.093	234.281 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070983.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:22
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:15:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.401 min
Delta R.T.: 0.000 min
Response: 115146396
Conc: 25.24 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC250



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 136161324
Conc: 24.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 83994234
Conc: 25.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 142468913
Conc: 25.91 ng/ml



#16 AR-1242-1

R.T.: 4.481 min
Delta R.T.: 0.000 min
Response: 31044472
Conc: 256.29 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC250



#16 AR-1242-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 41346745
Conc: 260.41 ng/ml



#17 AR-1242-2

R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 49429930
Conc: 260.09 ng/ml



#17 AR-1242-2

R.T.: 3.794 min
Delta R.T.: 0.000 min
Response: 62193156
Conc: 259.42 ng/ml



#18 AR-1242-3

R.T.: 4.557 min
Delta R.T.: 0.000 min
Response: 30938677
Conc: 261.61 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC250



#18 AR-1242-3

R.T.: 3.954 min
Delta R.T.: 0.000 min
Response: 33857327
Conc: 259.16 ng/ml



#19 AR-1242-4

R.T.: 4.647 min
Delta R.T.: 0.000 min
Response: 25299238
Conc: 260.00 ng/ml



#19 AR-1242-4

R.T.: 4.040 min
Delta R.T.: 0.000 min
Response: 32086157
Conc: 272.81 ng/ml



#20 AR-1242-5

R.T.: 5.352 min
Delta R.T.: 0.000 min
Response: 25666453
Conc: 325.09 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 4.534 min
Delta R.T.: 0.000 min
Response: 41818817
Conc: 234.28 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070984.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:37
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:20:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	23512581	27180813	5.122	4.883
2)	SA Decachlor...	8.496	7.534	16433954	30524389	5.033	5.432
	Target Compounds						
16)	L4 AR-1242-1	4.481	3.777	5511662	8960987	46.335	57.042m
17)	L4 AR-1242-2	4.499	3.793	9668594	12395166	50.697	51.291m
18)	L4 AR-1242-3	4.556	3.953	6071735	6826795	51.068	51.803m
19)	L4 AR-1242-4	4.646	4.039	5427928	5120094	54.521	44.579m
20)	L4 AR-1242-5	5.317	4.534	3890240	7835249	49.488m	45.137m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070984.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:37
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:20:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 23512581
Conc: 5.12 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 27180813
Conc: 4.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 16433954
Conc: 5.03 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 30524389
Conc: 5.43 ng/ml



#16 AR-1242-1

R.T.: 4.481 min
Delta R.T.: 0.000 min
Response: 5511662
Conc: 46.34 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#16 AR-1242-1

R.T.: 3.777 min
Delta R.T.: 0.000 min
Response: 8960987
Conc: 57.04 ng/ml m



#17 AR-1242-2

R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 9668594
Conc: 50.70 ng/ml



#17 AR-1242-2

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 12395166
Conc: 51.29 ng/ml m



#18 AR-1242-3

R.T.: 4.556 min
Delta R.T.: 0.000 min
Response: 6071735
Conc: 51.07 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#18 AR-1242-3

R.T.: 3.953 min
Delta R.T.: 0.000 min
Response: 6826795
Conc: 51.80 ng/ml m



#19 AR-1242-4

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 5427928
Conc: 54.52 ng/ml



#19 AR-1242-4

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 5120094
Conc: 44.58 ng/ml m



#20 AR-1242-5

R.T.: 5.317 min
Delta R.T.: 0.000 min
Response: 3890240
Conc: 49.49 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#20 AR-1242-5

R.T.: 4.534 min
Delta R.T.: 0.000 min
Response: 7835249
Conc: 45.14 ng/ml m

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070985.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:51
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:40:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	454.7E6	549.9E6	97.923	98.007
2)	SA Decachlor...	8.496	7.534	317.7E6	535.7E6	96.574	96.631
	Target Compounds						
21)	L5 AR-1248-1	4.480	3.778	89951061	118.1E6	950.710	962.148
22)	L5 AR-1248-2	4.743	4.002	126.8E6	173.0E6	954.654	954.823
23)	L5 AR-1248-3	4.930	4.039	153.0E6	167.6E6	958.744	969.281
24)	L5 AR-1248-4	5.318	4.197	167.1E6	207.4E6	954.987	947.004
25)	L5 AR-1248-5	5.353	4.572	163.7E6	211.1E6	1106.316	928.819

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070985.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:51
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:40:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 454650898
Conc: 97.92 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC1000



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 549926827
Conc: 98.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 317734844
Conc: 96.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 535728433
Conc: 96.63 ng/ml



#21 AR-1248-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 89951061
Conc: 950.71 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC1000



#21 AR-1248-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 118147108
Conc: 962.15 ng/ml



#22 AR-1248-2

R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 126849870
Conc: 954.65 ng/ml



#22 AR-1248-2

R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 172958749
Conc: 954.82 ng/ml



#23 AR-1248-3

R.T.: 4.930 min
Delta R.T.: 0.000 min
Response: 152980867
Conc: 958.74 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC1000



#23 AR-1248-3

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 167593672
Conc: 969.28 ng/ml



#24 AR-1248-4

R.T.: 5.318 min
Delta R.T.: 0.000 min
Response: 167077290
Conc: 954.99 ng/ml



#24 AR-1248-4

R.T.: 4.197 min
Delta R.T.: 0.000 min
Response: 207432918
Conc: 947.00 ng/ml



#25 AR-1248-5

R.T.: 5.353 min
Delta R.T.: 0.000 min
Response: 163667698
Conc: 1106.32 ng/m

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 4.572 min
Delta R.T.: 0.000 min
Response: 211078475
Conc: 928.82 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:06
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:46:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	344.9E6	417.8E6	74.527	74.644
2)	SA Decachlor...	8.495	7.533	243.6E6	409.6E6	74.366	74.245
	Target Compounds						
21)	L5 AR-1248-1	4.480	3.777	69602185	90481437	740.364	741.181
22)	L5 AR-1248-2	4.744	4.002	98823740	133.3E6	745.811	740.453
23)	L5 AR-1248-3	4.931	4.039	118.2E6	129.7E6	743.646	750.086
24)	L5 AR-1248-4	5.318	4.197	129.8E6	160.6E6	744.762	738.847
25)	L5 AR-1248-5	5.352	4.571	126.1E6	162.5E6	815.290	726.458

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:06
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:46:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 344936586
Conc: 74.53 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 417847054
Conc: 74.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.495 min
Delta R.T.: 0.000 min
Response: 243638918
Conc: 74.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 409558512
Conc: 74.25 ng/ml



#21 AR-1248-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 69602185
Conc: 740.36 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750



#21 AR-1248-1

R.T.: 3.777 min
Delta R.T.: 0.000 min
Response: 90481437
Conc: 741.18 ng/ml



#22 AR-1248-2

R.T.: 4.744 min
Delta R.T.: 0.000 min
Response: 98823740
Conc: 745.81 ng/ml



#22 AR-1248-2

R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 133279046
Conc: 740.45 ng/ml



#23 AR-1248-3

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 118158493
Conc: 743.65 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750



#23 AR-1248-3

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 129701077
Conc: 750.09 ng/ml



#24 AR-1248-4

R.T.: 5.318 min
Delta R.T.: 0.000 min
Response: 129844545
Conc: 744.76 ng/ml



#24 AR-1248-4

R.T.: 4.197 min
Delta R.T.: 0.000 min
Response: 160643613
Conc: 738.85 ng/ml



#25 AR-1248-5

R.T.: 5.352 min
Delta R.T.: 0.000 min
Response: 126102213
Conc: 815.29 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 4.571 min
Delta R.T.: 0.000 min
Response: 162540024
Conc: 726.46 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070987.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:20
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:37:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	237.0E6	286.1E6	50.000	50.000
2)	SA Decachlor...	8.496	7.532	170.1E6	286.5E6	50.000	50.000
	Target Compounds						
21)	L5 AR-1248-1	4.480	3.778	49639121	63721566	500.000	500.000
22)	L5 AR-1248-2	4.744	4.002	69450264	94662929	500.000	500.000
23)	L5 AR-1248-3	4.931	4.039	83073425	89108340	500.000	500.000
24)	L5 AR-1248-4	5.318	4.197	91413720	115.3E6	500.000	500.000
25)	L5 AR-1248-5	5.351	4.571	66105541	121.7E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070987.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:20
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:37:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 236970659
Conc: 50.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC500



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 286146810
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 170138327
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.532 min
Delta R.T.: 0.000 min
Response: 286540483
Conc: 50.00 ng/ml



#21 AR-1248-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 49639121
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC500



#21 AR-1248-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 63721566
Conc: 500.00 ng/ml



#22 AR-1248-2

R.T.: 4.744 min
Delta R.T.: 0.000 min
Response: 69450264
Conc: 500.00 ng/ml



#22 AR-1248-2

R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 94662929
Conc: 500.00 ng/ml



#23 AR-1248-3

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 83073425
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC500



#23 AR-1248-3

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 89108340
Conc: 500.00 ng/ml



#24 AR-1248-4

R.T.: 5.318 min
Delta R.T.: 0.000 min
Response: 91413720
Conc: 500.00 ng/ml



#24 AR-1248-4

R.T.: 4.197 min
Delta R.T.: 0.000 min
Response: 115324829
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 5.351 min
Delta R.T.: 0.000 min
Response: 66105541
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 4.571 min
Delta R.T.: 0.000 min
Response: 121715396
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:35
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:50:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.778	117.1E6	141.6E6	25.232	25.223
2)	SA Decachlor...	8.497	7.534	85869763	146.1E6	25.897	26.099
	Target Compounds						
21)	L5 AR-1248-1	4.481	3.778	24349976	33050222	256.699	265.233
22)	L5 AR-1248-2	4.744	4.002	35564341	49025433	263.550	266.409
23)	L5 AR-1248-3	4.931	4.039	41870274	45515698	260.002	259.790
24)	L5 AR-1248-4	5.318	4.196	46678299	58029144	263.071	262.398m
25)	L5 AR-1248-5	5.351	4.570	30347346	62706574	207.360	273.971m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:35
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:50:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.401 min
Delta R.T.: 0.000 min
Response: 117145709
Conc: 25.23 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 141612310
Conc: 25.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 85869763
Conc: 25.90 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 146109381
Conc: 26.10 ng/ml



#21 AR-1248-1

R.T.: 4.481 min
Delta R.T.: 0.000 min
Response: 24349976
Conc: 256.70 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#21 AR-1248-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 33050222
Conc: 265.23 ng/ml



#22 AR-1248-2

R.T.: 4.744 min
Delta R.T.: 0.000 min
Response: 35564341
Conc: 263.55 ng/ml



#22 AR-1248-2

R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 49025433
Conc: 266.41 ng/ml



#23 AR-1248-3

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 41870274
Conc: 260.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#23 AR-1248-3

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 45515698
Conc: 259.79 ng/ml



#24 AR-1248-4

R.T.: 5.318 min
Delta R.T.: 0.000 min
Response: 46678299
Conc: 263.07 ng/ml



#24 AR-1248-4

R.T.: 4.196 min
Delta R.T.: 0.000 min
Response: 58029144
Conc: 262.40 ng/ml m



#25 AR-1248-5

R.T.: 5.351 min
Delta R.T.: 0.000 min
Response: 30347346
Conc: 207.36 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#25 AR-1248-5

R.T.: 4.570 min
Delta R.T.: 0.000 min
Response: 62706574
Conc: 273.97 ng/ml m

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:50
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:31:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:30:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	22591108	27044968	4.892	4.853
2)	SA Decachlor...	8.497	7.534	16769582	30654485	5.046	5.373
	Target Compounds						
21)	L5 AR-1248-1	4.481	3.778	3394569	6701586	37.943	52.980 #
22)	L5 AR-1248-2	4.743	4.002	7123103	10336252	52.204	54.816
23)	L5 AR-1248-3	4.931	4.039	8138224	9873376	50.428	54.957
24)	L5 AR-1248-4	5.318	4.197	7886075	11863237	45.455	52.883
25)	L5 AR-1248-5	5.352	4.571	7959257	12107881	53.447	52.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:50
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:31:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:30:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 22591108
Conc: 4.89 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC050



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 27044968
Conc: 4.85 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 16769582
Conc: 5.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 30654485
Conc: 5.37 ng/ml



#21 AR-1248-1

R.T.: 4.481 min
Delta R.T.: 0.000 min
Response: 3394569
Conc: 37.94 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC050



#21 AR-1248-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 6701586
Conc: 52.98 ng/ml



#22 AR-1248-2

R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 7123103
Conc: 52.20 ng/ml



#22 AR-1248-2

R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 10336252
Conc: 54.82 ng/ml



#23 AR-1248-3

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 8138224
Conc: 50.43 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC050



#23 AR-1248-3

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 9873376
Conc: 54.96 ng/ml



#24 AR-1248-4

R.T.: 5.318 min
Delta R.T.: 0.000 min
Response: 7886075
Conc: 45.45 ng/ml



#24 AR-1248-4

R.T.: 4.197 min
Delta R.T.: 0.000 min
Response: 11863237
Conc: 52.88 ng/ml



#25 AR-1248-5

R.T.: 5.352 min
Delta R.T.: 0.000 min
Response: 7959257
Conc: 53.45 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC050



#25 AR-1248-5

R.T.: 4.571 min
Delta R.T.: 0.000 min
Response: 12107881
Conc: 52.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070990.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:04
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:53:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:52:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	445.0E6	531.2E6	98.289	98.401
2)	SA Decachlor...	8.496	7.533	313.6E6	536.1E6	97.450	96.785
	Target Compounds						
26)	L6 AR-1254-1	5.292	4.533	186.3E6	326.1E6	963.814	971.106
27)	L6 AR-1254-2	5.507	4.680	272.5E6	283.8E6	950.975	964.208
28)	L6 AR-1254-3	5.855	5.060	294.5E6	455.8E6	960.789	969.392
29)	L6 AR-1254-4	6.135	5.286	212.6E6	291.4E6	961.211	971.453
30)	L6 AR-1254-5	6.543	5.690	229.1E6	417.5E6	971.418	974.299

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070990.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:04
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:53:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:52:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 444955148
Conc: 98.29 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC1000



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 531185695
Conc: 98.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 313599352
Conc: 97.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 536076382
Conc: 96.79 ng/ml



#26 AR-1254-1

R.T.: 5.292 min
Delta R.T.: 0.000 min
Response: 186317660
Conc: 963.81 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC1000



#26 AR-1254-1

R.T.: 4.533 min
Delta R.T.: 0.000 min
Response: 326099704
Conc: 971.11 ng/ml



#27 AR-1254-2

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 272465681
Conc: 950.97 ng/ml



#27 AR-1254-2

R.T.: 4.680 min
Delta R.T.: 0.000 min
Response: 283781194
Conc: 964.21 ng/ml



#28 AR-1254-3

R.T.: 5.855 min
Delta R.T.: 0.000 min
Response: 294493285
Conc: 960.79 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC1000



#28 AR-1254-3

R.T.: 5.060 min
Delta R.T.: 0.000 min
Response: 455841488
Conc: 969.39 ng/ml



#29 AR-1254-4

R.T.: 6.135 min
Delta R.T.: 0.000 min
Response: 212636724
Conc: 961.21 ng/ml



#29 AR-1254-4

R.T.: 5.286 min
Delta R.T.: 0.000 min
Response: 291449906
Conc: 971.45 ng/ml



#30 AR-1254-5

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 229095618
Conc: 971.42 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC1000



#30 AR-1254-5

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 417470468
Conc: 974.30 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070991.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:19
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:59:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:59:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.778	343.8E6	408.9E6	75.627	75.492
2)	SA Decachlor...	8.496	7.534	242.8E6	418.3E6	75.295	75.344
	Target Compounds						
26)	L6 AR-1254-1	5.293	4.533	148.7E6	256.0E6	762.565	758.137
27)	L6 AR-1254-2	5.507	4.680	225.9E6	226.8E6	775.187	763.504
28)	L6 AR-1254-3	5.854	5.060	240.1E6	358.2E6	771.902	757.733
29)	L6 AR-1254-4	6.134	5.286	174.5E6	227.3E6	775.397	755.024
30)	L6 AR-1254-5	6.544	5.690	180.8E6	320.3E6	761.088	748.361

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070991.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:19
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:59:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:59:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.401 min
Delta R.T.: 0.000 min
Response: 343803158
Conc: 75.63 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 408857170
Conc: 75.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 242779345
Conc: 75.29 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 418279983
Conc: 75.34 ng/ml



#26 AR-1254-1

R.T.: 5.293 min
Delta R.T.: 0.000 min
Response: 148659073
Conc: 762.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750



#26 AR-1254-1

R.T.: 4.533 min
Delta R.T.: 0.000 min
Response: 255973065
Conc: 758.14 ng/ml



#27 AR-1254-2

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 225893559
Conc: 775.19 ng/ml



#27 AR-1254-2

R.T.: 4.680 min
Delta R.T.: 0.000 min
Response: 226752135
Conc: 763.50 ng/ml



#28 AR-1254-3

R.T.: 5.854 min
Delta R.T.: 0.000 min
Response: 240102864
Conc: 771.90 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750



#28 AR-1254-3

R.T.: 5.060 min
Delta R.T.: 0.000 min
Response: 358158828
Conc: 757.73 ng/ml



#29 AR-1254-4

R.T.: 6.134 min
Delta R.T.: 0.000 min
Response: 174485728
Conc: 775.40 ng/ml



#29 AR-1254-4

R.T.: 5.286 min
Delta R.T.: 0.000 min
Response: 227279383
Conc: 755.02 ng/ml



#30 AR-1254-5

R.T.: 6.544 min
Delta R.T.: 0.000 min
Response: 180828718
Conc: 761.09 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750



#30 AR-1254-5

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 320309594
Conc: 748.36 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:33
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:49:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:49:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	230.2E6	274.2E6	50.000	50.000
2)	SA Decachlor...	8.497	7.534	165.0E6	285.8E6	50.000	50.000
	Target Compounds						
26)	L6 AR-1254-1	5.293	4.532	100.2E6	172.8E6	500.000	500.000
27)	L6 AR-1254-2	5.507	4.680	150.3E6	152.4E6	500.000	500.000
28)	L6 AR-1254-3	5.855	5.060	159.3E6	242.3E6	500.000	500.000
29)	L6 AR-1254-4	6.136	5.286	114.9E6	154.3E6	500.000	500.000
30)	L6 AR-1254-5	6.543	5.689	121.3E6	219.7E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:33
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:49:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:49:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 230224891
Conc: 50.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC500



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 274224468
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 165005727
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 285844731
Conc: 50.00 ng/ml



#26 AR-1254-1

R.T.: 5.293 min
Delta R.T.: 0.000 min
Response: 100154125
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC500



#26 AR-1254-1

R.T.: 4.532 min
Delta R.T.: 0.000 min
Response: 172752640
Conc: 500.00 ng/ml



#27 AR-1254-2

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 150279116
Conc: 500.00 ng/ml



#27 AR-1254-2

R.T.: 4.680 min
Delta R.T.: 0.000 min
Response: 152424780
Conc: 500.00 ng/ml



#28 AR-1254-3

R.T.: 5.855 min
Delta R.T.: 0.000 min
Response: 159265361
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC500



#28 AR-1254-3

R.T.: 5.060 min
Delta R.T.: 0.000 min
Response: 242313593
Conc: 500.00 ng/ml



#29 AR-1254-4

R.T.: 6.136 min
Delta R.T.: 0.000 min
Response: 114899065
Conc: 500.00 ng/ml



#29 AR-1254-4

R.T.: 5.286 min
Delta R.T.: 0.000 min
Response: 154289356
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 121288407
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 5.689 min
Delta R.T.: 0.000 min
Response: 219747553
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070993.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:48
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:02:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:02:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.778	115.6E6	138.1E6	25.327	25.369
2)	SA Decachlor...	8.497	7.534	85811055	149.3E6	26.191	26.396
	Target Compounds						
26)	L6 AR-1254-1	5.293	4.533	55605487	94496803	275.527	271.759
27)	L6 AR-1254-2	5.507	4.679	79670910	81557409	267.151	268.017
28)	L6 AR-1254-3	5.855	5.060	83776868	125.4E6	264.224	261.310
29)	L6 AR-1254-4	6.135	5.286	59407243	79268444	260.355	259.866
30)	L6 AR-1254-5	6.544	5.690	62353894	113.1E6	259.216	260.512

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070993.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:48
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:02:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:02:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.401 min
Delta R.T.: 0.000 min
Response: 115639387
Conc: 25.33 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 138074617
Conc: 25.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 85811055
Conc: 26.19 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 149322341
Conc: 26.40 ng/ml



#26 AR-1254-1

R.T.: 5.293 min
Delta R.T.: 0.000 min
Response: 55605487
Conc: 275.53 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250



#26 AR-1254-1

R.T.: 4.533 min
Delta R.T.: 0.000 min
Response: 94496803
Conc: 271.76 ng/ml



#27 AR-1254-2

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 79670910
Conc: 267.15 ng/ml



#27 AR-1254-2

R.T.: 4.679 min
Delta R.T.: 0.000 min
Response: 81557409
Conc: 268.02 ng/ml



#28 AR-1254-3

R.T.: 5.855 min
Delta R.T.: 0.000 min
Response: 83776868
Conc: 264.22 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250



#28 AR-1254-3

R.T.: 5.060 min
Delta R.T.: 0.000 min
Response: 125404807
Conc: 261.31 ng/ml



#29 AR-1254-4

R.T.: 6.135 min
Delta R.T.: 0.000 min
Response: 59407243
Conc: 260.35 ng/ml



#29 AR-1254-4

R.T.: 5.286 min
Delta R.T.: 0.000 min
Response: 79268444
Conc: 259.87 ng/ml



#30 AR-1254-5

R.T.: 6.544 min
Delta R.T.: 0.000 min
Response: 62353894
Conc: 259.22 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 113088081
Conc: 260.51 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070994.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:15
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:31:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:31:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.777	18509464	21190782	4.213	4.074
2)	SA Decachlor...	8.497	7.533	13621454	26282104	4.302	4.713
	Target Compounds						
26)	L6 AR-1254-1	5.294	4.533	8706585	14758492	44.358	43.766
27)	L6 AR-1254-2	5.507	4.679	12975824	13428771	44.670	45.191
28)	L6 AR-1254-3	5.856	5.060	13388451	20784652	43.581	44.501
29)	L6 AR-1254-4	6.135	5.286	9249550	12805792	42.131	43.372
30)	L6 AR-1254-5	6.544	5.690	9864278	18267509	42.538	43.458

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070994.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:15
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:31:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:31:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 18509464
Conc: 4.21 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC050



#1 Tetrachloro-m-xylene

R.T.: 2.777 min
Delta R.T.: 0.000 min
Response: 21190782
Conc: 4.07 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 13621454
Conc: 4.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 26282104
Conc: 4.71 ng/ml



#26 AR-1254-1

R.T.: 5.294 min
Delta R.T.: 0.000 min
Response: 8706585
Conc: 44.36 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC050



#26 AR-1254-1

R.T.: 4.533 min
Delta R.T.: 0.000 min
Response: 14758492
Conc: 43.77 ng/ml



#27 AR-1254-2

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 12975824
Conc: 44.67 ng/ml



#27 AR-1254-2

R.T.: 4.679 min
Delta R.T.: 0.000 min
Response: 13428771
Conc: 45.19 ng/ml



#28 AR-1254-3

R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 13388451
Conc: 43.58 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC050



#28 AR-1254-3

R.T.: 5.060 min
Delta R.T.: 0.000 min
Response: 20784652
Conc: 44.50 ng/ml



#29 AR-1254-4

R.T.: 6.135 min
Delta R.T.: 0.000 min
Response: 9249550
Conc: 42.13 ng/ml



#29 AR-1254-4

R.T.: 5.286 min
Delta R.T.: 0.000 min
Response: 12805792
Conc: 43.37 ng/ml



#30 AR-1254-5

R.T.: 6.544 min
Delta R.T.: 0.000 min
Response: 9864278
Conc: 42.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC050



#30 AR-1254-5

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 18267509
Conc: 43.46 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070995.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:30
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:45:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:45:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	233.7E6	278.1E6	50.000	50.000
2)	SA Decachlor...	8.497	7.533	168.1E6	302.5E6	50.000	50.000
	Target Compounds						
36)	L8 AR-1262-1	6.841	5.733	137.0E6	226.2E6	500.000	500.000
37)	L8 AR-1262-2	7.146	5.979	249.2E6	213.9E6	500.000	500.000
38)	L8 AR-1262-3	7.419	6.500	164.3E6	166.4E6	500.000	500.000
39)	L8 AR-1262-4	7.492	6.561	126.2E6	308.4E6	500.000	500.000
40)	L8 AR-1262-5	7.988	7.051	82174920	151.5E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070995.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:30
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:45:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:45:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 233674057
Conc: 50.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1262ICC500



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 278097352
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 168129535
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 302467313
Conc: 50.00 ng/ml



#36 AR-1262-1

R.T.: 6.841 min
Delta R.T.: 0.000 min
Response: 137009240
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1262ICC500



#36 AR-1262-1

R.T.: 5.733 min
Delta R.T.: 0.000 min
Response: 226243354
Conc: 500.00 ng/ml



#37 AR-1262-2

R.T.: 7.146 min
Delta R.T.: 0.000 min
Response: 249238238
Conc: 500.00 ng/ml



#37 AR-1262-2

R.T.: 5.979 min
Delta R.T.: 0.000 min
Response: 213850677
Conc: 500.00 ng/ml



#38 AR-1262-3

R.T.: 7.419 min
Delta R.T.: 0.000 min
Response: 164329640
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1262ICC500



#38 AR-1262-3

R.T.: 6.500 min
Delta R.T.: 0.000 min
Response: 166353751
Conc: 500.00 ng/ml



#39 AR-1262-4

R.T.: 7.492 min
Delta R.T.: 0.000 min
Response: 126169859
Conc: 500.00 ng/ml



#39 AR-1262-4

R.T.: 6.561 min
Delta R.T.: 0.000 min
Response: 308409736
Conc: 500.00 ng/ml



#40 AR-1262-5

R.T.: 7.988 min
Delta R.T.: 0.000 min
Response: 82174920
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1262ICC500



#40 AR-1262-5

R.T.: 7.051 min
Delta R.T.: 0.000 min
Response: 151493238
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070996.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:44
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:44:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	473.5E6	558.2E6	98.121	98.662
2)	SA Decachlor...	8.496	7.533	575.8E6	1040.1E6	97.627	97.468
	Target Compounds						
41)	L9 AR-1268-1	7.418	6.501	555.6E6	931.5E6	975.616	977.850
42)	L9 AR-1268-2	7.494	6.563	510.3E6	858.9E6	977.260	976.287
43)	L9 AR-1268-3	7.676	6.763	416.5E6	746.0E6	976.432	977.977
44)	L9 AR-1268-4	7.988	7.051	179.4E6	330.2E6	972.218	977.102
45)	L9 AR-1268-5	8.271	7.318	1283.2E6	2374.7E6	985.329	987.978

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070996.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:44
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:44:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 473487442
Conc: 98.12 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC1000



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 558165948
Conc: 98.66 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 575751429
Conc: 97.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 1040108591
Conc: 97.47 ng/ml



#41 AR-1268-1

R.T.: 7.418 min
Delta R.T.: 0.000 min
Response: 555593291
Conc: 975.62 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC1000



#41 AR-1268-1

R.T.: 6.501 min
Delta R.T.: 0.000 min
Response: 931455367
Conc: 977.85 ng/ml



#42 AR-1268-2

R.T.: 7.494 min
Delta R.T.: 0.000 min
Response: 510325576
Conc: 977.26 ng/ml



#42 AR-1268-2

R.T.: 6.563 min
Delta R.T.: 0.000 min
Response: 858872879
Conc: 976.29 ng/ml



#43 AR-1268-3

R.T.: 7.676 min
Delta R.T.: 0.000 min
Response: 416538407
Conc: 976.43 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC1000



#43 AR-1268-3

R.T.: 6.763 min
Delta R.T.: 0.000 min
Response: 746016914
Conc: 977.98 ng/ml



#44 AR-1268-4

R.T.: 7.988 min
Delta R.T.: 0.000 min
Response: 179373252
Conc: 972.22 ng/ml



#44 AR-1268-4

R.T.: 7.051 min
Delta R.T.: 0.000 min
Response: 330162118
Conc: 977.10 ng/ml



#45 AR-1268-5

R.T.: 8.271 min
Delta R.T.: 0.000 min
Response: 1283167757
Conc: 985.33 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC1000



#45 AR-1268-5

R.T.: 7.318 min
Delta R.T.: 0.000 min
Response: 2374681378
Conc: 987.98 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:59
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:47:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	365.3E6	429.7E6	75.472	75.638
2)	SA Decachlor...	8.496	7.533	443.5E6	805.2E6	75.137	75.305
	Target Compounds						
41)	L9 AR-1268-1	7.417	6.501	429.3E6	721.3E6	752.594	754.792
42)	L9 AR-1268-2	7.495	6.564	393.0E6	668.4E6	751.776	756.465
43)	L9 AR-1268-3	7.677	6.763	320.6E6	577.3E6	751.101	754.528
44)	L9 AR-1268-4	7.988	7.051	139.1E6	254.6E6	752.792	752.326
45)	L9 AR-1268-5	8.270	7.318	981.0E6	1816.7E6	752.199	753.877

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:59
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:47:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 365345228
Conc: 75.47 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 429738765
Conc: 75.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 443523459
Conc: 75.14 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 805246413
Conc: 75.31 ng/ml



#41 AR-1268-1

R.T.: 7.417 min
Delta R.T.: 0.000 min
Response: 429329141
Conc: 752.59 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750



#41 AR-1268-1

R.T.: 6.501 min
Delta R.T.: 0.000 min
Response: 721284734
Conc: 754.79 ng/ml



#42 AR-1268-2

R.T.: 7.495 min
Delta R.T.: 0.000 min
Response: 393043003
Conc: 751.78 ng/ml



#42 AR-1268-2

R.T.: 6.564 min
Delta R.T.: 0.000 min
Response: 668369020
Conc: 756.47 ng/ml



#43 AR-1268-3

R.T.: 7.677 min
Delta R.T.: 0.000 min
Response: 320649152
Conc: 751.10 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750



#43 AR-1268-3

R.T.: 6.763 min
Delta R.T.: 0.000 min
Response: 577309161
Conc: 754.53 ng/ml



#44 AR-1268-4

R.T.: 7.988 min
Delta R.T.: 0.000 min
Response: 139148391
Conc: 752.79 ng/ml



#44 AR-1268-4

R.T.: 7.051 min
Delta R.T.: 0.000 min
Response: 254605180
Conc: 752.33 ng/ml



#45 AR-1268-5

R.T.: 8.270 min
Delta R.T.: 0.000 min
Response: 981007723
Conc: 752.20 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750



#45 AR-1268-5

R.T.: 7.318 min
Delta R.T.: 0.000 min
Response: 1816697371
Conc: 753.88 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070998.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:13
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:41:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.778	245.8E6	286.7E6	50.000	50.000
2)	SA Decachlor...	8.497	7.533	301.9E6	547.1E6	50.000	50.000
	Target Compounds						
41)	L9 AR-1268-1	7.418	6.501	291.7E6	486.8E6	500.000	500.000
42)	L9 AR-1268-2	7.496	6.564	267.0E6	450.3E6	500.000	500.000
43)	L9 AR-1268-3	7.677	6.763	218.3E6	389.8E6	500.000	500.000
44)	L9 AR-1268-4	7.988	7.051	94812460	172.8E6	500.000	500.000
45)	L9 AR-1268-5	8.271	7.318	660.7E6	1216.2E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070998.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:13
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:41:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.401 min
Delta R.T.: 0.000 min
Response: 245811296
Conc: 50.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC500



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 286653182
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 301868295
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 547076079
Conc: 50.00 ng/ml



#41 AR-1268-1

R.T.: 7.418 min
Delta R.T.: 0.000 min
Response: 291682733
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC500



#41 AR-1268-1

R.T.: 6.501 min
Delta R.T.: 0.000 min
Response: 486826354
Conc: 500.00 ng/ml



#42 AR-1268-2

R.T.: 7.496 min
Delta R.T.: 0.000 min
Response: 267037614
Conc: 500.00 ng/ml



#42 AR-1268-2

R.T.: 6.564 min
Delta R.T.: 0.000 min
Response: 450297458
Conc: 500.00 ng/ml



#43 AR-1268-3

R.T.: 7.677 min
Delta R.T.: 0.000 min
Response: 218323049
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC500



#43 AR-1268-3

R.T.: 6.763 min
Delta R.T.: 0.000 min
Response: 389807807
Conc: 500.00 ng/ml



#44 AR-1268-4

R.T.: 7.988 min
Delta R.T.: 0.000 min
Response: 94812460
Conc: 500.00 ng/ml



#44 AR-1268-4

R.T.: 7.051 min
Delta R.T.: 0.000 min
Response: 172818331
Conc: 500.00 ng/ml



#45 AR-1268-5

R.T.: 8.271 min
Delta R.T.: 0.000 min
Response: 660690181
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC500



#45 AR-1268-5

R.T.: 7.318 min
Delta R.T.: 0.000 min
Response: 1216236053
Conc: 500.00 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:28
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:50:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	124.1E6	145.7E6	25.476	25.482
2)	SA Decachlor...	8.496	7.533	154.8E6	281.7E6	25.909	25.997
	Target Compounds						
41)	L9 AR-1268-1	7.418	6.500	149.5E6	243.6E6	258.947	253.674
42)	L9 AR-1268-2	7.495	6.563	136.5E6	226.7E6	258.257	254.883
43)	L9 AR-1268-3	7.677	6.763	110.8E6	198.0E6	257.143	256.485
44)	L9 AR-1268-4	7.988	7.051	48417432	89098499	258.848	259.826
45)	L9 AR-1268-5	8.270	7.317	331.6E6	607.0E6	253.196	251.413

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:28
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:50:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 124113048
Conc: 25.48 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC250



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 145711594
Conc: 25.48 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 154811101
Conc: 25.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 281733949
Conc: 26.00 ng/ml



#41 AR-1268-1

R.T.: 7.418 min
Delta R.T.: 0.000 min
Response: 149503784
Conc: 258.95 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC250



#41 AR-1268-1

R.T.: 6.500 min
Delta R.T.: 0.000 min
Response: 243605830
Conc: 253.67 ng/ml



#42 AR-1268-2

R.T.: 7.495 min
Delta R.T.: 0.000 min
Response: 136525117
Conc: 258.26 ng/ml



#42 AR-1268-2

R.T.: 6.563 min
Delta R.T.: 0.000 min
Response: 226675271
Conc: 254.88 ng/ml



#43 AR-1268-3

R.T.: 7.677 min
Delta R.T.: 0.000 min
Response: 110831298
Conc: 257.14 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC250



#43 AR-1268-3

R.T.: 6.763 min
Delta R.T.: 0.000 min
Response: 197955027
Conc: 256.49 ng/ml



#44 AR-1268-4

R.T.: 7.988 min
Delta R.T.: 0.000 min
Response: 48417432
Conc: 258.85 ng/ml



#44 AR-1268-4

R.T.: 7.051 min
Delta R.T.: 0.000 min
Response: 89098499
Conc: 259.83 ng/ml



#45 AR-1268-5

R.T.: 8.270 min
Delta R.T.: 0.000 min
Response: 331628529
Conc: 253.20 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 7.317 min
Delta R.T.: 0.000 min
Response: 606999218
Conc: 251.41 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071000.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:43
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:16:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:16:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	24662420	28347182	5.050	4.966
2)	SA Decachlor...	8.497	7.533	31115212	59480232	5.165	5.383
	Target Compounds						
41)	L9 AR-1268-1	7.418	6.500	30187398	51549310	51.812	52.901
42)	L9 AR-1268-2	7.495	6.563	26994207	46812925	50.847	52.089
43)	L9 AR-1268-3	7.676	6.763	22122982	41071128	51.057	52.539
44)	L9 AR-1268-4	7.989	7.050	9507758	17547775	50.662	50.933
45)	L9 AR-1268-5	8.271	7.317	65947884	122.2E6	50.280	50.497

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071000.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:43
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:16:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:16:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 24662420
Conc: 5.05 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC050



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 28347182
Conc: 4.97 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 31115212
Conc: 5.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 59480232
Conc: 5.38 ng/ml



#41 AR-1268-1

R.T.: 7.418 min
Delta R.T.: 0.000 min
Response: 30187398
Conc: 51.81 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC050



#41 AR-1268-1

R.T.: 6.500 min
Delta R.T.: 0.000 min
Response: 51549310
Conc: 52.90 ng/ml



#42 AR-1268-2

R.T.: 7.495 min
Delta R.T.: 0.000 min
Response: 26994207
Conc: 50.85 ng/ml



#42 AR-1268-2

R.T.: 6.563 min
Delta R.T.: 0.000 min
Response: 46812925
Conc: 52.09 ng/ml



#43 AR-1268-3

R.T.: 7.676 min
Delta R.T.: 0.000 min
Response: 22122982
Conc: 51.06 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC050



#43 AR-1268-3

R.T.: 6.763 min
Delta R.T.: 0.000 min
Response: 41071128
Conc: 52.54 ng/ml



#44 AR-1268-4

R.T.: 7.989 min
Delta R.T.: 0.000 min
Response: 9507758
Conc: 50.66 ng/ml



#44 AR-1268-4

R.T.: 7.050 min
Delta R.T.: 0.000 min
Response: 17547775
Conc: 50.93 ng/ml



#45 AR-1268-5

R.T.: 8.271 min
Delta R.T.: 0.000 min
Response: 65947884
Conc: 50.28 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC050



#45 AR-1268-5

R.T.: 7.317 min
Delta R.T.: 0.000 min
Response: 122222176
Conc: 50.50 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:57
Operator : YP\AJ
Sample : PQ101025ICV
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:29:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:24:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	237.3E6	265.1E6	54.558	52.183
2)	SA Decachlor...	8.497	7.533	168.1E6	302.1E6	57.872	59.825
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.777	84536347	121.7E6	515.013	543.285m
4)	L1 AR-1016-2	4.499	3.792	137.9E6	157.0E6	537.936	461.544m
5)	L1 AR-1016-3	4.556	3.953	86370882	89605925	542.809	492.722m
6)	L1 AR-1016-4	4.646	4.001	72647096	77616175	559.128	493.075m
7)	L1 AR-1016-5	4.931	4.196	71678763	95334887	556.500	496.945m
31)	L7 AR-1260-1	6.019	5.188	124.6E6	179.0E6	531.360m	496.698m
32)	L7 AR-1260-2	6.275	5.377	157.7E6	216.4E6	569.990m	491.786m
33)	L7 AR-1260-3	6.627	5.518	115.0E6	195.1E6	576.989	470.083
34)	L7 AR-1260-4	6.840	5.979	126.1E6	186.2E6	586.767	517.623
35)	L7 AR-1260-5	7.146	6.224	259.3E6	433.0E6	597.827	515.349

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:57
Operator : YP\AJ
Sample : PQ101025ICV
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:29:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:24:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 237342010
Conc: 54.56 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 265088435
Conc: 52.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.001 min
Response: 168058544
Conc: 57.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 302090515
Conc: 59.83 ng/ml



#3 AR-1016-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 84536347
Conc: 515.01 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#3 AR-1016-1

R.T.: 3.777 min
Delta R.T.: -0.001 min
Response: 121697010
Conc: 543.29 ng/ml m



#4 AR-1016-2

R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 137870017
Conc: 537.94 ng/ml



#4 AR-1016-2

R.T.: 3.792 min
Delta R.T.: -0.001 min
Response: 156981599
Conc: 461.54 ng/ml m



#5 AR-1016-3

R.T.: 4.556 min
Delta R.T.: 0.000 min
Response: 86370882
Conc: 542.81 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#5 AR-1016-3

R.T.: 3.953 min
Delta R.T.: 0.000 min
Response: 89605925
Conc: 492.72 ng/ml m



#6 AR-1016-4

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 72647096
Conc: 559.13 ng/ml



#6 AR-1016-4

R.T.: 4.001 min
Delta R.T.: 0.000 min
Response: 77616175
Conc: 493.07 ng/ml m



#7 AR-1016-5

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 71678763
Conc: 556.50 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#7 AR-1016-5

R.T.: 4.196 min
Delta R.T.: 0.000 min
Response: 95334887
Conc: 496.95 ng/ml m



#31 AR-1260-1

R.T.: 6.019 min
Delta R.T.: 0.000 min
Response: 124588657
Conc: 531.36 ng/ml m



#31 AR-1260-1

R.T.: 5.188 min
Delta R.T.: -0.002 min
Response: 179013037
Conc: 496.70 ng/ml m



#32 AR-1260-2

R.T.: 6.275 min
Delta R.T.: 0.000 min
Response: 157664225
Conc: 569.99 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#32 AR-1260-2

R.T.: 5.377 min
Delta R.T.: 0.000 min
Response: 216382734
Conc: 491.79 ng/ml m



#33 AR-1260-3

R.T.: 6.627 min
Delta R.T.: 0.000 min
Response: 114998797
Conc: 576.99 ng/ml



#33 AR-1260-3

R.T.: 5.518 min
Delta R.T.: 0.000 min
Response: 195084285
Conc: 470.08 ng/ml



#34 AR-1260-4

R.T.: 6.840 min
Delta R.T.: 0.000 min
Response: 126086390
Conc: 586.77 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#34 AR-1260-4

R.T.: 5.979 min
Delta R.T.: 0.000 min
Response: 186221665
Conc: 517.62 ng/ml



#35 AR-1260-5

R.T.: 7.146 min
Delta R.T.: 0.001 min
Response: 259323595
Conc: 597.83 ng/ml



#35 AR-1260-5

R.T.: 6.224 min
Delta R.T.: 0.000 min
Response: 433038147
Conc: 515.35 ng/ml

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 17:41
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:19:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:16:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	255.6E6	301.0E6	55.692	54.076
2)	SA Decachlor...	8.496	7.534	182.2E6	328.4E6	55.793	58.440
	Target Compounds						
16)	L4 AR-1242-1	4.480	3.778	67446984	88802083	567.010	545.258
17)	L4 AR-1242-2	4.499	3.793	106.5E6	130.9E6	558.407	542.137
18)	L4 AR-1242-3	4.556	3.952	65846450	69528801	553.814	527.451m
19)	L4 AR-1242-4	4.646	4.039	55324188	66246241	555.710	578.219
20)	L4 AR-1242-5	5.349	4.534	35965359	88786139	456.866	509.855

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 17:41
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:19:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:16:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 255633420
Conc: 55.69 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 301010208
Conc: 54.08 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 182166881
Conc: 55.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 328407221
Conc: 58.44 ng/ml



#16 AR-1242-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 67446984
Conc: 567.01 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#16 AR-1242-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 88802083
Conc: 545.26 ng/ml



#17 AR-1242-2

R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 106495313
Conc: 558.41 ng/ml



#17 AR-1242-2

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 130858283
Conc: 542.14 ng/ml



#18 AR-1242-3

R.T.: 4.556 min
Delta R.T.: 0.000 min
Response: 65846450
Conc: 553.81 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#18 AR-1242-3

R.T.: 3.952 min
Delta R.T.: -0.002 min
Response: 69528801
Conc: 527.45 ng/ml m



#19 AR-1242-4

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 55324188
Conc: 555.71 ng/ml



#19 AR-1242-4

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 66246241
Conc: 578.22 ng/ml



#20 AR-1242-5

R.T.: 5.349 min
Delta R.T.: -0.002 min
Response: 35965359
Conc: 456.87 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#20 AR-1242-5

R.T.: 4.534 min
Delta R.T.: 0.000 min
Response: 88786139
Conc: 509.86 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 18:25
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:26:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:23:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	253.0E6	292.6E6	54.792	52.505
2)	SA Decachlor...	8.497	7.534	179.8E6	324.3E6	54.114	56.854
	Target Compounds						
21)	L5 AR-1248-1	4.480	3.777	47223685	65282112	527.847	516.093
22)	L5 AR-1248-2	4.743	4.002	73173417	97020820	536.277	514.526
23)	L5 AR-1248-3	4.931	4.039	87732864	87534945	543.629	487.239
24)	L5 AR-1248-4	5.317	4.196	98470443	117.2E6	567.576m	522.551m
25)	L5 AR-1248-5	5.349	4.571	59958766	126.5E6	402.630m	543.081 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 18:25
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:26:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:23:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 253021754
Conc: 54.79 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 292628939
Conc: 52.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 179847567
Conc: 54.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.001 min
Response: 324345066
Conc: 56.85 ng/ml



#21 AR-1248-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 47223685
Conc: 527.85 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#21 AR-1248-1

R.T.: 3.777 min
Delta R.T.: 0.000 min
Response: 65282112
Conc: 516.09 ng/ml



#22 AR-1248-2

R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 73173417
Conc: 536.28 ng/ml



#22 AR-1248-2

R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 97020820
Conc: 514.53 ng/ml



#23 AR-1248-3

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 87732864
Conc: 543.63 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#23 AR-1248-3

R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 87534945
Conc: 487.24 ng/ml



#24 AR-1248-4

R.T.: 5.317 min
Delta R.T.: 0.000 min
Response: 98470443
Conc: 567.58 ng/ml m



#24 AR-1248-4

R.T.: 4.196 min
Delta R.T.: 0.000 min
Response: 117224408
Conc: 522.55 ng/ml m



#25 AR-1248-5

R.T.: 5.349 min
Delta R.T.: -0.002 min
Response: 59958766
Conc: 402.63 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025



#25 AR-1248-5

R.T.: 4.571 min
Delta R.T.: 0.000 min
Response: 126452467
Conc: 543.08 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071004.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 19:08
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:34:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:30:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	247.0E6	287.2E6	56.214	55.217
2)	SA Decachlor...	8.498	7.533	175.2E6	316.3E6	55.323	56.717
	Target Compounds						
26)	L6 AR-1254-1	5.293	4.533	115.3E6	191.1E6	587.461	566.643
27)	L6 AR-1254-2	5.507	4.680	167.9E6	162.6E6	578.152	547.137
28)	L6 AR-1254-3	5.855	5.060	176.3E6	253.6E6	573.790	542.980
29)	L6 AR-1254-4	6.134	5.286	124.7E6	161.4E6	568.016	546.628
30)	L6 AR-1254-5	6.544	5.690	130.6E6	228.4E6	563.253	543.396

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071004.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 19:08
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:34:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:30:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 246956460
Conc: 56.21 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1254



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 287227626
Conc: 55.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.498 min
Delta R.T.: 0.000 min
Response: 175150667
Conc: 55.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 316297966
Conc: 56.72 ng/ml



#26 AR-1254-1

R.T.: 5.293 min
Delta R.T.: 0.000 min
Response: 115305910
Conc: 587.46 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1254



#26 AR-1254-1

R.T.: 4.533 min
Delta R.T.: 0.000 min
Response: 191078779
Conc: 566.64 ng/ml



#27 AR-1254-2

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 167943478
Conc: 578.15 ng/ml



#27 AR-1254-2

R.T.: 4.680 min
Delta R.T.: 0.000 min
Response: 162584096
Conc: 547.14 ng/ml



#28 AR-1254-3

R.T.: 5.855 min
Delta R.T.: 0.000 min
Response: 176272565
Conc: 573.79 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1254



#28 AR-1254-3

R.T.: 5.060 min
Delta R.T.: 0.000 min
Response: 253607279
Conc: 542.98 ng/ml



#29 AR-1254-4

R.T.: 6.134 min
Delta R.T.: -0.001 min
Response: 124702483
Conc: 568.02 ng/ml



#29 AR-1254-4

R.T.: 5.286 min
Delta R.T.: 0.000 min
Response: 161392628
Conc: 546.63 ng/ml



#30 AR-1254-5

R.T.: 6.544 min
Delta R.T.: 0.000 min
Response: 130615840
Conc: 563.25 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1254



#30 AR-1254-5

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 228416064
Conc: 543.40 ng/ml

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071005.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 19:52
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:39:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:37:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.777	256.6E6	296.1E6	52.549	51.872
2)	SA Decachlor...	8.497	7.533	317.0E6	560.4E6	52.621	50.721
	Target Compounds						
41)	L9 AR-1268-1	7.418	6.500	307.3E6	498.7E6	527.367	511.825
42)	L9 AR-1268-2	7.495	6.563	283.4E6	464.0E6	533.785	516.248
43)	L9 AR-1268-3	7.677	6.763	233.9E6	395.9E6	539.709	506.489
44)	L9 AR-1268-4	7.989	7.051	100.4E6	175.1E6	534.727	508.315
45)	L9 AR-1268-5	8.271	7.318	696.8E6	1236.0E6	531.266	510.673

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071005.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 19:52
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:39:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:37:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.401 min
Delta R.T.: 0.000 min
Response: 256641615
Conc: 52.55 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1268



#1 Tetrachloro-m-xylene

R.T.: 2.777 min
Delta R.T.: 0.000 min
Response: 296113428
Conc: 51.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 317031582
Conc: 52.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 560415145
Conc: 50.72 ng/ml



#41 AR-1268-1

R.T.: 7.418 min
Delta R.T.: 0.000 min
Response: 307261040
Conc: 527.37 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1268



#41 AR-1268-1

R.T.: 6.500 min
Delta R.T.: 0.000 min
Response: 498745928
Conc: 511.83 ng/ml



#42 AR-1268-2

R.T.: 7.495 min
Delta R.T.: 0.000 min
Response: 283380161
Conc: 533.78 ng/ml



#42 AR-1268-2

R.T.: 6.563 min
Delta R.T.: 0.000 min
Response: 463960839
Conc: 516.25 ng/ml



#43 AR-1268-3

R.T.: 7.677 min
Delta R.T.: 0.000 min
Response: 233855867
Conc: 539.71 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1268



#43 AR-1268-3

R.T.: 6.763 min
Delta R.T.: 0.000 min
Response: 395934354
Conc: 506.49 ng/ml



#44 AR-1268-4

R.T.: 7.989 min
Delta R.T.: 0.000 min
Response: 100352602
Conc: 534.73 ng/ml



#44 AR-1268-4

R.T.: 7.051 min
Delta R.T.: 0.000 min
Response: 175126997
Conc: 508.32 ng/ml



#45 AR-1268-5

R.T.: 8.271 min
Delta R.T.: 0.000 min
Response: 696811272
Conc: 531.27 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1268



#45 AR-1268-5

R.T.: 7.318 min
Delta R.T.: 0.000 min
Response: 1236017444
Conc: 510.67 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Continuing Calib Date: 10/20/2025 **Initial Calibration Date(s):** 09/23/2025 09/23/2025

Continuing Calib Time: 11:13 **Initial Calibration Time(s):** 08:45 17:14

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	5.81	5.81	5.71	5.91	0.00
Aroclor-1016-2	(2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-3	(3)	5.89	5.89	5.79	5.99	0.00
Aroclor-1016-4	(4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1016-5	(5)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-1	(1)	7.40	7.40	7.30	7.50	0.00
Aroclor-1260-2	(2)	7.65	7.65	7.55	7.75	0.00
Aroclor-1260-3	(3)	8.01	8.01	7.91	8.11	0.00
Aroclor-1260-4	(4)	8.24	8.24	8.14	8.34	0.00
Aroclor-1260-5	(5)	8.56	8.56	8.46	8.66	0.00
Tetrachloro-m-xylene		4.65	4.66	4.56	4.76	0.01
Decachlorobiphenyl		10.44	10.43	10.33	10.53	-0.01

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Continuing Calib Date: 10/20/2025 **Initial Calibration Date(s):** 09/23/2025 09/23/2025

Continuing Calib Time: 11:13 **Initial Calibration Time(s):** 08:45 17:14

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.89	4.89	4.79	4.99	0.00
Aroclor-1016-2	(2)	4.94	4.95	4.85	5.05	0.01
Aroclor-1016-3	(3)	5.06	5.07	4.97	5.17	0.01
Aroclor-1016-4	(4)	5.10	5.11	5.01	5.21	0.01
Aroclor-1016-5	(5)	5.32	5.33	5.23	5.43	0.01
Aroclor-1260-1	(1)	6.34	6.36	6.26	6.46	0.02
Aroclor-1260-2	(2)	6.53	6.54	6.44	6.64	0.01
Aroclor-1260-3	(3)	6.69	6.70	6.60	6.80	0.02
Aroclor-1260-4	(4)	7.15	7.17	7.07	7.27	0.02
Aroclor-1260-5	(5)	7.39	7.40	7.30	7.50	0.01
Tetrachloro-m-xylene		3.79	3.80	3.70	3.90	0.01
Decachlorobiphenyl		8.79	8.80	8.70	8.90	0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/20/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP075868.D **Time Analyzed:** 11:13

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.806	5.708	5.908	472.530	500.000	-5.5
Aroclor-1016-2	5.827	5.729	5.929	502.350	500.000	0.5
Aroclor-1016-3	5.890	5.792	5.992	485.700	500.000	-2.9
Aroclor-1016-4	5.987	5.889	6.089	509.420	500.000	1.9
Aroclor-1016-5	6.279	6.181	6.381	520.380	500.000	4.1
Aroclor-1260-1	7.397	7.299	7.499	482.450	500.000	-3.5
Aroclor-1260-2	7.650	7.551	7.751	470.250	500.000	-6.0
Aroclor-1260-3	8.008	7.909	8.109	501.460	500.000	0.3
Aroclor-1260-4	8.236	8.137	8.337	492.140	500.000	-1.6
Aroclor-1260-5	8.563	8.464	8.664	502.900	500.000	0.6
Decachlorobiphenyl	10.435	10.334	10.534	40.440	50.000	-19.1
Tetrachloro-m-xylene	4.653	4.555	4.755	47.380	50.000	-5.2

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/20/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP075868.D **Time Analyzed:** 11:13

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.885	4.793	4.993	540.620	500.000	8.1
Aroclor-1016-2	4.941	4.851	5.051	571.750	500.000	14.4
Aroclor-1016-3	5.064	4.971	5.171	520.270	500.000	4.1
Aroclor-1016-4	5.101	5.011	5.211	573.190	500.000	14.6
Aroclor-1016-5	5.316	5.225	5.425	541.560	500.000	8.3
Aroclor-1260-1	6.344	6.255	6.455	542.970	500.000	8.6
Aroclor-1260-2	6.532	6.441	6.641	549.380	500.000	9.9
Aroclor-1260-3	6.685	6.595	6.795	527.020	500.000	5.4
Aroclor-1260-4	7.153	7.065	7.265	484.910	500.000	-3.0
Aroclor-1260-5	7.394	7.304	7.504	458.580	500.000	-8.3
Decachlorobiphenyl	8.791	8.703	8.903	43.820	50.000	-12.4
Tetrachloro-m-xylene	3.789	3.696	3.896	51.660	50.000	3.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102025\
Data File : PP075868.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 11:13
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 23:52:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.653	3.789	61991095	315.2E6	47.376	51.660
2)	SA Decachlor...	10.435	8.791	40344705	264.3E6	40.444	43.819
	Target Compounds						
3)	L1 AR-1016-1	5.806	4.885	22730879	303.3E6	472.529	540.617m
4)	L1 AR-1016-2	5.827	4.941	36032526	146.9E6	502.345	571.746m
5)	L1 AR-1016-3	5.890	5.064	22600057	80781382	485.701	520.269
6)	L1 AR-1016-4	5.987	5.101	18153915	84241158	509.417	573.194m
7)	L1 AR-1016-5	6.279	5.316	17547729	88481741	520.375	541.563m
31)	L7 AR-1260-1	7.397	6.344	29004625	220.4E6	482.450	542.971
32)	L7 AR-1260-2	7.650	6.532	34233516	316.3E6	470.254	549.376
33)	L7 AR-1260-3	8.008	6.685	27151845	225.6E6	501.464	527.025
34)	L7 AR-1260-4	8.236	7.153	27824244	209.1E6	492.141	484.908
35)	L7 AR-1260-5	8.563	7.394	55222145	482.4E6	502.896	458.583

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102025\
Data File : PP075868.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 11:13
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 23:52:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.653 min
Delta R.T.: -0.002 min
Response: 61991095
Conc: 47.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#1 Tetrachloro-m-xylene

R.T.: 3.789 min
Delta R.T.: -0.007 min
Response: 315170942
Conc: 51.66 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.435 min
Delta R.T.: 0.000 min
Response: 40344705
Conc: 40.44 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.791 min
Delta R.T.: -0.013 min
Response: 264337884
Conc: 43.82 ng/ml



#3 AR-1016-1

R.T.: 5.806 min
Delta R.T.: -0.002 min
Response: 22730879
Conc: 472.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#3 AR-1016-1

R.T.: 4.885 min
Delta R.T.: -0.008 min
Response: 303293339
Conc: 540.62 ng/ml m



#4 AR-1016-2

R.T.: 5.827 min
Delta R.T.: -0.002 min
Response: 36032526
Conc: 502.35 ng/ml



#4 AR-1016-2

R.T.: 4.941 min
Delta R.T.: -0.010 min
Response: 146940045
Conc: 571.75 ng/ml m



#5 AR-1016-3

R.T.: 5.890 min
Delta R.T.: -0.002 min
Response: 22600057
Conc: 485.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#5 AR-1016-3

R.T.: 5.064 min
Delta R.T.: -0.007 min
Response: 80781382
Conc: 520.27 ng/ml



#6 AR-1016-4

R.T.: 5.987 min
Delta R.T.: -0.002 min
Response: 18153915
Conc: 509.42 ng/ml



#6 AR-1016-4

R.T.: 5.101 min
Delta R.T.: -0.010 min
Response: 84241158
Conc: 573.19 ng/ml m



#7 AR-1016-5

R.T.: 6.279 min
Delta R.T.: -0.002 min
Response: 17547729
Conc: 520.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#7 AR-1016-5

R.T.: 5.316 min
Delta R.T.: -0.009 min
Response: 88481741
Conc: 541.56 ng/ml m



#31 AR-1260-1

R.T.: 7.397 min
Delta R.T.: -0.002 min
Response: 29004625
Conc: 482.45 ng/ml



#31 AR-1260-1

R.T.: 6.344 min
Delta R.T.: -0.011 min
Response: 220411286
Conc: 542.97 ng/ml



#32 AR-1260-2

R.T.: 7.650 min
Delta R.T.: -0.002 min
Response: 34233516
Conc: 470.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#32 AR-1260-2

R.T.: 6.532 min
Delta R.T.: -0.009 min
Response: 316346977
Conc: 549.38 ng/ml



#33 AR-1260-3

R.T.: 8.008 min
Delta R.T.: -0.001 min
Response: 27151845
Conc: 501.46 ng/ml



#33 AR-1260-3

R.T.: 6.685 min
Delta R.T.: -0.010 min
Response: 225611434
Conc: 527.02 ng/ml



#34 AR-1260-4

R.T.: 8.236 min
Delta R.T.: -0.001 min
Response: 27824244
Conc: 492.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#34 AR-1260-4

R.T.: 7.153 min
Delta R.T.: -0.011 min
Response: 209104061
Conc: 484.91 ng/ml



#35 AR-1260-5

R.T.: 8.563 min
Delta R.T.: 0.000 min
Response: 55222145
Conc: 502.90 ng/ml



#35 AR-1260-5

R.T.: 7.394 min
Delta R.T.: -0.010 min
Response: 482423846
Conc: 458.58 ng/ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Continuing Calib Date: 10/20/2025 **Initial Calibration Date(s):** 09/23/2025 09/23/2025

Continuing Calib Time: 16:20 **Initial Calibration Time(s):** 08:45 17:14

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	5.81	5.81	5.71	5.91	0.00
Aroclor-1016-2	(2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-3	(3)	5.89	5.89	5.79	5.99	0.00
Aroclor-1016-4	(4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1016-5	(5)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-1	(1)	7.40	7.40	7.30	7.50	0.00
Aroclor-1260-2	(2)	7.65	7.65	7.55	7.75	0.00
Aroclor-1260-3	(3)	8.01	8.01	7.91	8.11	0.00
Aroclor-1260-4	(4)	8.24	8.24	8.14	8.34	0.00
Aroclor-1260-5	(5)	8.56	8.56	8.46	8.66	0.00
Tetrachloro-m-xylene		4.65	4.66	4.56	4.76	0.01
Decachlorobiphenyl		10.43	10.43	10.33	10.53	0.00

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Continuing Calib Date: 10/20/2025 **Initial Calibration Date(s):** 09/23/2025 09/23/2025

Continuing Calib Time: 16:20 **Initial Calibration Time(s):** 08:45 17:14

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.89	4.89	4.79	4.99	0.00
Aroclor-1016-2	(2)	4.94	4.95	4.85	5.05	0.01
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.11	5.11	5.01	5.21	0.00
Aroclor-1016-5	(5)	5.32	5.33	5.23	5.43	0.01
Aroclor-1260-1	(1)	6.35	6.36	6.26	6.46	0.01
Aroclor-1260-2	(2)	6.53	6.54	6.44	6.64	0.01
Aroclor-1260-3	(3)	6.69	6.70	6.60	6.80	0.01
Aroclor-1260-4	(4)	7.16	7.17	7.07	7.27	0.01
Aroclor-1260-5	(5)	7.40	7.40	7.30	7.50	0.00
Tetrachloro-m-xylene		3.79	3.80	3.70	3.90	0.01
Decachlorobiphenyl		8.79	8.80	8.70	8.90	0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/20/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP075881.D **Time Analyzed:** 16:20

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.807	5.708	5.908	499.980	500.000	0.0
Aroclor-1016-2	5.828	5.729	5.929	503.740	500.000	0.7
Aroclor-1016-3	5.891	5.792	5.992	505.040	500.000	1.0
Aroclor-1016-4	5.988	5.889	6.089	527.610	500.000	5.5
Aroclor-1016-5	6.280	6.181	6.381	521.200	500.000	4.2
Aroclor-1260-1	7.398	7.299	7.499	464.050	500.000	-7.2
Aroclor-1260-2	7.651	7.551	7.751	450.930	500.000	-9.8
Aroclor-1260-3	8.009	7.909	8.109	475.600	500.000	-4.9
Aroclor-1260-4	8.238	8.137	8.337	484.630	500.000	-3.1
Aroclor-1260-5	8.563	8.464	8.664	488.580	500.000	-2.3
Decachlorobiphenyl	10.433	10.334	10.534	39.770	50.000	-20.5
Tetrachloro-m-xylene	4.653	4.555	4.755	46.220	50.000	-7.6

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/20/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP075881.D **Time Analyzed:** 16:20

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.888	4.793	4.993	580.620	500.000	16.1
Aroclor-1016-2	4.943	4.851	5.051	632.020	500.000	26.4
Aroclor-1016-3	5.065	4.971	5.171	570.020	500.000	14.0
Aroclor-1016-4	5.106	5.011	5.211	566.310	500.000	13.3
Aroclor-1016-5	5.320	5.225	5.425	555.760	500.000	11.2
Aroclor-1260-1	6.348	6.255	6.455	483.300	500.000	-3.3
Aroclor-1260-2	6.534	6.441	6.641	518.490	500.000	3.7
Aroclor-1260-3	6.688	6.595	6.795	495.820	500.000	-0.8
Aroclor-1260-4	7.157	7.065	7.265	428.480	500.000	-14.3
Aroclor-1260-5	7.397	7.304	7.504	417.660	500.000	-16.5
Decachlorobiphenyl	8.792	8.703	8.903	41.820	50.000	-16.4
Tetrachloro-m-xylene	3.791	3.696	3.896	53.080	50.000	6.2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102025\
Data File : PP075881.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 16:20
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 23:58:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.653	3.791	60483863	323.8E6	46.224m	53.077
2)	SA Decachlor...	10.433	8.792	39669709	252.3E6	39.767m	41.823
	Target Compounds						
3)	L1 AR-1016-1	5.807	4.888	24051195	325.7E6	499.975	580.615
4)	L1 AR-1016-2	5.828	4.943	36132755	162.4E6	503.743	632.016m#
5)	L1 AR-1016-3	5.891	5.065	23500084	88505488	505.044	570.016
6)	L1 AR-1016-4	5.988	5.106	18802297	83229809	527.611	566.313
7)	L1 AR-1016-5	6.280	5.320	17575524	90801347	521.199	555.760
31)	L7 AR-1260-1	7.398	6.348	27898456	196.2E6	464.050	483.303
32)	L7 AR-1260-2	7.651	6.534	32826990	298.6E6	450.933	518.488
33)	L7 AR-1260-3	8.009	6.688	25751600	212.3E6	475.603	495.819
34)	L7 AR-1260-4	8.238	7.157	27399510	184.8E6	484.628	428.480
35)	L7 AR-1260-5	8.563	7.397	53649867	439.4E6	488.578	417.657

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102025\
Data File : PP075881.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 16:20
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 23:58:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.653 min
Delta R.T.: -0.002 min
Response: 60483863
Conc: 46.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#1 Tetrachloro-m-xylene

R.T.: 3.791 min
Delta R.T.: -0.004 min
Response: 323818182
Conc: 53.08 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.433 min
Delta R.T.: -0.002 min
Response: 39669709
Conc: 39.77 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.792 min
Delta R.T.: -0.012 min
Response: 252296517
Conc: 41.82 ng/ml



#3 AR-1016-1

R.T.: 5.807 min
Delta R.T.: 0.000 min
Response: 24051195
Conc: 499.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#3 AR-1016-1

R.T.: 4.888 min
Delta R.T.: -0.004 min
Response: 325732929
Conc: 580.62 ng/ml



#4 AR-1016-2

R.T.: 5.828 min
Delta R.T.: 0.000 min
Response: 36132755
Conc: 503.74 ng/ml



#4 AR-1016-2

R.T.: 4.943 min
Delta R.T.: -0.008 min
Response: 162429563
Conc: 632.02 ng/ml m



#5 AR-1016-3

R.T.: 5.891 min
Delta R.T.: 0.000 min
Response: 23500084
Conc: 505.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#5 AR-1016-3

R.T.: 5.065 min
Delta R.T.: -0.006 min
Response: 88505488
Conc: 570.02 ng/ml



#6 AR-1016-4

R.T.: 5.988 min
Delta R.T.: -0.001 min
Response: 18802297
Conc: 527.61 ng/ml



#6 AR-1016-4

R.T.: 5.106 min
Delta R.T.: -0.005 min
Response: 83229809
Conc: 566.31 ng/ml



#7 AR-1016-5

R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 17575524
Conc: 521.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#7 AR-1016-5

R.T.: 5.320 min
Delta R.T.: -0.005 min
Response: 90801347
Conc: 555.76 ng/ml



#31 AR-1260-1

R.T.: 7.398 min
Delta R.T.: 0.000 min
Response: 27898456
Conc: 464.05 ng/ml



#31 AR-1260-1

R.T.: 6.348 min
Delta R.T.: -0.007 min
Response: 196189925
Conc: 483.30 ng/ml



#32 AR-1260-2

R.T.: 7.651 min
Delta R.T.: 0.000 min
Response: 32826990
Conc: 450.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#32 AR-1260-2

R.T.: 6.534 min
Delta R.T.: -0.007 min
Response: 298561079
Conc: 518.49 ng/ml



#33 AR-1260-3

R.T.: 8.009 min
Delta R.T.: 0.000 min
Response: 25751600
Conc: 475.60 ng/ml



#33 AR-1260-3

R.T.: 6.688 min
Delta R.T.: -0.007 min
Response: 212252787
Conc: 495.82 ng/ml



#34 AR-1260-4

R.T.: 8.238 min
Delta R.T.: 0.000 min
Response: 27399510
Conc: 484.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#34 AR-1260-4

R.T.: 7.157 min
Delta R.T.: -0.008 min
Response: 184770745
Conc: 428.48 ng/ml



#35 AR-1260-5

R.T.: 8.563 min
Delta R.T.: 0.000 min
Response: 53649867
Conc: 488.58 ng/ml



#35 AR-1260-5

R.T.: 7.397 min
Delta R.T.: -0.007 min
Response: 439370400
Conc: 417.66 ng/ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Continuing Calib Date: 10/20/2025 **Initial Calibration Date(s):** 10/10/2025 10/10/2025

Continuing Calib Time: 10:18 **Initial Calibration Time(s):** 08:57 16:43

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.49	4.48	4.38	4.58	-0.01
Aroclor-1016-2	(2)	4.51	4.50	4.40	4.60	-0.01
Aroclor-1016-3	(3)	4.57	4.56	4.46	4.66	-0.01
Aroclor-1016-4	(4)	4.66	4.65	4.55	4.75	-0.01
Aroclor-1016-5	(5)	4.94	4.93	4.83	5.03	-0.01
Aroclor-1260-1	(1)	6.03	6.02	5.92	6.12	-0.01
Aroclor-1260-2	(2)	6.29	6.28	6.18	6.38	-0.01
Aroclor-1260-3	(3)	6.64	6.63	6.53	6.73	-0.01
Aroclor-1260-4	(4)	6.85	6.84	6.74	6.94	-0.01
Aroclor-1260-5	(5)	7.16	7.15	7.05	7.25	-0.01
Tetrachloro-m-xylene		3.41	3.40	3.30	3.50	-0.01
Decachlorobiphenyl		8.51	8.50	8.40	8.60	-0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Continuing Calib Date: 10/20/2025 **Initial Calibration Date(s):** 10/10/2025 10/10/2025

Continuing Calib Time: 10:18 **Initial Calibration Time(s):** 08:57 16:43

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1016-2	(2)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-3	(3)	3.95	3.95	3.85	4.05	0.00
Aroclor-1016-4	(4)	4.00	4.00	3.90	4.10	0.00
Aroclor-1016-5	(5)	4.20	4.20	4.10	4.30	0.00
Aroclor-1260-1	(1)	5.19	5.19	5.09	5.29	0.00
Aroclor-1260-2	(2)	5.38	5.38	5.28	5.48	0.00
Aroclor-1260-3	(3)	5.52	5.52	5.42	5.62	0.00
Aroclor-1260-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1260-5	(5)	6.23	6.22	6.12	6.32	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.54	7.53	7.43	7.63	-0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/20/2025

Lab Sample No.: AR1660CCC500 **Data File :** PQ071008.D **Time Analyzed:** 10:18

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.492	4.380	4.580	479.650	500.000	-4.1
Aroclor-1016-2	4.510	4.399	4.599	508.480	500.000	1.7
Aroclor-1016-3	4.567	4.456	4.656	536.740	500.000	7.3
Aroclor-1016-4	4.658	4.546	4.746	515.560	500.000	3.1
Aroclor-1016-5	4.942	4.831	5.031	511.360	500.000	2.3
Aroclor-1260-1	6.032	5.919	6.119	501.670	500.000	0.3
Aroclor-1260-2	6.288	6.175	6.375	472.070	500.000	-5.6
Aroclor-1260-3	6.638	6.526	6.726	523.450	500.000	4.7
Aroclor-1260-4	6.853	6.740	6.940	530.680	500.000	6.1
Aroclor-1260-5	7.158	7.045	7.245	493.240	500.000	-1.4
Decachlorobiphenyl	8.512	8.396	8.596	53.300	50.000	6.6
Tetrachloro-m-xylene	3.409	3.300	3.500	53.000	50.000	6.0

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/20/2025

Lab Sample No.: AR1660CCC500 **Data File :** PQ071008.D **Time Analyzed:** 10:18

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.778	3.678	3.878	419.270	500.000	-16.1
Aroclor-1016-2	3.793	3.693	3.893	465.010	500.000	-7.0
Aroclor-1016-3	3.953	3.853	4.053	468.600	500.000	-6.3
Aroclor-1016-4	4.001	3.902	4.102	489.820	500.000	-2.0
Aroclor-1016-5	4.196	4.097	4.297	467.260	500.000	-6.5
Aroclor-1260-1	5.189	5.089	5.289	468.620	500.000	-6.3
Aroclor-1260-2	5.379	5.278	5.478	452.340	500.000	-9.5
Aroclor-1260-3	5.520	5.419	5.619	466.410	500.000	-6.7
Aroclor-1260-4	5.981	5.879	6.079	464.170	500.000	-7.2
Aroclor-1260-5	6.225	6.124	6.324	437.130	500.000	-12.6
Decachlorobiphenyl	7.536	7.433	7.633	54.320	50.000	8.6
Tetrachloro-m-xylene	2.776	2.678	2.878	50.390	50.000	0.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071008.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 10:18
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:21:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.409	2.776	230.6E6	256.0E6	52.997	50.389
2)	SA Decachlor...	8.512	7.536	154.8E6	274.3E6	53.301	54.319
	Target Compounds						
3)	L1 AR-1016-1	4.492	3.778	78731379	93916390	479.648	419.266
4)	L1 AR-1016-2	4.510	3.793	130.3E6	158.2E6	508.477m	465.009
5)	L1 AR-1016-3	4.567	3.953	85404633	85218692	536.736	468.598
6)	L1 AR-1016-4	4.658	4.001	66986154	77104461	515.559	489.824
7)	L1 AR-1016-5	4.942	4.196	65865182	89640872	511.364	467.264
31)	L7 AR-1260-1	6.032	5.189	117.6E6	168.9E6	501.671	468.625
32)	L7 AR-1260-2	6.288	5.379	130.6E6	199.0E6	472.068	452.345
33)	L7 AR-1260-3	6.638	5.520	104.3E6	193.6E6	523.451	466.408
34)	L7 AR-1260-4	6.853	5.981	114.0E6	167.0E6	530.677	464.171
35)	L7 AR-1260-5	7.158	6.225	214.0E6	367.3E6	493.241	437.132

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071008.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 10:18
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:21:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.409 min
Delta R.T.: 0.009 min
Response: 230552166
Conc: 53.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025



#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.002 min
Response: 255977124
Conc: 50.39 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.512 min
Delta R.T.: 0.016 min
Response: 154786652
Conc: 53.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.536 min
Delta R.T.: 0.003 min
Response: 274287288
Conc: 54.32 ng/ml



#3 AR-1016-1

R.T.: 4.492 min
Delta R.T.: 0.011 min
Response: 78731379
Conc: 479.65 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025



#3 AR-1016-1

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 93916390
Conc: 419.27 ng/ml



#4 AR-1016-2

R.T.: 4.510 min
Delta R.T.: 0.011 min
Response: 130319752
Conc: 508.48 ng/ml m



#4 AR-1016-2

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 158160387
Conc: 465.01 ng/ml



#5 AR-1016-3

R.T.: 4.567 min
Delta R.T.: 0.011 min
Response: 85404633
Conc: 536.74 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025



#5 AR-1016-3

R.T.: 3.953 min
Delta R.T.: 0.000 min
Response: 85218692
Conc: 468.60 ng/ml



#6 AR-1016-4

R.T.: 4.658 min
Delta R.T.: 0.012 min
Response: 66986154
Conc: 515.56 ng/ml



#6 AR-1016-4

R.T.: 4.001 min
Delta R.T.: 0.000 min
Response: 77104461
Conc: 489.82 ng/ml



#7 AR-1016-5

R.T.: 4.942 min
Delta R.T.: 0.011 min
Response: 65865182
Conc: 511.36 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025



#7 AR-1016-5

R.T.: 4.196 min
Delta R.T.: 0.000 min
Response: 89640872
Conc: 467.26 ng/ml



#31 AR-1260-1

R.T.: 6.032 min
Delta R.T.: 0.012 min
Response: 117627257
Conc: 501.67 ng/ml



#31 AR-1260-1

R.T.: 5.189 min
Delta R.T.: 0.000 min
Response: 168895136
Conc: 468.62 ng/ml



#32 AR-1260-2

R.T.: 6.288 min
Delta R.T.: 0.013 min
Response: 130578290
Conc: 472.07 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025



#32 AR-1260-2

R.T.: 5.379 min
Delta R.T.: 0.000 min
Response: 199028852
Conc: 452.34 ng/ml



#33 AR-1260-3

R.T.: 6.638 min
Delta R.T.: 0.012 min
Response: 104328129
Conc: 523.45 ng/ml



#33 AR-1260-3

R.T.: 5.520 min
Delta R.T.: 0.000 min
Response: 193558906
Conc: 466.41 ng/ml



#34 AR-1260-4

R.T.: 6.853 min
Delta R.T.: 0.013 min
Response: 114033516
Conc: 530.68 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/25/2025



#34 AR-1260-4

R.T.: 5.981 min
Delta R.T.: 0.001 min
Response: 166991486
Conc: 464.17 ng/ml



#35 AR-1260-5

R.T.: 7.158 min
Delta R.T.: 0.014 min
Response: 213956682
Conc: 493.24 ng/ml



#35 AR-1260-5

R.T.: 6.225 min
Delta R.T.: 0.000 min
Response: 367314499
Conc: 437.13 ng/ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Continuing Calib Date: 10/20/2025 **Initial Calibration Date(s):** 10/10/2025 10/10/2025

Continuing Calib Time: 16:45 **Initial Calibration Time(s):** 08:57 16:43

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1016-2	(2)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.46	4.66	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.55	4.75	0.00
Aroclor-1016-5	(5)	4.93	4.93	4.83	5.03	0.00
Aroclor-1260-1	(1)	6.02	6.02	5.92	6.12	0.00
Aroclor-1260-2	(2)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	6.84	6.84	6.74	6.94	0.00
Aroclor-1260-5	(5)	7.15	7.15	7.05	7.25	0.00
Tetrachloro-m-xylene		3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl		8.50	8.50	8.40	8.60	0.00

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

Continuing Calib Date: 10/20/2025 **Initial Calibration Date(s):** 10/10/2025 10/10/2025

Continuing Calib Time: 16:45 **Initial Calibration Time(s):** 08:57 16:43

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1016-2	(2)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-3	(3)	3.95	3.95	3.85	4.05	0.00
Aroclor-1016-4	(4)	4.00	4.00	3.90	4.10	0.00
Aroclor-1016-5	(5)	4.20	4.20	4.10	4.30	0.00
Aroclor-1260-1	(1)	5.19	5.19	5.09	5.29	0.00
Aroclor-1260-2	(2)	5.38	5.38	5.28	5.48	0.00
Aroclor-1260-3	(3)	5.52	5.52	5.42	5.62	0.00
Aroclor-1260-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1260-5	(5)	6.22	6.22	6.12	6.32	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.53	7.53	7.43	7.63	0.00

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/20/2025

Lab Sample No.: AR1660CCC500 **Data File :** PQ071023.D **Time Analyzed:** 16:45

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.480	4.380	4.580	550.910	500.000	10.2
Aroclor-1016-2	4.499	4.399	4.599	545.480	500.000	9.1
Aroclor-1016-3	4.557	4.456	4.656	557.230	500.000	11.4
Aroclor-1016-4	4.647	4.546	4.746	552.890	500.000	10.6
Aroclor-1016-5	4.931	4.831	5.031	543.570	500.000	8.7
Aroclor-1260-1	6.021	5.919	6.119	524.350	500.000	4.9
Aroclor-1260-2	6.276	6.175	6.375	521.190	500.000	4.2
Aroclor-1260-3	6.628	6.526	6.726	548.540	500.000	9.7
Aroclor-1260-4	6.842	6.740	6.940	553.860	500.000	10.8
Aroclor-1260-5	7.147	7.045	7.245	548.590	500.000	9.7
Decachlorobiphenyl	8.499	8.396	8.596	57.120	50.000	14.2
Tetrachloro-m-xylene	3.400	3.300	3.500	54.370	50.000	8.7

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02

Lab Code: ACE **SDG NO.:** Q3375

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/20/2025

Lab Sample No.: AR1660CCC500 **Data File :** PQ071023.D **Time Analyzed:** 16:45

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.777	3.678	3.878	486.880	500.000	-2.6
Aroclor-1016-2	3.793	3.693	3.893	486.950	500.000	-2.6
Aroclor-1016-3	3.953	3.853	4.053	501.250	500.000	0.3
Aroclor-1016-4	4.002	3.902	4.102	507.110	500.000	1.4
Aroclor-1016-5	4.196	4.097	4.297	503.750	500.000	0.8
Aroclor-1260-1	5.189	5.089	5.289	495.440	500.000	-0.9
Aroclor-1260-2	5.378	5.278	5.478	487.970	500.000	-2.4
Aroclor-1260-3	5.519	5.419	5.619	493.830	500.000	-1.2
Aroclor-1260-4	5.979	5.879	6.079	490.320	500.000	-1.9
Aroclor-1260-5	6.224	6.124	6.324	470.860	500.000	-5.8
Decachlorobiphenyl	7.533	7.433	7.633	56.890	50.000	13.8
Tetrachloro-m-xylene	2.777	2.678	2.878	53.850	50.000	7.7

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071023.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 16:45
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:26:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.777	236.5E6	273.6E6	54.375	53.850
2)	SA Decachlor...	8.499	7.533	165.9E6	287.3E6	57.123	56.888
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.777	90428397	109.1E6	550.909	486.881
4)	L1 AR-1016-2	4.499	3.793	139.8E6	165.6E6	545.484	486.951
5)	L1 AR-1016-3	4.557	3.953	88664831	91157343	557.226	501.253
6)	L1 AR-1016-4	4.647	4.002	71836079	79825790	552.886	507.112
7)	L1 AR-1016-5	4.931	4.196	70013376	96640183	543.570	503.749
31)	L7 AR-1260-1	6.021	5.189	122.9E6	178.6E6	524.350	495.440
32)	L7 AR-1260-2	6.276	5.378	144.2E6	214.7E6	521.190	487.967
33)	L7 AR-1260-3	6.628	5.519	109.3E6	204.9E6	548.544	493.831
34)	L7 AR-1260-4	6.842	5.979	119.0E6	176.4E6	553.860	490.321
35)	L7 AR-1260-5	7.147	6.224	238.0E6	395.7E6	548.594	470.864

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071023.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 16:45
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:26:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 236544916
Conc: 54.37 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500



#1 Tetrachloro-m-xylene

R.T.: 2.777 min
Delta R.T.: 0.000 min
Response: 273560736
Conc: 53.85 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.499 min
Delta R.T.: 0.003 min
Response: 165882939
Conc: 57.12 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 287257298
Conc: 56.89 ng/ml



#3 AR-1016-1

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 90428397
Conc: 550.91 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500



#3 AR-1016-1

R.T.: 3.777 min
Delta R.T.: -0.001 min
Response: 109062239
Conc: 486.88 ng/ml



#4 AR-1016-2

R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 139804578
Conc: 545.48 ng/ml



#4 AR-1016-2

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 165623258
Conc: 486.95 ng/ml



#5 AR-1016-3

R.T.: 4.557 min
Delta R.T.: 0.000 min
Response: 88664831
Conc: 557.23 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500



#5 AR-1016-3

R.T.: 3.953 min
Delta R.T.: 0.000 min
Response: 91157343
Conc: 501.25 ng/ml



#6 AR-1016-4

R.T.: 4.647 min
Delta R.T.: 0.001 min
Response: 71836079
Conc: 552.89 ng/ml



#6 AR-1016-4

R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 79825790
Conc: 507.11 ng/ml



#7 AR-1016-5

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 70013376
Conc: 543.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500



#7 AR-1016-5

R.T.: 4.196 min
Delta R.T.: 0.000 min
Response: 96640183
Conc: 503.75 ng/ml



#31 AR-1260-1

R.T.: 6.021 min
Delta R.T.: 0.002 min
Response: 122944925
Conc: 524.35 ng/ml



#31 AR-1260-1

R.T.: 5.189 min
Delta R.T.: 0.000 min
Response: 178559449
Conc: 495.44 ng/ml



#32 AR-1260-2

R.T.: 6.276 min
Delta R.T.: 0.001 min
Response: 144165799
Conc: 521.19 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500



#32 AR-1260-2

R.T.: 5.378 min
Delta R.T.: 0.000 min
Response: 214702514
Conc: 487.97 ng/ml



#33 AR-1260-3

R.T.: 6.628 min
Delta R.T.: 0.002 min
Response: 109329482
Conc: 548.54 ng/ml



#33 AR-1260-3

R.T.: 5.519 min
Delta R.T.: 0.000 min
Response: 204939761
Conc: 493.83 ng/ml



#34 AR-1260-4

R.T.: 6.842 min
Delta R.T.: 0.002 min
Response: 119015206
Conc: 553.86 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500



#34 AR-1260-4

R.T.: 5.979 min
Delta R.T.: 0.000 min
Response: 176399367
Conc: 490.32 ng/ml



#35 AR-1260-5

R.T.: 7.147 min
Delta R.T.: 0.002 min
Response: 237967462
Conc: 548.59 ng/ml



#35 AR-1260-5

R.T.: 6.224 min
Delta R.T.: 0.000 min
Response: 395658630
Conc: 470.86 ng/ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

Client: AECOM **SDG No.:** Q3375

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 6069. **Instrument ID:** ECD_P

GC Column: ZB-MR1 **ID:** 0.32 (mm) **Inst. Calib. Date(s):** 09/23/2025 09/23/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES, AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
I.BLK	I.BLK	09/23/2025	08:29	PP075246.D	10.44	4.66
AR1660ICC1000	AR1660ICC1000	09/23/2025	08:45	PP075247.D	10.43	4.65
AR1660ICC750	AR1660ICC750	09/23/2025	09:01	PP075248.D	10.44	4.66
AR1660ICC500	AR1660ICC500	09/23/2025	09:18	PP075249.D	10.43	4.66
AR1660ICC250	AR1660ICC250	09/23/2025	09:34	PP075250.D	10.43	4.66
AR1660ICC050	AR1660ICC050	09/23/2025	10:11	PP075251.D	10.45	4.66
AR1221ICC500	AR1221ICC500	09/23/2025	10:27	PP075252.D	10.44	4.66
AR1232ICC500	AR1232ICC500	09/23/2025	10:43	PP075253.D	10.44	4.66
AR1242ICC1000	AR1242ICC1000	09/23/2025	11:00	PP075254.D	10.44	4.66
AR1242ICC750	AR1242ICC750	09/23/2025	11:16	PP075255.D	10.44	4.66
AR1242ICC500	AR1242ICC500	09/23/2025	11:32	PP075256.D	10.44	4.66
AR1242ICC250	AR1242ICC250	09/23/2025	11:48	PP075257.D	10.44	4.66
AR1242ICC050	AR1242ICC050	09/23/2025	12:05	PP075258.D	10.44	4.66
AR1248ICC1000	AR1248ICC1000	09/23/2025	12:37	PP075259.D	10.43	4.66
AR1248ICC750	AR1248ICC750	09/23/2025	12:53	PP075260.D	10.44	4.66
AR1248ICC500	AR1248ICC500	09/23/2025	13:10	PP075261.D	10.44	4.66
AR1248ICC250	AR1248ICC250	09/23/2025	13:26	PP075262.D	10.44	4.66
AR1248ICC050	AR1248ICC050	09/23/2025	13:42	PP075263.D	10.44	4.66
AR1254ICC1000	AR1254ICC1000	09/23/2025	14:15	PP075264.D	10.44	4.66
AR1254ICC750	AR1254ICC750	09/23/2025	14:31	PP075265.D	10.44	4.66
AR1254ICC500	AR1254ICC500	09/23/2025	14:47	PP075266.D	10.44	4.66
AR1254ICC250	AR1254ICC250	09/23/2025	15:04	PP075267.D	10.44	4.66
AR1254ICC050	AR1254ICC050	09/23/2025	15:20	PP075268.D	10.43	4.66
AR1262ICC500	AR1262ICC500	09/23/2025	15:52	PP075269.D	10.44	4.66
AR1268ICC1000	AR1268ICC1000	09/23/2025	16:09	PP075270.D	10.44	4.66
AR1268ICC750	AR1268ICC750	09/23/2025	16:25	PP075271.D	10.44	4.66
AR1268ICC500	AR1268ICC500	09/23/2025	16:41	PP075272.D	10.44	4.66
AR1268ICC250	AR1268ICC250	09/23/2025	16:58	PP075273.D	10.44	4.66
AR1268ICC050	AR1268ICC050	09/23/2025	17:14	PP075274.D	10.44	4.66
AR1660CCC500	AR1660CCC500	10/20/2025	11:13	PP075868.D	10.44	4.65
I.BLK	I.BLK	10/20/2025	12:41	PP075872.D	10.43	4.65
OR-03-101725MS	Q3383-01MS	10/20/2025	15:15	PP075879.D	10.44	4.65
OR-03-101725MSD	Q3383-01MSD	10/20/2025	15:31	PP075880.D	10.43	4.65
AR1660CCC500	AR1660CCC500	10/20/2025	16:20	PP075881.D	10.43	4.65
I.BLK	I.BLK	10/20/2025	17:25	PP075885.D	10.43	4.65
I.BLK	I.BLK	10/10/2025	08:42	PQ070972.D	8.50	3.40
AR1660ICC1000	AR1660ICC1000	10/10/2025	08:57	PQ070973.D	8.50	3.40
AR1660ICC750	AR1660ICC750	10/10/2025	09:11	PQ070974.D	8.50	3.40
AR1660ICC500	AR1660ICC500	10/10/2025	09:26	PQ070975.D	8.50	3.40
AR1660ICC250	AR1660ICC250	10/10/2025	09:40	PQ070976.D	8.50	3.40
AR1660ICC050	AR1660ICC050	10/10/2025	09:55	PQ070977.D	8.50	3.40
AR1221ICC500	AR1221ICC500	10/10/2025	10:09	PQ070978.D	8.50	3.40

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

AR1232ICC500	AR1232ICC500	10/10/2025	10:24	PQ070979.D	8.50	3.40
AR1242ICC1000	AR1242ICC1000	10/10/2025	10:38	PQ070980.D	8.50	3.40
AR1242ICC750	AR1242ICC750	10/10/2025	10:53	PQ070981.D	8.50	3.40
AR1242ICC500	AR1242ICC500	10/10/2025	11:08	PQ070982.D	8.50	3.40
AR1242ICC250	AR1242ICC250	10/10/2025	11:22	PQ070983.D	8.50	3.40
AR1242ICC050	AR1242ICC050	10/10/2025	11:37	PQ070984.D	8.50	3.40
AR1248ICC1000	AR1248ICC1000	10/10/2025	11:51	PQ070985.D	8.50	3.40
AR1248ICC750	AR1248ICC750	10/10/2025	12:06	PQ070986.D	8.50	3.40
AR1248ICC500	AR1248ICC500	10/10/2025	12:20	PQ070987.D	8.50	3.40
AR1248ICC250	AR1248ICC250	10/10/2025	12:35	PQ070988.D	8.50	3.40
AR1248ICC050	AR1248ICC050	10/10/2025	12:50	PQ070989.D	8.50	3.40
AR1254ICC1000	AR1254ICC1000	10/10/2025	13:04	PQ070990.D	8.50	3.40
AR1254ICC750	AR1254ICC750	10/10/2025	13:19	PQ070991.D	8.50	3.40
AR1254ICC500	AR1254ICC500	10/10/2025	13:33	PQ070992.D	8.50	3.40
AR1254ICC250	AR1254ICC250	10/10/2025	13:48	PQ070993.D	8.50	3.40
AR1254ICC050	AR1254ICC050	10/10/2025	15:15	PQ070994.D	8.50	3.40
AR1262ICC500	AR1262ICC500	10/10/2025	15:30	PQ070995.D	8.50	3.40
AR1268ICC1000	AR1268ICC1000	10/10/2025	15:44	PQ070996.D	8.50	3.40
AR1268ICC750	AR1268ICC750	10/10/2025	15:59	PQ070997.D	8.50	3.40
AR1268ICC500	AR1268ICC500	10/10/2025	16:13	PQ070998.D	8.50	3.40
AR1268ICC250	AR1268ICC250	10/10/2025	16:28	PQ070999.D	8.50	3.40
AR1268ICC050	AR1268ICC050	10/10/2025	16:43	PQ071000.D	8.50	3.40
AR1660CCC500	AR1660CCC500	10/20/2025	10:18	PQ071008.D	8.51	3.41
I.BLK	I.BLK	10/20/2025	12:36	PQ071012.D	8.52	3.41
PB170162BL	PB170162BL	10/20/2025	13:20	PQ071013.D	8.52	3.42
PB170162BS	PB170162BS	10/20/2025	13:35	PQ071014.D	8.51	3.40
PR132-S02-074099-20251016	Q3375-01	10/20/2025	14:19	PQ071017.D	8.50	3.40
PR132-S04-000013-20251016	Q3375-02	10/20/2025	14:33	PQ071018.D	8.50	3.40
PR132-S05-000057-20251016	Q3375-03	10/20/2025	14:48	PQ071019.D	8.50	3.40
PR132-S02-010074-20251016	Q3375-04	10/20/2025	15:02	PQ071020.D	8.50	3.40
PR132-DP02-20251016	Q3375-05	10/20/2025	15:47	PQ071021.D	8.52	3.42
AR1660CCC500	AR1660CCC500	10/20/2025	16:45	PQ071023.D	8.50	3.40
I.BLK	I.BLK	10/20/2025	17:44	PQ071027.D	8.50	3.40

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

Client: AECOM **SDG No.:** Q3375

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 6069. **Instrument ID:** ECD_P

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Inst. Calib. Date(s):** 09/23/2025 09/23/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES, AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
I.BLK	I.BLK	09/23/2025	08:29	PP075246.D	8.81	3.80
AR1660ICC1000	AR1660ICC1000	09/23/2025	08:45	PP075247.D	8.81	3.80
AR1660ICC750	AR1660ICC750	09/23/2025	09:01	PP075248.D	8.80	3.80
AR1660ICC500	AR1660ICC500	09/23/2025	09:18	PP075249.D	8.80	3.80
AR1660ICC250	AR1660ICC250	09/23/2025	09:34	PP075250.D	8.80	3.80
AR1660ICC050	AR1660ICC050	09/23/2025	10:11	PP075251.D	8.81	3.80
AR1221ICC500	AR1221ICC500	09/23/2025	10:27	PP075252.D	8.81	3.80
AR1232ICC500	AR1232ICC500	09/23/2025	10:43	PP075253.D	8.81	3.80
AR1242ICC1000	AR1242ICC1000	09/23/2025	11:00	PP075254.D	8.81	3.80
AR1242ICC750	AR1242ICC750	09/23/2025	11:16	PP075255.D	8.81	3.80
AR1242ICC500	AR1242ICC500	09/23/2025	11:32	PP075256.D	8.81	3.80
AR1242ICC250	AR1242ICC250	09/23/2025	11:48	PP075257.D	8.81	3.80
AR1242ICC050	AR1242ICC050	09/23/2025	12:05	PP075258.D	8.80	3.80
AR1248ICC1000	AR1248ICC1000	09/23/2025	12:37	PP075259.D	8.80	3.80
AR1248ICC750	AR1248ICC750	09/23/2025	12:53	PP075260.D	8.81	3.80
AR1248ICC500	AR1248ICC500	09/23/2025	13:10	PP075261.D	8.80	3.80
AR1248ICC250	AR1248ICC250	09/23/2025	13:26	PP075262.D	8.81	3.80
AR1248ICC050	AR1248ICC050	09/23/2025	13:42	PP075263.D	8.80	3.80
AR1254ICC1000	AR1254ICC1000	09/23/2025	14:15	PP075264.D	8.81	3.80
AR1254ICC750	AR1254ICC750	09/23/2025	14:31	PP075265.D	8.81	3.80
AR1254ICC500	AR1254ICC500	09/23/2025	14:47	PP075266.D	8.81	3.80
AR1254ICC250	AR1254ICC250	09/23/2025	15:04	PP075267.D	8.81	3.80
AR1254ICC050	AR1254ICC050	09/23/2025	15:20	PP075268.D	8.81	3.80
AR1262ICC500	AR1262ICC500	09/23/2025	15:52	PP075269.D	8.81	3.80
AR1268ICC1000	AR1268ICC1000	09/23/2025	16:09	PP075270.D	8.81	3.80
AR1268ICC750	AR1268ICC750	09/23/2025	16:25	PP075271.D	8.81	3.80
AR1268ICC500	AR1268ICC500	09/23/2025	16:41	PP075272.D	8.81	3.80
AR1268ICC250	AR1268ICC250	09/23/2025	16:58	PP075273.D	8.80	3.80
AR1268ICC050	AR1268ICC050	09/23/2025	17:14	PP075274.D	8.81	3.80
AR1660CCC500	AR1660CCC500	10/20/2025	11:13	PP075868.D	8.79	3.79
I.BLK	I.BLK	10/20/2025	12:41	PP075872.D	8.79	3.79
OR-03-101725MS	Q3383-01MS	10/20/2025	15:15	PP075879.D	8.79	3.79
OR-03-101725MSD	Q3383-01MSD	10/20/2025	15:31	PP075880.D	8.79	3.79
AR1660CCC500	AR1660CCC500	10/20/2025	16:20	PP075881.D	8.79	3.79
I.BLK	I.BLK	10/20/2025	17:25	PP075885.D	8.79	3.79
I.BLK	I.BLK	10/10/2025	08:42	PQ070972.D	7.53	2.78
AR1660ICC1000	AR1660ICC1000	10/10/2025	08:57	PQ070973.D	7.53	2.78
AR1660ICC750	AR1660ICC750	10/10/2025	09:11	PQ070974.D	7.53	2.78
AR1660ICC500	AR1660ICC500	10/10/2025	09:26	PQ070975.D	7.53	2.78
AR1660ICC250	AR1660ICC250	10/10/2025	09:40	PQ070976.D	7.53	2.78
AR1660ICC050	AR1660ICC050	10/10/2025	09:55	PQ070977.D	7.53	2.78
AR1221ICC500	AR1221ICC500	10/10/2025	10:09	PQ070978.D	7.53	2.78

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

AR1232ICC500	AR1232ICC500	10/10/2025	10:24	PQ070979.D	7.53	2.78
AR1242ICC1000	AR1242ICC1000	10/10/2025	10:38	PQ070980.D	7.53	2.78
AR1242ICC750	AR1242ICC750	10/10/2025	10:53	PQ070981.D	7.53	2.78
AR1242ICC500	AR1242ICC500	10/10/2025	11:08	PQ070982.D	7.53	2.78
AR1242ICC250	AR1242ICC250	10/10/2025	11:22	PQ070983.D	7.53	2.78
AR1242ICC050	AR1242ICC050	10/10/2025	11:37	PQ070984.D	7.53	2.78
AR1248ICC1000	AR1248ICC1000	10/10/2025	11:51	PQ070985.D	7.53	2.78
AR1248ICC750	AR1248ICC750	10/10/2025	12:06	PQ070986.D	7.53	2.78
AR1248ICC500	AR1248ICC500	10/10/2025	12:20	PQ070987.D	7.53	2.78
AR1248ICC250	AR1248ICC250	10/10/2025	12:35	PQ070988.D	7.53	2.78
AR1248ICC050	AR1248ICC050	10/10/2025	12:50	PQ070989.D	7.53	2.78
AR1254ICC1000	AR1254ICC1000	10/10/2025	13:04	PQ070990.D	7.53	2.78
AR1254ICC750	AR1254ICC750	10/10/2025	13:19	PQ070991.D	7.53	2.78
AR1254ICC500	AR1254ICC500	10/10/2025	13:33	PQ070992.D	7.53	2.78
AR1254ICC250	AR1254ICC250	10/10/2025	13:48	PQ070993.D	7.53	2.78
AR1254ICC050	AR1254ICC050	10/10/2025	15:15	PQ070994.D	7.53	2.78
AR1262ICC500	AR1262ICC500	10/10/2025	15:30	PQ070995.D	7.53	2.78
AR1268ICC1000	AR1268ICC1000	10/10/2025	15:44	PQ070996.D	7.53	2.78
AR1268ICC750	AR1268ICC750	10/10/2025	15:59	PQ070997.D	7.53	2.78
AR1268ICC500	AR1268ICC500	10/10/2025	16:13	PQ070998.D	7.53	2.78
AR1268ICC250	AR1268ICC250	10/10/2025	16:28	PQ070999.D	7.53	2.78
AR1268ICC050	AR1268ICC050	10/10/2025	16:43	PQ071000.D	7.53	2.78
AR1660CCC500	AR1660CCC500	10/20/2025	10:18	PQ071008.D	7.54	2.78
I.BLK	I.BLK	10/20/2025	12:36	PQ071012.D	7.54	2.78
PB170162BL	PB170162BL	10/20/2025	13:20	PQ071013.D	7.54	2.78
PB170162BS	PB170162BS	10/20/2025	13:35	PQ071014.D	7.54	2.78
PR132-S02-074099-20251016	Q3375-01	10/20/2025	14:19	PQ071017.D	7.53	2.78
PR132-S04-000013-20251016	Q3375-02	10/20/2025	14:33	PQ071018.D	7.53	2.78
PR132-S05-000057-20251016	Q3375-03	10/20/2025	14:48	PQ071019.D	7.53	2.78
PR132-S02-010074-20251016	Q3375-04	10/20/2025	15:02	PQ071020.D	7.53	2.78
PR132-DP02-20251016	Q3375-05	10/20/2025	15:47	PQ071021.D	7.54	2.78
AR1660CCC500	AR1660CCC500	10/20/2025	16:45	PQ071023.D	7.53	2.78
I.BLK	I.BLK	10/20/2025	17:44	PQ071027.D	7.53	2.78

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

OR-03-101725MS

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3375

Lab Sample ID: Q3383-01MS Date(s) Analyzed: 10/20/2025 10/20/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PP075879.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	5.805	5.755	5.855	184		
	2	5.827	5.777	5.877	194		
	3	5.89	5.84	5.94	184		
COLUMN 1	4	5.987	5.937	6.037	199		
	5	6.279	6.229	6.329	182	189	
	1	4.887	4.837	4.937	223		
	2	4.945	4.895	4.995	230		
	3	5.065	5.015	5.115	200		
	4	5.106	5.056	5.156	218		
COLUMN 2	5	5.326	5.276	5.376	429	260	31.63
Aroclor-1260	1	7.398	7.348	7.448	198		
	2	7.651	7.601	7.701	188		
	3	8.009	7.959	8.059	165		
COLUMN 1	4	8.236	8.186	8.286	194		
	5	8.563	8.513	8.613	175	184	
	1	6.348	6.298	6.398	176		
	2	6.534	6.484	6.584	187		
	3	6.685	6.635	6.735	207		
	4	7.157	7.107	7.207	145		
COLUMN 2	5	7.396	7.346	7.446	149	173	6.16

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

OR-03-101725MSD

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3375

Lab Sample ID: Q3383-01MSD Date(s) Analyzed: 10/20/2025 10/20/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PP075880.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	5.804	5.754	5.854	196		
	2	5.825	5.775	5.875	190		
	3	5.888	5.838	5.938	182		
COLUMN 1	4	5.985	5.935	6.035	193		
	5	6.277	6.227	6.327	176	187	
	1	4.886	4.836	4.936	203		
	2	4.943	4.893	4.993	207		
	3	5.064	5.014	5.114	187		
	4	5.104	5.054	5.154	198		
COLUMN 2	5	5.324	5.274	5.374	426	244	26.45
Aroclor-1260	1	7.396	7.346	7.446	191		
	2	7.649	7.599	7.699	182		
	3	8.007	7.957	8.057	158		
COLUMN 1	4	8.234	8.184	8.284	186		
	5	8.562	8.512	8.612	170	177	
	1	6.347	6.297	6.397	166		
	2	6.533	6.483	6.583	178		
	3	6.684	6.634	6.734	205		
	4	7.154	7.104	7.204	147		
COLUMN 2	5	7.395	7.345	7.445	140	167	5.81

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170162BS

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3375

Lab Sample ID: PB170162BS Date(s) Analyzed: 10/20/2025 10/20/2025

Instrument ID (1): ECD_Q Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071014.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	4.483	4.433	4.533	175		
	2	4.502	4.452	4.552	172		
	3	4.56	4.51	4.61	174		
COLUMN 1	4	4.65	4.6	4.7	174		
	5	4.935	4.885	4.985	166	172	
	1	3.777	3.727	3.827	153		
	2	3.793	3.743	3.843	153		
	3	3.952	3.902	4.002	156		
	4	4.001	3.951	4.051	154		
COLUMN 2	5	4.196	4.146	4.246	151	153	11.69
Aroclor-1260	1	6.024	5.974	6.074	175		
	2	6.28	6.23	6.33	173		
	3	6.631	6.581	6.681	152		
COLUMN 1	4	6.845	6.795	6.895	163		
	5	7.152	7.102	7.202	161	165	
	1	5.188	5.138	5.238	160		
	2	5.378	5.328	5.428	157		
	3	5.519	5.469	5.569	161		
	4	5.979	5.929	6.029	139		
COLUMN 2	5	6.224	6.174	6.274	135	151	8.86

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PR132-S05-000057-2025

Lab Name: Alliance | Contract: AECO02

Lab Code: ACE | SDG NO.: Q3375

Lab Sample ID: Q3375-03 | Date(s) Analyzed: 10/20/2025 10/20/2025

Instrument ID (1): ECD_Q | Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm) | GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071019.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1242	1	4.482	4.432	4.532	33.9		
	2	4.5	4.45	4.55	49.8		
	3	4.557	4.507	4.607	26.2		
COLUMN 1	4	4.646	4.596	4.696	51.1		
	5	5.353	5.303	5.403	65.1	45.2	
	1	3.777	3.727	3.827	37.2		
	2	3.792	3.742	3.842	43.1		
	3	3.952	3.902	4.002	42.5		
	4	4.038	3.988	4.088	65.6		
COLUMN 2	5	4.533	4.483	4.583	67.3	51.1	12.25
Aroclor-1254	1	5.292	5.242	5.342	27.6		
	2	5.504	5.454	5.554	24.1		
	3	5.856	5.806	5.906	20.9		
COLUMN 1	4	6.136	6.086	6.186	14.7		
	5	6.545	6.495	6.595	16.7	20.8	
	1	4.532	4.482	4.582	31.8		
	2	4.678	4.628	4.728	22.2		
	3	5.059	5.009	5.109	17.0		
	4	5.286	5.236	5.336	17.5		
COLUMN 2	5	5.689	5.639	5.739	17.4	21.2	1.9

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PR132-DP02-20251016

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3375

Lab Sample ID: Q3375-05 Date(s) Analyzed: 10/20/2025 10/20/2025

Instrument ID (1): ECD_Q Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071021.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1242	1	4.497	4.447	4.547	45.8		
	2	4.516	4.466	4.566	63.5		
	3	4.574	4.524	4.624	41.0		
COLUMN 1	4	4.662	4.612	4.712	76.7		
	5	5.37	5.32	5.42	82.7	61.9	
	1	3.775	3.725	3.825	42.0		
	2	3.791	3.741	3.841	44.7		
	3	3.951	3.901	4.001	51.1		
	4	4.036	3.986	4.086	76.1		
COLUMN 2	5	4.532	4.482	4.582	86.5	60.1	2.95
Aroclor-1254	1	5.31	5.26	5.36	30.4		
	2	5.521	5.471	5.571	29.0		
	3	5.876	5.826	5.926	22.8		
COLUMN 1	4	6.153	6.103	6.203	13.2		
	5	6.563	6.513	6.613	17.5	22.6	
	1	4.531	4.481	4.581	40.0		
	2	4.677	4.627	4.727	26.5		
	3	5.06	5.01	5.11	21.8		
	4	5.285	5.235	5.335	19.4		
COLUMN 2	5	5.69	5.64	5.74	18.8	25.3	11.27

QC SAMPLE DATA

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	
Client Sample ID:	PB170162BL	SDG No.:	Q3375
Lab Sample ID:	PB170162BL	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.01 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date 10/20/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	17.0	U	1	3.90	17.0	ug/kg	10/20/25 13:20	PB170162
11104-28-2	Aroclor-1221	17.0	U	1	4.00	17.0	ug/kg	10/20/25 13:20	PB170162
11141-16-5	Aroclor-1232	17.0	U	1	3.70	17.0	ug/kg	10/20/25 13:20	PB170162
53469-21-9	Aroclor-1242	17.0	U	1	4.00	17.0	ug/kg	10/20/25 13:20	PB170162
12672-29-6	Aroclor-1248	17.0	U	1	5.90	17.0	ug/kg	10/20/25 13:20	PB170162
11097-69-1	Aroclor-1254	17.0	U	1	3.20	17.0	ug/kg	10/20/25 13:20	PB170162
37324-23-5	Aroclor-1262	17.0	U	1	5.00	17.0	ug/kg	10/20/25 13:20	PB170162
11100-14-4	Aroclor-1268	17.0	U	1	3.60	17.0	ug/kg	10/20/25 13:20	PB170162
11096-82-5	Aroclor-1260	17.0	U	1	3.20	17.0	ug/kg	10/20/25 13:20	PB170162
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.7			32 - 144	108%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.1			32 - 175	126%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071013.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 13:20
Operator : YP\AJ
Sample : PB170162BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PB170162BL

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 08:56:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.415	2.775	94367263	109.7E6	21.692	21.586
2)	SA Decachlor...	8.523	7.537	73015121	125.0E6	25.143m	24.753
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071013.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 13:20
Operator : YP\AJ
Sample : PB170162BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PB170162BL

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 08:56:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.415 min
Delta R.T.: 0.015 min
Response: 94367263
Conc: 21.69 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PB170162BL

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/23/2025
Supervised By :mohammad ahmed 10/25/2025



#1 Tetrachloro-m-xylene

R.T.: 2.775 min
Delta R.T.: -0.003 min
Response: 109656163
Conc: 21.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.523 min
Delta R.T.: 0.023 min
Response: 73015121
Conc: 25.14 ng/ml m



#2 Decachlorobiphenyl

R.T.: 7.537 min
Delta R.T.: 0.004 min
Response: 124989389
Conc: 24.75 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	09/23/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	09/23/25
Client Sample ID:	PIBLK-PP075246.D	SDG No.:	Q3375
Lab Sample ID:	I.BLK-PP075246.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.50	U	1	0.097	0.50	ug/L	09/23/25	pp092325
11104-28-2	Aroclor-1221	0.50	U	1	0.13	0.50	ug/L	09/23/25	pp092325
11141-16-5	Aroclor-1232	0.50	U	1	0.096	0.50	ug/L	09/23/25	pp092325
53469-21-9	Aroclor-1242	0.50	U	1	0.12	0.50	ug/L	09/23/25	pp092325
12672-29-6	Aroclor-1248	0.50	U	1	0.071	0.50	ug/L	09/23/25	pp092325
11097-69-1	Aroclor-1254	0.50	U	1	0.094	0.50	ug/L	09/23/25	pp092325
11096-82-5	Aroclor-1260	0.50	U	1	0.081	0.50	ug/L	09/23/25	pp092325
37324-23-5	Aroclor-1262	0.50	U	1	0.14	0.50	ug/L	09/23/25	pp092325
11100-14-4	Aroclor-1268	0.50	U	1	0.11	0.50	ug/L	09/23/25	pp092325
SURROGATES									
877-09-8	Tetrachloro-m-xylene	17.9			60 - 140	90%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.4			60 - 140	97%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075246.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 08:29
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:44:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.657	3.797	26080915	109.2E6	19.932	17.901
2)	SA Decachlor...	10.437	8.805	19354075	120.9E6	19.402	20.045
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075246.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 08:29
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:44:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.657 min
Delta R.T.: 0.002 min
Response: 26080915
Conc: 19.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK



#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 109213871
Conc: 17.90 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.003 min
Response: 19354075
Conc: 19.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.805 min
Delta R.T.: 0.001 min
Response: 120921024
Conc: 20.04 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/20/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/20/25
Client Sample ID:	PIBLK-PP075872.D	SDG No.:	Q3375
Lab Sample ID:	I.BLK-PP075872.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.50	U	1	0.097	0.50	ug/L	10/20/25	pp102025
11104-28-2	Aroclor-1221	0.50	U	1	0.13	0.50	ug/L	10/20/25	pp102025
11141-16-5	Aroclor-1232	0.50	U	1	0.096	0.50	ug/L	10/20/25	pp102025
53469-21-9	Aroclor-1242	0.50	U	1	0.12	0.50	ug/L	10/20/25	pp102025
12672-29-6	Aroclor-1248	0.50	U	1	0.071	0.50	ug/L	10/20/25	pp102025
11097-69-1	Aroclor-1254	0.50	U	1	0.094	0.50	ug/L	10/20/25	pp102025
11096-82-5	Aroclor-1260	0.50	U	1	0.081	0.50	ug/L	10/20/25	pp102025
37324-23-5	Aroclor-1262	0.50	U	1	0.14	0.50	ug/L	10/20/25	pp102025
11100-14-4	Aroclor-1268	0.50	U	1	0.11	0.50	ug/L	10/20/25	pp102025
SURROGATES									
877-09-8	Tetrachloro-m-xylene	22.7			60 - 140	113%	SPK: 20		
2051-24-3	Decachlorobiphenyl	20.6			60 - 140	103%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102025\
Data File : PP075872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 12:41
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 23:54:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.654	3.791	30217386	138.4E6	23.093	22.691
2)	SA Decachlor...	10.433	8.792	20534994	132.6E6	20.585	21.985
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102025\
Data File : PP075872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 12:41
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 23:54:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.654 min
Delta R.T.: -0.001 min
Response: 30217386
Conc: 23.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK



#1 Tetrachloro-m-xylene

R.T.: 3.791 min
Delta R.T.: -0.005 min
Response: 138432593
Conc: 22.69 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.433 min
Delta R.T.: -0.001 min
Response: 20534994
Conc: 20.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.792 min
Delta R.T.: -0.012 min
Response: 132625761
Conc: 21.99 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/20/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/20/25
Client Sample ID:	PIBLK-PP075885.D	SDG No.:	Q3375
Lab Sample ID:	I.BLK-PP075885.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.50	U	1	0.097	0.50	ug/L	10/20/25	pp102025
11104-28-2	Aroclor-1221	0.50	U	1	0.13	0.50	ug/L	10/20/25	pp102025
11141-16-5	Aroclor-1232	0.50	U	1	0.096	0.50	ug/L	10/20/25	pp102025
53469-21-9	Aroclor-1242	0.50	U	1	0.12	0.50	ug/L	10/20/25	pp102025
12672-29-6	Aroclor-1248	0.50	U	1	0.071	0.50	ug/L	10/20/25	pp102025
11097-69-1	Aroclor-1254	0.50	U	1	0.094	0.50	ug/L	10/20/25	pp102025
11096-82-5	Aroclor-1260	0.50	U	1	0.081	0.50	ug/L	10/20/25	pp102025
37324-23-5	Aroclor-1262	0.50	U	1	0.14	0.50	ug/L	10/20/25	pp102025
11100-14-4	Aroclor-1268	0.50	U	1	0.11	0.50	ug/L	10/20/25	pp102025
SURROGATES									
877-09-8	Tetrachloro-m-xylene	22.7			60 - 140	114%	SPK: 20		
2051-24-3	Decachlorobiphenyl	20.0			60 - 140	100%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102025\
Data File : PP075885.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 17:25
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 00:00:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.650	3.787	29732305	143.2E6	22.723	23.470
2)	SA Decachlor...	10.430	8.787	19954961	125.8E6	20.004	20.848
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102025\
Data File : PP075885.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 17:25
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 00:00:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.650 min
Delta R.T.: -0.005 min
Response: 29732305
Conc: 22.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK



#1 Tetrachloro-m-xylene

R.T.: 3.787 min
Delta R.T.: -0.009 min
Response: 143186304
Conc: 23.47 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.430 min
Delta R.T.: -0.004 min
Response: 19954961
Conc: 20.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.787 min
Delta R.T.: -0.016 min
Response: 125768040
Conc: 20.85 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/10/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/10/25
Client Sample ID:	PIBLK-PQ070972.D	SDG No.:	Q3375
Lab Sample ID:	I.BLK-PQ070972.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.50	U	1	0.097	0.50	ug/L	10/10/25	pq101025
11104-28-2	Aroclor-1221	0.50	U	1	0.13	0.50	ug/L	10/10/25	pq101025
11141-16-5	Aroclor-1232	0.50	U	1	0.096	0.50	ug/L	10/10/25	pq101025
53469-21-9	Aroclor-1242	0.50	U	1	0.12	0.50	ug/L	10/10/25	pq101025
12672-29-6	Aroclor-1248	0.50	U	1	0.071	0.50	ug/L	10/10/25	pq101025
11097-69-1	Aroclor-1254	0.50	U	1	0.094	0.50	ug/L	10/10/25	pq101025
11096-82-5	Aroclor-1260	0.50	U	1	0.081	0.50	ug/L	10/10/25	pq101025
37324-23-5	Aroclor-1262	0.50	U	1	0.14	0.50	ug/L	10/10/25	pq101025
11100-14-4	Aroclor-1268	0.50	U	1	0.11	0.50	ug/L	10/10/25	pq101025
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.3			60 - 140	91%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.0			60 - 140	95%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070972.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 08:42
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:32:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:24:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	79509611	97022264	18.277	19.099
2)	SA Decachlor...	8.496	7.533	55246960	97962147	19.025	19.400
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070972.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 08:42
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:32:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:24:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 79509611
Conc: 18.28 ng/ml

Instrument :
ECD_Q
ClientSampleId :
I.BLK



#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 97022264
Conc: 19.10 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 55246960
Conc: 19.02 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 97962147
Conc: 19.40 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/20/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/20/25
Client Sample ID:	PIBLK-PQ071012.D	SDG No.:	Q3375
Lab Sample ID:	I.BLK-PQ071012.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.50	U	1	0.097	0.50	ug/L	10/20/25	PQ102025
11104-28-2	Aroclor-1221	0.50	U	1	0.13	0.50	ug/L	10/20/25	PQ102025
11141-16-5	Aroclor-1232	0.50	U	1	0.096	0.50	ug/L	10/20/25	PQ102025
53469-21-9	Aroclor-1242	0.50	U	1	0.12	0.50	ug/L	10/20/25	PQ102025
12672-29-6	Aroclor-1248	0.50	U	1	0.071	0.50	ug/L	10/20/25	PQ102025
11097-69-1	Aroclor-1254	0.50	U	1	0.094	0.50	ug/L	10/20/25	PQ102025
11096-82-5	Aroclor-1260	0.50	U	1	0.081	0.50	ug/L	10/20/25	PQ102025
37324-23-5	Aroclor-1262	0.50	U	1	0.14	0.50	ug/L	10/20/25	PQ102025
11100-14-4	Aroclor-1268	0.50	U	1	0.11	0.50	ug/L	10/20/25	PQ102025
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.7			60 - 140	94%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.0			60 - 140	110%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071012.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 12:36
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:23:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.414	2.775	81508294	96011109	18.736	18.900
2)	SA Decachlor...	8.519	7.538	64193337	111.3E6	22.105m	22.033
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071012.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 12:36
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:23:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.414 min
Delta R.T.: 0.014 min
Response: 81508294
Conc: 18.74 ng/ml

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025



#1 Tetrachloro-m-xylene

R.T.: 2.775 min
Delta R.T.: -0.003 min
Response: 96011109
Conc: 18.90 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.519 min
Delta R.T.: 0.023 min
Response: 64193337
Conc: 22.11 ng/ml m



#2 Decachlorobiphenyl

R.T.: 7.538 min
Delta R.T.: 0.005 min
Response: 111257012
Conc: 22.03 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/20/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/20/25
Client Sample ID:	PIBLK-PQ071027.D	SDG No.:	Q3375
Lab Sample ID:	I.BLK-PQ071027.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.50	U	1	0.097	0.50	ug/L	10/20/25	PQ102025
11104-28-2	Aroclor-1221	0.50	U	1	0.13	0.50	ug/L	10/20/25	PQ102025
11141-16-5	Aroclor-1232	0.50	U	1	0.096	0.50	ug/L	10/20/25	PQ102025
53469-21-9	Aroclor-1242	0.50	U	1	0.12	0.50	ug/L	10/20/25	PQ102025
12672-29-6	Aroclor-1248	0.50	U	1	0.071	0.50	ug/L	10/20/25	PQ102025
11097-69-1	Aroclor-1254	0.50	U	1	0.094	0.50	ug/L	10/20/25	PQ102025
11096-82-5	Aroclor-1260	0.50	U	1	0.081	0.50	ug/L	10/20/25	PQ102025
37324-23-5	Aroclor-1262	0.50	U	1	0.14	0.50	ug/L	10/20/25	PQ102025
11100-14-4	Aroclor-1268	0.50	U	1	0.11	0.50	ug/L	10/20/25	PQ102025
SURROGATES									
877-09-8	Tetrachloro-m-xylene	19.6			60 - 140	98%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.3			60 - 140	111%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071027.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 17:44
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:27:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.399	2.777	85148503	99903769	19.573	19.666
2)	SA Decachlor...	8.499	7.533	65560662	112.5E6	22.576	22.287
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071027.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 17:44
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:27:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.399 min
Delta R.T.: 0.000 min
Response: 85148503
Conc: 19.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :
I.BLK



#1 Tetrachloro-m-xylene

R.T.: 2.777 min
Delta R.T.: -0.001 min
Response: 99903769
Conc: 19.67 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.499 min
Delta R.T.: 0.003 min
Response: 65560662
Conc: 22.58 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 112538152
Conc: 22.29 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	
Client Sample ID:	PB170162BS	SDG No.:	Q3375
Lab Sample ID:	PB170162BS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.03 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date 10/20/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	172		1	3.90	17.0	ug/kg	10/20/25 13:35	PB170162
11104-28-2	Aroclor-1221	17.0	U	1	4.00	17.0	ug/kg	10/20/25 13:35	PB170162
11141-16-5	Aroclor-1232	17.0	U	1	3.70	17.0	ug/kg	10/20/25 13:35	PB170162
53469-21-9	Aroclor-1242	17.0	U	1	4.00	17.0	ug/kg	10/20/25 13:35	PB170162
12672-29-6	Aroclor-1248	17.0	U	1	5.90	17.0	ug/kg	10/20/25 13:35	PB170162
11097-69-1	Aroclor-1254	17.0	U	1	3.20	17.0	ug/kg	10/20/25 13:35	PB170162
37324-23-5	Aroclor-1262	17.0	U	1	5.00	17.0	ug/kg	10/20/25 13:35	PB170162
11100-14-4	Aroclor-1268	17.0	U	1	3.60	17.0	ug/kg	10/20/25 13:35	PB170162
11096-82-5	Aroclor-1260	165		1	3.20	17.0	ug/kg	10/20/25 13:35	PB170162
SURROGATES									
877-09-8	Tetrachloro-m-xylene	24.0			32 - 144	120%	SPK: 20		
2051-24-3	Decachlorobiphenyl	27.3			32 - 175	137%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071014.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 13:35
Operator : YP\AJ
Sample : PB170162BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PB170162BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:24:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.402	2.777	104.4E6	114.4E6	24.000	22.523
2)	SA Decachlor...	8.505	7.535	79385261	138.0E6	27.337	27.337
	Target Compounds						
3)	L1 AR-1016-1	4.483	3.777	86469235	102.9E6	526.789	459.319
4)	L1 AR-1016-2	4.502	3.793	132.1E6	156.1E6	515.323	458.874
5)	L1 AR-1016-3	4.560	3.952	82994793	85348928	521.592	469.314
6)	L1 AR-1016-4	4.650	4.001	67743632	72823533	521.389	462.628
7)	L1 AR-1016-5	4.935	4.196	64092788	86936492	497.604	453.167
31)	L7 AR-1260-1	6.024	5.188	123.2E6	173.7E6	525.452	481.878
32)	L7 AR-1260-2	6.280	5.378	144.1E6	207.9E6	520.774	472.457
33)	L7 AR-1260-3	6.631	5.519	90911766	200.2E6	456.136	482.502
34)	L7 AR-1260-4	6.845	5.979	105.4E6	150.2E6	490.559	417.581
35)	L7 AR-1260-5	7.152	6.224	210.2E6	341.0E6	484.494	405.845

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102025\
Data File : PQ071014.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 13:35
Operator : YP\AJ
Sample : PB170162BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PB170162BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 07:24:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 3.402 min
Delta R.T.: 0.002 min
Response: 104405794
Conc: 24.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PB170162BS



#1 Tetrachloro-m-xylene

R.T.: 2.777 min
Delta R.T.: -0.001 min
Response: 114418975
Conc: 22.52 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.505 min
Delta R.T.: 0.009 min
Response: 79385261
Conc: 27.34 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.535 min
Delta R.T.: 0.002 min
Response: 138038584
Conc: 27.34 ng/ml



#3 AR-1016-1

R.T.: 4.483 min
Delta R.T.: 0.003 min
Response: 86469235
Conc: 526.79 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PB170162BS



#3 AR-1016-1

R.T.: 3.777 min
Delta R.T.: -0.001 min
Response: 102888318
Conc: 459.32 ng/ml



#4 AR-1016-2

R.T.: 4.502 min
Delta R.T.: 0.003 min
Response: 132074415
Conc: 515.32 ng/ml



#4 AR-1016-2

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 156073537
Conc: 458.87 ng/ml



#5 AR-1016-3

R.T.: 4.560 min
Delta R.T.: 0.003 min
Response: 82994793
Conc: 521.59 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PB170162BS



#5 AR-1016-3

R.T.: 3.952 min
Delta R.T.: -0.001 min
Response: 85348928
Conc: 469.31 ng/ml



#6 AR-1016-4

R.T.: 4.650 min
Delta R.T.: 0.004 min
Response: 67743632
Conc: 521.39 ng/ml



#6 AR-1016-4

R.T.: 4.001 min
Delta R.T.: 0.000 min
Response: 72823533
Conc: 462.63 ng/ml



#7 AR-1016-5

R.T.: 4.935 min
Delta R.T.: 0.004 min
Response: 64092788
Conc: 497.60 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PB170162BS



#7 AR-1016-5

R.T.: 4.196 min
Delta R.T.: -0.001 min
Response: 86936492
Conc: 453.17 ng/ml



#31 AR-1260-1

R.T.: 6.024 min
Delta R.T.: 0.005 min
Response: 123203349
Conc: 525.45 ng/ml



#31 AR-1260-1

R.T.: 5.188 min
Delta R.T.: 0.000 min
Response: 173671721
Conc: 481.88 ng/ml



#32 AR-1260-2

R.T.: 6.280 min
Delta R.T.: 0.005 min
Response: 144050828
Conc: 520.77 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PB170162BS



#32 AR-1260-2

R.T.: 5.378 min
Delta R.T.: 0.000 min
Response: 207878210
Conc: 472.46 ng/ml



#33 AR-1260-3

R.T.: 6.631 min
Delta R.T.: 0.005 min
Response: 90911766
Conc: 456.14 ng/ml



#33 AR-1260-3

R.T.: 5.519 min
Delta R.T.: 0.000 min
Response: 200237916
Conc: 482.50 ng/ml



#34 AR-1260-4

R.T.: 6.845 min
Delta R.T.: 0.005 min
Response: 105412977
Conc: 490.56 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PB170162BS



#34 AR-1260-4

R.T.: 5.979 min
Delta R.T.: 0.000 min
Response: 150230243
Conc: 417.58 ng/ml



#35 AR-1260-5

R.T.: 7.152 min
Delta R.T.: 0.007 min
Response: 210162363
Conc: 484.49 ng/ml



#35 AR-1260-5

R.T.: 6.224 min
Delta R.T.: 0.000 min
Response: 341024268
Conc: 405.85 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/17/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/17/25
Client Sample ID:	OR-03-101725MS	SDG No.:	Q3375
Lab Sample ID:	Q3383-01MS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	92
Sample Wt/Vol:	30.08 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date 10/20/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	260		1	4.30	18.4	ug/kg	10/20/25 15:15	PB170162
11104-28-2	Aroclor-1221	18.4	U	1	4.40	18.4	ug/kg	10/20/25 15:15	PB170162
11141-16-5	Aroclor-1232	18.4	U	1	4.00	18.4	ug/kg	10/20/25 15:15	PB170162
53469-21-9	Aroclor-1242	18.4	U	1	4.30	18.4	ug/kg	10/20/25 15:15	PB170162
12672-29-6	Aroclor-1248	18.4	U	1	6.40	18.4	ug/kg	10/20/25 15:15	PB170162
11097-69-1	Aroclor-1254	18.4	U	1	3.50	18.4	ug/kg	10/20/25 15:15	PB170162
37324-23-5	Aroclor-1262	18.4	U	1	5.40	18.4	ug/kg	10/20/25 15:15	PB170162
11100-14-4	Aroclor-1268	18.4	U	1	3.90	18.4	ug/kg	10/20/25 15:15	PB170162
11096-82-5	Aroclor-1260	184		1	3.50	18.4	ug/kg	10/20/25 15:15	PB170162
SURROGATES									
877-09-8	Tetrachloro-m-xylene	28.0			32 - 144	140%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.4			32 - 175	112%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102025\
Data File : PP075879.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 15:15
Operator : YP\AJ
Sample : Q3383-01MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MS

Manual Integrations APPROVED
Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 23:57:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.653	3.792	36650479	161.3E6	28.010m	26.436m
2)	SA Decachlor...	10.435	8.793	22391920	135.2E6	22.447	22.420
	Target Compounds						
3)	L1 AR-1016-1	5.805	4.887	24539847	347.0E6	510.133m	618.441
4)	L1 AR-1016-2	5.827	4.945	38458218	163.5E6	536.163m	636.283
5)	L1 AR-1016-3	5.890	5.065	23702187	85990689	509.387m	553.820
6)	L1 AR-1016-4	5.987	5.106	19610242	88641745	550.283m	603.136
7)	L1 AR-1016-5	6.279	5.326	16939259	194.0E6	502.331m	1187.150 #
31)	L7 AR-1260-1	7.398	6.348	32871064	198.1E6	546.762	488.042
32)	L7 AR-1260-2	7.651	6.534	37968091	297.5E6	521.554	516.710
33)	L7 AR-1260-3	8.009	6.685	24651060	244.8E6	455.277	571.898m#
34)	L7 AR-1260-4	8.236	7.157	30374457	173.2E6	537.248m	401.651 #
35)	L7 AR-1260-5	8.563	7.396	53247355	433.4E6	484.912	411.939

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102025\
Data File : PP075879.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 15:15
Operator : YP\AJ
Sample : Q3383-01MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 23:57:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.653 min
Delta R.T.: -0.002 min
Response: 36650479
Conc: 28.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#1 Tetrachloro-m-xylene

R.T.: 3.792 min
Delta R.T.: -0.004 min
Response: 161284967
Conc: 26.44 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.435 min
Delta R.T.: 0.000 min
Response: 22391920
Conc: 22.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.793 min
Delta R.T.: -0.010 min
Response: 135248926
Conc: 22.42 ng/ml



#3 AR-1016-1

R.T.: 5.805 min
Delta R.T.: -0.002 min
Response: 24539847
Conc: 510.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#3 AR-1016-1

R.T.: 4.887 min
Delta R.T.: -0.006 min
Response: 346953780
Conc: 618.44 ng/ml



#4 AR-1016-2

R.T.: 5.827 min
Delta R.T.: -0.002 min
Response: 38458218
Conc: 536.16 ng/ml m



#4 AR-1016-2

R.T.: 4.945 min
Delta R.T.: -0.006 min
Response: 163526203
Conc: 636.28 ng/ml



#5 AR-1016-3

R.T.: 5.890 min
Delta R.T.: -0.002 min
Response: 23702187
Conc: 509.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#5 AR-1016-3

R.T.: 5.065 min
Delta R.T.: -0.006 min
Response: 85990689
Conc: 553.82 ng/ml



#6 AR-1016-4

R.T.: 5.987 min
Delta R.T.: -0.002 min
Response: 19610242
Conc: 550.28 ng/ml m



#6 AR-1016-4

R.T.: 5.106 min
Delta R.T.: -0.005 min
Response: 88641745
Conc: 603.14 ng/ml



#7 AR-1016-5

R.T.: 6.279 min
Delta R.T.: -0.002 min
Response: 16939259
Conc: 502.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#7 AR-1016-5

R.T.: 5.326 min
Delta R.T.: 0.001 min
Response: 193959087
Conc: 1187.15 ng/ml



#31 AR-1260-1

R.T.: 7.398 min
Delta R.T.: 0.000 min
Response: 32871064
Conc: 546.76 ng/ml



#31 AR-1260-1

R.T.: 6.348 min
Delta R.T.: -0.007 min
Response: 198113318
Conc: 488.04 ng/ml



#32 AR-1260-2

R.T.: 7.651 min
Delta R.T.: 0.000 min
Response: 37968091
Conc: 521.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#32 AR-1260-2

R.T.: 6.534 min
Delta R.T.: -0.007 min
Response: 297537048
Conc: 516.71 ng/ml



#33 AR-1260-3

R.T.: 8.009 min
Delta R.T.: 0.000 min
Response: 24651060
Conc: 455.28 ng/ml



#33 AR-1260-3

R.T.: 6.685 min
Delta R.T.: -0.010 min
Response: 244821152
Conc: 571.90 ng/ml m



#34 AR-1260-4

R.T.: 8.236 min
Delta R.T.: -0.001 min
Response: 30374457
Conc: 537.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#34 AR-1260-4

R.T.: 7.157 min
Delta R.T.: -0.008 min
Response: 173201521
Conc: 401.65 ng/ml



#35 AR-1260-5

R.T.: 8.563 min
Delta R.T.: 0.000 min
Response: 53247355
Conc: 484.91 ng/ml



#35 AR-1260-5

R.T.: 7.396 min
Delta R.T.: -0.007 min
Response: 433355156
Conc: 411.94 ng/ml

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/17/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/17/25
Client Sample ID:	OR-03-101725MSD	SDG No.:	Q3375
Lab Sample ID:	Q3383-01MSD	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	92
Sample Wt/Vol:	30.07 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date 10/20/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	244		1	4.30	18.4	ug/kg	10/20/25 15:31	PB170162
11104-28-2	Aroclor-1221	18.4	U	1	4.40	18.4	ug/kg	10/20/25 15:31	PB170162
11141-16-5	Aroclor-1232	18.4	U	1	4.00	18.4	ug/kg	10/20/25 15:31	PB170162
53469-21-9	Aroclor-1242	18.4	U	1	4.30	18.4	ug/kg	10/20/25 15:31	PB170162
12672-29-6	Aroclor-1248	18.4	U	1	6.40	18.4	ug/kg	10/20/25 15:31	PB170162
11097-69-1	Aroclor-1254	18.4	U	1	3.50	18.4	ug/kg	10/20/25 15:31	PB170162
37324-23-5	Aroclor-1262	18.4	U	1	5.40	18.4	ug/kg	10/20/25 15:31	PB170162
11100-14-4	Aroclor-1268	18.4	U	1	3.90	18.4	ug/kg	10/20/25 15:31	PB170162
11096-82-5	Aroclor-1260	177		1	3.50	18.4	ug/kg	10/20/25 15:31	PB170162
SURROGATES									
877-09-8	Tetrachloro-m-xylene	28.4			32 - 144	142%	SPK: 20		
2051-24-3	Decachlorobiphenyl	21.8			32 - 175	109%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102025\
Data File : PP075880.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 15:31
Operator : YP\AJ
Sample : Q3383-01MSD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MSD

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 23:58:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.653	3.789	37147413	115.5E6	28.390	18.926m#
2)	SA Decachlor...	10.433	8.790	21750647	128.6E6	21.804	21.323
	Target Compounds						
3)	L1 AR-1016-1	5.804	4.886	26019155	315.2E6	540.885m	561.814
4)	L1 AR-1016-2	5.825	4.943	37734613	147.0E6	526.075m	572.145
5)	L1 AR-1016-3	5.888	5.064	23479000	80176784	504.591m	516.376
6)	L1 AR-1016-4	5.985	5.104	19024806	80655198	533.855m	548.794
7)	L1 AR-1016-5	6.277	5.324	16406275	192.7E6	486.525m	1179.626 #
31)	L7 AR-1260-1	7.396	6.347	31688258	186.5E6	527.088	459.423
32)	L7 AR-1260-2	7.649	6.533	36612040	284.3E6	502.926	493.795
33)	L7 AR-1260-3	8.007	6.684	23669956	242.5E6	437.158	566.518m#
34)	L7 AR-1260-4	8.234	7.154	29025293	175.9E6	513.384m	407.883
35)	L7 AR-1260-5	8.562	7.395	51776195	407.0E6	471.514	386.850

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102025\
Data File : PP075880.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 15:31
Operator : YP\AJ
Sample : Q3383-01MSD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MSD

Manual Integrations APPROVED
Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 23:58:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.653 min
Delta R.T.: -0.002 min
Response: 37147413
Conc: 28.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#1 Tetrachloro-m-xylene

R.T.: 3.789 min
Delta R.T.: -0.007 min
Response: 115464507
Conc: 18.93 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.433 min
Delta R.T.: -0.002 min
Response: 21750647
Conc: 21.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.790 min
Delta R.T.: -0.013 min
Response: 128630164
Conc: 21.32 ng/ml



#3 AR-1016-1

R.T.: 5.804 min
Delta R.T.: -0.004 min
Response: 26019155
Conc: 540.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MSD

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#3 AR-1016-1

R.T.: 4.886 min
Delta R.T.: -0.007 min
Response: 315185195
Conc: 561.81 ng/ml



#4 AR-1016-2

R.T.: 5.825 min
Delta R.T.: -0.004 min
Response: 37734613
Conc: 526.07 ng/ml m



#4 AR-1016-2

R.T.: 4.943 min
Delta R.T.: -0.008 min
Response: 147042454
Conc: 572.14 ng/ml



#5 AR-1016-3

R.T.: 5.888 min
Delta R.T.: -0.004 min
Response: 23479000
Conc: 504.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MSD

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#5 AR-1016-3

R.T.: 5.064 min
Delta R.T.: -0.007 min
Response: 80176784
Conc: 516.38 ng/ml



#6 AR-1016-4

R.T.: 5.985 min
Delta R.T.: -0.004 min
Response: 19024806
Conc: 533.85 ng/ml m



#6 AR-1016-4

R.T.: 5.104 min
Delta R.T.: -0.007 min
Response: 80655198
Conc: 548.79 ng/ml



#7 AR-1016-5

R.T.: 6.277 min
Delta R.T.: -0.004 min
Response: 16406275
Conc: 486.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MSD

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#7 AR-1016-5

R.T.: 5.324 min
Delta R.T.: -0.001 min
Response: 192729880
Conc: 1179.63 ng/ml



#31 AR-1260-1

R.T.: 7.396 min
Delta R.T.: -0.002 min
Response: 31688258
Conc: 527.09 ng/ml



#31 AR-1260-1

R.T.: 6.347 min
Delta R.T.: -0.008 min
Response: 186495934
Conc: 459.42 ng/ml



#32 AR-1260-2

R.T.: 7.649 min
Delta R.T.: -0.002 min
Response: 36612040
Conc: 502.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MSD

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#32 AR-1260-2

R.T.: 6.533 min
Delta R.T.: -0.008 min
Response: 284342129
Conc: 493.80 ng/ml



#33 AR-1260-3

R.T.: 8.007 min
Delta R.T.: -0.002 min
Response: 23669956
Conc: 437.16 ng/ml



#33 AR-1260-3

R.T.: 6.684 min
Delta R.T.: -0.011 min
Response: 242517720
Conc: 566.52 ng/ml m



#34 AR-1260-4

R.T.: 8.234 min
Delta R.T.: -0.003 min
Response: 29025293
Conc: 513.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
OR-03-101725MSD

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#34 AR-1260-4

R.T.: 7.154 min
Delta R.T.: -0.010 min
Response: 175888915
Conc: 407.88 ng/ml



#35 AR-1260-5

R.T.: 8.562 min
Delta R.T.: -0.002 min
Response: 51776195
Conc: 471.51 ng/ml



#35 AR-1260-5

R.T.: 7.395 min
Delta R.T.: -0.008 min
Response: 406962039
Conc: 386.85 ng/ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp092325	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC250	PP075250.D	AR-1260-4	yogesh	9/24/2025 8:09:44 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	AR-1016-3	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	AR-1016-5	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	AR-1260-1	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	AR-1260-3 #2	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	AR-1260-4	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	AR-1260-5 #2	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	Decachlorobiphenyl	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICC1000	PP075254.D	AR-1242-4 #2	yogesh	9/24/2025 8:09:49 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICC750	PP075255.D	AR-1242-4 #2	yogesh	9/24/2025 8:09:51 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICC500	PP075256.D	AR-1242-4 #2	yogesh	9/24/2025 8:09:53 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICC250	PP075257.D	AR-1242-4 #2	yogesh	9/24/2025 8:09:55 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICC050	PP075258.D	AR-1242-5	yogesh	9/24/2025 8:09:57 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp092325	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242ICC050	PP075258.D	Tetrachloro-m-xylene	yogesh	9/24/2025 8:09:57 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC1000	PP075259.D	AR-1248-3 #2	yogesh	9/24/2025 8:09:59 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC750	PP075260.D	AR-1248-3 #2	yogesh	9/24/2025 8:10:01 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC500	PP075261.D	AR-1248-3 #2	yogesh	9/24/2025 8:10:03 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC250	PP075262.D	AR-1248-3 #2	yogesh	9/24/2025 8:10:05 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC050	PP075263.D	AR-1248-3 #2	yogesh	9/24/2025 8:10:06 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC050	PP075263.D	AR-1248-5	yogesh	9/24/2025 8:10:06 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC050	PP075263.D	Tetrachloro-m-xylene	yogesh	9/24/2025 8:10:06 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICC1000	PP075264.D	AR-1254-1	yogesh	9/24/2025 8:10:09 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICC750	PP075265.D	AR-1254-1	yogesh	9/24/2025 8:10:11 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICC500	PP075266.D	AR-1254-1	yogesh	9/24/2025 8:10:12 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICC250	PP075267.D	AR-1254-1	yogesh	9/24/2025 8:10:15 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICC050	PP075268.D	AR-1254-1	yogesh	9/24/2025 8:10:16 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp092325	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1268ICC050	PP075274.D	AR-1268-1	yogesh	9/24/2025 8:10:18 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1268ICC050	PP075274.D	AR-1268-3 #2	yogesh	9/24/2025 8:10:18 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1268ICC050	PP075274.D	AR-1268-4 #2	yogesh	9/24/2025 8:10:18 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1268ICC050	PP075274.D	AR-1268-5 #2	yogesh	9/24/2025 8:10:18 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1268ICC050	PP075274.D	Tetrachloro-m-xylene #2	yogesh	9/24/2025 8:10:18 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
PP092325ICV500	PP075275.D	Decachlorobiphenyl #2	yogesh	9/24/2025 8:10:21 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICV500	PP075276.D	AR-1242-4 #2	yogesh	9/24/2025 8:10:24 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICV500	PP075276.D	Decachlorobiphenyl #2	yogesh	9/24/2025 8:10:24 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICV500	PP075277.D	AR-1248-3 #2	yogesh	9/24/2025 8:10:26 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICV500	PP075278.D	AR-1254-1	yogesh	9/24/2025 8:10:28 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICV500	PP075278.D	Decachlorobiphenyl #2	yogesh	9/24/2025 8:10:28 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1268ICV500	PP075279.D	Decachlorobiphenyl #2	yogesh	9/24/2025 8:10:30 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp092325	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp102025	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP075868.D	AR-1016-1 #2	yogesh	10/23/2025 3:43:59 PM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1660CCC500	PP075868.D	AR-1016-2 #2	yogesh	10/23/2025 3:43:59 PM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1660CCC500	PP075868.D	AR-1016-4 #2	yogesh	10/23/2025 3:43:59 PM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1660CCC500	PP075868.D	AR-1016-5 #2	yogesh	10/23/2025 3:43:59 PM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1242CCC500	PP075869.D	AR-1242-1 #2	yogesh	10/21/2025 11:09:50 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1242CCC500	PP075869.D	AR-1242-2 #2	yogesh	10/21/2025 11:09:50 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1242CCC500	PP075869.D	AR-1242-4 #2	yogesh	10/21/2025 11:09:50 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1248CCC500	PP075870.D	AR-1248-2 #2	yogesh	10/21/2025 11:09:53 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1248CCC500	PP075870.D	AR-1248-3 #2	yogesh	10/21/2025 11:09:53 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1248CCC500	PP075870.D	AR-1248-5 #2	yogesh	10/21/2025 11:09:53 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1248CCC500	PP075870.D	Tetrachloro-m-xylene #2	yogesh	10/21/2025 11:09:53 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1254CCC500	PP075871.D	AR-1254-1	yogesh	10/21/2025 11:09:56 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1254CCC500	PP075871.D	Tetrachloro-m-xylene #2	yogesh	10/21/2025 11:09:56 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp102025	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3383-01MS	PP075879.D	AR-1016-1	yogesh	10/21/2025 11:10:02 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MS	PP075879.D	AR-1016-2	yogesh	10/21/2025 11:10:02 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MS	PP075879.D	AR-1016-3	yogesh	10/21/2025 11:10:02 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MS	PP075879.D	AR-1016-4	yogesh	10/21/2025 11:10:02 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MS	PP075879.D	AR-1016-5	yogesh	10/21/2025 11:10:02 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MS	PP075879.D	AR-1260-3 #2	yogesh	10/21/2025 11:10:02 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MS	PP075879.D	AR-1260-4	yogesh	10/21/2025 11:10:02 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MS	PP075879.D	Tetrachloro-m-xylene	yogesh	10/21/2025 11:10:02 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MS	PP075879.D	Tetrachloro-m-xylene #2	yogesh	10/21/2025 11:10:02 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MSD	PP075880.D	AR-1016-1	yogesh	10/21/2025 11:10:05 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MSD	PP075880.D	AR-1016-2	yogesh	10/21/2025 11:10:05 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MSD	PP075880.D	AR-1016-3	yogesh	10/21/2025 11:10:05 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MSD	PP075880.D	AR-1016-4	yogesh	10/21/2025 11:10:05 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp102025	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3383-01MSD	PP075880.D	AR-1016-5	yogesh	10/21/2025 11:10:05 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MSD	PP075880.D	AR-1260-3 #2	yogesh	10/21/2025 11:10:05 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MSD	PP075880.D	AR-1260-4	yogesh	10/21/2025 11:10:05 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
Q3383-01MSD	PP075880.D	Tetrachloro-m-xylene #2	yogesh	10/21/2025 11:10:05 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1660CCC500	PP075881.D	AR-1016-2 #2	yogesh	10/23/2025 3:44:00 PM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1660CCC500	PP075881.D	Decachlorobiphenyl	yogesh	10/23/2025 3:44:00 PM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1660CCC500	PP075881.D	Tetrachloro-m-xylene	yogesh	10/23/2025 3:44:00 PM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1242CCC500	PP075882.D	AR-1242-1 #2	yogesh	10/21/2025 11:10:10 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1242CCC500	PP075882.D	AR-1242-2 #2	yogesh	10/21/2025 11:10:10 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1242CCC500	PP075882.D	AR-1242-3 #2	yogesh	10/21/2025 11:10:10 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1242CCC500	PP075882.D	AR-1242-4 #2	yogesh	10/21/2025 11:10:10 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1242CCC500	PP075882.D	Tetrachloro-m-xylene #2	yogesh	10/21/2025 11:10:10 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1248CCC500	PP075883.D	AR-1248-1 #2	yogesh	10/21/2025 11:10:31 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp102025	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1248CCC500	PP075883.D	AR-1248-2 #2	yogesh	10/21/2025 11:10:31 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1248CCC500	PP075883.D	AR-1248-3 #2	yogesh	10/21/2025 11:10:31 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1248CCC500	PP075883.D	AR-1248-4 #2	yogesh	10/21/2025 11:10:31 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1248CCC500	PP075883.D	AR-1248-5 #2	yogesh	10/21/2025 11:10:31 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1248CCC500	PP075883.D	Tetrachloro-m-xylene #2	yogesh	10/21/2025 11:10:31 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1254CCC500	PP075884.D	AR-1254-1	yogesh	10/21/2025 11:10:12 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1254CCC500	PP075884.D	AR-1254-1 #2	yogesh	10/21/2025 11:10:12 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software
AR1254CCC500	PP075884.D	Tetrachloro-m-xylene #2	yogesh	10/21/2025 11:10:12 AM	mohammad	10/24/2025 1:00:41	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pq101025	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC050	PQ070977.D	AR-1260-1	yogesh	10/13/2025 8:00:35 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1660ICC050	PQ070977.D	AR-1260-2	yogesh	10/13/2025 8:00:35 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-1 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-2 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-3 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-4 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-5	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-5 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICC250	PQ070988.D	AR-1248-4 #2	yogesh	10/13/2025 8:00:39 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICC250	PQ070988.D	AR-1248-5 #2	yogesh	10/13/2025 8:00:39 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1016-1 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1016-2 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1016-3 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pq101025	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PQ101025ICV	PQ071001.D	AR-1016-4 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1016-5 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1260-1	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1260-1 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1260-2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1260-2 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICV500	PQ071002.D	AR-1242-3 #2	yogesh	10/13/2025 8:00:43 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICV500	PQ071003.D	AR-1248-4	yogesh	10/13/2025 8:00:45 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICV500	PQ071003.D	AR-1248-4 #2	yogesh	10/13/2025 8:00:45 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICV500	PQ071003.D	AR-1248-5	yogesh	10/13/2025 8:00:45 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PQ102025	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PQ071008.D	AR-1016-2	yogesh	10/21/2025 8:51:45 AM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
I.BLK	PQ071012.D	Decachlorobiphenyl	yogesh	10/24/2025 3:08:45 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
PB170162BL	PQ071013.D	Decachlorobiphenyl	yogesh	10/23/2025 3:43:24 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-03	PQ071019.D	AR-1254-1 #2	yogesh	10/21/2025 8:51:38 AM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-03	PQ071019.D	AR-1254-2 #2	yogesh	10/21/2025 8:51:38 AM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-03	PQ071019.D	AR-1254-3 #2	yogesh	10/21/2025 8:51:38 AM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-05	PQ071021.D	AR-1242-1	yogesh	10/23/2025 3:43:26 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-05	PQ071021.D	AR-1242-1 #2	yogesh	10/23/2025 3:43:26 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-05	PQ071021.D	AR-1242-2	yogesh	10/23/2025 3:43:26 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-05	PQ071021.D	AR-1242-2 #2	yogesh	10/23/2025 3:43:26 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-05	PQ071021.D	AR-1242-3 #2	yogesh	10/23/2025 3:43:26 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-05	PQ071021.D	AR-1242-4 #2	yogesh	10/23/2025 3:43:26 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-05	PQ071021.D	AR-1242-5	yogesh	10/23/2025 3:43:26 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PQ102025	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3375-05	PQ071021.D	AR-1254-1 #2	yogesh	10/23/2025 3:43:26 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-05	PQ071021.D	AR-1254-2 #2	yogesh	10/23/2025 3:43:26 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-05	PQ071021.D	AR-1254-3	yogesh	10/23/2025 3:43:26 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software
Q3375-05	PQ071021.D	AR-1254-5	yogesh	10/23/2025 3:43:26 PM	mohammad	10/25/2025 7:26:09	Peak Integrated by Software

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP092325

Review By	yogesh	Review On	9/23/2025 12:35:13 PM		
Supervise By	mohammad	Supervise On	9/25/2025 2:25:37 AM		
SubDirectory	PP092325	HP Acquire Method		HP Processing Method	PP092325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP075245.D	23 Sep 2025 08:13	YP\AJ	Ok
2	I.BLK	PP075246.D	23 Sep 2025 08:29	YP\AJ	Ok
3	AR1660ICC1000	PP075247.D	23 Sep 2025 08:45	YP\AJ	Ok
4	AR1660ICC750	PP075248.D	23 Sep 2025 09:01	YP\AJ	Ok
5	AR1660ICC500	PP075249.D	23 Sep 2025 09:18	YP\AJ	Ok
6	AR1660ICC250	PP075250.D	23 Sep 2025 09:34	YP\AJ	Ok,M
7	AR1660ICC050	PP075251.D	23 Sep 2025 10:11	YP\AJ	Ok,M
8	AR1221ICC500	PP075252.D	23 Sep 2025 10:27	YP\AJ	Ok
9	AR1232ICC500	PP075253.D	23 Sep 2025 10:43	YP\AJ	Ok
10	AR1242ICC1000	PP075254.D	23 Sep 2025 11:00	YP\AJ	Ok,M
11	AR1242ICC750	PP075255.D	23 Sep 2025 11:16	YP\AJ	Ok,M
12	AR1242ICC500	PP075256.D	23 Sep 2025 11:32	YP\AJ	Ok,M
13	AR1242ICC250	PP075257.D	23 Sep 2025 11:48	YP\AJ	Ok,M
14	AR1242ICC050	PP075258.D	23 Sep 2025 12:05	YP\AJ	Ok,M
15	AR1248ICC1000	PP075259.D	23 Sep 2025 12:37	YP\AJ	Ok,M
16	AR1248ICC750	PP075260.D	23 Sep 2025 12:53	YP\AJ	Ok,M
17	AR1248ICC500	PP075261.D	23 Sep 2025 13:10	YP\AJ	Ok,M
18	AR1248ICC250	PP075262.D	23 Sep 2025 13:26	YP\AJ	Ok,M
19	AR1248ICC050	PP075263.D	23 Sep 2025 13:42	YP\AJ	Ok,M
20	AR1254ICC1000	PP075264.D	23 Sep 2025 14:15	YP\AJ	Ok,M
21	AR1254ICC750	PP075265.D	23 Sep 2025 14:31	YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP092325

Review By	yogesh	Review On	9/23/2025 12:35:13 PM		
Supervise By	mohammad	Supervise On	9/25/2025 2:25:37 AM		
SubDirectory	PP092325	HP Acquire Method		HP Processing Method	PP092325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	AR1254ICC500	PP075266.D	23 Sep 2025 14:47	YP\AJ	Ok,M
23	AR1254ICC250	PP075267.D	23 Sep 2025 15:04	YP\AJ	Ok,M
24	AR1254ICC050	PP075268.D	23 Sep 2025 15:20	YP\AJ	Ok,M
25	AR1262ICC500	PP075269.D	23 Sep 2025 15:52	YP\AJ	Ok
26	AR1268ICC1000	PP075270.D	23 Sep 2025 16:09	YP\AJ	Ok
27	AR1268ICC750	PP075271.D	23 Sep 2025 16:25	YP\AJ	Ok
28	AR1268ICC500	PP075272.D	23 Sep 2025 16:41	YP\AJ	Ok
29	AR1268ICC250	PP075273.D	23 Sep 2025 16:58	YP\AJ	Ok
30	AR1268ICC050	PP075274.D	23 Sep 2025 17:14	YP\AJ	Ok,M
31	PP092325ICV500	PP075275.D	23 Sep 2025 17:46	YP\AJ	Ok,M
32	AR1242ICV500	PP075276.D	23 Sep 2025 18:35	YP\AJ	Ok,M
33	AR1248ICV500	PP075277.D	23 Sep 2025 19:08	YP\AJ	Ok,M
34	AR1254ICV500	PP075278.D	23 Sep 2025 19:40	YP\AJ	Ok,M
35	AR1268ICV500	PP075279.D	23 Sep 2025 19:57	YP\AJ	Ok,M
36	DDT ANALOG	PP075280.D	23 Sep 2025 20:29	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102025

Review By	yogesh	Review On	10/20/2025 1:46:56 PM		
Supervise By	mohammad	Supervise On	10/24/2025 1:00:41 AM		
SubDirectory	PP102025	HP Acquire Method		HP Processing Method	PP092325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP075867.D	20 Oct 2025 09:49	YP\AJ	Ok
2	AR1660CCC500	PP075868.D	20 Oct 2025 11:13	YP\AJ	Ok,M
3	AR1242CCC500	PP075869.D	20 Oct 2025 11:53	YP\AJ	Ok,M
4	AR1248CCC500	PP075870.D	20 Oct 2025 12:09	YP\AJ	Ok,M
5	AR1254CCC500	PP075871.D	20 Oct 2025 12:25	YP\AJ	Ok,M
6	I.BLK	PP075872.D	20 Oct 2025 12:41	YP\AJ	Ok
7	Q3307-06DL	PP075873.D	20 Oct 2025 12:58	YP\AJ	Ok,M
8	Q3381-03	PP075874.D	20 Oct 2025 13:53	YP\AJ	Ok
9	Q3381-05	PP075875.D	20 Oct 2025 14:10	YP\AJ	Ok
10	Q3381-07	PP075876.D	20 Oct 2025 14:26	YP\AJ	Ok
11	Q3382-01	PP075877.D	20 Oct 2025 14:42	YP\AJ	Ok
12	Q3383-01	PP075878.D	20 Oct 2025 14:58	YP\AJ	Ok,M
13	Q3383-01MS	PP075879.D	20 Oct 2025 15:15	YP\AJ	Ok,M
14	Q3383-01MSD	PP075880.D	20 Oct 2025 15:31	YP\AJ	Ok,M
15	AR1660CCC500	PP075881.D	20 Oct 2025 16:20	YP\AJ	Ok,M
16	AR1242CCC500	PP075882.D	20 Oct 2025 16:36	YP\AJ	Ok,M
17	AR1248CCC500	PP075883.D	20 Oct 2025 16:52	YP\AJ	Ok,M
18	AR1254CCC500	PP075884.D	20 Oct 2025 17:09	YP\AJ	Ok,M
19	I.BLK	PP075885.D	20 Oct 2025 17:25	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM		
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM		
SubDirectory	PQ101025	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PQ070971.D	10 Oct 2025 08:27	YP\AJ	Ok
2	I.BLK	PQ070972.D	10 Oct 2025 08:42	YP\AJ	Ok
3	AR1660ICC1000	PQ070973.D	10 Oct 2025 08:57	YP\AJ	Ok
4	AR1660ICC750	PQ070974.D	10 Oct 2025 09:11	YP\AJ	Ok
5	AR1660ICC500	PQ070975.D	10 Oct 2025 09:26	YP\AJ	Ok
6	AR1660ICC250	PQ070976.D	10 Oct 2025 09:40	YP\AJ	Ok
7	AR1660ICC050	PQ070977.D	10 Oct 2025 09:55	YP\AJ	Ok,M
8	AR1221ICC500	PQ070978.D	10 Oct 2025 10:09	YP\AJ	Ok
9	AR1232ICC500	PQ070979.D	10 Oct 2025 10:24	YP\AJ	Ok
10	AR1242ICC1000	PQ070980.D	10 Oct 2025 10:38	YP\AJ	Ok
11	AR1242ICC750	PQ070981.D	10 Oct 2025 10:53	YP\AJ	Ok
12	AR1242ICC500	PQ070982.D	10 Oct 2025 11:08	YP\AJ	Ok
13	AR1242ICC250	PQ070983.D	10 Oct 2025 11:22	YP\AJ	Ok
14	AR1242ICC050	PQ070984.D	10 Oct 2025 11:37	YP\AJ	Ok,M
15	AR1248ICC1000	PQ070985.D	10 Oct 2025 11:51	YP\AJ	Ok
16	AR1248ICC750	PQ070986.D	10 Oct 2025 12:06	YP\AJ	Ok
17	AR1248ICC500	PQ070987.D	10 Oct 2025 12:20	YP\AJ	Ok
18	AR1248ICC250	PQ070988.D	10 Oct 2025 12:35	YP\AJ	Ok,M
19	AR1248ICC050	PQ070989.D	10 Oct 2025 12:50	YP\AJ	Ok
20	AR1254ICC1000	PQ070990.D	10 Oct 2025 13:04	YP\AJ	Ok
21	AR1254ICC750	PQ070991.D	10 Oct 2025 13:19	YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM		
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM		
SubDirectory	PQ101025	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	AR1254ICC500	PQ070992.D	10 Oct 2025 13:33	YP\AJ	Ok
23	AR1254ICC250	PQ070993.D	10 Oct 2025 13:48	YP\AJ	Ok
24	AR1254ICC050	PQ070994.D	10 Oct 2025 15:15	YP\AJ	Ok
25	AR1262ICC500	PQ070995.D	10 Oct 2025 15:30	YP\AJ	Ok
26	AR1268ICC1000	PQ070996.D	10 Oct 2025 15:44	YP\AJ	Ok
27	AR1268ICC750	PQ070997.D	10 Oct 2025 15:59	YP\AJ	Ok
28	AR1268ICC500	PQ070998.D	10 Oct 2025 16:13	YP\AJ	Ok
29	AR1268ICC250	PQ070999.D	10 Oct 2025 16:28	YP\AJ	Ok
30	AR1268ICC050	PQ071000.D	10 Oct 2025 16:43	YP\AJ	Ok
31	PQ101025ICV	PQ071001.D	10 Oct 2025 16:57	YP\AJ	Ok,M
32	AR1242ICV500	PQ071002.D	10 Oct 2025 17:41	YP\AJ	Ok,M
33	AR1248ICV500	PQ071003.D	10 Oct 2025 18:25	YP\AJ	Ok,M
34	AR1254ICV500	PQ071004.D	10 Oct 2025 19:08	YP\AJ	Ok
35	AR1268ICV500	PQ071005.D	10 Oct 2025 19:52	YP\AJ	Ok
36	DDT ANALOG	PQ071006.D	10 Oct 2025 20:36	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ102025

Review By	yogesh	Review On	10/20/2025 1:25:53 PM		
Supervise By	mohammad	Supervise On	10/25/2025 7:26:09 AM		
SubDirectory	PQ102025	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PQ071007.D	20 Oct 2025 09:06	YP\AJ	Ok
2	AR1660CCC500	PQ071008.D	20 Oct 2025 10:18	YP\AJ	Ok,M
3	AR1242CCC500	PQ071009.D	20 Oct 2025 10:34	YP\AJ	Ok
4	AR1248CCC500	PQ071010.D	20 Oct 2025 10:52	YP\AJ	Ok
5	AR1254CCC500	PQ071011.D	20 Oct 2025 11:11	YP\AJ	Ok
6	I.BLK	PQ071012.D	20 Oct 2025 12:36	YP\AJ	Ok,M
7	PB170162BL	PQ071013.D	20 Oct 2025 13:20	YP\AJ	Ok,M
8	PB170162BS	PQ071014.D	20 Oct 2025 13:35	YP\AJ	Ok
9	Q3374-01	PQ071015.D	20 Oct 2025 13:49	YP\AJ	Ok,M
10	Q3374-05	PQ071016.D	20 Oct 2025 14:04	YP\AJ	Ok
11	Q3375-01	PQ071017.D	20 Oct 2025 14:19	YP\AJ	Ok
12	Q3375-02	PQ071018.D	20 Oct 2025 14:33	YP\AJ	Ok
13	Q3375-03	PQ071019.D	20 Oct 2025 14:48	YP\AJ	Ok,M
14	Q3375-04	PQ071020.D	20 Oct 2025 15:02	YP\AJ	Ok
15	Q3375-05	PQ071021.D	20 Oct 2025 15:47	YP\AJ	Ok,M
16	Q3381-01	PQ071022.D	20 Oct 2025 16:01	YP\AJ	Ok
17	AR1660CCC500	PQ071023.D	20 Oct 2025 16:45	YP\AJ	Ok
18	AR1242CCC500	PQ071024.D	20 Oct 2025 17:00	YP\AJ	Ok
19	AR1248CCC500	PQ071025.D	20 Oct 2025 17:14	YP\AJ	Ok
20	AR1254CCC500	PQ071026.D	20 Oct 2025 17:29	YP\AJ	Ok
21	I.BLK	PQ071027.D	20 Oct 2025 17:44	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP092325

Review By	yogesh	Review On	9/23/2025 12:35:13 PM		
Supervise By	mohammad	Supervise On	9/25/2025 2:25:37 AM		
SubDirectory	PP092325	HP Acquire Method		HP Processing Method	PP092325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075245.D	23 Sep 2025 08:13		YP\AJ	Ok
2	I.BLK	I.BLK	PP075246.D	23 Sep 2025 08:29		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP075247.D	23 Sep 2025 08:45		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP075248.D	23 Sep 2025 09:01		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP075249.D	23 Sep 2025 09:18		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP075250.D	23 Sep 2025 09:34		YP\AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PP075251.D	23 Sep 2025 10:11		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP075252.D	23 Sep 2025 10:27		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP075253.D	23 Sep 2025 10:43		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP075254.D	23 Sep 2025 11:00		YP\AJ	Ok,M
11	AR1242ICC750	AR1242ICC750	PP075255.D	23 Sep 2025 11:16		YP\AJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP075256.D	23 Sep 2025 11:32		YP\AJ	Ok,M
13	AR1242ICC250	AR1242ICC250	PP075257.D	23 Sep 2025 11:48		YP\AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP075258.D	23 Sep 2025 12:05		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP075259.D	23 Sep 2025 12:37		YP\AJ	Ok,M
16	AR1248ICC750	AR1248ICC750	PP075260.D	23 Sep 2025 12:53		YP\AJ	Ok,M
17	AR1248ICC500	AR1248ICC500	PP075261.D	23 Sep 2025 13:10		YP\AJ	Ok,M
18	AR1248ICC250	AR1248ICC250	PP075262.D	23 Sep 2025 13:26		YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP092325

Review By	yogesh	Review On	9/23/2025 12:35:13 PM		
Supervise By	mohammad	Supervise On	9/25/2025 2:25:37 AM		
SubDirectory	PP092325	HP Acquire Method		HP Processing Method	PP092325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	AR1248ICC050	AR1248ICC050	PP075263.D	23 Sep 2025 13:42		YP\AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP075264.D	23 Sep 2025 14:15		YP\AJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP075265.D	23 Sep 2025 14:31		YP\AJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP075266.D	23 Sep 2025 14:47		YP\AJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP075267.D	23 Sep 2025 15:04		YP\AJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP075268.D	23 Sep 2025 15:20		YP\AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP075269.D	23 Sep 2025 15:52		YP\AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP075270.D	23 Sep 2025 16:09		YP\AJ	Ok
27	AR1268ICC750	AR1268ICC750	PP075271.D	23 Sep 2025 16:25		YP\AJ	Ok
28	AR1268ICC500	AR1268ICC500	PP075272.D	23 Sep 2025 16:41		YP\AJ	Ok
29	AR1268ICC250	AR1268ICC250	PP075273.D	23 Sep 2025 16:58		YP\AJ	Ok
30	AR1268ICC050	AR1268ICC050	PP075274.D	23 Sep 2025 17:14		YP\AJ	Ok,M
31	PP092325ICV500	ICVPP092325	PP075275.D	23 Sep 2025 17:46		YP\AJ	Ok,M
32	AR1242ICV500	ICVPP092325AR1242	PP075276.D	23 Sep 2025 18:35		YP\AJ	Ok,M
33	AR1248ICV500	ICVPP092325AR1248	PP075277.D	23 Sep 2025 19:08		YP\AJ	Ok,M
34	AR1254ICV500	ICVPP092325AR1254	PP075278.D	23 Sep 2025 19:40		YP\AJ	Ok,M
35	AR1268ICV500	ICVPP092325AR1268	PP075279.D	23 Sep 2025 19:57		YP\AJ	Ok,M
36	DDT ANALOG	DDT ANALOG	PP075280.D	23 Sep 2025 20:29		YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102025

Review By	yogesh	Review On	10/20/2025 1:46:56 PM		
Supervise By	mohammad	Supervise On	10/24/2025 1:00:41 AM		
SubDirectory	PP102025	HP Acquire Method		HP Processing Method	PP092325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075867.D	20 Oct 2025 09:49		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP075868.D	20 Oct 2025 11:13		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP075869.D	20 Oct 2025 11:53		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP075870.D	20 Oct 2025 12:09		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP075871.D	20 Oct 2025 12:25		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP075872.D	20 Oct 2025 12:41		YP\AJ	Ok
7	Q3307-06DL	S-38(0-0.5)DL	PP075873.D	20 Oct 2025 12:58	AR1260+1254 Hit	YP\AJ	Ok,M
8	Q3381-03	VNJ-245	PP075874.D	20 Oct 2025 13:53	TCMX high in 2nd column	YP\AJ	Ok
9	Q3381-05	RBR200059	PP075875.D	20 Oct 2025 14:10	TCMX high in 2nd column	YP\AJ	Ok
10	Q3381-07	VNJ-255	PP075876.D	20 Oct 2025 14:26	TCMX high in 2nd column	YP\AJ	Ok
11	Q3382-01	CENTRAL-AVE-TP	PP075877.D	20 Oct 2025 14:42		YP\AJ	Ok
12	Q3383-01	OR-03-101725	PP075878.D	20 Oct 2025 14:58		YP\AJ	Ok,M
13	Q3383-01MS	OR-03-101725MS	PP075879.D	20 Oct 2025 15:15		YP\AJ	Ok,M
14	Q3383-01MSD	OR-03-101725MSD	PP075880.D	20 Oct 2025 15:31		YP\AJ	Ok,M
15	AR1660CCC500	AR1660CCC500	PP075881.D	20 Oct 2025 16:20	DCB Low in 1st column	YP\AJ	Ok,M
16	AR1242CCC500	AR1242CCC500	PP075882.D	20 Oct 2025 16:36		YP\AJ	Ok,M
17	AR1248CCC500	AR1248CCC500	PP075883.D	20 Oct 2025 16:52		YP\AJ	Ok,M
18	AR1254CCC500	AR1254CCC500	PP075884.D	20 Oct 2025 17:09		YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102025

Review By	yogesh	Review On	10/20/2025 1:46:56 PM		
Supervise By	mohammad	Supervise On	10/24/2025 1:00:41 AM		
SubDirectory	PP102025	HP Acquire Method		HP Processing Method	PP092325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	I.BLK	I.BLK	PP075885.D	20 Oct 2025 17:25		YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM		
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM		
SubDirectory	PQ101025	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ070971.D	10 Oct 2025 08:27		YP\AJ	Ok
2	I.BLK	I.BLK	PQ070972.D	10 Oct 2025 08:42		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PQ070973.D	10 Oct 2025 08:57		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PQ070974.D	10 Oct 2025 09:11		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PQ070975.D	10 Oct 2025 09:26		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PQ070976.D	10 Oct 2025 09:40		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PQ070977.D	10 Oct 2025 09:55		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PQ070978.D	10 Oct 2025 10:09		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PQ070979.D	10 Oct 2025 10:24		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PQ070980.D	10 Oct 2025 10:38		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PQ070981.D	10 Oct 2025 10:53		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PQ070982.D	10 Oct 2025 11:08		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PQ070983.D	10 Oct 2025 11:22		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PQ070984.D	10 Oct 2025 11:37		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PQ070985.D	10 Oct 2025 11:51		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PQ070986.D	10 Oct 2025 12:06		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PQ070987.D	10 Oct 2025 12:20		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PQ070988.D	10 Oct 2025 12:35		YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM		
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM		
SubDirectory	PQ101025	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	AR1248ICC050	AR1248ICC050	PQ070989.D	10 Oct 2025 12:50		YP\AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PQ070990.D	10 Oct 2025 13:04		YP\AJ	Ok
21	AR1254ICC750	AR1254ICC750	PQ070991.D	10 Oct 2025 13:19		YP\AJ	Ok
22	AR1254ICC500	AR1254ICC500	PQ070992.D	10 Oct 2025 13:33		YP\AJ	Ok
23	AR1254ICC250	AR1254ICC250	PQ070993.D	10 Oct 2025 13:48		YP\AJ	Ok
24	AR1254ICC050	AR1254ICC050	PQ070994.D	10 Oct 2025 15:15		YP\AJ	Ok
25	AR1262ICC500	AR1262ICC500	PQ070995.D	10 Oct 2025 15:30		YP\AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PQ070996.D	10 Oct 2025 15:44		YP\AJ	Ok
27	AR1268ICC750	AR1268ICC750	PQ070997.D	10 Oct 2025 15:59		YP\AJ	Ok
28	AR1268ICC500	AR1268ICC500	PQ070998.D	10 Oct 2025 16:13		YP\AJ	Ok
29	AR1268ICC250	AR1268ICC250	PQ070999.D	10 Oct 2025 16:28		YP\AJ	Ok
30	AR1268ICC050	AR1268ICC050	PQ071000.D	10 Oct 2025 16:43		YP\AJ	Ok
31	PQ101025ICV	ICVPQ101025	PQ071001.D	10 Oct 2025 16:57		YP\AJ	Ok,M
32	AR1242ICV500	ICVPQ101025AR1242	PQ071002.D	10 Oct 2025 17:41		YP\AJ	Ok,M
33	AR1248ICV500	ICVPQ101025AR1248	PQ071003.D	10 Oct 2025 18:25		YP\AJ	Ok,M
34	AR1254ICV500	ICVPQ101025AR1254	PQ071004.D	10 Oct 2025 19:08		YP\AJ	Ok
35	AR1268ICV500	ICVPQ101025AR1268	PQ071005.D	10 Oct 2025 19:52		YP\AJ	Ok
36	DDT ANALOG	DDT ANALOG	PQ071006.D	10 Oct 2025 20:36		YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ102025

Review By	yogesh	Review On	10/20/2025 1:25:53 PM		
Supervise By	mohammad	Supervise On	10/25/2025 7:26:09 AM		
SubDirectory	PQ102025	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ071007.D	20 Oct 2025 09:06		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PQ071008.D	20 Oct 2025 10:18		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PQ071009.D	20 Oct 2025 10:34		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PQ071010.D	20 Oct 2025 10:52		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PQ071011.D	20 Oct 2025 11:11		YP\AJ	Ok
6	I.BLK	I.BLK	PQ071012.D	20 Oct 2025 12:36		YP\AJ	Ok,M
7	PB170162BL	PB170162BL	PQ071013.D	20 Oct 2025 13:20		YP\AJ	Ok,M
8	PB170162BS	PB170162BS	PQ071014.D	20 Oct 2025 13:35		YP\AJ	Ok
9	Q3374-01	TP-4-WC	PQ071015.D	20 Oct 2025 13:49		YP\AJ	Ok,M
10	Q3374-05	TP-5-WC	PQ071016.D	20 Oct 2025 14:04		YP\AJ	Ok
11	Q3375-01	PR132-S02-074099-20	PQ071017.D	20 Oct 2025 14:19		YP\AJ	Ok
12	Q3375-02	PR132-S04-000013-20	PQ071018.D	20 Oct 2025 14:33		YP\AJ	Ok
13	Q3375-03	PR132-S05-000057-20	PQ071019.D	20 Oct 2025 14:48	AR1242 & AR1254 hits	YP\AJ	Ok,M
14	Q3375-04	PR132-S02-010074-20	PQ071020.D	20 Oct 2025 15:02		YP\AJ	Ok
15	Q3375-05	PR132-DP02-20251016	PQ071021.D	20 Oct 2025 15:47	AR1242 & AR1254 hits	YP\AJ	Ok,M
16	Q3381-01	VNJ-222	PQ071022.D	20 Oct 2025 16:01		YP\AJ	Ok
17	AR1660CCC500	AR1660CCC500	PQ071023.D	20 Oct 2025 16:45		YP\AJ	Ok
18	AR1242CCC500	AR1242CCC500	PQ071024.D	20 Oct 2025 17:00		YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ102025

Review By	yogesh	Review On	10/20/2025 1:25:53 PM		
Supervise By	mohammad	Supervise On	10/25/2025 7:26:09 AM		
SubDirectory	PQ102025	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	AR1248CCC500	AR1248CCC500	PQ071025.D	20 Oct 2025 17:14		YP\AJ	Ok
20	AR1254CCC500	AR1254CCC500	PQ071026.D	20 Oct 2025 17:29		YP\AJ	Ok
21	I.BLK	I.BLK	PQ071027.D	20 Oct 2025 17:44		YP\AJ	Ok

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: rubina
Date: 10/20/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/17/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN #1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 09:25
Out Date: 10/18/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC: LB137576

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3373-05	SVOC-GPC-BLANK	1	1.00	1.00	2.00	2.00	100.0	
Q3373-06	PEST-GPC-BLANK	2	1.00	1.00	2.00	2.00	100.0	
Q3373-07	PEST-GPC-BLANK-SPIKE	3	1.00	1.00	2.00	2.00	100.0	
Q3373-08	SVOC-GPC2-BLANK	4	1.00	1.00	2.00	2.00	100.0	
Q3373-09	PEST-GPC2-BLANK	5	1.00	1.00	2.00	2.00	100.0	
Q3373-10	PEST -GPC2-BLANK-SPIKE	6	1.00	1.00	2.00	2.00	100.0	
Q3374-01	TP-4-WC	7	1.15	10.69	11.84	10.25	85.1	
Q3374-02	TP-4-WC-EPH	8	1.11	10.86	11.97	11.38	94.6	
Q3374-03	TP-4-WC-VOC	9	1.15	10.45	11.6	10.31	87.7	
Q3374-05	TP-5-WC	10	1.13	10.30	11.43	10.72	93.1	
Q3374-06	TP-5-WC-EPH	11	1.17	10.34	11.51	9.95	84.9	
Q3374-07	TP-5-WC-VOC	12	1.18	10.94	12.12	11.19	91.5	
Q3375-01	PR132-S02-074099-20251016	13	1.12	10.21	11.33	8.19	69.2	
Q3375-02	PR132-S04-000013-20251016	14	1.15	10.84	11.99	9.25	74.7	
Q3375-03	PR132-S05-000057-20251016	15	1.16	10.58	11.74	8.22	66.7	
Q3375-04	PR132-S02-010074-20251016	16	1.14	10.34	11.48	6.86	55.3	
Q3375-05	PR132-DP02-074099-20251016	17	1.15	10.84	11.99	8.38	66.7	
Q3376-01	PR132-S05-057082-20251016	18	1.13	10.84	11.97	6.93	53.5	
Q3376-02	PR132-DP01-20251016	19	1.16	10.07	11.23	6.56	53.6	
Q3376-03	PR132-COMP2-20251016	20	1.11	10.13	11.24	6.45	52.7	
Q3376-04	PR132-S12-047072-20251016	21	1.15	10.70	11.85	6.64	51.3	
Q3376-05	PR132-S12-020047-20251016	22	1.13	10.49	11.62	8.14	66.8	
Q3381-01	VNJ-222	23	1.18	10.38	11.56	10.46	89.4	
Q3381-03	VNJ-245	24	1.18	10.64	11.82	9.86	81.6	
Q3381-05	RBR200059	25	1.11	10.71	11.82	10.6	88.6	

PERCENT SOLID

Supervisor: Iwona
Analyst: rubina
Date: 10/20/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/17/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN #1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 09:25
Out Date: 10/18/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC: LB137576

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3381-07	VNJ-255	26	1.16	10.11	11.27	10.51	92.5	
Q3382-01	CENTRAL-AVE-TP	27	1.14	10.73	11.87	10.25	84.9	
Q3382-02	CENTRAL-AVE-TP-EPH-2	28	1.16	10.81	11.97	9.78	79.7	
Q3383-01	OR-03-101725	29	1.13	10.60	11.73	10.88	92.0	
Q3383-02	OR-03-101725-E2	30	1.14	10.91	12.05	11.07	91.0	
Q3386-01	237	38	1.15	10.68	11.83	11.2	94.1	
Q3386-02	237-E2	39	1.15	10.85	12.00	11.2	92.6	
Q3387-01	BC289050-1-1	40	1.00	1.00	2.00	2.00	100.0	PILC
Q3387-02	BC289050-1-2	41	1.00	1.00	2.00	2.00	100.0	PILC
Q3387-03	BC289050-2-1	42	1.00	1.00	2.00	2.00	100.0	PILC
Q3387-04	BC289050-2-2	43	1.00	1.00	2.00	2.00	100.0	PILC
Q3388-01	461	44	1.00	1.00	2.00	2.00	100.0	oily-debris
Q3389-01	KMH472J-1-1	45	1.00	1.00	2.00	2.00	100.0	PILC
Q3389-02	KMH472J-1-2	46	1.00	1.00	2.00	2.00	100.0	PILC
Q3390-01	4175	47	1.00	1.00	2.00	2.00	100.0	debris
Q3390-03	340	48	1.18	10.76	11.94	10.34	85.1	
Q3390-04	340-E2	49	1.16	10.88	12.04	9.11	73.1	
Q3391-01	RCA	31	1.00	1.00	2.00	2.00	100.0	Concreate sample
Q3392-01	MH-20	32	1.18	10.77	11.95	10.9	90.3	
Q3392-02	MH-20-EPH	33	1.16	11.00	12.16	11.07	90.1	
Q3392-03	MH-20-VOC	34	1.18	10.89	12.07	10.69	87.3	
Q3392-05	MH-19	35	1.16	10.50	11.66	10.34	87.4	
Q3392-06	MH-19-EPH	36	1.12	10.87	11.99	10.41	85.5	
Q3392-07	MH-19-VOC	37	1.16	10.50	11.66	10.37	87.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

Q3375-PCB

VR 137576

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-101725 **WorkList ID :** 192523 **Department :** Wet-Chemistry **Date :** 10-17-2025 10:20:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3373-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/10/2025	Chemtech -SO
Q3373-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/10/2025	Chemtech -SO
Q3373-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/10/2025	Chemtech -SO
Q3373-08	SVOC-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/10/2025	Chemtech -SO
Q3373-09	PEST-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/10/2025	Chemtech -SO
Q3373-10	PEST -GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/10/2025	Chemtech -SO
Q3374-01	TP-4-WC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/16/2025	Chemtech -SO
Q3374-02	TP-4-WC-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/16/2025	Chemtech -SO
Q3374-03	TP-4-WC-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/16/2025	Chemtech -SO
Q3374-05	TP-5-WC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/17/2025	Chemtech -SO
Q3374-06	TP-5-WC-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/17/2025	Chemtech -SO
Q3374-07	TP-5-WC-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/17/2025	Chemtech -SO
Q3375-01	PR132-S02-074099-20251016	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/16/2025	Chemtech -SO
Q3375-02	PR132-S04-000013-20251016	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/16/2025	Chemtech -SO
Q3375-03	PR132-S05-000057-20251016	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/16/2025	Chemtech -SO
Q3375-04	PR132-S02-010074-20251016	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/16/2025	Chemtech -SO
Q3375-05	PR132-DP02-074099-20251016	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/16/2025	Chemtech -SO
Q3376-01	PR132-S05-057082-20251016	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/16/2025	Chemtech -SO
Q3376-02	PR132-DP01-20251016	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/16/2025	Chemtech -SO
Q3376-03	PR132-COMP2-20251016	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/16/2025	Chemtech -SO
Q3376-04	PR132-S12-047072-20251016	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/16/2025	Chemtech -SO

631 of 732

Date/Time 10/17/25 16:00

Raw Sample Received by: BB WC)

Raw Sample Relinquished by: CP (CSM)

Date/Time 10/17/25 17:35

Raw Sample Received by: CP (CSM)

Raw Sample Relinquished by: BB WC)

19 18 17 16 15 14 13 12 11 10 9 8 7 6 5 4 3 2 1

Q3375-PCB

WORKLIST(Hardcopy Internal Chain)

VB 137576

WorkList Name : %1-101725 **WorkList ID :** 192523 **Department :** Wet-Chemistry **Date :** 10-17-2025 10:20:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3376-05	PR132-S12-020047-20251016	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/16/2025	Chemtech -SO
Q3381-01	VNJ-222	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3381-03	VNJ-245	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3381-05	RBR200059	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3381-07	VNJ-255	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3382-01	CENTRAL-AVE-TP	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3382-02	CENTRAL-AVE-TP-EPH-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/17/2025	Chemtech -SO
Q3383-01	OR-03-101725	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/17/2025	Chemtech -SO
Q3383-02	OR-03-101725-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	10/17/2025	Chemtech -SO
Q3386-01	237	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	10/17/2025	Chemtech -SO
Q3386-02	237-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/17/2025	Chemtech -SO
Q3387-01	BC289050-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/17/2025	Chemtech -SO
Q3387-02	BC289050-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3387-03	BC289050-2-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3387-04	BC289050-2-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3388-01	461	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3389-01	KMH472J-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3389-02	KMH472J-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/17/2025	Chemtech -SO
Q3390-01	4175	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/17/2025	Chemtech -SO
Q3390-03	340	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/17/2025	Chemtech -SO
Q3390-04	340-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03		10/17/2025	Chemtech -SO

Date/Time 10/17/25 16:00

Raw Sample Received by: JB WC

Raw Sample Relinquished by: CR (SM)

Date/Time 10/17/25 17:35

Raw Sample Received by: CR (SM)

Raw Sample Relinquished by: JB WC

WORKLIST(Hardcopy Internal Chain)

VB 137576

WorkList Name : %1-101725 **WorkList ID :** 192523 **Department :** Wet-Chemistry **Date :** 10-17-2025 10:20:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3391-01	RCA	Solid	Percent Solids	Cool 4 deg C	YANN01	D31	09/18/2025	Chemtech -SO
Q3392-01	MH-20	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3392-02	MH-20-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3392-03	MH-20-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3392-05	MH-19	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3392-06	MH-19-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO
Q3392-07	MH-19-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/17/2025	Chemtech -SO

Date/Time 10/17/25 16:00

Raw Sample Received by: TB WC

Raw Sample Relinquished by: CP CSM

Date/Time 10/17/25 17:35

Raw Sample Received by: CP CSM

Raw Sample Relinquished by: TB WC

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PB170162

SOP ID: M3541-ASE Extraction-15

Clean Up SOP #: Acid Cleanup

Matrix : Solid

Weigh By: EH

Balance check: EH

Balance ID: EX-SC-2

pH Strip Lot#: N/A

Extraction By: EH

Filter By: RS

pH Meter ID: N/A

Hood ID: 3,7

Extraction Start Date : 10/20/2025

Extraction Start Time : 09:00

Extraction End Date : 10/20/2025

Extraction End Time : 12:25

Concentration By: EH

Supervisor By : RUPESH

Extraction Method: [ ] Seperatory Funnel [ ] Continious Liquid/Liquid [ ] Sonication [ ] Waste Dilution [√] Soxhlet

Standared Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24930
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2643
Baked Na2SO4	N/A	EP2652
Hexane	N/A	E3978
H2SO4 1:1	N/A	EP2610
Sand	N/A	E3951
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721. Q3388-01 & Q3390-01 used Limited volume as samples are oily debries.

KD Bath ID: N/A

KD Bath Temperature: N/A

Envap ID: NEVAP-02

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/20/25	RS (Ext lab)	R. Best/PCB Lab
12:30	**Preparation Group**	**Analysis Group**

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PrepBatch ID : PB170162

Analytical Method: M3541-ASE Extraction-15

Concentration Date: 10/20/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170162BL	ABLK162	PCB	30.01	N/A	Evelyn	RUPESH	10			U1-1
PB170162BS	ALCS162	PCB	30.03	N/A	Evelyn	RUPESH	10			2
Q3374-01	TP-4-WC	PCB	30.06	N/A	Evelyn	RUPESH	10	E		3
Q3374-05	TP-5-WC	PCB	30.09	N/A	Evelyn	RUPESH	10	E		4
Q3375-01	PR132-S02-074099-2025 1016	PCB	30.02	N/A	Evelyn	RUPESH	10	E		5
Q3375-02	PR132-S04-000013-2025 1016	PCB	30.05	N/A	Evelyn	RUPESH	10	E		6
Q3375-03	PR132-S05-000057-2025 1016	PCB	30.08	N/A	Evelyn	RUPESH	10	E		U2-1
Q3375-04	PR132-S02-010074-2025 1016	PCB	30.04	N/A	Evelyn	RUPESH	10	E		2
Q3375-05	PR132-DP02-074099-202 51016	PCB	30.07	N/A	Evelyn	RUPESH	10	E		3
Q3381-01	VNJ-222	PCB	30.10	N/A	Evelyn	RUPESH	10	E		4
Q3381-03	VNJ-245	PCB	30.03	N/A	Evelyn	RUPESH	10	E		5
Q3381-05	RBR200059	PCB	30.06	N/A	Evelyn	RUPESH	10	E		6
Q3381-07	VNJ-255	PCB	30.09	N/A	Evelyn	RUPESH	10	E		U3-1
Q3382-01	CENTRAL-AVE-TP	PCB	30.02	N/A	Evelyn	RUPESH	10	E		2
Q3383-01	OR-03-101725	PCB	30.05	N/A	Evelyn	RUPESH	10	E		3
Q3383-01MS	OR-03-101725MS	PCB	30.08	N/A	Evelyn	RUPESH	10	E		4
Q3383-01MSD	OR-03-101725MSD	PCB	30.07	N/A	Evelyn	RUPESH	10	E		5
Q3386-01	237	PCB	30.04	N/A	Evelyn	RUPESH	10	E		6
Q3388-01	461	PCB	5.06	N/A	Evelyn	RUPESH	10	E	Oily Debris	U5-1
Q3390-01	4175	PCB	5.09	N/A	Evelyn	RUPESH	10	E	Oily Debris	2
Q3390-03	340	PCB	30.03	N/A	Evelyn	RUPESH	10	E		3
Q3392-01	MH-20	PCB	30.07	N/A	Evelyn	RUPESH	10	E		4
Q3392-05	MH-19	PCB	30.09	N/A	Evelyn	RUPESH	10	E		5

10/20

* **Extracts relinquished on the same date as received.**

Q3375 PCB (1762 100)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3381 **WorkList ID :** 192552 **Department :** Extraction **Date :** 10-20-2025 08:52:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3374-01	TP-4-WC	Solid	PCB	Cool 4 deg C	PSEG03	D41	10/16/2025	8082A
Q3374-05	TP-5-WC	Solid	PCB	Cool 4 deg C	PSEG03	D41	10/17/2025	8082A
Q3375-01	PR132-S02-074099-20251016	Solid	PCB	Cool 4 deg C	AECO02	D31	10/16/2025	8082A
Q3375-02	PR132-S04-000013-20251016	Solid	PCB	Cool 4 deg C	AECO02	D31	10/16/2025	8082A
Q3375-03	PR132-S05-000057-20251016	Solid	PCB	Cool 4 deg C	AECO02	D31	10/16/2025	8082A
Q3375-04	PR132-S02-010074-20251016	Solid	PCB	Cool 4 deg C	AECO02	D31	10/16/2025	8082A
Q3375-05	PR132-DP02-074099-2025101€	Solid	PCB	Cool 4 deg C	AECO02	D31	10/16/2025	8082A
Q3381-01	VNJ-222	Solid	PCB	Cool 4 deg C	PSEG03	D31	10/17/2025	8082A
Q3381-03	VNJ-245	Solid	PCB	Cool 4 deg C	PSEG03	D31	10/17/2025	8082A
Q3381-05	RBR200059	Solid	PCB	Cool 4 deg C	PSEG03	D31	10/17/2025	8082A
Q3381-07	VNJ-255	Solid	PCB	Cool 4 deg C	PSEG03	D31	10/17/2025	8082A
Q3382-01	CENTRAL-AVE-TP	Solid	PCB	Cool 4 deg C	PSEG03	D41	10/17/2025	8082A
Q3383-01	OR-03-101725	Solid	PCB	Cool 4 deg C	PSEG05	D41	10/17/2025	8082A
Q3386-01	237	Solid	PCB	Cool 4 deg C	PSEG03	D41	10/17/2025	8082A
Q3388-01	461	Solid	PCB	Cool 4 deg C	PSEG03	D31	10/17/2025	8082A
Q3390-01	4175	Solid	PCB	Cool 4 deg C	PSEG03	D41	10/17/2025	8082A
Q3390-03	340	Solid	PCB	Cool 4 deg C	PSEG03		10/17/2025	8082A
Q3392-01	MH-20	Solid	PCB	Cool 4 deg C	PSEG03	D31	10/17/2025	8082A
Q3392-05	MH-19	Solid	PCB	Cool 4 deg C	PSEG03	D31	10/17/2025	8082A

Date/Time 10/20/25 8:55
Raw Sample Received by: EH (Ext-196)
Raw Sample Relinquished by:

Date/Time 10/20/25 9:30
Raw Sample Received by:
Raw Sample Relinquished by: EH (Ext-196)

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID : Q3375

Test : PCB

Prepbatch ID : PB170162,

Sequence ID/Qc Batch ID: pp102025,PQ102025,

Standard ID :
EP2610,EP2643,EP2652,PP24329,PP24330,PP24381,PP24663,PP24805,PP24806,PP24807,PP24808,PP24809,PP24810,PP24811,PP24812,PP24814,PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862,PP24930,

Chemical ID :
E3804,E3875,E3944,E3951,E3962,E3963,E3971,E3972,E3978,M6157,P11522,P11592,P12700,P12703,P12952,P12953,P12957,P13356,P13373,P13375,P13590,P13592,P13695,P13697,P13706,P13786,P13789,P13833,P13879,P13881,P13885,W3112,W3177,W3234,

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025

FROM 1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
230	1:1ACETONE/HEXANE	EP2643	09/16/2025	03/16/2026	Evelyn Huang	None	None	Riteshkumar Patel 09/16/2025

FROM 4000.00000ml of E3971 + 4000.00000ml of E3972 = Final Quantity: 8000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3923	Baked Sodium Sulfate	EP2652	10/10/2025	01/28/2026	Evelyn Huang	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 10/10/2025

FROM 4000.00000gram of E3875 = Final Quantity: 4000.000 gram

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
465	200 PPB Pest/PCB Surrogate Spike	PP24663	06/24/2025	12/24/2025	Abdul Mirza	None	None	Yogesh Patel 07/21/2025

FROM 1.00000ml of P13786 + 999.00000ml of E3944 = Final Quantity: 1000.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
84	Pest/PCB Surrogate Stock 20 PPM	PP24805	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13789 + 9.00000ml of E3962 = Final Quantity: 10.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
213	AR1221 1000 PPB WORKING SOLUTION	PP24806	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P13375 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
214	AR1232 1000 PPB WORKING SOLUTION	PP24807	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P13879 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
215	AR1242 1000 PPB WORKING STD	PP24808	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P13706 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
216	AR1248 1000 PPB WORKING STD	PP24809	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P13695 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
217	AR1254 1000 PPB WORKING STD	PP24810	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P11592 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1529	AR1262 1000 PPB Working Solution	PP24811	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P13881 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1532	AR1268 1000 PPB Working Solution	PP24812	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P13885 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
404	AR1660 100 PPM Stock Solution 2nd Source	PP24814	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P12952 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
405	AR1660 1000/100 PPB ICV STD	PP24815	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 98.50000ml of E3962 + 0.50000ml of PP24805 + 1.00000ml of PP24814 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24816	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13373 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24817	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P12700 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24818	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13590 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24819	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13592 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
2193	AR1254 100 PPM STOCK SOLUTION	PP24820	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13833 + 9.00000ml of E3963 = Final Quantity: 10.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3757	AR1262 1000 PPB Working Solution second source	PP24821	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P12703 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24822	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P11522 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
203	AR1660 750 PPB STD	PP24823	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
204	AR1660 500 PPB STD	PP24824	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
205	AR1660 250 PPB STD	PP24825	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
206	AR1660 50 PPB STD	PP24826	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24824 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1079	AR1221 750 PPB STD	PP24827	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24806 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
222	AR1221 500 PPB STD	PP24828	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24806 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1080	AR1221 250 PPB STD	PP24829	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24806 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1081	AR1221 50 PPB STD	PP24830	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24828 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1063	AR1232 750 PPB STD	PP24831	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24807 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
223	AR1232 500 PPB STD	PP24832	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24807 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1064	AR1232 250 PPB STD	PP24833	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24807 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1065	AR1232 50 PPB STD	PP24834	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24832 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1067	AR1242 750 PPB STD	PP24835	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24808 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
224	AR1242 500 PPB STD	PP24836	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24808 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1068	AR1242 250 PPB STD	PP24837	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24808 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1069	AR1242 50 PPB STD	PP24838	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24836 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1075	AR1248 750 PPB STD	PP24839	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24809 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
225	AR1248 500 PPB STD	PP24840	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24809 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1076	AR1248 250 PPB STD	PP24841	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24809 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1077	AR1248 50 PPB STD	PP24842	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24840 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1071	AR1254 750 PPB STD	PP24843	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24810 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
226	AR1254 500 PPB STD	PP24844	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24810 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1072	AR1254 250 PPB STD	PP24845	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24810 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1073	AR1254 50 PPB STD	PP24846	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24844 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3753	AR1262 750 PPB STD	PP24847	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24811 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1530	AR1262 500 PPB STD	PP24848	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24811 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3754	AR1262 250 PPB STD	PP24849	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24811 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3755	AR1262 50 PPB STD	PP24850	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24848 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3820	AR1268 750 PPB STD	PP24851	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24812 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1533	AR1268 500 PPB STD	PP24852	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24812 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3821	AR1268 250 PPB STD	PP24853	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24812 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3822	AR1268 50 PPB STD	PP24854	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24852 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
406	AR1660 500 PPB ICV	PP24855	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24815 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3790	AR1221 500 PPB ICV(AGILENT)	PP24856	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24816 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1888	AR1232 500 PPB ICV	PP24857	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24817 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1891	AR1242 500 PPB ICV	PP24858	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24818 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1892	AR1248 500 PPB ICV	PP24859	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24819 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1894	AR1254 500 PPB ICV	PP24860	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3758	AR1262 500 PPB STD ICV	PP24861	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24821 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3823	AR1268 500 PPB STD ICV	PP24862	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24822 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24930	09/18/2025	03/18/2026	Abdul Mirza	None	None	Yogesh Patel 09/19/2025

FROM 0.50000ml of P12953 + 99.50000ml of E3972 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	02/18/2026	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	01/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	06/20/2025 / RUPESH	05/14/2025 / RUPESH	E3944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	25A2756718	12/31/2028	07/09/2025 / RUPESH	04/28/2020 / RUPESH	E3951

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	08/05/2025 / RUPESH	07/30/2025 / RUPESH	E3962

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	02/15/2026	08/15/2025 / RUPESH	08/07/2025 / RUPESH	E3963

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	09/15/2025 / Evelyn	09/04/2025 / Riteshkumar	E3971

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	09/16/2025 / Evelyn	09/04/2025 / Riteshkumar	E3972

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362006	04/13/2026	10/13/2025 / RUPESH	10/08/2025 / Riteshkumar	E3978

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0175403	02/18/2026	08/18/2025 / yogesh	03/18/2022 / Abdul	P11592

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards,Inc	91867 / Aroclor 1232 100 ug/mL	020823	02/18/2026	08/18/2025 / yogesh	08/07/2023 / Ankita	P12700

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards,Inc	x9166 / Aroclor 1262 100 ug/mL	060523	02/18/2026	08/18/2025 / yogesh	08/07/2023 / Ankita	P12703

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	02/18/2026	08/18/2025 / yogesh	12/20/2023 / Yogesh	P12952

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	03/18/2026	09/18/2025 / Abdul	12/20/2023 / Yogesh	P12953

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Arochlor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0207511	02/18/2026	08/18/2025 / yogesh	05/03/2024 / Abdul	P13375

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	02/18/2026	08/18/2025 / yogesh	10/14/2024 / Ankita	P13590

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	02/18/2026	08/18/2025 / yogesh	10/14/2024 / Ankita	P13592

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	A0214733	02/18/2026	08/18/2025 / yogesh	10/17/2024 / yogesh	P13695

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	A0210232	02/18/2026	08/18/2025 / yogesh	10/17/2024 / yogesh	P13706

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	12/24/2025	06/24/2025 / Abdul	11/19/2024 / Ankita	P13786

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	02/18/2026	08/18/2025 / yogesh	11/19/2024 / Ankita	P13789

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	a0217391	02/18/2026	08/18/2025 / yogesh	12/09/2024 / Ankita	P13833

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13879

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0217041	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13881

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0217264	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13885

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	07/28/2025 / jignesh	07/25/2025 / jignesh	W3234

Acetone
CMOS

avantor™ | J.T.Baker®

Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.1
Titrable Base (µeq/g)	≤ 0.5	0.1
Water (H_2O)	≤ 0.5 %	0.1 %
Solubility in H_2O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO_4)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities - Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities - Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities - Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities - Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities - Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities - Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities - Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities - Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities - Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

>>> Continued on page 2 >>>

Recd by RP on 9/25/24

E 3804

Acetone
CMOS

avantor™ | J.T.Baker®

Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

Acetone
CMOS

avantor™ | J.T.Baker®

Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC

Michelle Bales
Sr. Manager, Quality Assurance

PRODUCTOS QUIMICOS MONTERREY, S.A. DE C.V.

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE **RMB3375**)	FORMULA :	Na_2SO_4
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na_2SO_4)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO_4)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreing matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date:2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3944

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Certificate of Analysis

Material	BDH9274-2.5KG
Material Description	BDH SAND STDD OTTAWA W+I 2.5KG
Grade	NOT APPLICABLE
Batch	25A2756718
Reassay Date	12/31/2028
CAS Number	14808-60-7
Molecular Formula	SiO2
Molecular Mass	60.09
Date of Manufacture	12/05/2024
Storage	Room Temperature

Characteristics	Specifications	Measured Values
Appearance	Beige granules.	Beige granules.
Moisture	<= 0.1 %	0.1 %
Particle Size 30-40 mesh	>= 80 %	99 %
CUSTOMER PART # BDH9274-2.5KG		

Received on 7/9/25.

E3951

Internal ID #: 793

Signature	Additional Information
We certify that this batch conforms to the specifications listed above.	Analysis may have been rounded to significant digits in specification limits
This document has been electronically produced and is valid without a signature.	Product meets analytical specifications of the grades listed.
Leona Edwardson, Quality Control Sr. Manager - Solon VWR Chemicals, LLC. 28600 Fountain Parkway, Solon OH 44139 USA	

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date:2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	5
Assay (Total Saturated C_6 Isomers) (byGC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, correctedfor water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.1 ppm
Substances Darkened by H_2SO_4	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Recieved on 7/30/25

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3962

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

PO: PO2-2575 PRODUCT CODE: SHIP DATE: 6/20/25

Acetone,
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date:2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Recieved on 8/6/25

E 3963

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

Page 1 of 1

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date:2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	5
Assay (Total Saturated C_6 Isomers) (byGC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, correctedfor water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.1 ppm
Substances Darkened by H_2SO_4	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3971

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date:2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3972

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9262-03
Batch No.: 25C0362006
Manufactured Date: 2025-01-29
Expiration Date:2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	4
Assay (Total Saturated C_6 Isomers) (byGC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, correctedfor water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Darkened by H_2SO_4	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3978

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

M615 MB

avantor™ | J.T.Baker

Material No.: 9673-33
Batch No.: 24I1262013
Manufactured Date: 2024-08-07
Retest Date:2029-08-06
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H_2SO_4)	95.0 – 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	<1 ppm
ACS – Substances Reducing Permanganate(as SO2)	<= 2 ppm	<2 ppm
Ammonium (NH_4)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO_3)	<= 0.2 ppm	0.1 ppm
Phosphate (PO_4)	<= 0.5 ppm	<0.1 ppm
Trace Impurities – Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities – Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities – Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities – Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities – Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities – Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities – Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities – Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities – Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities – Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities – Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities – Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities – Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities – Silver (Ag)	<= 1.0 ppb	<1.0 ppb

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™ | J.T.Baker

Material No.: 9673-33
Batch No.: 24I1262013

Test	Specification	Result
Trace Impurities - Sodium (Na)	<= 500.0 ppb	<5.0 ppb
Trace Impurities - Strontium (Sr)	<= 5.0 ppb	<1.0 ppb
Trace Impurities - Tin (Sn)	<= 5.0 ppb	1.1 ppb
Trace Impurities - Zinc (Zn)	<= 5.0 ppb	<1.0 ppb

For Laboratory,Research,or Manufacturing Use

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

J Croak

Jamie Croak
Director Quality Operations, Bioscience Production

P11518 → P11522 AJ 02/21/22

Agilent

Certificate of Analysis

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1 **Lot Issue Date:** 09-Feb-2021

Lot Number: 0006587800 **Expiration Date:** 31-Mar-2029

Description:
This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:
The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:
This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:
This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:
Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:
Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:
The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:
If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

M Bourgeois

Monica Bourgeois
QMS Representative

ANAB ACCREDITED REFERENCE MATERIAL PRODUCER
ISO 17034 Cert No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026
Page: 1 of 1
www.agilent.com/quality/
CSD-QA-015.1

ANAB ACCREDITED TESTING LABORATORY
ISO 17025 Cert No. AT-1937

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0175403

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Aroclor 1254	1,000.7 µg/mL	+/- 5.9437 µg/mL	Gravimetric
	CAS # 11097-69-1 (Lot 124-191-B)		+/- 31.7284 µg/mL	Unstressed
	Purity ----%		+/- 41.4406 µg/mL	Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P11588 ↓ P11592 | 5

AR 04/30/2022

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Cathleen Soltis - Mix Technician

Date Mixed: 15-Aug-2021 **Balance:** 1128360905

Alexis Shelow - Operations Tech I

Date Passed: 17-Aug-2021

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

P11588 | (5)
1
P11592

AR
04/30/22

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Refe

CERTIFIED WEIGHT REPORT

Part Number:	91867
Lot Number:	020823
Description:	WP 037 - Aroclor 1232 PCB Technical Mixture
Expiration Date:	020833
Recommended Storage:	Ambient (20 °C)
Nominal Concentration (µg/mL):	100
NIST Test ID#:	6UTB

Solvent/ Acet

Weight(s) shown below were combined and diluted to (mL): 100.0 — 5E-05 Balance Uncertainty; 0.057 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Targ Weight (g
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection =1.5µL, Range=3



Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Referen

CERTIFIED WEIGHT REPORT

Part Number:	X9166	**Solvent(s):**	
Lot Number:	060523	Methanol	E
Description:	Aroclor 1262 Appendix IX Compound		
Expiration Date:	060533		
Recommended Storage:	Ambient (20 °C)		
Nominal Concentration (µg/mL):	100		
NIST Test ID#:	6UTB	5E-05	Balance Uncertainty
Volume(s) shown below were combined and diluted to (mL):	20.0	0.002	Flask Uncertainty

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)
1. Aroclor 1262	70444	080322	0.10	2.00	0.017

Method GC7MSD-7.M: Column:(30m X 0.25mm ID X 0.25µm film thickn
min.), Rate = 8°C/min., Injector B= 200°C, Detector B = 290°C.



- The certified value is the concentration calculated from gravimetric and volumetric me
- Standards are prepared gravimetrically using balances that are calibrated with weight
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appro
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number:	20064
Lot Number:	022023
Description:	CLP PCB'S - Aroclor Mix
	Aroclors 1016 & 1260
Expiration Date:	022033
Recommended Storage:	Ambient (20 °C)
Nominal Concentration (µg/mL):	1000
NIST Test ID#:	6UTB

Solvent(s): Hexane **Lot#** 273615

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 200.0 0.010 Flask Uncertainty

Formulated By: Benson Chan	022023 DATE
Reviewed By: Pedro L. Rentas	022023 DATE

P12946 ↓ P12955 12/20/23

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11096-82-5	0.5mg/m3	orl-rat 1315mg/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection =1.5µL, Range=3



Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number:	20064
Lot Number:	022023
Description:	CLP PCB'S - Aroclor Mix Aroclors 1016 & 1260
Expiration Date:	022033
Recommended Storage:	Ambient (20 °C)
Nominal Concentration (µg/mL):	1000
NIST Test ID#:	6UTB

Solvent(s):	Lot#
Hexane	273615

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 200.0 0.010 Flask Uncertainty

Formulated By:	Benson Chan	022023 DATE
Reviewed By:	Pedro L. Rentas	022023 DATE

P12946 → P12955 12/20

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11096-82-5	0.5mg/m3	orl-rat 1315mg/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection =1.5µL, Range=3



Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 99139
Lot Number: 121823
Description: Aroclor 1254

Solvent(s): Iso-octane **Lot#** 82227

Expiration Date: 121833
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 100
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.003 Flask Uncertainty

Volume(s) shown below were combined and diluted to (mL): 20.0

Formulated By:	Anthony Mahoney	121823 DATE
Reviewed By:	Pedro L. Rentas	121823 DATE

P12956 & P12957 Y-P 12/20/23

Note: Aroclor 1254 is a mix of isomers.

	Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Initial Conc.(µg/mL)	Final Conc.(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1.	Aroclor 1254	79100	121823	0.10	2.00	0.017	1003.3	100.1	1.8	11097-69-1	0.5mg/m3 (skin)	orl-rat 1295mg/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection =1.5µL, Range=3

Peak No.	Name	FID RT (min.)
1	Total Aroclor 1254	18.12



RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

P13348 (10)
↓
P13357

04/25/2024

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
hellum-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

P13348 (10)

↓

P13357

04/25/2025

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0207511

Description : Aroclor® 1221 Standard

Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	13192500	----%	1,002.7 µg/mL	+/- 55.6424

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13374 (2)
P 13375 1

05/6/2024

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente - Operations Technician I **Date Mixed:** 06-Feb-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC **Date Passed:** 08-Feb-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

P133774 (2)
↓
P133775

05/6/2024

ISO 17034

Agilent
Trusted Answers

Reference Material Certificate

Product Name:	Aroclor 1242 Standard	**Lot Number:**	0006665550
Product Number:	PP-312-1	**Lot Issue Date:**	08-Feb-2022
Storage Conditions:	Store at Room Temperature (15° to 30°C).	**Expiration Date:**	31-Jan-2027

Component Name	CERTIFIED VALUES Concentration	Expanded Uncertainty	CAS#	Analyte Lot
Aroclor 1242	100.4 ±	0.5 µg/mL	053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:
This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:
The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:
This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:
Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:
Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:
This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:
The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589 ↓ P13590

AJ 10/14/24

CSD-QA-015.1

Maintenance of Certification:
If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative

Agilent
Trusted Answers

ISO 17034

Reference Material Certificate

Product Information Sheet

Product Name: Aroclor 1248 Standard
Product Number: PP-342-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006726317
Lot Issue Date: 27-Jan-2023
Expiration Date: 28-Feb-2031

Component Name	Concentration		Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3	±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:
This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:
The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:
This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:
Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:
Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:
This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:
The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AJ
10/14/2024

Agilent

Trusted Answers

Maintenance of Certification:
If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative

ANAB
ANSI National Accreditation Board
ACCREDITED
REFERENCE MATERIAL PRODUCER

ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015 Quality Management System. Cert# 951215321

www.agilent.com/quality/
CSD-QA-015.1

ISO 17025

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32010 **Lot No.:** A0214733

Description : Aroclor® 1248 Standard

Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul

P13695
P13696
Y.P.
10/17/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	15242800	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Rebecca Gingerich - Operations Tech II

Date Mixed: 02-Aug-2024 **Balance Serial #** C322230531

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Aug-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix

Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697 → P13701 } Y.P. 10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024 **Balance Serial #** B442140311

Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32009 **Lot No.:** A0210232

Description : Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

P13706
P13707
Y.P. 10/17/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	----%	1,000.7 µg/mL	+/- 55.5295

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Amanda Miller - Operations Tech III - ARM QC

Date Mixed: 11-Apr-2024 **Balance Serial #** B442140311

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 17-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785 ↓ P19789 AJ 11/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785 ↓ P19789 AJ 11/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832
↓
P13834

AJ
12/09/24

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye - Operations Tech I **Date Mixed:** 02-Oct-2024 **Balance Serial #** C322230531

Jennifer Pollino - Operations Tech III - ARM QC **Date Passed:** 07-Oct-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008 **Lot No.:** A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	----%	1,007.0 µg/mL	+/- 55.8810

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13878
↓
P13880

AJ
01/28/25

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024 **Balance Serial #** C322230531

Brittany Federinko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32409 **Lot No.:** A0217041

Description : Aroclor® 1262 Standard

Aroclor® 1262 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	----%	1,000.0 µg/mL	+/- 55.4925

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13881

AJ
01/28/25

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Rebecca Gingerich - Operations Tech II

Date Mixed: 25-Sep-2024 **Balance Serial #** C322230531

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 30-Sep-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ilac-MRA
A2LA ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA
A2LA ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410 **Lot No.:** A0217264

Description : Aroclor® 1268 Standard

Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	12353400	----%	1,009.8 µg/mL	+/- 56.0364

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13884
↓
P13886

AJ
01/28/25

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 500:1

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 01-Oct-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 02-Oct-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™ | J.T.Baker®

W3147
W3147 update. 02/03/2025 JP

Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C_6 Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H_2SO_4	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

JCroak

Jamie Croak
Director Quality Operations, Bioscience Production

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™ | J.T.Baker

W3234 JB
Update: 07/28/2025

Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date:2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	5
Assay (Total Saturated C_6 Isomers) (byGC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, correctedfor water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.1 ppm
Substances Darkened by H_2SO_4	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

SHIPPING DOCUMENTS

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

ALLIANCE PROJECT NO.
QUOTE NO. Q3377
COC Number 2045166Q3375

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: AECOM
ADDRESS: 605 3rd Avenue,
CITY: New York STATE: NY ZIP: 10158
ATTENTION: Rob Forstner
PHONE: 212-377-8721 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Port Richmond 132
PROJECT NO.: 60693795 LOCATION: Staten Island
PROJECT MANAGER: Rob Forstner
e-mail: robert.forstner@aecom.com
PHONE: 212-377-8721 FAX:

CLIENT BILLING INFORMATION

BILL TO: AECOM PO#:
ADDRESS: 605 3rd Ave
CITY New York STATE: NY ZIP: 10158
ATTENTION: Rob Forstner PHONE: 212-377-8721

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

- ☐ Level 1 (Results Only)
- ☐ Level 2 (Results + QC)
- ☐ Level 3 (Results + QC + Raw Data)
- ☐ Level 4 (QC + Full Raw Data)
- ☐ NJ Reduced
- ☐ NYS ASP A
- ☒ NYS ASP B
- ☐ Other _____
- ☐ EDD FORMAT _____

ANALYSIS

1: SVOCs 2: Pesticides-TCL 3: Mercury-Metals 4: PCBs/Hexachrome 5: Dioxin 6: Grain Size 7: TOC 8: VOCs 9:

PRESERVATIVES

Method

COMMENTS

← Specify Preservatives
A-HCl D-NaOH
B-HN03 E-ICE
C-H2SO4 F-OTHER

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP	SAMPLE TYPE GRAB	SAMPLE COLLECTION DATE	SAMPLE COLLECTION TIME	# OF BOTTLES	1	2	3	4	5	6	7	8	9
1.	PR132-SO2-074099-20251016	soil		X	10/16	840	11	X	X	X	X	X	X	X	X	
2.	PR132-SO4-000013-20251016	soil		X	10/16	945	11	X	X	X	X	X	X	X	X	
3.	~~PR132-SO4~~	~~soil~~		~~X~~	~~10/16~~											
4.	PR132-SO5-000057-20251016	soil		X	10/16	1110	11	X	X	X	X	X	X	X	X	
5.	PR132-SO2-010074-20251016	soil		X	10/16	850	11	X	X	X	X	X	X	X	X	
6.	PR132-DP02-20251016	soil		X	10/16	1113	11	X	X	X	X	X	X	X	X	
7.	Temp Blank	water													X	
8.																
9.																
10.																

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:
1. Jim Aldersen	10/16/25 1650	1. 10-16-25 1650
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:
2.		2.
RELINQUISHED BY SAMPLER:	DATE/TIME: 1805 10-16-25	RECEIVED BY: 3.

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 3.2 °C
Comments:

Page _____ of

CLIENT: ☐ Hand Delivered ☐ Other _____

Shipment Complete ☐ YES ☐ NO

Copyright © 2024

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT YELLOW - ALLIANCE COPY PINK - SAMPLER COPY

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3375 AECO02
Client Name : AECOM
Client Contact : Rob Forstner
Invoice Name : AECOM
Invoice Contact : Rob Forstner

Order Date : 10/17/2025 10:32:00 AM
Project Name : AE1-CTY 3.2.Z-PR-DOCK
Receive DateTime : 10/16/2025 6:15:00 PM
Purchase Order :

Project Mgr :
Report Type : NYS ASP B
EDD Type : EQUIS
Hard Copy Date :
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3375-01	PR132-S02-074099-20251016	Solid	10/16/2025	08:40					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q3375-02	PR132-S04-000013-20251016	Solid	10/16/2025	09:45					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q3375-03	PR132-S05-000057-20251016	Solid	10/16/2025	10:10					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q3375-04	PR132-S02-010074-20251016	Solid	10/16/2025	08:50					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q3375-05	PR132-DP02-~~074099~~-20251016	Solid	10/16/2025	11:13					
					VOC-TCLVOA-10		8260D		10 Bus. Days

Relinquished By : CL
Date / Time : 10/17/25 13:35

Received By : Sam
Date / Time : 10/17/25 15:35

RJ# 6
F+2

Storage Area : VOA Refridgerator Room