

Matrix Spike Summary

Client:	AECOM	SDG No.:	Q3375
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-171	Sample ID:	Q3375-05
Client ID:	PR132-DP02-20251016MS	Percent Solids for Spike Sample:	66.7

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/Kg	75-125	31100	OR	30000	OR	1000	1	110		10/20/2025

Matrix Spike Summary

Client:	AECOM	SDG No.:	Q3375
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-171	Sample ID:	Q3375-05
Client ID:	PR132-DP02-20251016MSD	Percent Solids for Spike Sample:	66.7

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/Kg	75-125	31200	OR	30000	OR	1000	1	120		10/20/2025

Matrix Spike Summary

Client:	AECOM	SDG No.:	Q3375
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-171	Sample ID:	Q3403-01
Client ID:	MH-18MS	Percent Solids for Spike Sample:	91.1

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	75-125	1360		0.076	U	1410	40	96		10/21/2025

Matrix Spike Summary

Client:	AECOM	SDG No.:	Q3375
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-171	Sample ID:	Q3403-01
Client ID:	MH-18MS	Percent Solids for Spike Sample:	91.1

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	85-115	40.4		0.076	U	43.9	2	92		10/21/2025

Matrix Spike Summary

Client:	AECOM	SDG No.:	Q3375
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-171	Sample ID:	Q3403-01
Client ID:	MH-18MS	Percent Solids for Spike Sample:	91.1

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	75-125	38.1		0.076	U	43.9	2	87		10/21/2025