



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3375

Contract: AECO02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB79	Mercury	0.076	+/-0.2	U	0.20	CV	10/20/2025	14:23	LB137584

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3375

Contract: AECO02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB57	Mercury	0.076	+/-0.2	U	0.20	CV	10/20/2025	14:28	LB137584
CCB58	Mercury	0.076	+/-0.2	U	0.20	CV	10/20/2025	15:08	LB137584
CCB59	Mercury	0.076	+/-0.2	U	0.20	CV	10/20/2025	15:43	LB137584
CCB60	Mercury	0.076	+/-0.2	U	0.20	CV	10/20/2025	16:20	LB137584
CCB61	Mercury	0.076	+/-0.2	U	0.20	CV	10/21/2025	09:00	LB137584

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3375

Contract: AECO02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB83	Mercury	0.076	+/-0.2	U	0.20	CV	10/22/2025	11:31	LB137610

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3375

Contract: AECO02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB71	Mercury	0.076	+/-0.2	U	0.20	CV	10/22/2025	11:36	LB137610
CCB72	Mercury	0.076	+/-0.2	U	0.20	CV	10/22/2025	12:16	LB137610
CCB73	Mercury	0.076	+/-0.2	U	0.20	CV	10/22/2025	12:54	LB137610
CCB74	Mercury	0.076	+/-0.2	U	0.20	CV	10/22/2025	13:26	LB137610
CCB75	Mercury	0.076	+/-0.2	U	0.20	CV	10/22/2025	14:13	LB137610
CCB76	Mercury	0.076	+/-0.2	U	0.20	CV	10/22/2025	14:44	LB137610
CCB77	Mercury	0.076	+/-0.2	U	0.20	CV	10/22/2025	15:02	LB137610

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3375

Contract: AECO02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	1.94	+/-10	U	20.0	P	10/22/2025	13:16	LB137619
	Antimony	0.11	+/-1	U	2.00	P	10/22/2025	13:16	LB137619
	Arsenic	0.089	+/-0.5	U	1.00	P	10/22/2025	13:16	LB137619
	Barium	0.21	+/-5	U	10.0	P	10/22/2025	13:16	LB137619
	Beryllium	0.32	+/-0.5	U	1.00	P	10/22/2025	13:16	LB137619
	Cadmium	0.34	+/-0.5	U	1.00	P	10/22/2025	13:16	LB137619
	Calcium	45.7	+/-250	U	500	P	10/22/2025	13:16	LB137619
	Chromium	0.21	+/-1	U	2.00	P	10/22/2025	13:16	LB137619
	Cobalt	0.070	+/-0.5	U	1.00	P	10/22/2025	13:16	LB137619
	Copper	0.30	+/-1	U	2.00	P	10/22/2025	13:16	LB137619
	Iron	7.81	+/-25	U	50.0	P	10/22/2025	13:16	LB137619
	Lead	0.21	+/-0.5	U	1.00	P	10/22/2025	13:16	LB137619
	Magnesium	19.5	+/-250	U	500	P	10/22/2025	13:16	LB137619
	Manganese	0.43	+/-0.5	U	1.00	P	10/22/2025	13:16	LB137619
	Nickel	0.27	+/-0.5	U	1.00	P	10/22/2025	13:16	LB137619
	Potassium	36.4	+/-250	U	500	P	10/22/2025	13:16	LB137619
	Selenium	2.90	+/-2.5	U	5.00	P	10/22/2025	13:16	LB137619
	Silver	0.060	+/-0.5	U	1.00	P	10/22/2025	13:16	LB137619
	Sodium	128	+/-250	U	500	P	10/22/2025	13:16	LB137619
	Thallium	0.060	+/-0.5	U	1.00	P	10/22/2025	13:16	LB137619
	Vanadium	0.077	+/-2.5	U	5.00	P	10/22/2025	13:16	LB137619
	Zinc	1.25	+/-2.5	U	5.00	P	10/22/2025	13:16	LB137619

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3375

Contract: AECO02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	1.94	+/-10	U	20.0	P	10/22/2025	13:30	LB137619
	Antimony	0.17	+/-1	J	2.00	P	10/22/2025	13:30	LB137619
	Arsenic	0.089	+/-0.5	U	1.00	P	10/22/2025	13:30	LB137619
	Barium	0.21	+/-5	U	10.0	P	10/22/2025	13:30	LB137619
	Beryllium	0.32	+/-0.5	U	1.00	P	10/22/2025	13:30	LB137619
	Cadmium	0.34	+/-0.5	U	1.00	P	10/22/2025	13:30	LB137619
	Calcium	45.7	+/-250	U	500	P	10/22/2025	13:30	LB137619
	Chromium	0.21	+/-1	U	2.00	P	10/22/2025	13:30	LB137619
	Cobalt	0.070	+/-0.5	U	1.00	P	10/22/2025	13:30	LB137619
	Copper	0.30	+/-1	U	2.00	P	10/22/2025	13:30	LB137619
	Iron	7.81	+/-25	U	50.0	P	10/22/2025	13:30	LB137619
	Lead	0.21	+/-0.5	U	1.00	P	10/22/2025	13:30	LB137619
	Magnesium	19.5	+/-250	U	500	P	10/22/2025	13:30	LB137619
	Manganese	0.43	+/-0.5	U	1.00	P	10/22/2025	13:30	LB137619
	Nickel	0.27	+/-0.5	U	1.00	P	10/22/2025	13:30	LB137619
	Potassium	36.4	+/-250	U	500	P	10/22/2025	13:30	LB137619
	Selenium	2.90	+/-2.5	U	5.00	P	10/22/2025	13:30	LB137619
	Silver	0.060	+/-0.5	U	1.00	P	10/22/2025	13:30	LB137619
	Sodium	128	+/-250	U	500	P	10/22/2025	13:30	LB137619
	Thallium	0.080	+/-0.5	J	1.00	P	10/22/2025	13:30	LB137619
	Vanadium	0.077	+/-2.5	U	5.00	P	10/22/2025	13:30	LB137619
	Zinc	1.25	+/-2.5	U	5.00	P	10/22/2025	13:30	LB137619
CCB02	Aluminum	1.94	+/-10	U	20.0	P	10/22/2025	14:18	LB137619
	Antimony	0.11	+/-1	U	2.00	P	10/22/2025	14:18	LB137619
	Arsenic	0.089	+/-0.5	U	1.00	P	10/22/2025	14:18	LB137619
	Barium	0.21	+/-5	U	10.0	P	10/22/2025	14:18	LB137619
	Beryllium	0.32	+/-0.5	U	1.00	P	10/22/2025	14:18	LB137619
	Cadmium	0.34	+/-0.5	U	1.00	P	10/22/2025	14:18	LB137619
	Calcium	45.7	+/-250	U	500	P	10/22/2025	14:18	LB137619
	Chromium	0.21	+/-1	U	2.00	P	10/22/2025	14:18	LB137619
	Cobalt	0.070	+/-0.5	U	1.00	P	10/22/2025	14:18	LB137619
	Copper	0.30	+/-1	U	2.00	P	10/22/2025	14:18	LB137619
	Iron	7.81	+/-25	U	50.0	P	10/22/2025	14:18	LB137619
	Lead	0.21	+/-0.5	U	1.00	P	10/22/2025	14:18	LB137619
	Magnesium	19.5	+/-250	U	500	P	10/22/2025	14:18	LB137619
	Manganese	0.43	+/-0.5	U	1.00	P	10/22/2025	14:18	LB137619
	Nickel	0.27	+/-0.5	U	1.00	P	10/22/2025	14:18	LB137619
	Potassium	36.4	+/-250	U	500	P	10/22/2025	14:18	LB137619
	Selenium	2.90	+/-2.5	U	5.00	P	10/22/2025	14:18	LB137619

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3375

Contract: AECO02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	0.060	+/-0.5	U	1.00	P	10/22/2025	14:18	LB137619
	Sodium	128	+/-250	U	500	P	10/22/2025	14:18	LB137619
	Thallium	0.060	+/-0.5	U	1.00	P	10/22/2025	14:18	LB137619
	Vanadium	0.077	+/-2.5	U	5.00	P	10/22/2025	14:18	LB137619
	Zinc	1.25	+/-2.5	U	5.00	P	10/22/2025	14:18	LB137619
CCB03	Aluminum	1.94	+/-10	U	20.0	P	10/22/2025	15:05	LB137619
	Antimony	0.11	+/-1	U	2.00	P	10/22/2025	15:05	LB137619
	Arsenic	0.089	+/-0.5	U	1.00	P	10/22/2025	15:05	LB137619
	Barium	0.21	+/-5	U	10.0	P	10/22/2025	15:05	LB137619
	Beryllium	0.32	+/-0.5	U	1.00	P	10/22/2025	15:05	LB137619
	Cadmium	0.34	+/-0.5	U	1.00	P	10/22/2025	15:05	LB137619
	Calcium	45.7	+/-250	U	500	P	10/22/2025	15:05	LB137619
	Chromium	0.21	+/-1	U	2.00	P	10/22/2025	15:05	LB137619
	Cobalt	0.070	+/-0.5	U	1.00	P	10/22/2025	15:05	LB137619
	Copper	0.30	+/-1	U	2.00	P	10/22/2025	15:05	LB137619
	Iron	7.81	+/-25	U	50.0	P	10/22/2025	15:05	LB137619
	Lead	0.21	+/-0.5	U	1.00	P	10/22/2025	15:05	LB137619
	Magnesium	19.5	+/-250	U	500	P	10/22/2025	15:05	LB137619
	Manganese	0.43	+/-0.5	U	1.00	P	10/22/2025	15:05	LB137619
	Nickel	0.27	+/-0.5	U	1.00	P	10/22/2025	15:05	LB137619
	Potassium	36.4	+/-250	U	500	P	10/22/2025	15:05	LB137619
	Selenium	2.90	+/-2.5	U	5.00	P	10/22/2025	15:05	LB137619
	Silver	0.060	+/-0.5	U	1.00	P	10/22/2025	15:05	LB137619
	Sodium	128	+/-250	U	500	P	10/22/2025	15:05	LB137619
	Thallium	0.060	+/-0.5	U	1.00	P	10/22/2025	15:05	LB137619
	Vanadium	0.077	+/-2.5	U	5.00	P	10/22/2025	15:05	LB137619
	Zinc	1.25	+/-2.5	U	5.00	P	10/22/2025	15:05	LB137619
CCB04	Aluminum	1.94	+/-10	U	20.0	P	10/22/2025	16:00	LB137619
	Antimony	0.11	+/-1	U	2.00	P	10/22/2025	16:00	LB137619
	Arsenic	0.089	+/-0.5	U	1.00	P	10/22/2025	16:00	LB137619
	Barium	0.21	+/-5	U	10.0	P	10/22/2025	16:00	LB137619
	Beryllium	0.32	+/-0.5	U	1.00	P	10/22/2025	16:00	LB137619
	Cadmium	0.34	+/-0.5	U	1.00	P	10/22/2025	16:00	LB137619
	Calcium	45.7	+/-250	U	500	P	10/22/2025	16:00	LB137619
	Chromium	0.21	+/-1	U	2.00	P	10/22/2025	16:00	LB137619
	Cobalt	0.070	+/-0.5	U	1.00	P	10/22/2025	16:00	LB137619
	Copper	0.30	+/-1	U	2.00	P	10/22/2025	16:00	LB137619
	Iron	7.81	+/-25	U	50.0	P	10/22/2025	16:00	LB137619
	Lead	0.21	+/-0.5	U	1.00	P	10/22/2025	16:00	LB137619

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3375

Contract: AECO02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	19.5	+/-250	U	500	P	10/22/2025	16:00	LB137619
	Manganese	0.43	+/-0.5	U	1.00	P	10/22/2025	16:00	LB137619
	Nickel	0.27	+/-0.5	U	1.00	P	10/22/2025	16:00	LB137619
	Potassium	36.4	+/-250	U	500	P	10/22/2025	16:00	LB137619
	Selenium	2.90	+/-2.5	U	5.00	P	10/22/2025	16:00	LB137619
	Silver	0.060	+/-0.5	U	1.00	P	10/22/2025	16:00	LB137619
	Sodium	128	+/-250	U	500	P	10/22/2025	16:00	LB137619
	Thallium	0.060	+/-0.5	U	1.00	P	10/22/2025	16:00	LB137619
	Vanadium	0.077	+/-2.5	U	5.00	P	10/22/2025	16:00	LB137619
	Zinc	1.25	+/-2.5	U	5.00	P	10/22/2025	16:00	LB137619
CCB05	Aluminum	1.94	+/-10	U	20.0	P	10/22/2025	16:55	LB137619
	Antimony	0.11	+/-1	U	2.00	P	10/22/2025	16:55	LB137619
	Arsenic	0.089	+/-0.5	U	1.00	P	10/22/2025	16:55	LB137619
	Barium	0.21	+/-5	U	10.0	P	10/22/2025	16:55	LB137619
	Beryllium	0.32	+/-0.5	U	1.00	P	10/22/2025	16:55	LB137619
	Cadmium	0.34	+/-0.5	U	1.00	P	10/22/2025	16:55	LB137619
	Calcium	45.7	+/-250	U	500	P	10/22/2025	16:55	LB137619
	Chromium	0.21	+/-1	U	2.00	P	10/22/2025	16:55	LB137619
	Cobalt	0.070	+/-0.5	U	1.00	P	10/22/2025	16:55	LB137619
	Copper	0.30	+/-1	U	2.00	P	10/22/2025	16:55	LB137619
	Iron	7.81	+/-25	U	50.0	P	10/22/2025	16:55	LB137619
	Lead	0.21	+/-0.5	U	1.00	P	10/22/2025	16:55	LB137619
	Magnesium	19.5	+/-250	U	500	P	10/22/2025	16:55	LB137619
	Manganese	0.43	+/-0.5	U	1.00	P	10/22/2025	16:55	LB137619
	Nickel	0.27	+/-0.5	U	1.00	P	10/22/2025	16:55	LB137619
	Potassium	36.4	+/-250	U	500	P	10/22/2025	16:55	LB137619
	Selenium	2.90	+/-2.5	U	5.00	P	10/22/2025	16:55	LB137619
	Silver	0.060	+/-0.5	U	1.00	P	10/22/2025	16:55	LB137619
	Sodium	182	+/-250	J	500	P	10/22/2025	16:55	LB137619
	Thallium	0.060	+/-0.5	U	1.00	P	10/22/2025	16:55	LB137619
	Vanadium	0.077	+/-2.5	U	5.00	P	10/22/2025	16:55	LB137619
	Zinc	1.25	+/-2.5	U	5.00	P	10/22/2025	16:55	LB137619
CCB06	Aluminum	1.94	+/-10	U	20.0	P	10/22/2025	18:10	LB137619
	Antimony	0.11	+/-1	U	2.00	P	10/22/2025	18:10	LB137619
	Arsenic	0.089	+/-0.5	U	1.00	P	10/22/2025	18:10	LB137619
	Barium	0.21	+/-5	U	10.0	P	10/22/2025	18:10	LB137619
	Beryllium	0.32	+/-0.5	U	1.00	P	10/22/2025	18:10	LB137619
	Cadmium	0.34	+/-0.5	U	1.00	P	10/22/2025	18:10	LB137619
	Calcium	45.7	+/-250	U	500	P	10/22/2025	18:10	LB137619

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3375

Contract: AECO02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	0.21	+/-1	U	2.00	P	10/22/2025	18:10	LB137619
	Cobalt	0.070	+/-0.5	U	1.00	P	10/22/2025	18:10	LB137619
	Copper	0.30	+/-1	U	2.00	P	10/22/2025	18:10	LB137619
	Iron	7.81	+/-25	U	50.0	P	10/22/2025	18:10	LB137619
	Lead	0.21	+/-0.5	U	1.00	P	10/22/2025	18:10	LB137619
	Magnesium	19.5	+/-250	U	500	P	10/22/2025	18:10	LB137619
	Manganese	0.43	+/-0.5	U	1.00	P	10/22/2025	18:10	LB137619
	Nickel	0.27	+/-0.5	U	1.00	P	10/22/2025	18:10	LB137619
	Potassium	36.4	+/-250	U	500	P	10/22/2025	18:10	LB137619
	Selenium	2.90	+/-2.5	U	5.00	P	10/22/2025	18:10	LB137619
	Silver	0.060	+/-0.5	U	1.00	P	10/22/2025	18:10	LB137619
	Sodium	128	+/-250	U	500	P	10/22/2025	18:10	LB137619
	Thallium	0.060	+/-0.5	U	1.00	P	10/22/2025	18:10	LB137619
	Vanadium	0.077	+/-2.5	U	5.00	P	10/22/2025	18:10	LB137619
	Zinc	1.25	+/-2.5	U	5.00	P	10/22/2025	18:10	LB137619
CCB07	Aluminum	1.94	+/-10	U	20.0	P	10/22/2025	18:49	LB137619
	Antimony	0.11	+/-1	U	2.00	P	10/22/2025	18:49	LB137619
	Arsenic	0.089	+/-0.5	U	1.00	P	10/22/2025	18:49	LB137619
	Barium	0.21	+/-5	U	10.0	P	10/22/2025	18:49	LB137619
	Beryllium	0.32	+/-0.5	U	1.00	P	10/22/2025	18:49	LB137619
	Cadmium	0.34	+/-0.5	U	1.00	P	10/22/2025	18:49	LB137619
	Calcium	45.7	+/-250	U	500	P	10/22/2025	18:49	LB137619
	Chromium	0.21	+/-1	U	2.00	P	10/22/2025	18:49	LB137619
	Cobalt	0.070	+/-0.5	U	1.00	P	10/22/2025	18:49	LB137619
	Copper	0.30	+/-1	U	2.00	P	10/22/2025	18:49	LB137619
	Iron	7.81	+/-25	U	50.0	P	10/22/2025	18:49	LB137619
	Lead	0.21	+/-0.5	U	1.00	P	10/22/2025	18:49	LB137619
	Magnesium	19.5	+/-250	U	500	P	10/22/2025	18:49	LB137619
	Manganese	0.43	+/-0.5	U	1.00	P	10/22/2025	18:49	LB137619
	Nickel	0.27	+/-0.5	U	1.00	P	10/22/2025	18:49	LB137619
	Potassium	36.4	+/-250	U	500	P	10/22/2025	18:49	LB137619
	Selenium	2.90	+/-2.5	U	5.00	P	10/22/2025	18:49	LB137619
	Silver	0.060	+/-0.5	U	1.00	P	10/22/2025	18:49	LB137619
	Sodium	128	+/-250	U	500	P	10/22/2025	18:49	LB137619
	Thallium	0.060	+/-0.5	U	1.00	P	10/22/2025	18:49	LB137619
	Vanadium	0.077	+/-2.5	U	5.00	P	10/22/2025	18:49	LB137619
	Zinc	1.25	+/-2.5	U	5.00	P	10/22/2025	18:49	LB137619
CCB08	Aluminum	1.94	+/-10	U	20.0	P	10/22/2025	19:52	LB137619
	Antimony	0.11	+/-1	U	2.00	P	10/22/2025	19:52	LB137619

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3375

Contract: AECO02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	0.089	+/-0.5	U	1.00	P	10/22/2025	19:52	LB137619
	Barium	0.21	+/-5	U	10.0	P	10/22/2025	19:52	LB137619
	Beryllium	0.32	+/-0.5	U	1.00	P	10/22/2025	19:52	LB137619
	Cadmium	0.34	+/-0.5	U	1.00	P	10/22/2025	19:52	LB137619
	Calcium	45.7	+/-250	U	500	P	10/22/2025	19:52	LB137619
	Chromium	0.21	+/-1	U	2.00	P	10/22/2025	19:52	LB137619
	Cobalt	0.070	+/-0.5	U	1.00	P	10/22/2025	19:52	LB137619
	Copper	0.30	+/-1	U	2.00	P	10/22/2025	19:52	LB137619
	Iron	7.81	+/-25	U	50.0	P	10/22/2025	19:52	LB137619
	Lead	0.21	+/-0.5	U	1.00	P	10/22/2025	19:52	LB137619
	Magnesium	19.5	+/-250	U	500	P	10/22/2025	19:52	LB137619
	Manganese	0.43	+/-0.5	U	1.00	P	10/22/2025	19:52	LB137619
	Nickel	0.27	+/-0.5	U	1.00	P	10/22/2025	19:52	LB137619
	Potassium	36.4	+/-250	U	500	P	10/22/2025	19:52	LB137619
	Selenium	2.90	+/-2.5	U	5.00	P	10/22/2025	19:52	LB137619
	Silver	0.060	+/-0.5	U	1.00	P	10/22/2025	19:52	LB137619
	Sodium	128	+/-250	U	500	P	10/22/2025	19:52	LB137619
	Thallium	0.060	+/-0.5	U	1.00	P	10/22/2025	19:52	LB137619
	Vanadium	0.077	+/-2.5	U	5.00	P	10/22/2025	19:52	LB137619
	Zinc	1.25	+/-2.5	U	5.00	P	10/22/2025	19:52	LB137619
CCB09	Aluminum	1.94	+/-10	U	20.0	P	10/22/2025	20:31	LB137619
	Antimony	0.18	+/-1	J	2.00	P	10/22/2025	20:31	LB137619
	Arsenic	0.089	+/-0.5	U	1.00	P	10/22/2025	20:31	LB137619
	Barium	0.21	+/-5	U	10.0	P	10/22/2025	20:31	LB137619
	Beryllium	0.32	+/-0.5	U	1.00	P	10/22/2025	20:31	LB137619
	Cadmium	0.34	+/-0.5	U	1.00	P	10/22/2025	20:31	LB137619
	Calcium	45.7	+/-250	U	500	P	10/22/2025	20:31	LB137619
	Chromium	0.21	+/-1	U	2.00	P	10/22/2025	20:31	LB137619
	Cobalt	0.070	+/-0.5	U	1.00	P	10/22/2025	20:31	LB137619
	Copper	0.30	+/-1	U	2.00	P	10/22/2025	20:31	LB137619
	Iron	7.81	+/-25	U	50.0	P	10/22/2025	20:31	LB137619
	Lead	0.21	+/-0.5	U	1.00	P	10/22/2025	20:31	LB137619
	Magnesium	19.5	+/-250	U	500	P	10/22/2025	20:31	LB137619
	Manganese	0.43	+/-0.5	U	1.00	P	10/22/2025	20:31	LB137619
	Nickel	0.27	+/-0.5	U	1.00	P	10/22/2025	20:31	LB137619
	Potassium	43.9	+/-250	J	500	P	10/22/2025	20:31	LB137619
	Selenium	2.90	+/-2.5	U	5.00	P	10/22/2025	20:31	LB137619
	Silver	0.060	+/-0.5	U	1.00	P	10/22/2025	20:31	LB137619
	Sodium	128	+/-250	U	500	P	10/22/2025	20:31	LB137619

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3375

Contract: AECO02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	0.060	+/-0.5	U	1.00	P	10/22/2025	20:31	LB137619
	Vanadium	0.077	+/-2.5	U	5.00	P	10/22/2025	20:31	LB137619
	Zinc	1.25	+/-2.5	U	5.00	P	10/22/2025	20:31	LB137619
CCB10	Aluminum	1.94	+/-10	U	20.0	P	10/22/2025	21:06	LB137619
	Antimony	0.19	+/-1	J	2.00	P	10/22/2025	21:06	LB137619
	Arsenic	0.089	+/-0.5	U	1.00	P	10/22/2025	21:06	LB137619
	Barium	0.21	+/-5	U	10.0	P	10/22/2025	21:06	LB137619
	Beryllium	0.32	+/-0.5	U	1.00	P	10/22/2025	21:06	LB137619
	Cadmium	0.34	+/-0.5	U	1.00	P	10/22/2025	21:06	LB137619
	Calcium	45.7	+/-250	U	500	P	10/22/2025	21:06	LB137619
	Chromium	0.21	+/-1	U	2.00	P	10/22/2025	21:06	LB137619
	Cobalt	0.070	+/-0.5	U	1.00	P	10/22/2025	21:06	LB137619
	Copper	0.30	+/-1	U	2.00	P	10/22/2025	21:06	LB137619
	Iron	7.81	+/-25	U	50.0	P	10/22/2025	21:06	LB137619
	Lead	0.21	+/-0.5	U	1.00	P	10/22/2025	21:06	LB137619
	Magnesium	19.5	+/-250	U	500	P	10/22/2025	21:06	LB137619
	Manganese	0.43	+/-0.5	U	1.00	P	10/22/2025	21:06	LB137619
	Nickel	0.27	+/-0.5	U	1.00	P	10/22/2025	21:06	LB137619
	Potassium	36.4	+/-250	U	500	P	10/22/2025	21:06	LB137619
	Selenium	2.90	+/-2.5	U	5.00	P	10/22/2025	21:06	LB137619
	Silver	0.060	+/-0.5	U	1.00	P	10/22/2025	21:06	LB137619
	Sodium	128	+/-250	U	500	P	10/22/2025	21:06	LB137619
	Thallium	0.060	+/-0.5	U	1.00	P	10/22/2025	21:06	LB137619
	Vanadium	0.077	+/-2.5	U	5.00	P	10/22/2025	21:06	LB137619
	Zinc	1.25	+/-2.5	U	5.00	P	10/22/2025	21:06	LB137619