

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: AECOM

Contract: AECO02

Lab Code: ACE

SDG No.: Q3375

Instrument ID: P7

Start Date : 10/22/2025

Run Number: LB137619

End Date : 10/22/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
S0	S0	1154	100		100		100		100		100	
S2	S2	1157	99		98		101		99		99	
S3	S3	1204	98		99		99		98		97	
S4	S4	1207	96		96		99		98		99	
S5	S5	1210	96		94		96		95		99	
S6	S6	1213	97		93		96		95		100	
S7	S7	1216	98		95		97		96		102	
S8	S8	1219	101		101		101		94		104	
ICV01	ICV01	1238	96		100		102		101		100	
LLICV01	LLICV01	1312	92		97		100		98		97	
ICB01	ICB01	1316	93		99		99		98		97	
ICSA01	ICSA01	1320	94		100		101		97		102	
ICSAB01	ICSAB01	1323	95		104		107		103		108	
CCV01	CCV01	1327	95		105		106		100		105	
CCB01	CCB01	1330	90		98		101		99		99	
CRI01	CRI01	1333	92		100		101		101		101	
ZZZZZZ	ZZZZZZ	1340										
ZZZZZZ	ZZZZZZ	1343										
ZZZZZZ	ZZZZZZ	1346										
ZZZZZZ	ZZZZZZ	1349										
ZZZZZZ	ZZZZZZ	1353										
ZZZZZZ	ZZZZZZ	1356										
ZZZZZZ	ZZZZZZ	1359										
ZZZZZZ	ZZZZZZ	1402										
ZZZZZZ	ZZZZZZ	1408										
CCV02	CCV02	1411	95		104		105		99		104	
CCB02	CCB02	1418	92		99		101		103		100	
ZZZZZZ	ZZZZZZ	1421										
ZZZZZZ	ZZZZZZ	1424										
ZZZZZZ	ZZZZZZ	1427										
ZZZZZZ	ZZZZZZ	1430										
ZZZZZZ	ZZZZZZ	1434										
ZZZZZZ	ZZZZZZ	1437										
ZZZZZZ	ZZZZZZ	1440										
ZZZZZZ	ZZZZZZ	1443										
ZZZZZZ	ZZZZZZ	1446										

Internal Standard %RI Limit: 70 - 130

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: AECOM

Contract: AECO02

Lab Code: ACE

SDG No.: Q3375

Instrument ID: P7

Start Date : 10/22/2025

Run Number: LB137619

End Date : 10/22/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	ZZZZZZ	1449										
CCV03	CCV03	1454	90		100		102		96		101	
CCB03	CCB03	1505	89		99		101		100		100	
ZZZZZZ	ZZZZZZ	1514										
ZZZZZZ	ZZZZZZ	1517										
PB170158BL	PB170158BL	1520	89		96		98		98		99	
PB170158BS	PB170158BS	1524	92		98		100		97		101	
ZZZZZZ	ZZZZZZ	1527										
ZZZZZZ	ZZZZZZ	1530										
Q3375-01	PR132-S02-0	1533	94		103		115		96		100	
Q3375-02	PR132-S04-0	1536	90		102		109		98		97	
Q3375-03	PR132-S05-0	1539	93		107		117		98		102	
Q3375-04	PR132-S02-0	1542	96		111		122		98		103	
CCV04	CCV04	1545	92		99		101		96		101	
CCB04	CCB04	1600	87		96		98		98		97	
Q3375-05	PR132-DP02-	1603	96		113		127		99		104	
ZZZZZZ	ZZZZZZ	1606										
ZZZZZZ	ZZZZZZ	1609										
ZZZZZZ	ZZZZZZ	1613										
ZZZZZZ	ZZZZZZ	1616										
ZZZZZZ	ZZZZZZ	1619										
Q3376-05L	PR132-S12-0	1625	89		100		103		100		101	
Q3376-05MSD	PR132-S12-0	1631	90		103		113		96		103	
Q3376-05DUP	PR132-S12-0	1640	93		106		116		99		104	
Q3376-05MS	PR132-S12-0	1643	95		107		118		98		105	
CCV05	CCV05	1646	94		102		103		99		105	
CCB05	CCB05	1655	87		95		98		98		100	
Q3376-05A	PR132-S12-0	1700	93		106		116		97		104	
ZZZZZZ	ZZZZZZ	1715										
ZZZZZZ	ZZZZZZ	1718										
ZZZZZZ	ZZZZZZ	1722										
ZZZZZZ	ZZZZZZ	1725										
ZZZZZZ	ZZZZZZ	1728										
ZZZZZZ	ZZZZZZ	1731										
ZZZZZZ	ZZZZZZ	1734										
ZZZZZZ	ZZZZZZ	1741										

Internal Standard %RI Limit: 70 - 130

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: AECOM

Contract: AECO02

Lab Code: ACE

SDG No.: Q3375

Instrument ID: P7

Start Date : 10/22/2025

Run Number: LB137619

End Date : 10/22/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	ZZZZZZ	1744										
CCV06	CCV06	1759	85		95		97		93		102	
CCB06	CCB06	1810	82		90		93		94		97	
ZZZZZZ	ZZZZZZ	1815										
ZZZZZZ	ZZZZZZ	1818										
ZZZZZZ	ZZZZZZ	1821										
ZZZZZZ	ZZZZZZ	1823										
ZZZZZZ	ZZZZZZ	1826										
ZZZZZZ	ZZZZZZ	1829										
CCV07	CCV07	1841	84		94		98		93		99	
CCB07	CCB07	1849	80		90		94		95		96	
ZZZZZZ	ZZZZZZ	1912										
ZZZZZZ	ZZZZZZ	1915										
ZZZZZZ	ZZZZZZ	1918										
ZZZZZZ	ZZZZZZ	1922										
ZZZZZZ	ZZZZZZ	1925										
ZZZZZZ	ZZZZZZ	1928										
ZZZZZZ	ZZZZZZ	1931										
ZZZZZZ	ZZZZZZ	1934										
ZZZZZZ	ZZZZZZ	1937										
Q3375-03	PR132-S05-0	1941	83		95		100		97		98	
CCV08	CCV08	1944	84		92		96		92		98	
CCB08	CCB08	1952	78		89		92		92		94	
Q3375-05	PR132-DP02-	1957	81		91		97		91		95	
ZZZZZZ	ZZZZZZ	2000										
ZZZZZZ	ZZZZZZ	2003										
ZZZZZZ	ZZZZZZ	2006										
ZZZZZZ	ZZZZZZ	2009										
ZZZZZZ	ZZZZZZ	2012										
ZZZZZZ	ZZZZZZ	2016										
ZZZZZZ	ZZZZZZ	2019										
ZZZZZZ	ZZZZZZ	2022										
ZZZZZZ	ZZZZZZ	2025										
CCV09	CCV09	2028	83		92		94		91		97	
CCB09	CCB09	2031	80		89		92		91		95	

Internal Standard %RI Limit: 70 - 130

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: AECOM

Contract: AECO02

Lab Code: ACE

SDG No.: Q3375

Instrument ID: P7

Start Date : 10/22/2025

Run Number: LB137619

End Date : 10/22/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
S0	S0	1154	100		100		100		100		100	
S2	S2	1157	99		99		100		98		99	
S3	S3	1204	98		97		97		98		97	
S4	S4	1207	96		97		94		97		99	
S5	S5	1210	95		95		93		96		97	
S6	S6	1213	96		95		95		99		100	
S7	S7	1216	97		95		94		100		101	
S8	S8	1219	105		104		94		103		105	
ICV01	ICV01	1238	101		102		100		101		102	
LLICV01	LLICV01	1312	98		99		97		98		99	
ICB01	ICB01	1316	99		101		99		99		100	
ICSA01	ICSA01	1320	101		100		97		101		103	
ICSAB01	ICSAB01	1323	104		102		100		104		105	
CCV01	CCV01	1327	103		102		97		103		105	
CCB01	CCB01	1330	98		100		99		99		100	
CRI01	CRI01	1333	99		99		99		101		101	
ZZZZZZ	ZZZZZZ	1340										
ZZZZZZ	ZZZZZZ	1343										
ZZZZZZ	ZZZZZZ	1346										
ZZZZZZ	ZZZZZZ	1349										
ZZZZZZ	ZZZZZZ	1353										
ZZZZZZ	ZZZZZZ	1356										
ZZZZZZ	ZZZZZZ	1359										
ZZZZZZ	ZZZZZZ	1402										
ZZZZZZ	ZZZZZZ	1408										
CCV02	CCV02	1411	104		104		98		105		105	
CCB02	CCB02	1418	100		102		102		103		104	
ZZZZZZ	ZZZZZZ	1421										
ZZZZZZ	ZZZZZZ	1424										
ZZZZZZ	ZZZZZZ	1427										
ZZZZZZ	ZZZZZZ	1430										
ZZZZZZ	ZZZZZZ	1434										
ZZZZZZ	ZZZZZZ	1437										
ZZZZZZ	ZZZZZZ	1440										
ZZZZZZ	ZZZZZZ	1443										
ZZZZZZ	ZZZZZZ	1446										

Internal Standard %RI Limit: 70 - 130

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: AECOM

Contract: AECO02

Lab Code: ACE

SDG No.: Q3375

Instrument ID: P7

Start Date : 10/22/2025

Run Number: LB137619

End Date : 10/22/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
ZZZZZZ	ZZZZZZ	1449										
CCV03	CCV03	1454	101		101		96		104		103	
CCB03	CCB03	1505	97		100		99		102		102	
ZZZZZZ	ZZZZZZ	1514										
ZZZZZZ	ZZZZZZ	1517										
PB170158BL	PB170158BL	1520	95		97		97		101		100	
PB170158BS	PB170158BS	1524	96		97		97		100		102	
ZZZZZZ	ZZZZZZ	1527										
ZZZZZZ	ZZZZZZ	1530										
Q3375-01	PR132-S02-0	1533	106		116		98		104		105	
Q3375-02	PR132-S04-0	1536	102		109		97		102		103	
Q3375-03	PR132-S05-0	1539	107		115		99		103		104	
Q3375-04	PR132-S02-0	1542	111		123		98		105		105	
CCV04	CCV04	1545	99		98		95		101		102	
CCB04	CCB04	1600	94		97		96		98		100	
Q3375-05	PR132-DP02-	1603	112		128		98		108		108	
ZZZZZZ	ZZZZZZ	1606										
ZZZZZZ	ZZZZZZ	1609										
ZZZZZZ	ZZZZZZ	1613										
ZZZZZZ	ZZZZZZ	1616										
ZZZZZZ	ZZZZZZ	1619										
Q3376-05L	PR132-S12-0	1625	102		104		100		103		103	
Q3376-05MSD	PR132-S12-0	1631	105		114		96		103		103	
Q3376-05DUP	PR132-S12-0	1640	107		119		101		108		108	
Q3376-05MS	PR132-S12-0	1643	108		118		99		107		107	
CCV05	CCV05	1646	103		102		99		107		108	
CCB05	CCB05	1655	94		98		98		101		102	
Q3376-05A	PR132-S12-0	1700	106		115		101		109		107	
ZZZZZZ	ZZZZZZ	1715										
ZZZZZZ	ZZZZZZ	1718										
ZZZZZZ	ZZZZZZ	1722										
ZZZZZZ	ZZZZZZ	1725										
ZZZZZZ	ZZZZZZ	1728										
ZZZZZZ	ZZZZZZ	1731										
ZZZZZZ	ZZZZZZ	1734										
ZZZZZZ	ZZZZZZ	1741										

Internal Standard %RI Limit: 70 - 130

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: AECOM

Contract: AECO02

Lab Code: ACE

SDG No.: Q3375

Instrument ID: P7

Start Date : 10/22/2025

Run Number: LB137619

End Date : 10/22/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
ZZZZZZ	ZZZZZZ	1744										
CCV06	CCV06	1759	96		96		93		102		102	
CCB06	CCB06	1810	89		93		93		97		97	
ZZZZZZ	ZZZZZZ	1815										
ZZZZZZ	ZZZZZZ	1818										
ZZZZZZ	ZZZZZZ	1821										
ZZZZZZ	ZZZZZZ	1823										
ZZZZZZ	ZZZZZZ	1826										
ZZZZZZ	ZZZZZZ	1829										
CCV07	CCV07	1841	95		95		92		99		100	
CCB07	CCB07	1849	90		94		95		97		96	
ZZZZZZ	ZZZZZZ	1912										
ZZZZZZ	ZZZZZZ	1915										
ZZZZZZ	ZZZZZZ	1918										
ZZZZZZ	ZZZZZZ	1922										
ZZZZZZ	ZZZZZZ	1925										
ZZZZZZ	ZZZZZZ	1928										
ZZZZZZ	ZZZZZZ	1931										
ZZZZZZ	ZZZZZZ	1934										
ZZZZZZ	ZZZZZZ	1937										
Q3375-03	PR132-S05-0	1941	96		100		96		100		100	
CCV08	CCV08	1944	93		95		91		100		100	
CCB08	CCB08	1952	89		93		93		95		95	
Q3375-05	PR132-DP02-	1957	91		97		92		97		97	
ZZZZZZ	ZZZZZZ	2000										
ZZZZZZ	ZZZZZZ	2003										
ZZZZZZ	ZZZZZZ	2006										
ZZZZZZ	ZZZZZZ	2009										
ZZZZZZ	ZZZZZZ	2012										
ZZZZZZ	ZZZZZZ	2016										
ZZZZZZ	ZZZZZZ	2019										
ZZZZZZ	ZZZZZZ	2022										
ZZZZZZ	ZZZZZZ	2025										
CCV09	CCV09	2028	92		93		90		98		98	
CCB09	CCB09	2031	89		94		93		97		96	

Internal Standard %RI Limit: 70 - 130