

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075868.D **Time Analyzed:** 11:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.806	5.708	5.908	472.530	500.000	-5.5
Aroclor-1016-2	5.827	5.729	5.929	502.350	500.000	0.5
Aroclor-1016-3	5.890	5.792	5.992	485.700	500.000	-2.9
Aroclor-1016-4	5.987	5.889	6.089	509.420	500.000	1.9
Aroclor-1016-5	6.279	6.181	6.381	520.380	500.000	4.1
Aroclor-1260-1	7.397	7.299	7.499	482.450	500.000	-3.5
Aroclor-1260-2	7.650	7.551	7.751	470.250	500.000	-6.0
Aroclor-1260-3	8.008	7.909	8.109	501.460	500.000	0.3
Aroclor-1260-4	8.236	8.137	8.337	492.140	500.000	-1.6
Aroclor-1260-5	8.563	8.464	8.664	502.900	500.000	0.6
Decachlorobiphenyl	10.435	10.334	10.534	40.440	50.000	-19.1
Tetrachloro-m-xylene	4.653	4.555	4.755	47.380	50.000	-5.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075868.D **Time Analyzed:** 11:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.885	4.793	4.993	540.620	500.000	8.1
Aroclor-1016-2	4.941	4.851	5.051	571.750	500.000	14.4
Aroclor-1016-3	5.064	4.971	5.171	520.270	500.000	4.1
Aroclor-1016-4	5.101	5.011	5.211	573.190	500.000	14.6
Aroclor-1016-5	5.316	5.225	5.425	541.560	500.000	8.3
Aroclor-1260-1	6.344	6.255	6.455	542.970	500.000	8.6
Aroclor-1260-2	6.532	6.441	6.641	549.380	500.000	9.9
Aroclor-1260-3	6.685	6.595	6.795	527.020	500.000	5.4
Aroclor-1260-4	7.153	7.065	7.265	484.910	500.000	-3.0
Aroclor-1260-5	7.394	7.304	7.504	458.580	500.000	-8.3
Decachlorobiphenyl	8.791	8.703	8.903	43.820	50.000	-12.4
Tetrachloro-m-xylene	3.789	3.696	3.896	51.660	50.000	3.3