

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075881.D **Time Analyzed:** 16:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.807	5.708	5.908	499.980	500.000	0.0
Aroclor-1016-2	5.828	5.729	5.929	503.740	500.000	0.7
Aroclor-1016-3	5.891	5.792	5.992	505.040	500.000	1.0
Aroclor-1016-4	5.988	5.889	6.089	527.610	500.000	5.5
Aroclor-1016-5	6.280	6.181	6.381	521.200	500.000	4.2
Aroclor-1260-1	7.398	7.299	7.499	464.050	500.000	-7.2
Aroclor-1260-2	7.651	7.551	7.751	450.930	500.000	-9.8
Aroclor-1260-3	8.009	7.909	8.109	475.600	500.000	-4.9
Aroclor-1260-4	8.238	8.137	8.337	484.630	500.000	-3.1
Aroclor-1260-5	8.563	8.464	8.664	488.580	500.000	-2.3
Decachlorobiphenyl	10.433	10.334	10.534	39.770	50.000	-20.5
Tetrachloro-m-xylene	4.653	4.555	4.755	46.220	50.000	-7.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075881.D **Time Analyzed:** 16:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.888	4.793	4.993	580.620	500.000	16.1
Aroclor-1016-2	4.943	4.851	5.051	632.020	500.000	26.4
Aroclor-1016-3	5.065	4.971	5.171	570.020	500.000	14.0
Aroclor-1016-4	5.106	5.011	5.211	566.310	500.000	13.3
Aroclor-1016-5	5.320	5.225	5.425	555.760	500.000	11.2
Aroclor-1260-1	6.348	6.255	6.455	483.300	500.000	-3.3
Aroclor-1260-2	6.534	6.441	6.641	518.490	500.000	3.7
Aroclor-1260-3	6.688	6.595	6.795	495.820	500.000	-0.8
Aroclor-1260-4	7.157	7.065	7.265	428.480	500.000	-14.3
Aroclor-1260-5	7.397	7.304	7.504	417.660	500.000	-16.5
Decachlorobiphenyl	8.792	8.703	8.903	41.820	50.000	-16.4
Tetrachloro-m-xylene	3.791	3.696	3.896	53.080	50.000	6.2