

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071008.D **Time Analyzed:** 10:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.492	4.380	4.580	479.650	500.000	-4.1
Aroclor-1016-2	4.510	4.399	4.599	508.480	500.000	1.7
Aroclor-1016-3	4.567	4.456	4.656	536.740	500.000	7.3
Aroclor-1016-4	4.658	4.546	4.746	515.560	500.000	3.1
Aroclor-1016-5	4.942	4.831	5.031	511.360	500.000	2.3
Aroclor-1260-1	6.032	5.919	6.119	501.670	500.000	0.3
Aroclor-1260-2	6.288	6.175	6.375	472.070	500.000	-5.6
Aroclor-1260-3	6.638	6.526	6.726	523.450	500.000	4.7
Aroclor-1260-4	6.853	6.740	6.940	530.680	500.000	6.1
Aroclor-1260-5	7.158	7.045	7.245	493.240	500.000	-1.4
Decachlorobiphenyl	8.512	8.396	8.596	53.300	50.000	6.6
Tetrachloro-m-xylene	3.409	3.300	3.500	53.000	50.000	6.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071008.D **Time Analyzed:** 10:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.778	3.678	3.878	419.270	500.000	-16.1
Aroclor-1016-2	3.793	3.693	3.893	465.010	500.000	-7.0
Aroclor-1016-3	3.953	3.853	4.053	468.600	500.000	-6.3
Aroclor-1016-4	4.001	3.902	4.102	489.820	500.000	-2.0
Aroclor-1016-5	4.196	4.097	4.297	467.260	500.000	-6.5
Aroclor-1260-1	5.189	5.089	5.289	468.620	500.000	-6.3
Aroclor-1260-2	5.379	5.278	5.478	452.340	500.000	-9.5
Aroclor-1260-3	5.520	5.419	5.619	466.410	500.000	-6.7
Aroclor-1260-4	5.981	5.879	6.079	464.170	500.000	-7.2
Aroclor-1260-5	6.225	6.124	6.324	437.130	500.000	-12.6
Decachlorobiphenyl	7.536	7.433	7.633	54.320	50.000	8.6
Tetrachloro-m-xylene	2.776	2.678	2.878	50.390	50.000	0.8