

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071023.D **Time Analyzed:** 16:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.480	4.380	4.580	550.910	500.000	10.2
Aroclor-1016-2	4.499	4.399	4.599	545.480	500.000	9.1
Aroclor-1016-3	4.557	4.456	4.656	557.230	500.000	11.4
Aroclor-1016-4	4.647	4.546	4.746	552.890	500.000	10.6
Aroclor-1016-5	4.931	4.831	5.031	543.570	500.000	8.7
Aroclor-1260-1	6.021	5.919	6.119	524.350	500.000	4.9
Aroclor-1260-2	6.276	6.175	6.375	521.190	500.000	4.2
Aroclor-1260-3	6.628	6.526	6.726	548.540	500.000	9.7
Aroclor-1260-4	6.842	6.740	6.940	553.860	500.000	10.8
Aroclor-1260-5	7.147	7.045	7.245	548.590	500.000	9.7
Decachlorobiphenyl	8.499	8.396	8.596	57.120	50.000	14.2
Tetrachloro-m-xylene	3.400	3.300	3.500	54.370	50.000	8.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071023.D **Time Analyzed:** 16:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.777	3.678	3.878	486.880	500.000	-2.6
Aroclor-1016-2	3.793	3.693	3.893	486.950	500.000	-2.6
Aroclor-1016-3	3.953	3.853	4.053	501.250	500.000	0.3
Aroclor-1016-4	4.002	3.902	4.102	507.110	500.000	1.4
Aroclor-1016-5	4.196	4.097	4.297	503.750	500.000	0.8
Aroclor-1260-1	5.189	5.089	5.289	495.440	500.000	-0.9
Aroclor-1260-2	5.378	5.278	5.478	487.970	500.000	-2.4
Aroclor-1260-3	5.519	5.419	5.619	493.830	500.000	-1.2
Aroclor-1260-4	5.979	5.879	6.079	490.320	500.000	-1.9
Aroclor-1260-5	6.224	6.124	6.324	470.860	500.000	-5.8
Decachlorobiphenyl	7.533	7.433	7.633	56.890	50.000	13.8
Tetrachloro-m-xylene	2.777	2.678	2.878	53.850	50.000	7.7