

Surrogate Summary

SDG No.: Q3375

Client: AECOM

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP075246.D	PIBLK-PP075246.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	19.4	97		60	140
		Tetrachloro-m-xyl	2	20	17.9	90		60	140
		Decachlorobiphen	2	20	20.0	100		60	140
I.BLK-PP075872.D	PIBLK-PP075872.D	Tetrachloro-m-xyl	1	20	23.1	115		60	140
		Decachlorobiphen	1	20	20.6	103		60	140
		Tetrachloro-m-xyl	2	20	22.7	113		60	140
		Decachlorobiphen	2	20	22.0	110		60	140
Q3383-01MS	OR-03-101725MS	Tetrachloro-m-xyl	1	20	28.0	140		32	144
		Decachlorobiphen	1	20	22.4	112		32	175
		Tetrachloro-m-xyl	2	20	26.4	132		32	144
		Decachlorobiphen	2	20	22.4	112		32	175
Q3383-01MSD	OR-03-101725MSD	Tetrachloro-m-xyl	1	20	28.4	142		32	144
		Decachlorobiphen	1	20	21.8	109		32	175
		Tetrachloro-m-xyl	2	20	18.9	95		32	144
		Decachlorobiphen	2	20	21.3	107		32	175
I.BLK-PP075885.D	PIBLK-PP075885.D	Tetrachloro-m-xyl	1	20	22.7	114		60	140
		Decachlorobiphen	1	20	20.0	100		60	140
		Tetrachloro-m-xyl	2	20	23.5	117		60	140
		Decachlorobiphen	2	20	20.9	104		60	140
I.BLK-PQ070972.D	PIBLK-PQ070972.D	Tetrachloro-m-xyl	1	20	18.3	91		60	140
		Decachlorobiphen	1	20	19.0	95		60	140
		Tetrachloro-m-xyl	2	20	19.1	96		60	140
		Decachlorobiphen	2	20	19.4	97		60	140
I.BLK-PQ071012.D	PIBLK-PQ071012.D	Tetrachloro-m-xyl	1	20	18.7	94		60	140
		Decachlorobiphen	1	20	22.1	111		60	140
		Tetrachloro-m-xyl	2	20	18.9	95		60	140
		Decachlorobiphen	2	20	22.0	110		60	140
PB170162BL	PB170162BL	Tetrachloro-m-xyl	1	20	21.7	108		32	144
		Decachlorobiphen	1	20	25.1	126		32	175
		Tetrachloro-m-xyl	2	20	21.6	108		32	144
		Decachlorobiphen	2	20	24.8	124		32	175
PB170162BS	PB170162BS	Tetrachloro-m-xyl	1	20	24.0	120		32	144
		Decachlorobiphen	1	20	27.3	137		32	175
		Tetrachloro-m-xyl	2	20	22.5	113		32	144
		Decachlorobiphen	2	20	27.3	137		32	175
Q3375-01	PR132-S02-074099-20251016	Tetrachloro-m-xyl	1	20	20.3	102		32	144
		Decachlorobiphen	1	20	15.7	78		32	175
		Tetrachloro-m-xyl	2	20	21.0	105		32	144
		Decachlorobiphen	2	20	15.1	75		32	175
Q3375-02	PR132-S04-000013-20251016	Tetrachloro-m-xyl	1	20	22.8	114		32	144

Surrogate Summary

SDG No.: Q3375

Client: AECOM

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3375-02	PR132-S04-000013-20251016	Decachlorobiphen	1	20	16.4	82		32	175
		Tetrachloro-m-xyl	2	20	23.8	119		32	144
		Decachlorobiphen	2	20	16.5	83		32	175
Q3375-03	PR132-S05-000057-20251016	Tetrachloro-m-xyl	1	20	18.0	90		32	144
		Decachlorobiphen	1	20	17.0	85		32	175
		Tetrachloro-m-xyl	2	20	18.5	92		32	144
		Decachlorobiphen	2	20	16.8	84		32	175
Q3375-04	PR132-S02-010074-20251016	Tetrachloro-m-xyl	1	20	21.2	106		32	144
		Decachlorobiphen	1	20	18.6	93		32	175
		Tetrachloro-m-xyl	2	20	21.5	108		32	144
		Decachlorobiphen	2	20	18.3	92		32	175
Q3375-05	PR132-DP02-20251016	Tetrachloro-m-xyl	1	20	20.1	100		32	144
		Decachlorobiphen	1	20	18.3	92		32	175
		Tetrachloro-m-xyl	2	20	19.8	99		32	144
		Decachlorobiphen	2	20	17.8	89		32	175
I.BLK-PQ071027.D	PIBLK-PQ071027.D	Tetrachloro-m-xyl	1	20	19.6	98		60	140
		Decachlorobiphen	1	20	22.6	113		60	140
		Tetrachloro-m-xyl	2	20	19.7	98		60	140
		Decachlorobiphen	2	20	22.3	111		60	140