



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3375</u>	Analytical Method:	<u>8082A</u>
Client:	<u>AECOM</u>	Datafile :	<u>PQ071014.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB170162BS (Column 1)	AR1016	166.5	172	ug/kg	103				71	120	
	AR1260	166.5	165	ug/kg	99				65	130	
PB170162BS (Column 2)	AR1016	166.5	153	ug/kg	92				71	120	
	AR1260	166.5	151	ug/kg	91				65	130	