

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/20/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090713.D **Time Analyzed:** 19:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.600	6.800	50.870	50.000	1.7
4,4'-DDE	6.190	6.090	6.290	50.630	50.000	1.3
4,4'-DDT	7.016	6.915	7.115	51.190	50.000	2.4
Aldrin	5.264	5.164	5.364	51.810	50.000	3.6
alpha-BHC	3.994	3.894	4.094	53.030	50.000	6.1
alpha-Chlordane	6.021	5.921	6.121	48.460	50.000	-3.1
beta-BHC	4.512	4.412	4.612	50.550	50.000	1.1
Decachlorobiphenyl	9.065	8.964	9.164	46.510	50.000	-7.0
delta-BHC	4.760	4.660	4.860	52.270	50.000	4.5
Dieldrin	6.341	6.240	6.440	51.090	50.000	2.2
Endosulfan I	6.068	5.968	6.168	50.490	50.000	1.0
Endosulfan II	6.781	6.681	6.881	49.930	50.000	-0.1
Endosulfan sulfate	7.144	7.045	7.245	49.480	50.000	-1.0
Endrin	6.568	6.468	6.668	50.890	50.000	1.8
Endrin aldehyde	6.910	6.810	7.010	50.140	50.000	0.3
Endrin ketone	7.625	7.525	7.725	49.830	50.000	-0.3
gamma-BHC (Lindane)	4.324	4.225	4.425	52.450	50.000	4.9
gamma-Chlordane	5.940	5.840	6.040	50.930	50.000	1.9
Heptachlor	4.923	4.823	5.023	52.090	50.000	4.2
Heptachlor epoxide	5.684	5.584	5.784	51.040	50.000	2.1
Methoxychlor	7.488	7.388	7.588	48.640	50.000	-2.7
Tetrachloro-m-xylene	3.544	3.445	3.645	50.600	50.000	1.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3375
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/20/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090713.D **Time Analyzed:** 19:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.920	5.821	6.021	48.360	50.000	-3.3
4,4'-DDE	5.365	5.266	5.466	48.900	50.000	-2.2
4,4'-DDT	6.173	6.074	6.274	50.220	50.000	0.4
Aldrin	4.360	4.261	4.461	49.500	50.000	-1.0
alpha-BHC	3.386	3.287	3.487	49.550	50.000	-0.9
alpha-Chlordane	5.181	5.081	5.281	49.160	50.000	-1.7
beta-BHC	4.018	3.919	4.119	49.120	50.000	-1.8
Decachlorobiphenyl	8.060	7.962	8.162	45.380	50.000	-9.2
delta-BHC	4.255	4.155	4.355	49.490	50.000	-1.0
Dieldrin	5.503	5.404	5.604	49.370	50.000	-1.3
Endosulfan I	5.238	5.138	5.338	49.000	50.000	-2.0
Endosulfan II	6.071	5.972	6.172	49.040	50.000	-1.9
Endosulfan sulfate	6.473	6.374	6.574	48.550	50.000	-2.9
Endrin	5.780	5.681	5.881	49.520	50.000	-1.0
Endrin aldehyde	6.249	6.150	6.350	48.770	50.000	-2.5
Endrin ketone	6.982	6.883	7.083	48.370	50.000	-3.3
gamma-BHC (Lindane)	3.722	3.623	3.823	49.650	50.000	-0.7
gamma-Chlordane	5.116	5.017	5.217	49.350	50.000	-1.3
Heptachlor	4.074	3.975	4.175	49.390	50.000	-1.2
Heptachlor epoxide	4.863	4.764	4.964	48.140	50.000	-3.7
Methoxychlor	6.744	6.645	6.845	49.130	50.000	-1.7
Tetrachloro-m-xylene	2.874	2.775	2.975	49.050	50.000	-1.9