

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24942,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD090646.D	17 Oct 2025 10:40		AR\AJ	Ok
2	I.BLK	I.BLK	PD090647.D	17 Oct 2025 10:53		AR\AJ	Ok
3	PEM	PEM	PD090648.D	17 Oct 2025 11:07		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD090649.D	17 Oct 2025 11:20		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PD090650.D	17 Oct 2025 11:34		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD090651.D	17 Oct 2025 11:48		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD090652.D	17 Oct 2025 12:01		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD090653.D	17 Oct 2025 12:15		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD090654.D	17 Oct 2025 12:28		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD090655.D	17 Oct 2025 12:42		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD090656.D	17 Oct 2025 12:56		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD090657.D	17 Oct 2025 13:09		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD090658.D	17 Oct 2025 13:23		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD090659.D	17 Oct 2025 13:36		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PD090660.D	17 Oct 2025 13:50		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD090661.D	17 Oct 2025 14:03		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD090662.D	17 Oct 2025 14:17		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD090663.D	17 Oct 2025 14:30		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD090664.D	17 Oct 2025 16:18		AR\AJ	Ok
20	PSTDICV050	ICVPD101725	PD090665.D	17 Oct 2025 16:46		AR\AJ	Ok
21	PCHLORICV500	ICVPD101725CHLOR	PD090666.D	17 Oct 2025 17:50		AR\AJ	Ok
22	PTOXICV500	ICVPD101725TOX	PD090667.D	17 Oct 2025 18:42		AR\AJ	Ok,M
23	I.BLK	I.BLK	PD090668.D	17 Oct 2025 19:50		AR\AJ	Ok
24	PEM	PEM	PD090669.D	17 Oct 2025 20:04		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PD090670.D	17 Oct 2025 20:18		AR\AJ	Ok
26	PB170142BL	PB170142BL	PD090671.D	17 Oct 2025 20:31		AR\AJ	Ok
27	PB170142BS	PB170142BS	PD090672.D	17 Oct 2025 20:45		AR\AJ	Ok
28	Q3366-01	TR-06-101625	PD090673.D	17 Oct 2025 20:59		AR\AJ	Ok,M
29	Q3366-03	TR-1-101625	PD090674.D	17 Oct 2025 21:12		AR\AJ	Ok,M
30	Q3367-01	BU-03-101625	PD090675.D	17 Oct 2025 21:26		AR\AJ	Ok,M
31	Q3369-01	MH-22	PD090676.D	17 Oct 2025 21:40		AR\AJ	Ok,M
32	Q3369-05	MH-21	PD090677.D	17 Oct 2025 21:53		AR\AJ	Ok,M
33	Q3363-01	PR132-SO1-085011-20	PD090678.D	17 Oct 2025 22:48		AR\AJ	Ok,M
34	Q3363-02	PR132-SO1-039085-20	PD090679.D	17 Oct 2025 23:02		AR\AJ	Ok,M
35	Q3363-03	PR132-SO3-028010-20	PD090680.D	17 Oct 2025 23:15		AR\AJ	Ok,M
36	Q3363-04	PR132-SO3-010013-20	PD090681.D	17 Oct 2025 23:29		AR\AJ	Ok,M
37	Q3363-04MS	PR132-SO3-010013-20	PD090682.D	17 Oct 2025 23:43		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q3363-04MSD	PR132-SO3-010013-20	PD090683.D	17 Oct 2025 23:56		AR\AJ	Ok,M
39	Q3363-05	COMP01-20251015	PD090684.D	18 Oct 2025 00:10		AR\AJ	Ok,M
40	Q3343-04MS	SOIL-ROLLOFF-WC-11	PD090685.D	18 Oct 2025 00:24		AR\AJ	Ok,M
41	Q3343-04MSD	SOIL-ROLLOFF-WC-11	PD090686.D	18 Oct 2025 00:37		AR\AJ	Ok
42	I.BLK	I.BLK	PD090687.D	18 Oct 2025 00:51		AR\AJ	Ok
43	PSTDCCC050	PSTDCCC050	PD090688.D	18 Oct 2025 02:13		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD102025

Review By	Abdul	Review On	10/21/2025 8:53:08 AM
Supervise By	mohammad	Supervise On	10/24/2025 1:05:58 AM
SubDirectory	PD102025	HP Acquire Method	HP Processing Method pd101725 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24942,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD090689.D	20 Oct 2025 09:51		AR\AJ	Ok
2	I.BLK	I.BLK	PD090690.D	20 Oct 2025 10:04		AR\AJ	Ok
3	PEM	PEM	PD090691.D	20 Oct 2025 10:18		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD090692.D	20 Oct 2025 10:54		AR\AJ	Ok,M
5	PB170163BL	PB170163BL	PD090693.D	20 Oct 2025 13:25		AR\AJ	Ok
6	PB170163BS	PB170163BS	PD090694.D	20 Oct 2025 14:12		AR\AJ	Ok,M
7	Q3390-03	340	PD090695.D	20 Oct 2025 14:33		AR\AJ	Ok
8	Q3386-01	237	PD090696.D	20 Oct 2025 14:47		AR\AJ	Ok,M
9	Q3381-01	VNJ-222	PD090697.D	20 Oct 2025 15:00		AR\AJ	Ok,M
10	Q3381-03	VNJ-245	PD090698.D	20 Oct 2025 15:14		AR\AJ	Ok,M
11	Q3381-05	RBR200059	PD090699.D	20 Oct 2025 15:28		AR\AJ	Ok,M
12	Q3381-07	VNJ-255	PD090700.D	20 Oct 2025 15:41		AR\AJ	Ok,M
13	Q3382-01	CENTRAL-AVE-TP	PD090701.D	20 Oct 2025 15:55		AR\AJ	Ok,M
14	Q3383-01	OR-03-101725	PD090702.D	20 Oct 2025 16:09		AR\AJ	Ok,M
15	I.BLK	I.BLK	PD090703.D	20 Oct 2025 16:22		AR\AJ	Ok
16	PSTDCCC050	PSTDCCC050	PD090704.D	20 Oct 2025 16:52		AR\AJ	Ok
17	Q3374-01	TP-4-WC	PD090705.D	20 Oct 2025 17:20		AR\AJ	Ok,M
18	Q3374-05	TP-5-WC	PD090706.D	20 Oct 2025 17:33		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD102025

Review By	Abdul	Review On	10/21/2025 8:53:08 AM
Supervise By	mohammad	Supervise On	10/24/2025 1:05:58 AM
SubDirectory	PD102025	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3374-05MS	TP-5-WCMS	PD090707.D	20 Oct 2025 17:47		AR\AJ	Ok,M
20	Q3374-05MSD	TP-5-WCMSD	PD090708.D	20 Oct 2025 18:01		AR\AJ	Ok,M
21	Q3392-01	MH-20	PD090709.D	20 Oct 2025 18:14		AR\AJ	Ok
22	Q3392-05	MH-19	PD090710.D	20 Oct 2025 18:28		AR\AJ	Ok
23	I.BLK	I.BLK	PD090711.D	20 Oct 2025 18:42		AR\AJ	Ok
24	PEM	PEM	PD090712.D	20 Oct 2025 18:55		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PD090713.D	20 Oct 2025 19:09		AR\AJ	Ok
26	Q3375-01	PR132-S02-074099-20	PD090714.D	20 Oct 2025 19:50		AR\AJ	Ok,M
27	Q3375-02	PR132-S04-000013-20	PD090715.D	20 Oct 2025 20:04		AR\AJ	Ok,M
28	Q3375-03	PR132-S05-000057-20	PD090716.D	20 Oct 2025 20:17		AR\AJ	Ok,M
29	Q3375-04	PR132-S02-010074-20	PD090717.D	20 Oct 2025 20:31		AR\AJ	Ok,M
30	Q3375-05	PR132-DP02-20251016	PD090718.D	20 Oct 2025 20:44		AR\AJ	Ok,M
31	I.BLK	I.BLK	PD090719.D	20 Oct 2025 20:58		AR\AJ	Ok
32	PSTDCCC050	PSTDCCC050	PD090720.D	20 Oct 2025 21:12		AR\AJ	Ok,M

M : Manual Integration