

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3375

Client ID : SSTDCCC040

Date Analyzed: 10/20/2025

Lab File ID: BG064547.D

Time Analyzed: 12:44

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	12980	7.817	58801	10.61	45965	14.44
UPPER LIMIT	25960	8.317	117602	11.107	91930	14.944
LOWER LIMIT	6490	7.317	29400.5	10.107	22982.5	13.944
EPA SAMPLE NO.						
01 PB170165BL	14412	7.81	55177	10.60	43937	14.45
02 PB170165BS	13636	7.82	56148	10.61	44549	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.