

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3375	Analytical Method:	SW8260D
Client:	AECOM	Datafile :	VW032381.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW1020SBSD01	Dichlorodifluoromethane	20	21.7	ug/Kg	109	9		64	136	20
	Chloromethane	20	20.0	ug/Kg	100	1		52	151	20
	Vinyl chloride	20	21.4	ug/Kg	107	4		56	148	20
	Bromomethane	20	21.7	ug/Kg	109	3		58	141	20
	Chloroethane	20	21.6	ug/Kg	108	1		69	130	20
	Trichlorofluoromethane	20	21.6	ug/Kg	108	11		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	20.6	ug/Kg	103	1		81	123	20
	1,1-Dichloroethene	20	21.6	ug/Kg	108	1		79	121	20
	Acetone	100	110	ug/Kg	110	13		40	171	20
	Carbon disulfide	20	21.6	ug/Kg	108	2		59	130	20
	Methyl tert-butyl Ether	20	21.3	ug/Kg	106	11		77	129	20
	Methyl Acetate	20	20.9	ug/Kg	104	17		69	149	20
	Methylene Chloride	20	19.5	ug/Kg	98	3		72	131	20
	trans-1,2-Dichloroethene	20	21.5	ug/Kg	108	3		80	123	20
	1,1-Dichloroethane	20	21.5	ug/Kg	108	4		82	123	20
	Cyclohexane	20	19.7	ug/Kg	99	3		76	122	20
	2-Butanone	100	110	ug/Kg	110	22	*	69	131	20
	Carbon Tetrachloride	20	22.3	ug/Kg	112	2		76	129	20
	cis-1,2-Dichloroethene	20	21.2	ug/Kg	106	4		82	123	20
	Bromochloromethane	20	21.5	ug/Kg	108	4		80	127	20
	Chloroform	20	22.5	ug/Kg	113	5		82	125	20
	1,1,1-Trichloroethane	20	22.4	ug/Kg	112	4		80	126	20
	Methylcyclohexane	20	19.4	ug/Kg	97	0		77	123	20
	Benzene	20	21.8	ug/Kg	109	5		84	121	20
	1,2-Dichloroethane	20	22.0	ug/Kg	110	7		81	126	20
	Trichloroethene	20	21.9	ug/Kg	110	8		83	122	20
	1,2-Dichloropropane	20	22.2	ug/Kg	111	6		83	122	20
	Bromodichloromethane	20	23.0	ug/Kg	115	4		82	123	20
	4-Methyl-2-Pentanone	100	110	ug/Kg	110	18		70	135	20
	Toluene	20	22.5	ug/Kg	113	8		83	122	20
	t-1,3-Dichloropropene	20	22.6	ug/Kg	113	11		78	124	20
	cis-1,3-Dichloropropene	20	22.1	ug/Kg	111	7		81	122	20
	1,1,2-Trichloroethane	20	22.6	ug/Kg	113	8		82	125	20
	2-Hexanone	100	110	ug/Kg	110	17		66	138	20
	Dibromochloromethane	20	23.9	ug/Kg	119	9		79	125	20
	1,2-Dibromoethane	20	22.8	ug/Kg	114	11		80	125	20
	Tetrachloroethene	20	21.7	ug/Kg	109	8		83	125	20
	Chlorobenzene	20	21.3	ug/Kg	106	5		84	122	20
	Ethyl Benzene	20	20.6	ug/Kg	103	0		82	124	20
	m/p-Xylenes	40	42.2	ug/Kg	106	3		83	124	20
	o-Xylene	20	20.7	ug/Kg	104	5		83	123	20
	Styrene	20	21.8	ug/Kg	109	3		82	124	20
	Bromoform	20	23.5	ug/Kg	117	8		75	127	20
	Isopropylbenzene	20	19.8	ug/Kg	99	0		82	124	20
	1,1,2,2-Tetrachloroethane	20	21.8	ug/Kg	109	14		77	127	20
	1,3-Dichlorobenzene	20	21.5	ug/Kg	108	6		83	122	20
	1,4-Dichlorobenzene	20	21.0	ug/Kg	105	3		84	121	20
	1,2-Dichlorobenzene	20	21.3	ug/Kg	106	6		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3375</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>AECOM</u>	Datafile :	<u>VW032381.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW1020SBSD01	1,2-Dibromo-3-Chloropropane	20	23.0	ug/Kg	115	22	*	66	134	20
	1,2,4-Trichlorobenzene	20	20.3	ug/Kg	102	6		78	127	20
	1,2,3-Trichlorobenzene	20	20.9	ug/Kg	104	9		70	137	20