

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3375
 Lab File ID: VW032378.D Date Analyzed: 10/20/2025
 Instrument ID: MSVOA_W Time Analyzed: 10:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	234766	13.556				
UPPER LIMIT	469532	14.056				
LOWER LIMIT	117383	13.056				
EPA SAMPLE NO.						
PR132-S02-074099-20251016	230143	13.56				
PR132-S04-000013-20251016	224932	13.56				
PR132-S05-000057-20251016	221559	13.56				
PR132-S02-010074-20251016	201731	13.56				
PR132-DP02-20251016	228053	13.56				
VW1020SBL01	188560	13.56				
VW1020SBS01	240746	13.56				
VW1020SBSD01	239754	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.