



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

AECOM

Project Name: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

Project # N/A

Order ID # Q3376

Test Name: Hexavalent Chromium,TOC

A. Number of Samples and Date of Receipt:

5 Solid samples were received on 10/16/2025.

1 Water sample was received on 10/16/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Hexavalent Chromium,TOC. This data package contains results for Hexavalent Chromium,TOC.

C. Analytical Techniques:

The analysis of Hexavalent Chromium was based on method 7196A, The analysis of TOC was based on method Lloyd Kahn.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike analysis met criteria for all compounds.

The Matrix Spike Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

Lab has used least representable sample weight for the sample Q3376-01, Q3376-02, Q3376-03, Q3376-04 and Q3376-05 for TOC analysis. Therefore Lab has reported the TOC result with "OR" qualifier.

As per method, aqueous sample for Hexavalent Chromium analysis should be filtered within 15 minutes of collection time. However, samples were not filtered as per requirement therefore Sample was filtered in the Lab within the same day of sample receipt and analyzed within 24 hours from collection.



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____