



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3376

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137578

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.504	0.5	101	90-110	10/17/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	10/17/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.501	0.5	100	90-110	10/17/2025

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3376

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137593

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TOC	mg/L	968	1000	97	90-110	07/22/2025
Sample ID: CCV1 TOC	mg/L	984	1000	98	90-110	10/21/2025
Sample ID: CCV2 TOC	mg/L	1070	1000	107	90-110	10/21/2025
Sample ID: CCV3 TOC	mg/L	1080	1000	108	90-110	10/21/2025
Sample ID: CCV4 TOC	mg/L	1070	1000	107	90-110	10/22/2025
Sample ID: CCV5 TOC	mg/L	985	1000	98	90-110	10/22/2025

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3376

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137601

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.495	0.5	99	90-110	10/21/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.496	0.5	99	90-110	10/21/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.491	0.5	98	90-110	10/21/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.496	0.5	99	90-110	10/21/2025
Sample ID: CCV4 Hexavalent Chromium	mg/L	0.489	0.5	98	90-110	10/21/2025