

LB137583


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

42862



A= 0.0000e+000

B= 2.3620e-004

C= 2.5826e-002

Rho= 0.9989390

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	-68	0.000	-68	0			
0.2	0.200	0.200	0.000	738	0.0 %	738	0			
2.5	2.500	2.505	0.005	10496	0.0 %	10496	0			
5.0	5.000	5.179	0.179	21819	0.0 %	21819				
7.5	7.500	7.156	-0.344	30186	0.0 %	30186				
10.0	10.000	10.150	0.150	42862	0.0 %	42862				

lb137583 instrument id : cv1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Oct 2025 11:28:11

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	-68	PPB	Std	0.0000	7470A	20 Oct 2025 11:41:18
S	0.2 - 1	52	-	738	PPB	Std	0.2000	7470A	20 Oct 2025 11:43:35
S	2.5 - 1	52.5	-	10496	PPB	Std	2.5000	7470A	20 Oct 2025 11:45:52
S	5.0 - 1	55	-	21819	PPB	Std	5.0000	7470A	20 Oct 2025 11:48:09
S	7.5 - 1	57.5	-	30186	PPB	Std	7.5000	7470A	20 Oct 2025 11:50:28
S	10.0 - 1	510	-	42862	PPB	Std	10.0000	7470A	20 Oct 2025 11:55:33
U	ICV78 - 1	ICV78	4.1981	17664	PPB	SMPL	-7470A		20 Oct 2025 12:01:15
U	ICB78 - 1	ICB78	-0.1702	-830	PPB	SMPL	-7470A		20 Oct 2025 12:03:31
U	CCV54 - 1	CCV54	5.1204	21569	PPB	SMPL	-7470A		20 Oct 2025 12:06:04
U	CCB54 - 1	CCB54	-0.0169	-181	PPB	SMPL	-7470A		20 Oct 2025 12:08:29
U	CRA - 1	CRA	0.2169	809	PPB	SMPL	-7470A		20 Oct 2025 12:10:48
U	HighStd - 1	HighStd	10.0001	42228	PPB	SMPL	-7470A		20 Oct 2025 12:13:11
U	ChkStd - 1	ChkStd	6.3414	26738	PPB	SMPL	-7470A		20 Oct 2025 12:18:13
U	PB170157BL - 1	PBW	-0.0566	-349	PPB	SMPL	-7470A		20 Oct 2025 12:20:30
U	PB170157BS - 1	LCSW	3.2653	13715	PPB	SMPL	-7470A		20 Oct 2025 12:22:53
U	Q3376-07 - 1	PR132-FB01-20251016	-0.0679	-397	PPB	SMPL	-7470A		20 Oct 2025 12:25:09
U	Q3390-05 - 1	4174	0.1281	433	PPB	SMPL	-7470A		20 Oct 2025 12:27:28
U	Q3390-06 - 1	4160	0.0537	118	PPB	SMPL	-7470A		20 Oct 2025 12:29:45
U	Q3390-07 - 1	4090	0.0483	95	PPB	SMPL	-7470A		20 Oct 2025 12:32:02
U	Q3390-07DUP - 1	4090DUP	0.0395	58	PPB	SMPL	-7470A		20 Oct 2025 12:34:18
U	CCV55 - 1	CCV55	5.2397	22074	PPB	SMPL	-7470A		20 Oct 2025 12:41:42
U	CCB55 - 1	CCB55	-0.0415	-285	PPB	SMPL	-7470A		20 Oct 2025 12:43:56
U	Q3390-07MS - 1	4090MS	2.9684	12458	PPB	SMPL	-7470A		20 Oct 2025 12:46:17
U	Q3390-07MSD - 1	4090MSD	2.7362	11475	PPB	SMPL	-7470A		20 Oct 2025 12:48:32
U	Q3390-071x5 - 1		-0.0051	-131	PPB	SMPL	-7470A		20 Oct 2025 12:50:50
U	Q3390-07a - 1		3.9817	16748	PPB	SMPL	-7470A		20 Oct 2025 12:53:09
U	CCV56 - 1	CCV56	4.8601	20467	PPB	SMPL	-7470A		20 Oct 2025 12:58:15
U	CCB56 - 1	CCB56	-0.0500	-321	PPB	SMPL	-7470A		20 Oct 2025 13:00:31