

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3376

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV78	Mercury	4.20	4.0	105	90 - 110	CV	10/20/2025	12:01	LB137583

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3376

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV54	Mercury	5.12	5.0	102	90 - 110	CV	10/20/2025	12:06	LB137583
CCV55	Mercury	5.24	5.0	105	90 - 110	CV	10/20/2025	12:41	LB137583
CCV56	Mercury	4.86	5.0	97	90 - 110	CV	10/20/2025	12:58	LB137583

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3376

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7550	8000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Antimony	3880	4000	97	90 - 110	P	10/21/2025	12:40	LB137613
	Arsenic	3800	4000	95	90 - 110	P	10/21/2025	12:40	LB137613
	Barium	7500	8000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Beryllium	194	200	97	90 - 110	P	10/21/2025	12:40	LB137613
	Cadmium	1890	2000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Calcium	18800	20000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Chromium	784	800	98	90 - 110	P	10/21/2025	12:40	LB137613
	Cobalt	1900	2000	95	90 - 110	P	10/21/2025	12:40	LB137613
	Copper	976	1000	98	90 - 110	P	10/21/2025	12:40	LB137613
	Iron	3800	4000	95	90 - 110	P	10/21/2025	12:40	LB137613
	Lead	3700	4000	92	90 - 110	P	10/21/2025	12:40	LB137613
	Magnesium	18800	20000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Manganese	1860	2000	93	90 - 110	P	10/21/2025	12:40	LB137613
	Nickel	1910	2000	95	90 - 110	P	10/21/2025	12:40	LB137613
	Potassium	18900	20000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Selenium	3850	4000	96	90 - 110	P	10/21/2025	12:40	LB137613
	Silver	956	1000	96	90 - 110	P	10/21/2025	12:40	LB137613
	Sodium	18300	20000	92	90 - 110	P	10/21/2025	12:40	LB137613
	Thallium	3840	4000	96	90 - 110	P	10/21/2025	12:40	LB137613
	Vanadium	1880	2000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Zinc	1920	2000	96	90 - 110	P	10/21/2025	12:40	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3376

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	110	100	110	80 - 120	P	10/21/2025	12:44	LB137613
	Antimony	52.1	50.0	104	80 - 120	P	10/21/2025	12:44	LB137613
	Arsenic	20.1	20.0	100	80 - 120	P	10/21/2025	12:44	LB137613
	Barium	95.0	100	95	80 - 120	P	10/21/2025	12:44	LB137613
	Beryllium	6.21	6.0	104	80 - 120	P	10/21/2025	12:44	LB137613
	Cadmium	5.68	6.0	95	80 - 120	P	10/21/2025	12:44	LB137613
	Calcium	1990	2000	100	80 - 120	P	10/21/2025	12:44	LB137613
	Chromium	9.98	10.0	100	80 - 120	P	10/21/2025	12:44	LB137613
	Cobalt	28.8	30.0	96	80 - 120	P	10/21/2025	12:44	LB137613
	Copper	21.5	20.0	107	80 - 120	P	10/21/2025	12:44	LB137613
	Iron	109	100	109	80 - 120	P	10/21/2025	12:44	LB137613
	Lead	11.0	12.0	92	80 - 120	P	10/21/2025	12:44	LB137613
	Magnesium	2060	2000	103	80 - 120	P	10/21/2025	12:44	LB137613
	Manganese	19.7	20.0	98	80 - 120	P	10/21/2025	12:44	LB137613
	Nickel	38.9	40.0	97	80 - 120	P	10/21/2025	12:44	LB137613
	Potassium	1980	2000	99	80 - 120	P	10/21/2025	12:44	LB137613
	Selenium	23.1	20.0	115	80 - 120	P	10/21/2025	12:44	LB137613
	Silver	10.3	10.0	103	80 - 120	P	10/21/2025	12:44	LB137613
	Sodium	1790	2000	90	80 - 120	P	10/21/2025	12:44	LB137613
	Thallium	41.7	40.0	104	80 - 120	P	10/21/2025	12:44	LB137613
	Vanadium	34.8	40.0	87	80 - 120	P	10/21/2025	12:44	LB137613
	Zinc	41.7	40.0	104	80 - 120	P	10/21/2025	12:44	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3376

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9780	10000	98	90 - 110	P	10/21/2025	13:45	LB137613
	Antimony	4830	5000	97	90 - 110	P	10/21/2025	13:45	LB137613
	Arsenic	4820	5000	96	90 - 110	P	10/21/2025	13:45	LB137613
	Barium	9660	10000	97	90 - 110	P	10/21/2025	13:45	LB137613
	Beryllium	242	250	97	90 - 110	P	10/21/2025	13:45	LB137613
	Cadmium	2420	2500	97	90 - 110	P	10/21/2025	13:45	LB137613
	Calcium	24200	25000	97	90 - 110	P	10/21/2025	13:45	LB137613
	Chromium	976	1000	98	90 - 110	P	10/21/2025	13:45	LB137613
	Cobalt	2410	2500	96	90 - 110	P	10/21/2025	13:45	LB137613
	Copper	1220	1250	98	90 - 110	P	10/21/2025	13:45	LB137613
	Iron	4750	5000	95	90 - 110	P	10/21/2025	13:45	LB137613
	Lead	4800	5000	96	90 - 110	P	10/21/2025	13:45	LB137613
	Magnesium	24300	25000	97	90 - 110	P	10/21/2025	13:45	LB137613
	Manganese	2410	2500	96	90 - 110	P	10/21/2025	13:45	LB137613
	Nickel	2410	2500	97	90 - 110	P	10/21/2025	13:45	LB137613
	Potassium	23800	25000	95	90 - 110	P	10/21/2025	13:45	LB137613
	Selenium	4820	5000	96	90 - 110	P	10/21/2025	13:45	LB137613
	Silver	1200	1250	96	90 - 110	P	10/21/2025	13:45	LB137613
	Sodium	23900	25000	96	90 - 110	P	10/21/2025	13:45	LB137613
	Thallium	5230	5000	104	90 - 110	P	10/21/2025	13:45	LB137613
	Vanadium	2430	2500	97	90 - 110	P	10/21/2025	13:45	LB137613
	Zinc	2430	2500	97	90 - 110	P	10/21/2025	13:45	LB137613
CCV02	Aluminum	9530	10000	95	90 - 110	P	10/21/2025	15:03	LB137613
	Antimony	4980	5000	100	90 - 110	P	10/21/2025	15:03	LB137613
	Arsenic	4830	5000	97	90 - 110	P	10/21/2025	15:03	LB137613
	Barium	9800	10000	98	90 - 110	P	10/21/2025	15:03	LB137613
	Beryllium	227	250	91	90 - 110	P	10/21/2025	15:03	LB137613
	Cadmium	2310	2500	92	90 - 110	P	10/21/2025	15:03	LB137613
	Calcium	23700	25000	95	90 - 110	P	10/21/2025	15:03	LB137613
	Chromium	934	1000	93	90 - 110	P	10/21/2025	15:03	LB137613
	Cobalt	2340	2500	94	90 - 110	P	10/21/2025	15:03	LB137613
	Copper	1210	1250	96	90 - 110	P	10/21/2025	15:03	LB137613
	Iron	4910	5000	98	90 - 110	P	10/21/2025	15:03	LB137613
	Lead	4660	5000	93	90 - 110	P	10/21/2025	15:03	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3376

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	23200	25000	93	90 - 110	P	10/21/2025	15:03	LB137613
	Manganese	2380	2500	95	90 - 110	P	10/21/2025	15:03	LB137613
	Nickel	2330	2500	93	90 - 110	P	10/21/2025	15:03	LB137613
	Potassium	24700	25000	99	90 - 110	P	10/21/2025	15:03	LB137613
	Selenium	4920	5000	98	90 - 110	P	10/21/2025	15:03	LB137613
	Silver	1200	1250	96	90 - 110	P	10/21/2025	15:03	LB137613
	Sodium	24800	25000	99	90 - 110	P	10/21/2025	15:03	LB137613
	Thallium	4950	5000	99	90 - 110	P	10/21/2025	15:03	LB137613
	Vanadium	2390	2500	96	90 - 110	P	10/21/2025	15:03	LB137613
	Zinc	2370	2500	95	90 - 110	P	10/21/2025	15:03	LB137613
CCV03	Aluminum	9580	10000	96	90 - 110	P	10/21/2025	15:56	LB137613
	Antimony	4900	5000	98	90 - 110	P	10/21/2025	15:56	LB137613
	Arsenic	4820	5000	96	90 - 110	P	10/21/2025	15:56	LB137613
	Barium	9710	10000	97	90 - 110	P	10/21/2025	15:56	LB137613
	Beryllium	233	250	93	90 - 110	P	10/21/2025	15:56	LB137613
	Cadmium	2330	2500	93	90 - 110	P	10/21/2025	15:56	LB137613
	Calcium	23800	25000	95	90 - 110	P	10/21/2025	15:56	LB137613
	Chromium	941	1000	94	90 - 110	P	10/21/2025	15:56	LB137613
	Cobalt	2360	2500	94	90 - 110	P	10/21/2025	15:56	LB137613
	Copper	1210	1250	96	90 - 110	P	10/21/2025	15:56	LB137613
	Iron	4850	5000	97	90 - 110	P	10/21/2025	15:56	LB137613
	Lead	4680	5000	94	90 - 110	P	10/21/2025	15:56	LB137613
	Magnesium	23400	25000	94	90 - 110	P	10/21/2025	15:56	LB137613
	Manganese	2380	2500	95	90 - 110	P	10/21/2025	15:56	LB137613
	Nickel	2350	2500	94	90 - 110	P	10/21/2025	15:56	LB137613
	Potassium	25200	25000	101	90 - 110	P	10/21/2025	15:56	LB137613
	Selenium	4870	5000	97	90 - 110	P	10/21/2025	15:56	LB137613
	Silver	1200	1250	96	90 - 110	P	10/21/2025	15:56	LB137613
	Sodium	25300	25000	101	90 - 110	P	10/21/2025	15:56	LB137613
	Thallium	5060	5000	101	90 - 110	P	10/21/2025	15:56	LB137613
	Vanadium	2400	2500	96	90 - 110	P	10/21/2025	15:56	LB137613
	Zinc	2390	2500	96	90 - 110	P	10/21/2025	15:56	LB137613
CCV04	Aluminum	9280	10000	93	90 - 110	P	10/21/2025	16:48	LB137613
	Antimony	4870	5000	98	90 - 110	P	10/21/2025	16:48	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3376

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4750	5000	95	90 - 110	P	10/21/2025	16:48	LB137613
	Barium	9490	10000	95	90 - 110	P	10/21/2025	16:48	LB137613
	Beryllium	232	250	93	90 - 110	P	10/21/2025	16:48	LB137613
	Cadmium	2300	2500	92	90 - 110	P	10/21/2025	16:48	LB137613
	Calcium	23200	25000	93	90 - 110	P	10/21/2025	16:48	LB137613
	Chromium	934	1000	93	90 - 110	P	10/21/2025	16:48	LB137613
	Cobalt	2330	2500	93	90 - 110	P	10/21/2025	16:48	LB137613
	Copper	1190	1250	96	90 - 110	P	10/21/2025	16:48	LB137613
	Iron	4840	5000	97	90 - 110	P	10/21/2025	16:48	LB137613
	Lead	4640	5000	93	90 - 110	P	10/21/2025	16:48	LB137613
	Magnesium	22900	25000	92	90 - 110	P	10/21/2025	16:48	LB137613
	Manganese	2350	2500	94	90 - 110	P	10/21/2025	16:48	LB137613
	Nickel	2330	2500	93	90 - 110	P	10/21/2025	16:48	LB137613
	Potassium	24100	25000	96	90 - 110	P	10/21/2025	16:48	LB137613
	Selenium	4830	5000	97	90 - 110	P	10/21/2025	16:48	LB137613
	Silver	1210	1250	97	90 - 110	P	10/21/2025	16:48	LB137613
	Sodium	24000	25000	96	90 - 110	P	10/21/2025	16:48	LB137613
	Thallium	4900	5000	98	90 - 110	P	10/21/2025	16:48	LB137613
	Vanadium	2330	2500	93	90 - 110	P	10/21/2025	16:48	LB137613
	Zinc	2400	2500	96	90 - 110	P	10/21/2025	16:48	LB137613
CCV05	Aluminum	9540	10000	95	90 - 110	P	10/21/2025	17:37	LB137613
	Antimony	5030	5000	101	90 - 110	P	10/21/2025	17:37	LB137613
	Arsenic	4900	5000	98	90 - 110	P	10/21/2025	17:37	LB137613
	Barium	9470	10000	95	90 - 110	P	10/21/2025	17:37	LB137613
	Beryllium	234	250	94	90 - 110	P	10/21/2025	17:37	LB137613
	Cadmium	2370	2500	95	90 - 110	P	10/21/2025	17:37	LB137613
	Calcium	23800	25000	95	90 - 110	P	10/21/2025	17:37	LB137613
	Chromium	964	1000	96	90 - 110	P	10/21/2025	17:37	LB137613
	Cobalt	2390	2500	96	90 - 110	P	10/21/2025	17:37	LB137613
	Copper	1220	1250	98	90 - 110	P	10/21/2025	17:37	LB137613
	Iron	4840	5000	97	90 - 110	P	10/21/2025	17:37	LB137613
	Lead	4780	5000	96	90 - 110	P	10/21/2025	17:37	LB137613
	Magnesium	23600	25000	94	90 - 110	P	10/21/2025	17:37	LB137613
	Manganese	2360	2500	94	90 - 110	P	10/21/2025	17:37	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3376

Contract: AE002

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2390	2500	96	90 - 110	P	10/21/2025	17:37	LB137613
	Potassium	23700	25000	95	90 - 110	P	10/21/2025	17:37	LB137613
	Selenium	5030	5000	100	90 - 110	P	10/21/2025	17:37	LB137613
	Silver	1210	1250	97	90 - 110	P	10/21/2025	17:37	LB137613
	Sodium	23100	25000	92	90 - 110	P	10/21/2025	17:37	LB137613
	Thallium	5030	5000	100	90 - 110	P	10/21/2025	17:37	LB137613
	Vanadium	2390	2500	96	90 - 110	P	10/21/2025	17:37	LB137613
	Zinc	2430	2500	97	90 - 110	P	10/21/2025	17:37	LB137613
CCV06	Aluminum	9220	10000	92	90 - 110	P	10/21/2025	18:27	LB137613
	Antimony	4900	5000	98	90 - 110	P	10/21/2025	18:27	LB137613
	Arsenic	4750	5000	95	90 - 110	P	10/21/2025	18:27	LB137613
	Barium	9370	10000	94	90 - 110	P	10/21/2025	18:27	LB137613
	Beryllium	226	250	90	90 - 110	P	10/21/2025	18:27	LB137613
	Cadmium	2270	2500	91	90 - 110	P	10/21/2025	18:27	LB137613
	Calcium	23000	25000	92	90 - 110	P	10/21/2025	18:27	LB137613
	Chromium	905	1000	90	90 - 110	P	10/21/2025	18:27	LB137613
	Cobalt	2310	2500	92	90 - 110	P	10/21/2025	18:27	LB137613
	Copper	1180	1250	95	90 - 110	P	10/21/2025	18:27	LB137613
	Iron	4730	5000	95	90 - 110	P	10/21/2025	18:27	LB137613
	Lead	4610	5000	92	90 - 110	P	10/21/2025	18:27	LB137613
	Magnesium	22800	25000	91	90 - 110	P	10/21/2025	18:27	LB137613
	Manganese	2330	2500	93	90 - 110	P	10/21/2025	18:27	LB137613
	Nickel	2300	2500	92	90 - 110	P	10/21/2025	18:27	LB137613
	Potassium	24100	25000	96	90 - 110	P	10/21/2025	18:27	LB137613
	Selenium	4870	5000	97	90 - 110	P	10/21/2025	18:27	LB137613
	Silver	1170	1250	94	90 - 110	P	10/21/2025	18:27	LB137613
	Sodium	24400	25000	98	90 - 110	P	10/21/2025	18:27	LB137613
	Thallium	4850	5000	97	90 - 110	P	10/21/2025	18:27	LB137613
	Vanadium	2320	2500	93	90 - 110	P	10/21/2025	18:27	LB137613
	Zinc	2320	2500	93	90 - 110	P	10/21/2025	18:27	LB137613
CCV07	Aluminum	9650	10000	96	90 - 110	P	10/21/2025	19:18	LB137613
	Antimony	4900	5000	98	90 - 110	P	10/21/2025	19:18	LB137613
	Arsenic	4770	5000	95	90 - 110	P	10/21/2025	19:18	LB137613
	Barium	9640	10000	96	90 - 110	P	10/21/2025	19:18	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3376

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	236	250	94	90 - 110	P	10/21/2025	19:18	LB137613
	Cadmium	2320	2500	93	90 - 110	P	10/21/2025	19:18	LB137613
	Calcium	24100	25000	96	90 - 110	P	10/21/2025	19:18	LB137613
	Chromium	945	1000	94	90 - 110	P	10/21/2025	19:18	LB137613
	Cobalt	2350	2500	94	90 - 110	P	10/21/2025	19:18	LB137613
	Copper	1200	1250	96	90 - 110	P	10/21/2025	19:18	LB137613
	Iron	4820	5000	96	90 - 110	P	10/21/2025	19:18	LB137613
	Lead	4680	5000	94	90 - 110	P	10/21/2025	19:18	LB137613
	Magnesium	23800	25000	95	90 - 110	P	10/21/2025	19:18	LB137613
	Manganese	2410	2500	96	90 - 110	P	10/21/2025	19:18	LB137613
	Nickel	2350	2500	94	90 - 110	P	10/21/2025	19:18	LB137613
	Potassium	23500	25000	94	90 - 110	P	10/21/2025	19:18	LB137613
	Selenium	4850	5000	97	90 - 110	P	10/21/2025	19:18	LB137613
	Silver	1190	1250	95	90 - 110	P	10/21/2025	19:18	LB137613
	Sodium	22700	25000	91	90 - 110	P	10/21/2025	19:18	LB137613
	Thallium	4820	5000	96	90 - 110	P	10/21/2025	19:18	LB137613
	Vanadium	2430	2500	97	90 - 110	P	10/21/2025	19:18	LB137613
	Zinc	2370	2500	95	90 - 110	P	10/21/2025	19:18	LB137613
CCV08	Aluminum	9520	10000	95	90 - 110	P	10/21/2025	20:03	LB137613
	Antimony	5000	5000	100	90 - 110	P	10/21/2025	20:03	LB137613
	Arsenic	4850	5000	97	90 - 110	P	10/21/2025	20:03	LB137613
	Barium	9610	10000	96	90 - 110	P	10/21/2025	20:03	LB137613
	Beryllium	234	250	94	90 - 110	P	10/21/2025	20:03	LB137613
	Cadmium	2330	2500	93	90 - 110	P	10/21/2025	20:03	LB137613
	Calcium	23700	25000	95	90 - 110	P	10/21/2025	20:03	LB137613
	Chromium	915	1000	92	90 - 110	P	10/21/2025	20:03	LB137613
	Cobalt	2360	2500	94	90 - 110	P	10/21/2025	20:03	LB137613
	Copper	1210	1250	97	90 - 110	P	10/21/2025	20:03	LB137613
	Iron	4830	5000	97	90 - 110	P	10/21/2025	20:03	LB137613
	Lead	4730	5000	94	90 - 110	P	10/21/2025	20:03	LB137613
	Magnesium	23600	25000	94	90 - 110	P	10/21/2025	20:03	LB137613
	Manganese	2410	2500	96	90 - 110	P	10/21/2025	20:03	LB137613
	Nickel	2350	2500	94	90 - 110	P	10/21/2025	20:03	LB137613
	Potassium	23100	25000	93	90 - 110	P	10/21/2025	20:03	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3376

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4970	5000	100	90 - 110	P	10/21/2025	20:03	LB137613
	Silver	1190	1250	95	90 - 110	P	10/21/2025	20:03	LB137613
	Sodium	23300	25000	93	90 - 110	P	10/21/2025	20:03	LB137613
	Thallium	5050	5000	101	90 - 110	P	10/21/2025	20:03	LB137613
	Vanadium	2400	2500	96	90 - 110	P	10/21/2025	20:03	LB137613
	Zinc	2340	2500	94	90 - 110	P	10/21/2025	20:03	LB137613