

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: AECOM

SDG No.: Q3376

Contract: AECO02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170157BS Mercury	ug/L	4.0	3.27		82	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: AECOM

SDG No.: Q3376

Contract: AECO02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170192BS							
Aluminum	ug/L	1000	920		92	80 - 120	P
Antimony	ug/L	400	383		96	80 - 120	P
Arsenic	ug/L	400	367		92	80 - 120	P
Barium	ug/L	100	89.5		90	80 - 120	P
Beryllium	ug/L	100	93.8		94	80 - 120	P
Cadmium	ug/L	100	86.0		86	80 - 120	P
Calcium	ug/L	500	467	J	93	80 - 120	P
Chromium	ug/L	200	186		93	80 - 120	P
Cobalt	ug/L	100	89.0		89	80 - 120	P
Copper	ug/L	150	145		97	80 - 120	P
Iron	ug/L	1500	1380		92	80 - 120	P
Lead	ug/L	500	436		87	80 - 120	P
Magnesium	ug/L	1000	919	J	92	80 - 120	P
Manganese	ug/L	100	94.2		94	80 - 120	P
Nickel	ug/L	250	224		90	80 - 120	P
Potassium	ug/L	5000	4350		87	80 - 120	P
Selenium	ug/L	1000	947		95	80 - 120	P
Silver	ug/L	37.5	33.4		89	80 - 120	P
Sodium	ug/L	1500	1400		93	80 - 120	P
Thallium	ug/L	1000	898		90	80 - 120	P
Vanadium	ug/L	150	137		91	80 - 120	P
Zinc	ug/L	100	95.3		95	80 - 120	P