

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3376
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071044.D **Time Analyzed:** 16:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.480	4.380	4.580	534.530	500.000	6.9
Aroclor-1016-2	4.499	4.399	4.599	541.920	500.000	8.4
Aroclor-1016-3	4.556	4.456	4.656	553.860	500.000	10.8
Aroclor-1016-4	4.646	4.546	4.746	553.000	500.000	10.6
Aroclor-1016-5	4.931	4.831	5.031	544.190	500.000	8.8
Aroclor-1260-1	6.021	5.919	6.119	531.460	500.000	6.3
Aroclor-1260-2	6.276	6.175	6.375	531.970	500.000	6.4
Aroclor-1260-3	6.627	6.526	6.726	554.240	500.000	10.8
Aroclor-1260-4	6.841	6.740	6.940	566.600	500.000	13.3
Aroclor-1260-5	7.147	7.045	7.245	560.470	500.000	12.1
Decachlorobiphenyl	8.499	8.396	8.596	56.520	50.000	13.0
Tetrachloro-m-xylene	3.399	3.300	3.500	54.110	50.000	8.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3376
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071044.D **Time Analyzed:** 16:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	455.190	500.000	-9.0
Aroclor-1016-2	3.792	3.693	3.893	458.730	500.000	-8.3
Aroclor-1016-3	3.952	3.853	4.053	475.870	500.000	-4.8
Aroclor-1016-4	4.001	3.902	4.102	481.420	500.000	-3.7
Aroclor-1016-5	4.196	4.097	4.297	472.600	500.000	-5.5
Aroclor-1260-1	5.188	5.089	5.289	482.970	500.000	-3.4
Aroclor-1260-2	5.377	5.278	5.478	467.330	500.000	-6.5
Aroclor-1260-3	5.519	5.419	5.619	471.320	500.000	-5.7
Aroclor-1260-4	5.978	5.879	6.079	473.940	500.000	-5.2
Aroclor-1260-5	6.223	6.124	6.324	471.650	500.000	-5.7
Decachlorobiphenyl	7.533	7.433	7.633	55.480	50.000	11.0
Tetrachloro-m-xylene	2.777	2.678	2.878	50.160	50.000	0.3