

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3376
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071134.D **Time Analyzed:** 13:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.479	4.380	4.580	526.520	500.000	5.3
Aroclor-1016-2	4.498	4.399	4.599	518.670	500.000	3.7
Aroclor-1016-3	4.555	4.456	4.656	529.350	500.000	5.9
Aroclor-1016-4	4.645	4.546	4.746	529.330	500.000	5.9
Aroclor-1016-5	4.931	4.831	5.031	530.670	500.000	6.1
Aroclor-1260-1	6.020	5.919	6.119	517.720	500.000	3.5
Aroclor-1260-2	6.276	6.175	6.375	527.440	500.000	5.5
Aroclor-1260-3	6.626	6.526	6.726	541.970	500.000	8.4
Aroclor-1260-4	6.840	6.740	6.940	552.700	500.000	10.5
Aroclor-1260-5	7.147	7.045	7.245	544.530	500.000	8.9
Decachlorobiphenyl	8.496	8.396	8.596	55.010	50.000	10.0
Tetrachloro-m-xylene	3.399	3.300	3.500	51.670	50.000	3.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3376
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071134.D **Time Analyzed:** 13:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	426.650	500.000	-14.7
Aroclor-1016-2	3.791	3.693	3.893	423.200	500.000	-15.4
Aroclor-1016-3	3.951	3.853	4.053	435.820	500.000	-12.8
Aroclor-1016-4	4.000	3.902	4.102	441.110	500.000	-11.8
Aroclor-1016-5	4.194	4.097	4.297	436.050	500.000	-12.8
Aroclor-1260-1	5.187	5.089	5.289	440.070	500.000	-12.0
Aroclor-1260-2	5.376	5.278	5.478	434.740	500.000	-13.1
Aroclor-1260-3	5.517	5.419	5.619	435.540	500.000	-12.9
Aroclor-1260-4	5.977	5.879	6.079	443.990	500.000	-11.2
Aroclor-1260-5	6.221	6.124	6.324	440.770	500.000	-11.8
Decachlorobiphenyl	7.531	7.433	7.633	51.870	50.000	3.7
Tetrachloro-m-xylene	2.776	2.678	2.878	46.000	50.000	-8.0