

Surrogate Summary

SDG No.: Q3376

Client: AECOM

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		57	171
		Tetrachloro-m-xyl	1	20	19.4	97		61	148
		Decachlorobiphen	2	20	20.3	102		57	171
		Tetrachloro-m-xyl	2	20	20.0	100		61	148
I.BLK-PD090722.D	PIBLK-PD090722.D	Decachlorobiphen	1	20	21.2	106		57	171
		Tetrachloro-m-xyl	1	20	21.3	106		61	148
		Decachlorobiphen	2	20	22.1	111		57	171
		Tetrachloro-m-xyl	2	20	22.1	110		61	148
PB170179BL	PB170179BL	Decachlorobiphen	1	20	21.6	108		20	144
		Tetrachloro-m-xyl	1	20	20.3	101		19	148
		Decachlorobiphen	2	20	23.1	115		20	144
		Tetrachloro-m-xyl	2	20	21.2	106		19	148
PB170179BS	PB170179BS	Decachlorobiphen	1	20	20.3	101		20	144
		Tetrachloro-m-xyl	1	20	16.8	84		19	148
		Decachlorobiphen	2	20	21.9	110		20	144
		Tetrachloro-m-xyl	2	20	17.1	86		19	148
Q3402-07MS	41-LBMS	Decachlorobiphen	1	20	7.89	39		20	144
		Tetrachloro-m-xyl	1	20	13.4	67		19	148
		Decachlorobiphen	2	20	8.23	41		20	144
		Tetrachloro-m-xyl	2	20	14.0	70		19	148
Q3402-07MSD	41-LBMSD	Decachlorobiphen	1	20	8.21	41		20	144
		Tetrachloro-m-xyl	1	20	13.1	65		19	148
		Decachlorobiphen	2	20	8.28	41		20	144
		Tetrachloro-m-xyl	2	20	14.0	70		19	148
I.BLK-PD090732.D	PIBLK-PD090732.D	Decachlorobiphen	1	20	20.8	104		57	171
		Tetrachloro-m-xyl	1	20	20.4	102		61	148
		Decachlorobiphen	2	20	22.7	114		57	171
		Tetrachloro-m-xyl	2	20	21.1	106		61	148
Q3376-01	PR132-S05-057082-20251016	Decachlorobiphen	1	20	13.4	67		20	144
		Tetrachloro-m-xyl	1	20	12.7	63		19	148
		Decachlorobiphen	2	20	13.9	69		20	144
		Tetrachloro-m-xyl	2	20	13.6	68		19	148
Q3376-02	PR132-DP01-20251016	Decachlorobiphen	1	20	11.0	55		20	144
		Tetrachloro-m-xyl	1	20	11.6	58		19	148
		Decachlorobiphen	2	20	10.7	54		20	144
		Tetrachloro-m-xyl	2	20	12.5	62		19	148
Q3376-03	PR132-COMP2-20251016	Decachlorobiphen	1	20	11.2	56		20	144
		Tetrachloro-m-xyl	1	20	12.6	63		19	148
		Decachlorobiphen	2	20	10.6	53		20	144
		Tetrachloro-m-xyl	2	20	13.7	69		19	148
Q3376-04	PR132-S12-047072-20251016	Decachlorobiphen	1	20	17.6	88		20	144

Surrogate Summary

SDG No.: Q3376

Client: AECOM

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3376-04	PR132-S12-047072-20251016	Tetrachloro-m-xyl	1	20	17.6	88		19	148
		Decachlorobiphen	2	20	18.6	93		20	144
		Tetrachloro-m-xyl	2	20	18.8	94		19	148
Q3376-05	PR132-S12-020047-20251016	Decachlorobiphen	1	20	14.0	70		20	144
		Tetrachloro-m-xyl	1	20	16.7	84		19	148
		Decachlorobiphen	2	20	14.4	72		20	144
		Tetrachloro-m-xyl	2	20	17.5	88		19	148
I.BLK-PD090744.D	PIBLK-PD090744.D	Decachlorobiphen	1	20	21.0	105		57	171
		Tetrachloro-m-xyl	1	20	20.1	101		61	148
		Decachlorobiphen	2	20	22.9	115		57	171
		Tetrachloro-m-xyl	2	20	20.8	104		61	148