

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3376

Client ID : SSTDCCC040

Date Analyzed: 10/22/2025

Lab File ID: BF144032.D

Time Analyzed: 10:25

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	112960	6.881	428129	8.16	220935	9.92
UPPER LIMIT	225920	7.381	856258	8.663	441870	10.422
LOWER LIMIT	56480	6.381	214065	7.663	110468	9.422
EPA SAMPLE NO.						
01 PB170180BL	133448	6.88	515654	8.16	278111	9.92
02 PB170180BS	119718	6.88	459704	8.16	246572	9.92
03 PR132-COMP2-20251016	91108	6.88	322317	8.16	162319	9.92
04 PR132-S12-047072-20251016	70898	6.88	250555	8.16	129776	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.