

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3376

Client ID : SSTDCCC040

Date Analyzed: 10/28/2025

Lab File ID: BG064600.D

Time Analyzed: 13:45

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	28654	8.219	112694	11.05	87446	14.84
UPPER LIMIT	57308	8.719	225388	11.545	174892	15.341
LOWER LIMIT	14327	7.719	56347	10.545	43723	14.341
EPA SAMPLE NO.						
01 PR132-S12-020047-20251016	37243	8.22	140181	11.04	106965	14.84

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.