

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100125\
 Data File : VW032320.D
 Acq On : 01 Oct 2025 16:04
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ICVVW100125

Quant Time: Oct 02 21:27:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:24:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.314	0.295	6.1	111	0.00
3 P	Chloromethane	0.484	0.484	0.0	115	0.00
4 C	Vinyl Chloride	0.578	0.556	3.8#	104	0.00
5 T	Bromomethane	0.425	0.402	5.4	104	0.00
6 T	Chloroethane	0.382	0.370	3.1	103	0.00
7 T	Trichlorofluoromethane	0.433	0.432	0.2	102	0.00
8 T	Diethyl Ether	0.375	0.372	0.8	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.519	0.501	3.5	107	-0.01
10 T	Methyl Iodide	0.741	0.702	5.3	100	0.00
11 T	Tert butyl alcohol	0.055	0.054	1.8	110	0.00
12 CM	1,1-Dichloroethene	0.564	0.561	0.5#	107	0.00
13 T	Acrolein	0.090	0.084	6.7	103	0.00
14 T	Allyl chloride	0.952	0.948	0.4	103	0.00
15 T	Acrylonitrile	0.195	0.202	-3.6	107	0.00
16 T	Acetone	0.202	0.185	8.4	100	0.00
17 T	Carbon Disulfide	1.538	1.522	1.0	105	0.00
18 T	Methyl Acetate	0.578	0.625	-8.1	99	0.00
19 T	Methyl tert-butyl Ether	1.110	1.140	-2.7	106	0.00
20 T	Methylene Chloride	0.838	0.677	19.2	106	0.00
21 T	trans-1,2-Dichloroethene	0.624	0.631	-1.1	106	0.00
22 T	Diisopropyl ether	1.979	2.027	-2.4	106	0.00
23 T	Vinyl Acetate	1.322	1.394	-5.4	110	0.00
24 P	1,1-Dichloroethane	1.203	1.201	0.2	105	0.00
25 T	2-Butanone	0.275	0.279	-1.5	106	0.00
26 T	2,2-Dichloropropane	0.615	0.597	2.9	103	0.00
27 T	cis-1,2-Dichloroethene	0.749	0.764	-2.0	106	0.00
28 T	Bromochloromethane	0.566	0.572	-1.1	104	0.00
29 T	Tetrahydrofuran	0.173	0.185	-6.9	107	0.00
30 C	Chloroform	1.220	1.244	-2.0#	108	0.00
31 T	Cyclohexane	1.027	0.989	3.7	108	0.00
32 T	1,1,1-Trichloroethane	0.864	0.864	0.0	105	0.00
33 S	1,2-Dichloroethane-d4	0.729	0.739	-1.4	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.324	0.318	1.9	106	0.00
36 T	1,1-Dichloropropene	0.457	0.454	0.7	106	0.00
37 T	Ethyl Acetate	0.304	0.316	-3.9	111	0.00
38 T	Carbon Tetrachloride	0.430	0.427	0.7	108	0.00
39 T	Methylcyclohexane	0.551	0.568	-3.1	112	0.00
40 TM	Benzene	1.432	1.398	2.4	106	0.00
41 T	Methacrylonitrile	0.164	0.188	-14.6	117	0.00
42 TM	1,2-Dichloroethane	0.469	0.457	2.6	106	0.00
43 T	Isopropyl Acetate	0.540	0.560	-3.7	110	0.00
44 TM	Trichloroethene	0.335	0.328	2.1	108	0.00
45 C	1,2-Dichloropropane	0.358	0.353	1.4#	106	0.00
46 T	Dibromomethane	0.220	0.219	0.5	106	0.00
47 T	Bromodichloromethane	0.503	0.504	-0.2	108	0.00
48 T	Methyl methacrylate	0.262	0.277	-5.7	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100125\
 Data File : VW032320.D
 Acq On : 01 Oct 2025 16:04
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ICVVW100125

Quant Time: Oct 02 21:27:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:24:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	106	0.00
50 S	Toluene-d8	1.195	1.189	0.5	102	0.00
51 T	4-Methyl-2-Pentanone	0.313	0.324	-3.5	106	0.00
52 CM	Toluene	0.892	0.893	-0.1#	107	0.00
53 T	t-1,3-Dichloropropene	0.488	0.500	-2.5	107	0.00
54 T	cis-1,3-Dichloropropene	0.562	0.564	-0.4	106	0.00
55 T	1,1,2-Trichloroethane	0.294	0.293	0.3	107	0.00
56 T	Ethyl methacrylate	0.411	0.431	-4.9	108	0.00
57 T	1,3-Dichloropropane	0.519	0.513	1.2	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.219	0.243	-11.0	103	0.00
59 T	2-Hexanone	0.220	0.231	-5.0	108	0.00
60 T	Dibromochloromethane	0.334	0.335	-0.3	105	0.00
61 T	1,2-Dibromoethane	0.287	0.285	0.7	105	0.00
62 S	4-Bromofluorobenzene	0.452	0.436	3.5	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.299	0.297	0.7	109	0.00
65 PM	Chlorobenzene	1.100	1.075	2.3	107	0.00
66 T	1,1,1,2-Tetrachloroethane	0.343	0.333	2.9	107	0.00
67 C	Ethyl Benzene	1.834	1.853	-1.0#	109	0.00
68 T	m/p-Xylenes	0.707	0.702	0.7	107	0.00
69 T	o-Xylene	0.661	0.668	-1.1	107	0.00
70 T	Styrene	1.155	1.193	-3.3	108	0.00
71 P	Bromoform	0.196	0.201	-2.6	109	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.421	3.563	-4.2	108	0.00
74 T	N-amyl acetate	1.091	1.172	-7.4	111	0.00
75 P	1,1,2,2-Tetrachloroethane	0.849	0.863	-1.6	107	0.00
76 T	1,2,3-Trichloropropane	0.658	0.669	-1.7	111	0.00
77 T	Bromobenzene	0.825	0.835	-1.2	103	0.00
78 T	n-propylbenzene	4.273	4.415	-3.3	108	0.00
79 T	2-Chlorotoluene	2.613	2.680	-2.6	107	0.00
80 T	1,3,5-Trimethylbenzene	2.958	3.098	-4.7	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.260	0.282	-8.5	106	0.00
82 T	4-Chlorotoluene	2.778	2.819	-1.5	107	0.00
83 T	tert-Butylbenzene	2.427	2.597	-7.0	111	0.00
84 T	1,2,4-Trimethylbenzene	3.008	3.099	-3.0	105	0.00
85 T	sec-Butylbenzene	3.709	3.840	-3.5	108	0.00
86 T	p-Isopropyltoluene	3.077	3.243	-5.4	109	0.00
87 T	1,3-Dichlorobenzene	1.670	1.687	-1.0	106	0.00
88 T	1,4-Dichlorobenzene	1.670	1.615	3.3	105	0.00
89 T	n-Butylbenzene	3.005	3.178	-5.8	110	0.00
90 T	Hexachloroethane	0.552	0.573	-3.8	110	0.00
91 T	1,2-Dichlorobenzene	1.509	1.473	2.4	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.149	0.155	-4.0	106	0.00
93 T	1,2,4-Trichlorobenzene	0.905	0.926	-2.3	112	0.00
94 T	Hexachlorobutadiene	0.414	0.385	7.0	107	0.00
95 T	Naphthalene	2.174	2.429	-11.7	111	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100125\
Data File : VW032320.D
Acq On : 01 Oct 2025 16:04
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ICVWV100125

Quant Time: Oct 02 21:27:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 02 21:24:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.853	0.877	-2.8	109	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6