

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3376</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>10/21/2025 09:05</u>
Lab File ID:	<u>VY023521.D</u>	Init. Calib. Date(s):	<u>10/07/2025 10/07/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>09:17 12:32</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.360	0.323		-10.28	20
Chloromethane	0.490	0.450	0.1	-8.16	20
Vinyl Chloride	0.639	0.605		-5.32	20
Bromomethane	0.536	0.513		-4.29	20
Chloroethane	0.435	0.447		2.76	20
Trichlorofluoromethane	0.887	0.883		-0.45	20
1,1,2-Trichlorotrifluoroethane	0.460	0.490		6.52	20
1,1-Dichloroethene	0.446	0.455		2.02	20
Acetone	0.100	0.113		13	20
Carbon Disulfide	1.277	1.260		-1.33	20
Methyl tert-butyl Ether	1.072	1.116		4.1	20
Methyl Acetate	0.261	0.303		16.09	20
Methylene Chloride	0.538	0.509		-5.39	20
trans-1,2-Dichloroethene	0.491	0.520		5.91	20
1,1-Dichloroethane	0.796	0.840	0.1	5.53	20
Cyclohexane	0.696	0.643		-7.61	20
2-Butanone	0.122	0.124		1.64	20
Carbon Tetrachloride	0.493	0.562		14	20
cis-1,2-Dichloroethene	0.567	0.614		8.29	20
Bromochloromethane	0.297	0.269		-9.43	20
Chloroform	0.875	0.960		9.71	20
1,1,1-Trichloroethane	0.795	0.859		8.05	20
Methylcyclohexane	0.530	0.568		7.17	20
Benzene	1.288	1.448		12.42	20
1,2-Dichloroethane	0.332	0.360		8.43	20
Trichloroethene	0.377	0.421		11.67	20
1,2-Dichloropropane	0.285	0.320		12.28	20
Bromodichloromethane	0.453	0.522		15.23	20
4-Methyl-2-Pentanone	0.166	0.186		12.05	20
Toluene	0.836	0.967		15.67	20
t-1,3-Dichloropropene	0.394	0.441		11.93	20
cis-1,3-Dichloropropene	0.465	0.524		12.69	20
1,1,2-Trichloroethane	0.242	0.285		17.77	20
2-Hexanone	0.119	0.135		13.44	20
Dibromochloromethane	0.339	0.403		18.88	20
1,2-Dibromoethane	0.231	0.269		16.45	20
Tetrachloroethene	0.516	0.575		11.43	20
Chlorobenzene	1.087	1.181	0.3	8.65	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3376</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>10/21/2025</u> <u>09:05</u>
Lab File ID:	<u>VY023521.D</u>	Init. Calib. Date(s):	<u>10/07/2025</u> <u>10/07/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>09:17</u> <u>12:32</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.760	1.961		11.42	20
m/p-Xylenes	0.706	0.799		13.17	20
o-Xylene	0.662	0.739		11.63	20
Styrene	1.089	1.261		15.79	20
Bromoform	0.240	0.269	0.1	12.08	20
Isopropylbenzene	3.295	3.640		10.47	20
1,1,2,2-Tetrachloroethane	0.539	0.571	0.3	5.94	20
1,3-Dichlorobenzene	1.706	1.887		10.61	20
1,4-Dichlorobenzene	1.685	1.813		7.6	20
1,2-Dichlorobenzene	1.496	1.634		9.23	20
1,2-Dibromo-3-Chloropropane	0.088	0.088		0	20
1,2,4-Trichlorobenzene	0.932	0.982		5.36	20
1,2,3-Trichlorobenzene	0.808	0.843		4.33	20
1,2-Dichloroethane-d4	0.418	0.408		-2.39	20
Dibromofluoromethane	0.307	0.331		7.82	20
Toluene-d8	1.144	1.218		6.47	20
4-Bromofluorobenzene	0.378	0.405		7.14	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.