

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_W\Data\VW100125\  
 Data File : VW032320.D  
 Acq On : 01 Oct 2025 16:04  
 Operator : SY/MD  
 Sample : VSTDICV050  
 Misc : 5.00g/5mL/MSVOA\_W/SOIL  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 ICVVW100125

Quant Time: Oct 02 21:27:08 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_W\Method\82W100125S.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 02 21:24:16 2025  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025  
 Supervised By :Semsettin Yesilyurt 10/03/2025

| Compound                     | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                        |                |      |                     |         |        |          |
| Internal Standards           |                |      |                     |         |        |          |
| 1) Pentafluorobenzene        | 7.959          | 168  | 286648              | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 8.855          | 114  | 542900              | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 11.629         | 117  | 498213              | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.556         | 152  | 235528              | 50.000  | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |                     |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 8.319          | 65   | 211714              | 50.638  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 63 - 155 |      | Recovery = 101.280% |         |        |          |
| 35) Dibromofluoromethane     | 7.904          | 113  | 172657              | 49.108  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 70 - 134 |      | Recovery = 98.220%  |         |        |          |
| 50) Toluene-d8               | 10.325         | 98   | 645322              | 49.731  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 74 - 123 |      | Recovery = 99.460%  |         |        |          |
| 62) 4-Bromofluorobenzene     | 12.617         | 95   | 236491              | 48.218  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 17 - 146 |      | Recovery = 96.440%  |         |        |          |
| Target Compounds             |                |      |                     |         |        |          |
|                              |                |      |                     |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 2.046          | 85   | 84688               | 47.081  | ug/l   | 96       |
| 3) Chloromethane             | 2.259          | 50   | 138637              | 49.950  | ug/l   | 96       |
| 4) Vinyl Chloride            | 2.405          | 62   | 159289              | 48.072  | ug/l   | 97       |
| 5) Bromomethane              | 2.826          | 94   | 115275              | 47.340  | ug/l   | 97       |
| 6) Chloroethane              | 2.972          | 64   | 106181              | 48.536  | ug/l   | 97       |
| 7) Trichlorofluoromethane    | 3.314          | 101  | 123819              | 49.907  | ug/l   | 98       |
| 8) Diethyl Ether             | 3.722          | 74   | 106705              | 49.634  | ug/l   | 89       |
| 9) 1,1,2-Trichlorotrifluo... | 4.094          | 101  | 143716              | 48.296  | ug/l   | 93       |
| 10) Methyl Iodide            | 4.320          | 142  | 201132              | 47.371  | ug/l   | 100      |
| 11) Tert butyl alcohol       | 5.228          | 59   | 77891               | 246.126 | ug/l   | 97       |
| 12) 1,1-Dichloroethene       | 4.082          | 96   | 160883              | 49.800  | ug/l   | 95       |
| 13) Acrolein                 | 3.930          | 56   | 119840              | 233.328 | ug/l   | 99       |
| 14) Allyl chloride           | 4.704          | 41   | 271835              | 49.818  | ug/l   | 88       |
| 15) Acrylonitrile            | 5.405          | 53   | 288801              | 258.222 | ug/l   | 100      |
| 16) Acetone                  | 4.161          | 43   | 264871              | 228.796 | ug/l   | 100      |
| 17) Carbon Disulfide         | 4.429          | 76   | 436417              | 49.511  | ug/l   | 100      |
| 18) Methyl Acetate           | 4.710          | 43   | 179271              | 54.058  | ug/l   | 96       |
| 19) Methyl tert-butyl Ether  | 5.466          | 73   | 326683              | 51.342  | ug/l   | 95       |
| 20) Methylene Chloride       | 4.954          | 84   | 194169              | 50.746  | ug/l   | 97       |
| 21) trans-1,2-Dichloroethene | 5.460          | 96   | 180820              | 50.555  | ug/l   | 98       |
| 22) Diisopropyl ether        | 6.344          | 45   | 581042              | 51.202  | ug/l   | 95       |
| 23) Vinyl Acetate            | 6.283          | 43   | 1998017             | 263.598 | ug/l   | 100      |
| 24) 1,1-Dichloroethane       | 6.246          | 63   | 344370              | 49.949  | ug/l   | 99       |
| 25) 2-Butanone               | 7.197          | 43   | 399441              | 253.341 | ug/l   | 97       |
| 26) 2,2-Dichloropropane      | 7.191          | 77   | 171137              | 48.528  | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 7.191          | 96   | 218943              | 50.954  | ug/l   | 94       |
| 28) Bromochloromethane       | 7.533          | 49   | 164051              | 50.575  | ug/l   | 85       |
| 29) Tetrahydrofuran          | 7.551          | 42   | 265382              | 267.648 | ug/l   | 94       |
| 30) Chloroform               | 7.691          | 83   | 356459              | 50.980  | ug/l   | 98       |
| 31) Cyclohexane              | 7.971          | 56   | 283574              | 48.179  | ug/l   | 99       |
| 32) 1,1,1-Trichloroethane    | 7.892          | 97   | 247681              | 49.989  | ug/l   | 98       |
| 36) 1,1-Dichloropropene      | 8.100          | 75   | 246414              | 49.661  | ug/l   | 98       |
| 37) Ethyl Acetate            | 7.276          | 43   | 171315              | 51.871  | ug/l   | 100      |
| 38) Carbon Tetrachloride     | 8.081          | 117  | 231677              | 49.617  | ug/l   | 96       |
| 39) Methylcyclohexane        | 9.343          | 83   | 308455              | 51.537  | ug/l   | 96       |
| 40) Benzene                  | 8.337          | 78   | 758892              | 48.794  | ug/l   | 96       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_W\Data\VW100125\  
 Data File : VW032320.D  
 Acq On : 01 Oct 2025 16:04  
 Operator : SY/MD  
 Sample : VSTDICV050  
 Misc : 5.00g/5mL/MSVOA\_W/SOIL  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 ICVVW100125

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025  
 Supervised By :Semsettin Yesilyurt 10/03/2025

Quant Time: Oct 02 21:27:08 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_W\Method\82W100125S.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 02 21:24:16 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 41) Methacrylonitrile         | 7.502  | 41   | 101959   | 57.284   | ug/l # | 85       |
| 42) 1,2-Dichloroethane        | 8.416  | 62   | 247842   | 48.701   | ug/l   | 99       |
| 43) Isopropyl Acetate         | 8.435  | 43   | 303946   | 51.877   | ug/l   | 100      |
| 44) Trichloroethene           | 9.105  | 130  | 178095   | 48.904   | ug/l   | 87       |
| 45) 1,2-Dichloropropane       | 9.374  | 63   | 191384   | 49.299   | ug/l   | 97       |
| 46) Dibromomethane            | 9.465  | 93   | 118979   | 49.841   | ug/l   | 94       |
| 47) Bromodichloromethane      | 9.654  | 83   | 273884   | 50.162   | ug/l   | 99       |
| 48) Methyl methacrylate       | 9.441  | 41   | 150648   | 52.884   | ug/l # | 91       |
| 49) 1,4-Dioxane               | 9.465  | 88   | 34528    | 1058.229 | ug/l   | 98       |
| 51) 4-Methyl-2-Pentanone      | 10.215 | 43   | 878208   | 258.790  | ug/l   | 100      |
| 52) Toluene                   | 10.392 | 92   | 484871   | 50.061   | ug/l   | 95       |
| 53) t-1,3-Dichloropropene     | 10.611 | 75   | 271391   | 51.234   | ug/l   | 100      |
| 54) cis-1,3-Dichloropropene   | 10.075 | 75   | 306358   | 50.209   | ug/l   | 97       |
| 55) 1,1,2-Trichloroethane     | 10.788 | 97   | 159106   | 49.869   | ug/l   | 99       |
| 56) Ethyl methacrylate        | 10.648 | 69   | 233853   | 52.367   | ug/l   | 93       |
| 57) 1,3-Dichloropropane       | 10.934 | 76   | 278336   | 49.394   | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 9.928  | 63   | 658757   | 277.181  | ug/l   | 96       |
| 59) 2-Hexanone                | 10.971 | 43   | 627586   | 262.215  | ug/l   | 99       |
| 60) Dibromochloromethane      | 11.136 | 129  | 181813   | 50.187   | ug/l   | 98       |
| 61) 1,2-Dibromoethane         | 11.239 | 107  | 154714   | 49.732   | ug/l   | 99       |
| 64) Tetrachloroethene         | 10.867 | 164  | 147899   | 49.611   | ug/l   | 96       |
| 65) Chlorobenzene             | 11.660 | 112  | 535619   | 48.862   | ug/l   | 98       |
| 66) 1,1,1,2-Tetrachloroethane | 11.733 | 131  | 166082   | 48.644   | ug/l   | 99       |
| 67) Ethyl Benzene             | 11.733 | 91   | 922981   | 50.513   | ug/l   | 100      |
| 68) m/p-Xylenes               | 11.837 | 106  | 699490   | 99.276   | ug/l   | 99       |
| 69) o-Xylene                  | 12.166 | 106  | 332905   | 50.523   | ug/l   | 100      |
| 70) Styrene                   | 12.178 | 104  | 594393   | 51.667   | ug/l   | 97       |
| 71) Bromoform                 | 12.349 | 173  | 100214   | 51.207   | ug/l # | 97       |
| 73) Isopropylbenzene          | 12.465 | 105  | 839135   | 52.076   | ug/l   | 100      |
| 74) N-amyl acetate            | 12.269 | 43   | 276033   | 53.697   | ug/l   | 99       |
| 75) 1,1,2,2-Tetrachloroethane | 12.715 | 83   | 203339   | 50.822   | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 12.763 | 75   | 157491m  | 50.813   | ug/l   |          |
| 77) Bromobenzene              | 12.745 | 156  | 196776   | 50.659   | ug/l   | 87       |
| 78) n-propylbenzene           | 12.800 | 91   | 1039770  | 51.657   | ug/l   | 98       |
| 79) 2-Chlorotoluene           | 12.891 | 91   | 631196   | 51.287   | ug/l   | 97       |
| 80) 1,3,5-Trimethylbenzene    | 12.940 | 105  | 729628   | 52.365   | ug/l   | 97       |
| 81) trans-1,4-Dichloro-2-b... | 12.513 | 75   | 66508    | 54.360   | ug/l   | 91       |
| 82) 4-Chlorotoluene           | 12.989 | 91   | 663842   | 50.734   | ug/l   | 100      |
| 83) tert-Butylbenzene         | 13.202 | 119  | 611748   | 53.510   | ug/l   | 97       |
| 84) 1,2,4-Trimethylbenzene    | 13.245 | 105  | 729966   | 51.522   | ug/l   | 99       |
| 85) sec-Butylbenzene          | 13.379 | 105  | 904532   | 51.774   | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 13.495 | 119  | 763912   | 52.697   | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 13.501 | 146  | 397346   | 50.518   | ug/l   | 97       |
| 88) 1,4-Dichlorobenzene       | 13.574 | 146  | 380433   | 48.372   | ug/l   | 97       |
| 89) n-Butylbenzene            | 13.818 | 91   | 748452   | 52.880   | ug/l   | 99       |
| 90) Hexachloroethane          | 14.086 | 117  | 134862   | 51.903   | ug/l   | 94       |
| 91) 1,2-Dichlorobenzene       | 13.867 | 146  | 347039   | 48.810   | ug/l   | 97       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.476 | 75   | 36524    | 52.139   | ug/l   | 88       |
| 93) 1,2,4-Trichlorobenzene    | 15.129 | 180  | 218198   | 51.178   | ug/l   | 93       |
| 94) Hexachlorobutadiene       | 15.232 | 225  | 90599    | 46.478   | ug/l   | 87       |
| 95) Naphthalene               | 15.360 | 128  | 572147   | 55.881   | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 15.549 | 180  | 206623   | 51.417   | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_W\Data\VW100125\  
Data File : VW032320.D  
Acq On : 01 Oct 2025 16:04  
Operator : SY/MD  
Sample : VSTDICV050  
Misc : 5.00g/5mL/MSVOA\_W/SOIL  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125

Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025

Quant Time: Oct 02 21:27:08 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_W\Method\82W100125S.M  
Quant Title : SW846 8260  
QLast Update : Thu Oct 02 21:24:16 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|


(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 7.959 min Scan# 996  
Delta R.T. -0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125



Tgt Ion:168 Resp: 286648  
Ion Ratio Lower Upper  
168 100  
99 68.6 49.3 73.9




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#2  
Dichlorodifluoromethane  
Concen: 47.081 ug/l  
RT: 2.046 min Scan# 26  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion: 85 Resp: 84688  
Ion Ratio Lower Upper  
85 100  
87 31.8 17.2 51.6





#3  
Chloromethane  
Concen: 49.950 ug/l  
RT: 2.259 min Scan# 61  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVW100125



Tgt Ion: 50 Resp: 13863  
Ion Ratio Lower Upper  
50 100  
52 30.2 26.1 39.1




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#4  
Vinyl Chloride  
Concen: 48.072 ug/l  
RT: 2.405 min Scan# 85  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion: 62 Resp: 159289  
Ion Ratio Lower Upper  
62 100  
64 32.7 24.8 37.2





#5  
Bromomethane  
Concen: 47.340 ug/l  
RT: 2.826 min Scan# 115275  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125



Tgt Ion: 94 Resp: 115275  
Ion Ratio Lower Upper  
94 100  
96 92.5 76.6 115.0




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#6  
Chloroethane  
Concen: 48.536 ug/l  
RT: 2.972 min Scan# 178  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion: 64 Resp: 106181  
Ion Ratio Lower Upper  
64 100  
66 34.4 26.2 39.4





#7  
 Trichlorofluoromethane  
 Concen: 49.907 ug/l  
 RT: 3.314 min Scan# 233  
 Delta R.T. 0.006 min  
 Lab File: VW032320.D  
 Acq: 01 Oct 2025 16:04




Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 ICVVW100125




Tgt Ion:101 Resp: 123819  
 Ion Ratio Lower Upper  
 101 100  
 103 61.1 47.8 71.8

Manual Integrations  
 APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
 Supervised By :Semsettin Yesilyurt 10/03/2025




#8  
 Diethyl Ether  
 Concen: 49.634 ug/l  
 RT: 3.722 min Scan# 301  
 Delta R.T. 0.006 min  
 Lab File: VW032320.D  
 Acq: 01 Oct 2025 16:04



Tgt Ion: 74 Resp: 106705  
 Ion Ratio Lower Upper  
 74 100  
 45 96.8 43.1 129.4











#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 48.296 ug/l  
RT: 4.094 min Scan# 362  
Delta R.T. -0.012 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04

Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVW100125

Tgt Ion:101 Resp: 143710  
Ion Ratio Lower Upper  
101 100  
85 43.7 37.1 55.7  
151 69.6 61.5 92.3




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#10  
Methyl Iodide  
Concen: 47.371 ug/l  
RT: 4.320 min Scan# 399  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion:142 Resp: 201132  
Ion Ratio Lower Upper  
142 100  
127 47.4 38.2 57.2  
141 15.3 12.2 18.2





#11  
Tert butyl alcohol  
Concen: 246.126 ug/l  
RT: 5.228 min Scan# 547  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVW100125



Tgt Ion: 59 Resp: 7789  
Ion Ratio Lower Upper  
59 100  
57 11.0 9.8 14.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#12  
1,1-Dichloroethene  
Concen: 49.800 ug/l  
RT: 4.082 min Scan# 360  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion: 96 Resp: 160883  
Ion Ratio Lower Upper  
96 100  
61 162.9 123.0 184.4  
98 61.4 48.5 72.7





#13  
Acrolein  
Concen: 233.328 ug/l  
RT: 3.930 min Scan# 335  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVW100125



Tgt Ion: 56 Resp: 119840  
Ion Ratio Lower Upper  
56 100  
55 70.0 56.4 84.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#14  
Allyl chloride  
Concen: 49.818 ug/l  
RT: 4.704 min Scan# 462  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion: 41 Resp: 271835  
Ion Ratio Lower Upper  
41 100  
39 67.6 64.1 96.1  
76 34.3 31.4 47.2





#15  
Acrylonitrile  
Concen: 258.222 ug/l  
RT: 5.405 min Scan# 51  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125



Tgt Ion: 53 Resp: 28880  
Ion Ratio Lower Upper  
53 100  
52 84.6 67.4 101.2  
51 37.6 30.3 45.5




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#16  
Acetone  
Concen: 228.796 ug/l  
RT: 4.161 min Scan# 373  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion: 43 Resp: 264871  
Ion Ratio Lower Upper  
43 100  
58 31.8 25.3 37.9





#17  
Carbon Disulfide  
Concen: 49.511 ug/l  
RT: 4.429 min Scan# 416  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125



Tgt Ion: 76 Resp: 43641  
Ion Ratio Lower Upper  
76 100  
78 8.9 7.0 10.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#18  
Methyl Acetate  
Concen: 54.058 ug/l  
RT: 4.710 min Scan# 463  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04

Tgt Ion: 43 Resp: 179271  
Ion Ratio Lower Upper  
43 100  
74 23.4 20.2 30.4









#19  
Methyl tert-butyl Ether  
Concen: 51.342 ug/l  
RT: 5.466 min Scan# 51  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04

Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125

Tgt Ion: 73 Resp: 32668  
Ion Ratio Lower Upper  
73 100  
57 22.7 16.2 24.4




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#20  
Methylene Chloride  
Concen: 50.746 ug/l  
RT: 4.954 min Scan# 503  
Delta R.T. -0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion: 84 Resp: 194169  
Ion Ratio Lower Upper  
84 100  
49 121.2 92.2 138.4  
51 37.1 31.0 46.6  
86 63.0 50.5 75.7





#21  
trans-1,2-Dichloroethene  
Concen: 50.555 ug/l  
RT: 5.460 min Scan# 586  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument : MSVOA\_W  
ClientSampleId : ICVVW100125



Tgt Ion: 96 Resp: 180820  
Ion Ratio Lower Upper  
96 100  
61 144.0 112.5 168.7  
98 67.4 54.2 81.4




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#22  
Diisopropyl ether  
Concen: 51.202 ug/l  
RT: 6.344 min Scan# 731  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion: 45 Resp: 581042  
Ion Ratio Lower Upper  
45 100  
43 60.9 51.4 77.2  
87 28.2 25.9 38.9  
59 11.3 10.1 15.1





#23  
 Vinyl Acetate  
 Concen: 263.598 ug/l  
 RT: 6.283 min Scan# 721  
 Delta R.T. -0.000 min  
 Lab File: VW032320.D  
 Acq: 01 Oct 2025 16:04




Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 ICVVW100125



Tgt Ion: 43 Resp: 199801  
 Ion Ratio Lower Upper  
 43 100  
 86 10.9 8.8 13.2




Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025  
 Supervised By :Semsettin Yesilyurt 10/03/2025




#24  
 1,1-Dichloroethane  
 Concen: 49.949 ug/l  
 RT: 6.246 min Scan# 715  
 Delta R.T. -0.000 min  
 Lab File: VW032320.D  
 Acq: 01 Oct 2025 16:04

Tgt Ion: 63 Resp: 344370  
 Ion Ratio Lower Upper  
 63 100  
 98 7.2 3.8 11.2  
 100 4.7 2.5 7.5








#25  
2-Butanone  
Concen: 253.341 ug/l  
RT: 7.197 min Scan# 81  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125



Tgt Ion: 43 Resp: 39944  
Ion Ratio Lower Upper  
43 100  
72 25.3 21.4 32.2




Manual Integrations  
**APPROVED**


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#26  
2,2-Dichloropropane  
Concen: 48.528 ug/l  
RT: 7.191 min Scan# 870  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion: 77 Resp: 171137  
Ion Ratio Lower Upper  
77 100  
97 24.0 11.9 35.5





#27  
 cis-1,2-Dichloroethene  
 Concen: 50.954 ug/l  
 RT: 7.191 min Scan# 870  
 Delta R.T. -0.000 min  
 Lab File: VW032320.D  
 Acq: 01 Oct 2025 16:04




Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 ICVVW100125




Tgt Ion: 96 Resp: 21894  
 Ion Ratio Lower Upper  
 96 100  
 61 143.0 0.0 268.0  
 98 63.9 0.0 130.8

Manual Integrations  
 APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
 Supervised By :Semsettin Yesilyurt 10/03/2025




#28  
 Bromochloromethane  
 Concen: 50.575 ug/l  
 RT: 7.533 min Scan# 926  
 Delta R.T. -0.000 min  
 Lab File: VW032320.D  
 Acq: 01 Oct 2025 16:04




Tgt Ion: 49 Resp: 164051  
 Ion Ratio Lower Upper  
 49 100  
 129 1.8 0.0 4.2  
 130 70.8 67.5 101.3





#29  
Tetrahydrofuran  
Concen: 267.648 ug/l  
RT: 7.551 min Scan# 91  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125



Tgt Ion: 42 Resp: 26538  
Ion Ratio Lower Upper  
42 100  
72 42.9 37.4 56.2  
71 40.4 36.0 54.0




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#30  
Chloroform  
Concen: 50.980 ug/l  
RT: 7.691 min Scan# 952  
Delta R.T. -0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion: 83 Resp: 356459  
Ion Ratio Lower Upper  
83 100  
85 65.4 51.2 76.8





#31  
Cyclohexane  
Concen: 48.179 ug/l  
RT: 7.971 min Scan# 998  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument : MSVOA\_W  
Client Sample Id : ICVVW100125




Tgt Ion: 56 Resp: 283574  
Ion Ratio Lower Upper  
56 100  
69 35.2 27.7 41.5  
84 92.2 74.6 112.0

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#32  
1,1,1-Trichloroethane  
Concen: 49.989 ug/l  
RT: 7.892 min Scan# 985  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Tgt Ion: 97 Resp: 247681  
Ion Ratio Lower Upper  
97 100  
99 65.0 50.9 76.3  
61 46.2 36.2 54.4





#33  
 1,2-Dichloroethane-d4  
 Concen: 50.638 ug/l  
 RT: 8.319 min Scan# 1055  
 Delta R.T. -0.000 min  
 Lab File: VW032320.D  
 Acq: 01 Oct 2025 16:04




Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 ICVVW100125



Tgt Ion: 65 Resp: 211714  
 Ion Ratio Lower Upper  
 65 100  
 67 52.7 0.0 99.6




Manual Integrations  
 APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
 Supervised By :Semsettin Yesilyurt 10/03/2025



#34  
 1,4-Difluorobenzene  
 Concen: 50.000 ug/l  
 RT: 8.855 min Scan# 1143  
 Delta R.T. -0.000 min  
 Lab File: VW032320.D  
 Acq: 01 Oct 2025 16:04


Tgt Ion:114 Resp: 542900  
 Ion Ratio Lower Upper  
 114 100  
 63 20.8 0.0 40.4  
 88 15.9 0.0 32.0





#35  
Dibromofluoromethane  
Concen: 49.108 ug/l  
RT: 7.904 min Scan# 91  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion: 113 Resp: 17265  
Ion Ratio Lower Upper  
113 100  
111 102.1 83.9 125.9  
192 15.3 14.6 21.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#36  
1,1-Dichloropropene  
Concen: 49.661 ug/l  
RT: 8.100 min Scan# 1019  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Tgt Ion: 75 Resp: 246414  
Ion Ratio Lower Upper  
75 100  
110 33.2 17.5 52.5  
77 30.9 24.1 36.1





#37  
Ethyl Acetate  
Concen: 51.871 ug/l  
RT: 7.276 min Scan# 884  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVW100125



Tgt Ion: 43 Resp: 17131  
Ion Ratio Lower Upper  
43 100  
61 13.9 11.3 16.9  
70 11.0 8.9 13.3




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#38  
Carbon Tetrachloride  
Concen: 49.617 ug/l  
RT: 8.081 min Scan# 1016  
Delta R.T. -0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04

Tgt Ion:117 Resp: 231677  
Ion Ratio Lower Upper  
117 100  
119 96.2 74.2 111.2  
121 28.3 25.6 38.4





#39

Methylcyclohexane

Concen: 51.537 ug/l

RT: 9.343 min Scan# 1223

Delta R.T. -0.000 min

Lab File: VW032320.D

Acq: 01 Oct 2025 16:04

Instrument :

MSVOA\_W

ClientSampleId :

ICVW100125

Tgt Ion: 83 Resp: 30845

Ion Ratio Lower Upper

83 100

55 75.8 58.4 87.6





98 45.4 39.4 59.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025

Supervised By :Semsettin Yesilyurt 10/03/2025



#40

Benzene

Concen: 48.794 ug/l

RT: 8.337 min Scan# 1058

Delta R.T. -0.000 min

Lab File: VW032320.D




Acq: 01 Oct 2025 16:04

Tgt Ion: 78 Resp: 758892


Ion Ratio Lower Upper

78 100

77 24.0 20.7 31.1








#41  
Methacrylonitrile  
Concen: 57.284 ug/l  
RT: 7.502 min Scan# 91  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125



Tgt Ion: 41 Resp: 101959  
Ion Ratio Lower Upper  
41 100  
39 52.3 50.9 76.3  
67 68.2 65.6 98.4  
52 28.7 29.1 43.7#




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#42  
1,2-Dichloroethane  
Concen: 48.701 ug/l  
RT: 8.416 min Scan# 1071  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04

Tgt Ion: 62 Resp: 247842  
Ion Ratio Lower Upper  
62 100  
98 9.4 0.0 18.0





#43

Isopropyl Acetate

Concen: 51.877 ug/l

RT: 8.435 min Scan# 1074

Delta R.T. -0.000 min

Lab File: VW032320.D

Acq: 01 Oct 2025 16:04

Instrument :

MSVOA\_W

ClientSampleId :

ICVWV100125





Tgt Ion: 43 Resp: 303940

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 43  | 100   |       |       |
| 61  | 29.1  | 23.3  | 34.9  |
| 87  | 13.4  | 11.1  | 16.7  |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#44

Trichloroethene

Concen: 48.904 ug/l

RT: 9.105 min Scan# 1184




Delta R.T. 0.006 min


Lab File: VW032320.D

Acq: 01 Oct 2025 16:04

Tgt Ion:130 Resp: 178095

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 130 | 100   |       |       |
| 95  | 96.2  | 0.0   | 220.0 |





#45

1,2-Dichloropropane

Concen: 49.299 ug/l

RT: 9.374 min Scan# 1228

Delta R.T. -0.000 min

Lab File: VW032320.D

Acq: 01 Oct 2025 16:04

Instrument :

MSVOA\_W

ClientSampleId :

ICVW100125

Tgt Ion: 63 Resp: 191384

Ion Ratio Lower Upper





63 100

65 31.9 24.3 36.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#46

Dibromomethane

Concen: 49.841 ug/l

RT: 9.465 min Scan# 1243

Delta R.T. -0.000 min

Lab File: VW032320.D

Acq: 01 Oct 2025 16:04




Tgt Ion: 93 Resp: 118979


Ion Ratio Lower Upper

93 100

95 83.5 67.4 101.2

174 84.9 75.7 113.5





#47

Bromodichloromethane

Concen: 50.162 ug/l

RT: 9.654 min Scan# 1274

Delta R.T. 0.006 min

Lab File: VW032320.D

Acq: 01 Oct 2025 16:04

Instrument :

MSVOA\_W

ClientSampleId :

ICVW100125





Tgt Ion: 83 Resp: 273884

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 85  | 62.0  | 48.7  | 73.1  |
| 127 | 7.6   | 6.4   | 9.6   |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#48

Methyl methacrylate

Concen: 52.884 ug/l

RT: 9.441 min Scan# 1239




Delta R.T. -0.000 min


Lab File: VW032320.D

Acq: 01 Oct 2025 16:04

Tgt Ion: 41 Resp: 150648


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 41  | 100   |       |       |
| 69  | 87.7  | 71.5  | 107.3 |
| 39  | 55.2  | 55.9  | 83.9# |





#49  
1,4-Dioxane  
Concen: 1058.229 ug/l  
RT: 9.465 min Scan# 1243  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVVW100125



Tgt Ion: 88 Resp: 34528  
Ion Ratio Lower Upper  
88 100  
43 25.4 18.0 27.0  
58 70.1 55.4 83.0




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#50  
Toluene-d8  
Concen: 49.731 ug/l  
RT: 10.325 min Scan# 1384  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion: 98 Resp: 645322  
Ion Ratio Lower Upper  
98 100  
100 66.3 51.4 77.2





#51  
4-Methyl-2-Pentanone  
Concen: 258.790 ug/l  
RT: 10.215 min Scan# 1366  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion: 43 Resp: 878208  
Ion Ratio Lower Upper  
43 100  
58 39.9 32.0 48.0

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#52  
Toluene  
Concen: 50.061 ug/l  
RT: 10.392 min Scan# 1395  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Tgt Ion: 92 Resp: 484871  
Ion Ratio Lower Upper  
92 100  
91 165.7 137.5 206.3





#53  
t-1,3-Dichloropropene  
Concen: 51.234 ug/l  
RT: 10.611 min Scan# 1431  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument : MSVOA\_W  
ClientSampleId : ICVVW100125



Tgt Ion: 75 Resp: 27139  
Ion Ratio Lower Upper  
75 100  
77 32.1 25.6 38.4




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#54  
cis-1,3-Dichloropropene  
Concen: 50.209 ug/l  
RT: 10.075 min Scan# 1343  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion: 75 Resp: 306358  
Ion Ratio Lower Upper  
75 100  
77 31.5 25.1 37.7  
39 50.4 43.2 64.8





#55  
1,1,2-Trichloroethane  
Concen: 49.869 ug/l  
RT: 10.788 min Scan# 1460  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion: 97 Resp: 159100  
Ion Ratio Lower Upper  
97 100  
83 88.9 71.0 106.6  
85 54.8 42.8 64.2  
99 63.5 51.0 76.6

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#56  
Ethyl methacrylate  
Concen: 52.367 ug/l  
RT: 10.648 min Scan# 1437  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04



Tgt Ion: 69 Resp: 233853  
Ion Ratio Lower Upper  
69 100  
41 70.4 53.2 79.8  
39 35.0 32.9 49.3








#57  
1,3-Dichloropropane  
Concen: 49.394 ug/l  
RT: 10.934 min Scan# 1484  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion: 76 Resp: 278330  
Ion Ratio Lower Upper  
76 100  
78 33.3 26.2 39.4

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#58  
2-Chloroethyl Vinyl ether  
Concen: 277.181 ug/l  
RT: 9.928 min Scan# 1319  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Tgt Ion: 63 Resp: 658757  
Ion Ratio Lower Upper  
63 100  
106 28.4 24.3 36.5





#59  
2-Hexanone  
Concen: 262.215 ug/l  
RT: 10.971 min Scan# 1490  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125



Tgt Ion: 43 Resp: 627580  
Ion Ratio Lower Upper  
43 100  
58 54.0 27.5 82.4




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#60  
Dibromochloromethane  
Concen: 50.187 ug/l  
RT: 11.136 min Scan# 1517  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion:129 Resp: 181813  
Ion Ratio Lower Upper  
129 100  
127 78.4 38.1 114.5





#61  
1,2-Dibromoethane  
Concen: 49.732 ug/l  
RT: 11.239 min Scan# 1534  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion:107 Resp: 154714  
Ion Ratio Lower Upper  
107 100  
109 94.4 74.5 111.7

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#62  
4-Bromofluorobenzene  
Concen: 48.218 ug/l  
RT: 12.617 min Scan# 1760  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Tgt Ion: 95 Resp: 236491  
Ion Ratio Lower Upper  
95 100  
174 64.6 0.0 145.6  
176 63.3 0.0 140.8





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.629 min Scan# 1159  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion:117 Resp: 498211  
Ion Ratio Lower Upper  
117 100  
82 59.1 42.7 64.1  
119 32.5 25.2 37.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#64  
Tetrachloroethene  
Concen: 49.611 ug/l  
RT: 10.867 min Scan# 1473  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04



Tgt Ion:164 Resp: 147899  
Ion Ratio Lower Upper  
164 100  
166 120.0 97.5 146.3  
129 92.9 71.2 106.8  
131 96.5 70.7 106.1









#65  
Chlorobenzene  
Concen: 48.862 ug/l  
RT: 11.660 min Scan# 1603  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04

Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125

Tgt Ion:112 Resp: 535619  
Ion Ratio Lower Upper  
112 100  
114 31.4 26.0 39.0




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#66  
1,1,1,2-Tetrachloroethane  
Concen: 48.644 ug/l  
RT: 11.733 min Scan# 1615  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion:131 Resp: 166082  
Ion Ratio Lower Upper  
131 100  
133 95.9 47.7 143.1  
119 66.2 32.8 98.3





#67  
Ethyl Benzene  
Concen: 50.513 ug/l  
RT: 11.733 min Scan# 1615  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion: 91 Resp: 92298  
Ion Ratio Lower Upper  
91 100  
106 31.3 25.3 37.9

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#68  
m/p-Xylenes  
Concen: 99.276 ug/l  
RT: 11.837 min Scan# 1632  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Tgt Ion:106 Resp: 699490  
Ion Ratio Lower Upper  
106 100  
91 207.4 164.2 246.2





#69  
o-Xylene  
Concen: 50.523 ug/l  
RT: 12.166 min Scan# 1686  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion:106 Resp: 332905  
Ion Ratio Lower Upper  
106 100  
91 218.0 108.7 326.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#70  
Styrene  
Concen: 51.667 ug/l  
RT: 12.178 min Scan# 1688  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04



Tgt Ion:104 Resp: 594393  
Ion Ratio Lower Upper  
104 100  
78 50.7 41.6 62.4  
103 52.9 44.8 67.2





#71

Bromoform

Concen: 51.207 ug/l

RT: 12.349 min Scan# 1716

Delta R.T. -0.000 min

Lab File: VW032320.D

Acq: 01 Oct 2025 16:04

Instrument :

MSVOA\_W

ClientSampleId :





ICVW100125

Tgt Ion:173 Resp: 100214  
 Ion Ratio Lower Upper  
 173 100  
 175 47.4 24.9 74.7  
 254 0.0 0.2 0.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025  
 Supervised By :Semsettin Yesilyurt 10/03/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.556 min Scan# 1914

Delta R.T. -0.000 min


Lab File: VW032320.D

Acq: 01 Oct 2025 16:04

Tgt Ion:152 Resp: 235528  
 Ion Ratio Lower Upper  
 152 100  
 115 66.0 31.8 95.4  
 150 173.3 0.0 343.0








#73  
Isopropylbenzene  
Concen: 52.076 ug/l  
RT: 12.465 min Scan# 1735  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion:105 Resp: 839135  
Ion Ratio Lower Upper  
105 100  
120 25.7 13.0 38.9

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#74  
N-ethyl acetate  
Concen: 53.697 ug/l  
RT: 12.269 min Scan# 1703  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Tgt Ion: 43 Resp: 276033  
Ion Ratio Lower Upper  
43 100  
70 44.2 35.3 52.9  
55 25.8 19.9 29.9  
61 25.1 19.6 29.4





#75  
1,1,2,2-Tetrachloroethane  
Concen: 50.822 ug/l  
RT: 12.715 min Scan# 1776  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument : MSVOA\_W  
ClientSampleId : ICVVW100125




Tgt Ion: 83 Resp: 203339  
Ion Ratio Lower Upper  
83 100  
131 9.9 5.1 15.4  
85 64.1 32.2 96.6

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#76  
1,2,3-Trichloropropane  
Concen: 50.813 ug/l m  
RT: 12.763 min Scan# 1784  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Tgt Ion: 75 Resp: 157491  
Ion Ratio Lower Upper  
75 100  
77 282.7 141.7 425.1





#77  
Bromobenzene  
Concen: 50.659 ug/l  
RT: 12.745 min Scan# 1781  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125



Tgt Ion:156 Resp: 196770  
Ion Ratio Lower Upper  
156 100  
77 226.3 98.6 295.7  
158 95.6 48.4 145.0




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#78  
n-propylbenzene  
Concen: 51.657 ug/l  
RT: 12.800 min Scan# 1790  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04

Tgt Ion: 91 Resp: 1039770  
Ion Ratio Lower Upper  
91 100  
120 22.5 11.8 35.4









#79  
2-Chlorotoluene  
Concen: 51.287 ug/l  
RT: 12.891 min Scan# 1805  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04

Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125

Tgt Ion: 91 Resp: 631190  
Ion Ratio Lower Upper  
91 100  
126 31.3 16.6 49.7




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#80  
1,3,5-Trimethylbenzene  
Concen: 52.365 ug/l  
RT: 12.940 min Scan# 1813  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04


Tgt Ion:105 Resp: 729628  
Ion Ratio Lower Upper  
105 100  
120 47.2 24.8 74.4





#81  
trans-1,4-Dichloro-2-butene  
Concen: 54.360 ug/l  
RT: 12.513 min Scan# 1743  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion: 75 Resp: 66508  
Ion Ratio Lower Upper  
75 100  
53 102.7 91.0 136.6  
89 45.6 39.4 59.2

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#82  
4-Chlorotoluene  
Concen: 50.734 ug/l  
RT: 12.989 min Scan# 1821  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Tgt Ion: 91 Resp: 663842  
Ion Ratio Lower Upper  
91 100  
126 32.1 16.1 48.2





#83  
tert-Butylbenzene  
Concen: 53.510 ug/l  
RT: 13.202 min Scan# 1856  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument : MSVOA\_W  
ClientSampleId : ICVVW100125




Tgt Ion:119 Resp: 611748  
Ion Ratio Lower Upper  
119 100  
91 70.3 33.8 101.3  
134 24.6 12.5 37.5

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#84  
1,2,4-Trimethylbenzene  
Concen: 51.522 ug/l  
RT: 13.245 min Scan# 1863  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Tgt Ion:105 Resp: 729966  
Ion Ratio Lower Upper  
105 100  
120 44.6 22.1 66.3





#85  
 sec-Butylbenzene  
 Concen: 51.774 ug/l  
 RT: 13.379 min Scan# 1885  
 Delta R.T. -0.000 min  
 Lab File: VW032320.D  
 Acq: 01 Oct 2025 16:04




Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 ICVVW100125




Tgt Ion:105 Resp: 904532  
 Ion Ratio Lower Upper  
 105 100  
 134 19.6 10.1 30.1

Manual Integrations  
 APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
 Supervised By :Semsettin Yesilyurt 10/03/2025




#86  
 p-Isopropyltoluene  
 Concen: 52.697 ug/l  
 RT: 13.495 min Scan# 1904  
 Delta R.T. -0.000 min  
 Lab File: VW032320.D  
 Acq: 01 Oct 2025 16:04



Tgt Ion:119 Resp: 763912  
 Ion Ratio Lower Upper  
 119 100  
 134 24.9 12.6 37.6  
 91 24.0 12.4 37.0





#87

1,3-Dichlorobenzene

Concen: 50.518 ug/l

RT: 13.501 min Scan# 1905

Delta R.T. 0.006 min

Lab File: VW032320.D


Acq: 01 Oct 2025 16:04

Instrument :

MSVOA\_W

ClientSampleId :

ICVW100125



Tgt Ion:146 Resp: 397340

Ion Ratio Lower Upper

146 100

111 43.0 20.6 61.9




148 64.0 30.6 92.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025

Supervised By :Semsettin Yesilyurt 10/03/2025



#88

1,4-Dichlorobenzene

Concen: 48.372 ug/l

RT: 13.574 min Scan# 1917

Delta R.T. -0.000 min

Lab File: VW032320.D

Acq: 01 Oct 2025 16:04




Tgt Ion:146 Resp: 380433

Ion Ratio Lower Upper


146 100

111 43.9 21.3 63.7

148 64.2 33.4 100.1








#89  
n-Butylbenzene  
Concen: 52.880 ug/l  
RT: 13.818 min Scan# 1957  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion: 91 Resp: 748452  
Ion Ratio Lower Upper  
91 100  
92 53.6 27.0 81.0  
134 24.8 13.1 39.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#90  
Hexachloroethane  
Concen: 51.903 ug/l  
RT: 14.086 min Scan# 2001  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Tgt Ion:117 Resp: 134862  
Ion Ratio Lower Upper  
117 100  
201 60.5 32.8 98.4





#91  
1,2-Dichlorobenzene  
Concen: 48.810 ug/l  
RT: 13.867 min Scan# 1965  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion:146 Resp: 347039  
Ion Ratio Lower Upper  
146 100  
111 46.4 23.4 70.0  
148 64.5 34.1 102.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 52.139 ug/l  
RT: 14.476 min Scan# 2065  
Delta R.T. -0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04



Tgt Ion: 75 Resp: 36524  
Ion Ratio Lower Upper  
75 100  
155 71.1 40.8 122.5  
157 91.4 51.8 155.4





#93

1,2,4-Trichlorobenzene

Concen: 51.178 ug/l

RT: 15.129 min Scan# 2171

Delta R.T. 0.006 min

Lab File: VW032320.D


Acq: 01 Oct 2025 16:04

Instrument :

MSVOA\_W

ClientSampleId :

ICVW100125



Tgt Ion:180 Resp: 218198

Ion Ratio Lower Upper

180 100

182 87.8 47.4 142.2




145 36.6 16.4 49.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025

Supervised By :Semsettin Yesilyurt 10/03/2025



#94

Hexachlorobutadiene

Concen: 46.478 ug/l

RT: 15.232 min Scan# 2189

Delta R.T. 0.006 min

Lab File: VW032320.D

Acq: 01 Oct 2025 16:04




Tgt Ion:225 Resp: 90599


Ion Ratio Lower Upper

225 100

223 70.3 29.8 89.3


227 70.0 30.5 91.5





#95  
Naphthalene  
Concen: 55.881 ug/l  
RT: 15.360 min Scan# 2210  
Delta R.T. -0.000 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04




Instrument :  
MSVOA\_W  
ClientSampleId :  
ICVWV100125




Tgt Ion:128 Resp: 57214  
Ion Ratio Lower Upper  
128 100  
127 12.7 10.6 16.0  
129 10.6 8.4 12.6

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025  
Supervised By :Semsettin Yesilyurt 10/03/2025



#96  
1,2,3-Trichlorobenzene  
Concen: 51.417 ug/l  
RT: 15.549 min Scan# 2241  
Delta R.T. 0.006 min  
Lab File: VW032320.D  
Acq: 01 Oct 2025 16:04



Tgt Ion:180 Resp: 206623  
Ion Ratio Lower Upper  
180 100  
182 97.4 47.7 143.1  
145 35.4 17.2 51.6

