

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
 Data File : VX048256.D
 Acq On : 21 Oct 2025 10:33
 Operator : JC/MD
 Sample : VX1021WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1021WBSD01

Quant Time: Oct 24 04:24:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	137604	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.738	114	224103	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.036	117	203364	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	101316	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.933	65	89758	46.523	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.040%
35) Dibromofluoromethane	5.367	113	73394	47.972	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.940%
50) Toluene-d8	8.628	98	256689	47.867	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.740%
62) 4-Bromofluorobenzene	11.061	95	96462	47.283	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.560%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.178	85	25428	17.976	ug/l	97
3) Chloromethane	1.300	50	22848	17.304	ug/l	97
4) Vinyl Chloride	1.379	62	20627	16.870	ug/l	94
5) Bromomethane	1.623	94	15738	18.966	ug/l	96
6) Chloroethane	1.696	64	13651	19.000	ug/l	92
7) Trichlorofluoromethane	1.898	101	42048	17.931	ug/l	97
8) Diethyl Ether	2.135	74	12151	18.416	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.337	101	26272	17.693	ug/l	94
10) Methyl Iodide	2.452	142	33915	18.168	ug/l	99
11) Tert butyl alcohol	2.934	59	14617	76.307	ug/l	96
12) 1,1-Dichloroethene	2.324	96	25810	17.866	ug/l	96
13) Acrolein	2.233	56	19717	100.924	ug/l	99
14) Allyl chloride	2.666	41	43150	17.990	ug/l	97
15) Acrylonitrile	3.050	53	60796	88.017	ug/l	99
16) Acetone	2.373	43	53983	81.562	ug/l	98
17) Carbon Disulfide	2.513	76	76828	16.874	ug/l	100
18) Methyl Acetate	2.696	43	23399	16.928	ug/l	91
19) Methyl tert-butyl Ether	3.105	73	88912	18.041	ug/l	98
20) Methylene Chloride	2.788	84	29832	18.498	ug/l	89
21) trans-1,2-Dichloroethene	3.086	96	29850	18.810	ug/l	87
22) Diisopropyl ether	3.745	45	93736	19.594	ug/l	88
23) Vinyl Acetate	3.708	43	360854	94.430	ug/l	96
24) 1,1-Dichloroethane	3.605	63	54330	18.782	ug/l	97
25) 2-Butanone	4.537	43	71780	85.682	ug/l	94
26) 2,2-Dichloropropane	4.464	77	46208	17.698	ug/l	100
27) cis-1,2-Dichloroethene	4.476	96	35291	19.056	ug/l	98
28) Bromochloromethane	4.879	49	25497	21.258	ug/l	85
29) Tetrahydrofuran	4.982	42	43202	86.106	ug/l	93
30) Chloroform	5.074	83	58812	19.030	ug/l	100
31) Cyclohexane	5.452	56	44037	18.388	ug/l	92
32) 1,1,1-Trichloroethane	5.373	97	52248	18.344	ug/l	97
36) 1,1-Dichloropropene	5.671	75	38187	17.561	ug/l	99
37) Ethyl Acetate	4.696	43	32528	18.047	ug/l	99
38) Carbon Tetrachloride	5.659	117	46209	17.224	ug/l	99
39) Methylcyclohexane	7.360	83	46866	18.249	ug/l	95
40) Benzene	6.025	78	118391	18.752	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
 Data File : VX048256.D
 Acq On : 21 Oct 2025 10:33
 Operator : JC/MD
 Sample : VX1021WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1021WBSD01

Quant Time: Oct 24 04:24:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.903	41	16538	17.472	ug/l	97
42) 1,2-Dichloroethane	6.068	62	43164	18.528	ug/l	100
43) Isopropyl Acetate	6.318	43	52017	17.569	ug/l	96
44) Trichloroethene	7.110	130	30449	18.265	ug/l	98
45) 1,2-Dichloropropane	7.409	63	29710	19.793	ug/l	99
46) Dibromomethane	7.561	93	21708	18.563	ug/l	94
47) Bromodichloromethane	7.805	83	47421	18.781	ug/l	95
48) Methyl methacrylate	7.677	41	26520	17.751	ug/l	95
49) 1,4-Dioxane	7.640	88	6381	323.294	ug/l #	86
51) 4-Methyl-2-Pentanone	8.555	43	151895	88.982	ug/l	96
52) Toluene	8.701	92	75829	18.906	ug/l	100
53) t-1,3-Dichloropropene	8.963	75	45008	18.497	ug/l	99
54) cis-1,3-Dichloropropene	8.348	75	48675	18.541	ug/l	95
55) 1,1,2-Trichloroethane	9.134	97	28743	19.271	ug/l	97
56) Ethyl methacrylate	9.098	69	39428	18.073	ug/l	94
57) 1,3-Dichloropropane	9.293	76	48237	18.933	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.226	63	97758	84.467	ug/l	97
59) 2-Hexanone	9.415	43	104512	87.804	ug/l	98
60) Dibromochloromethane	9.500	129	36743	18.328	ug/l	100
61) 1,2-Dibromoethane	9.591	107	29504	18.555	ug/l	99
64) Tetrachloroethene	9.256	164	26555	17.492	ug/l	97
65) Chlorobenzene	10.061	112	86348	18.875	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.146	131	30678	18.258	ug/l	99
67) Ethyl Benzene	10.177	91	142694	18.320	ug/l	97
68) m/p-Xylenes	10.286	106	108512	37.497	ug/l	100
69) o-Xylene	10.622	106	51598	18.468	ug/l	98
70) Styrene	10.640	104	91270	19.185	ug/l	98
71) Bromoform	10.780	173	23785	17.711	ug/l #	100
73) Isopropylbenzene	10.945	105	136638	18.464	ug/l	100
74) N-amyl acetate	10.829	43	45310	17.261	ug/l	95
75) 1,1,2,2-Tetrachloroethane	11.195	83	37185	18.401	ug/l	99
76) 1,2,3-Trichloropropane	11.219	75	38024	23.143	ug/l	82
77) Bromobenzene	11.183	156	34541	18.615	ug/l	94
78) n-propylbenzene	11.286	91	160675	18.753	ug/l	98
79) 2-Chlorotoluene	11.347	91	97491	18.767	ug/l	98
80) 1,3,5-Trimethylbenzene	11.432	105	114521	18.700	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.000	75	11571	17.134	ug/l	99
82) 4-Chlorotoluene	11.439	91	115393	18.713	ug/l	97
83) tert-Butylbenzene	11.695	119	117357	18.651	ug/l	95
84) 1,2,4-Trimethylbenzene	11.731	105	114561	18.745	ug/l	99
85) sec-Butylbenzene	11.871	105	139017	18.364	ug/l	100
86) p-Isopropyltoluene	11.993	119	118663	18.135	ug/l	100
87) 1,3-Dichlorobenzene	11.951	146	63730	18.219	ug/l	99
88) 1,4-Dichlorobenzene	12.024	146	65963	18.203	ug/l	99
89) n-Butylbenzene	12.316	91	106210	17.834	ug/l	98
90) Hexachloroethane	12.518	117	22528	17.435	ug/l	96
91) 1,2-Dichlorobenzene	12.316	146	61124	18.612	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.920	75	6932	16.899	ug/l	97
93) 1,2,4-Trichlorobenzene	13.566	180	37218	16.758	ug/l	98
94) Hexachlorobutadiene	13.706	225	14700	17.706	ug/l	98
95) Naphthalene	13.755	128	100059	17.005	ug/l	99
96) 1,2,3-Trichlorobenzene	13.944	180	35212	17.277	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
Data File : VX048256.D
Acq On : 21 Oct 2025 10:33
Operator : JC/MD
Sample : VX1021WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Quant Time: Oct 24 04:24:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 24 04:24:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.531 min Scan# 71
Delta R.T. -0.013 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion:168 Resp: 137604
Ion Ratio Lower Upper
168 100
99 60.1 46.4 69.6



#2
Dichlorodifluoromethane
Concen: 17.976 ug/l
RT: 1.178 min Scan# 15
Delta R.T. 0.006 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 85 Resp: 25428
Ion Ratio Lower Upper
85 100
87 32.2 15.3 45.9









#3
Chloromethane
Concen: 17.304 ug/l
RT: 1.300 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 50 Resp: 22848
Ion Ratio Lower Upper
50 100
52 34.4 25.9 38.9



#4
Vinyl Chloride
Concen: 16.870 ug/l
RT: 1.379 min Scan# 48
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 62 Resp: 20627
Ion Ratio Lower Upper
62 100
64 34.7 25.1 37.7









#5
Bromomethane
Concen: 18.966 ug/l
RT: 1.623 min Scan# 81
Delta R.T. 0.006 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 94 Resp: 15738
Ion Ratio Lower Upper
94 100
96 91.5 76.0 114.0



#6
Chloroethane
Concen: 19.000 ug/l
RT: 1.696 min Scan# 100
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 64 Resp: 13651
Ion Ratio Lower Upper
64 100
66 34.4 23.9 35.9









#7
 Trichlorofluoromethane
 Concen: 17.931 ug/l
 RT: 1.898 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX048256.D
 Acq: 21 Oct 2025 10:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1021WBSD01


Tgt Ion:101 Resp: 42048
 Ion Ratio Lower Upper
 101 100
 103 62.2 51.6 77.4



#8
 Diethyl Ether
 Concen: 18.416 ug/l
 RT: 2.135 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX048256.D
 Acq: 21 Oct 2025 10:33

Tgt Ion: 74 Resp: 12151
 Ion Ratio Lower Upper
 74 100
 45 93.4 50.0 149.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 17.693 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion:101 Resp: 26272
Ion Ratio Lower Upper
101 100
85 46.4 35.6 53.4
151 79.7 58.0 87.0



#10
Methyl Iodide
Concen: 18.168 ug/l
RT: 2.452 min Scan# 224
Delta R.T. -0.007 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion:142 Resp: 33915
Ion Ratio Lower Upper
142 100
127 47.7 38.5 57.7
141 14.2 11.6 17.4





#11

Tert butyl alcohol

Concen: 76.307 ug/l

RT: 2.934 min Scan# 305

Delta R.T. -0.007 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :





VX1021WBSD01

Tgt Ion: 59 Resp: 14617

Ion Ratio Lower Upper

59 100

57 8.7 8.2 12.4



#12

1,1-Dichloroethene

Concen: 17.866 ug/l

RT: 2.324 min Scan# 203

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33




Tgt Ion: 96 Resp: 25810


Ion Ratio Lower Upper

96 100

61 167.8 139.0 208.6

98 64.4 52.4 78.6





#13

Acrolein

Concen: 100.924 ug/l

RT: 2.233 min Scan# 188

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :





VX1021WBSD01

Tgt Ion: 56 Resp: 19717

Ion Ratio Lower Upper

56 100

55 71.3 56.1 84.1



#14

Allyl chloride

Concen: 17.990 ug/l

RT: 2.666 min Scan# 259

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33




Tgt Ion: 41 Resp: 43150


Ion Ratio Lower Upper

41 100

39 68.5 56.0 84.0

76 38.0 27.5 41.3









#15
Acrylonitrile
Concen: 88.017 ug/l
RT: 3.050 min Scan# 31
Delta R.T. -0.007 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 53 Resp: 60796
Ion Ratio Lower Upper
53 100
52 83.9 66.1 99.1
51 36.9 29.0 43.4



#16
Acetone
Concen: 81.562 ug/l
RT: 2.373 min Scan# 211
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 43 Resp: 53983
Ion Ratio Lower Upper
43 100
58 30.3 23.4 35.2









#17
Carbon Disulfide
Concen: 16.874 ug/l
RT: 2.513 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 76 Resp: 76828
Ion Ratio Lower Upper
76 100
78 8.9 7.1 10.7



#18
Methyl Acetate
Concen: 16.928 ug/l
RT: 2.696 min Scan# 264
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 43 Resp: 23399
Ion Ratio Lower Upper
43 100
74 26.0 17.5 26.3









#19
Methyl tert-butyl Ether
Concen: 18.041 ug/l
RT: 3.105 min Scan# 311
Delta R.T. -0.007 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 73 Resp: 88912
Ion Ratio Lower Upper
73 100
57 21.1 17.7 26.5



#20
Methylene Chloride
Concen: 18.498 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 84 Resp: 29832
Ion Ratio Lower Upper
84 100
49 110.8 104.1 156.1
51 35.6 31.3 46.9
86 66.7 51.3 76.9









#21
trans-1,2-Dichloroethene
Concen: 18.810 ug/l
RT: 3.086 min Scan# 31
Delta R.T. -0.007 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 96 Resp: 29850
Ion Ratio Lower Upper
96 100
61 121.5 111.9 167.9
98 56.2 51.2 76.8



#22
Diisopropyl ether
Concen: 19.594 ug/l
RT: 3.745 min Scan# 436
Delta R.T. -0.007 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 45 Resp: 93736
Ion Ratio Lower Upper
45 100
43 127.1 89.8 134.8
87 29.0 20.6 31.0
59 12.3 9.1 13.7









#23
 Vinyl Acetate
 Concen: 94.430 ug/l
 RT: 3.708 min Scan# 41
 Delta R.T. -0.007 min
 Lab File: VX048256.D
 Acq: 21 Oct 2025 10:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1021WBSD01


Tgt Ion: 43 Resp: 360854
 Ion Ratio Lower Upper
 43 100
 86 11.2 7.9 11.9



#24
 1,1-Dichloroethane
 Concen: 18.782 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX048256.D
 Acq: 21 Oct 2025 10:33

Tgt Ion: 63 Resp: 54330
 Ion Ratio Lower Upper
 63 100
 98 8.3 3.5 10.6
 100 4.8 2.1 6.3









#25
2-Butanone
Concen: 85.682 ug/l
RT: 4.537 min Scan# 51
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 43 Resp: 71780
Ion Ratio Lower Upper
43 100
72 28.1 19.9 29.9



#26
2,2-Dichloropropane
Concen: 17.698 ug/l
RT: 4.464 min Scan# 554
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 77 Resp: 46208
Ion Ratio Lower Upper
77 100
97 22.3 11.1 33.3









#27
 cis-1,2-Dichloroethene
 Concen: 19.056 ug/l
 RT: 4.476 min Scan# 557
 Delta R.T. -0.007 min
 Lab File: VX048256.D
 Acq: 21 Oct 2025 10:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1021WBSD01


Tgt Ion: 96 Resp: 35291
 Ion Ratio Lower Upper
 96 100
 61 143.6 0.0 294.2
 98 64.4 0.0 128.4



#28
 Bromochloromethane
 Concen: 21.258 ug/l
 RT: 4.879 min Scan# 622
 Delta R.T. -0.007 min
 Lab File: VX048256.D
 Acq: 21 Oct 2025 10:33

Tgt Ion: 49 Resp: 25497
 Ion Ratio Lower Upper
 49 100
 129 2.1 0.0 4.4
 130 86.8 59.3 88.9





#29

Tetrahydrofuran

Concen: 86.106 ug/l

RT: 4.982 min Scan# 61

Delta R.T. -0.007 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :

VX1021WBSD01





Tgt Ion: 42 Resp: 43202

Ion Ratio Lower Upper

42 100

72 48.4 35.3 52.9

71 45.3 32.2 48.4



#30

Chloroform

Concen: 19.030 ug/l

RT: 5.074 min Scan# 654

Delta R.T. -0.007 min

Lab File: VX048256.D




Acq: 21 Oct 2025 10:33


Tgt Ion: 83 Resp: 58812

Ion Ratio Lower Upper

83 100


85 62.7 50.1 75.1








#31
Cyclohexane
Concen: 18.388 ug/l
RT: 5.452 min Scan# 716
Delta R.T. -0.007 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01



Tgt Ion: 56 Resp: 44037




Ion	Ratio	Lower	Upper
56	100		
69	30.8	23.0	34.6
84	89.8	64.6	97.0




#32
1,1,1-Trichloroethane
Concen: 18.344 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 97 Resp: 52248

Ion	Ratio	Lower	Upper
97	100		
99	63.0	51.7	77.5
61	41.2	34.8	52.2









#33
 1,2-Dichloroethane-d4
 Concen: 46.523 ug/l
 RT: 5.933 min Scan# 795
 Delta R.T. -0.007 min
 Lab File: VX048256.D
 Acq: 21 Oct 2025 10:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1021WBSD01


Tgt Ion: 65 Resp: 89758
 Ion Ratio Lower Upper
 65 100
 67 51.0 0.0 105.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.738 min Scan# 927
 Delta R.T. -0.007 min
 Lab File: VX048256.D
 Acq: 21 Oct 2025 10:33

Tgt Ion: 114 Resp: 224103
 Ion Ratio Lower Upper
 114 100
 63 20.7 0.0 43.2
 88 16.6 0.0 33.6









#35
Dibromofluoromethane
Concen: 47.972 ug/l
RT: 5.367 min Scan# 704
Delta R.T. -0.007 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion:113 Resp: 73394
Ion Ratio Lower Upper
113 100
111 100.5 82.2 123.4
192 19.4 15.1 22.7



#36
1,1-Dichloropropene
Concen: 17.561 ug/l
RT: 5.671 min Scan# 752
Delta R.T. -0.013 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 75 Resp: 38187
Ion Ratio Lower Upper
75 100
110 36.3 17.7 53.1
77 31.0 24.3 36.5





#37

Ethyl Acetate

Concen: 18.047 ug/l

RT: 4.696 min Scan# 593

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :

VX1021WBSD01





Tgt Ion: 43 Resp: 32528

Ion Ratio Lower Upper

43 100

61 13.5 10.6 15.8

70 10.4 7.8 11.6



#38

Carbon Tetrachloride

Concen: 17.224 ug/l

RT: 5.659 min Scan# 750

Delta R.T. -0.007 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33




Tgt Ion: 117 Resp: 46209


Ion Ratio Lower Upper

117 100

119 95.8 77.2 115.8

121 32.7 24.8 37.2









#39
Methylcyclohexane
Concen: 18.249 ug/l
RT: 7.360 min Scan# 1029
Delta R.T. -0.007 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 83 Resp: 46866
Ion Ratio Lower Upper
83 100
55 76.6 66.2 99.2
98 50.5 38.9 58.3



#40
Benzene
Concen: 18.752 ug/l
RT: 6.025 min Scan# 810
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 78 Resp: 118391
Ion Ratio Lower Upper
78 100
77 22.3 19.3 28.9









#41
Methacrylonitrile
Concen: 17.472 ug/l
RT: 4.903 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01

Tgt Ion: 41 Resp: 16538




Ion	Ratio	Lower	Upper
41	100		
39	55.4	44.0	66.0
67	78.9	58.7	88.1
52	32.4	25.9	38.9




#42
1,2-Dichloroethane
Concen: 18.528 ug/l
RT: 6.068 min Scan# 817
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 62 Resp: 43164

Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	17.2





#43

Isopropyl Acetate

Concen: 17.569 ug/l

RT: 6.318 min Scan# 858

Delta R.T. -0.007 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :

VX1021WBSD01





Tgt Ion: 43 Resp: 52017

Ion Ratio Lower Upper

43 100

61 20.3 14.9 22.3

87 13.7 9.4 14.2



#44

Trichloroethene

Concen: 18.265 ug/l

RT: 7.110 min Scan# 988

Delta R.T. -0.000 min

Lab File: VX048256.D




Acq: 21 Oct 2025 10:33


Tgt Ion:130 Resp: 30449

Ion Ratio Lower Upper

130 100

95 98.3 0.0 200.6





#45

1,2-Dichloropropane

Concen: 19.793 ug/l

RT: 7.409 min Scan# 1037

Delta R.T. -0.007 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :





VX1021WBSD01

Tgt Ion: 63 Resp: 29710

Ion Ratio Lower Upper

63 100

65 30.9 24.1 36.1



#46

Dibromomethane

Concen: 18.563 ug/l

RT: 7.561 min Scan# 1062

Delta R.T. -0.007 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33




Tgt Ion: 93 Resp: 21708


Ion Ratio Lower Upper

93 100

95 81.8 65.9 98.9

174 99.5 71.7 107.5









#47
Bromodichloromethane
Concen: 18.781 ug/l
RT: 7.805 min Scan# 1102
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 83 Resp: 47421
Ion Ratio Lower Upper
83 100
85 69.3 52.2 78.4
127 8.3 7.3 10.9



#48
Methyl methacrylate
Concen: 17.751 ug/l
RT: 7.677 min Scan# 1081
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 41 Resp: 26520
Ion Ratio Lower Upper
41 100
69 86.1 63.4 95.0
39 59.0 47.1 70.7





#49

1,4-Dioxane

Concen: 323.294 ug/l

RT: 7.640 min Scan# 1076

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :

VX1021WBSD01





Tgt Ion: 88 Resp: 6381

Ion Ratio Lower Upper

88 100

43 21.2 35.0 52.4#

58 75.7 61.9 92.9



#50

Toluene-d8

Concen: 47.867 ug/l

RT: 8.628 min Scan# 1237

Delta R.T. -0.007 min

Lab File: VX048256.D




Acq: 21 Oct 2025 10:33


Tgt Ion: 98 Resp: 256689

Ion Ratio Lower Upper

98 100

100 66.4 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 88.982 ug/l
RT: 8.555 min Scan# 1225
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 43 Resp: 151895
Ion Ratio Lower Upper
43 100
58 39.5 29.5 44.3



#52
Toluene
Concen: 18.906 ug/l
RT: 8.701 min Scan# 1249
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 92 Resp: 75829
Ion Ratio Lower Upper
92 100
91 169.9 136.1 204.1





#53

t-1,3-Dichloropropene

Concen: 18.497 ug/l

RT: 8.963 min Scan# 1191

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :





VX1021WBSD01

Tgt Ion: 75 Resp: 45008

Ion Ratio Lower Upper

75 100

77 30.5 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 18.541 ug/l

RT: 8.348 min Scan# 1191

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33




Tgt Ion: 75 Resp: 48675


Ion Ratio Lower Upper

75 100

77 32.9 24.9 37.3

39 48.8 42.6 64.0









#55
 1,1,2-Trichloroethane
 Concen: 19.271 ug/l
 RT: 9.134 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX048256.D
 Acq: 21 Oct 2025 10:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1021WBSD01


Tgt Ion:	97	Resp:	28743
Ion	Ratio	Lower	Upper
97	100		
83	85.5	70.6	105.8
85	54.0	45.0	67.6
99	65.6	50.9	76.3



#56
 Ethyl methacrylate
 Concen: 18.073 ug/l
 RT: 9.098 min Scan# 1314
 Delta R.T. -0.000 min
 Lab File: VX048256.D
 Acq: 21 Oct 2025 10:33

Tgt Ion:	69	Resp:	39428
Ion	Ratio	Lower	Upper
69	100		
41	64.9	56.0	84.0
39	35.9	31.6	47.4





#57

1,3-Dichloropropane

Concen: 18.933 ug/l

RT: 9.293 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :





VX1021WBSD01

Tgt Ion: 76 Resp: 48237

Ion Ratio Lower Upper

76 100

78 32.2 25.4 38.0



#58

2-Chloroethyl Vinyl ether

Concen: 84.467 ug/l

RT: 8.226 min Scan# 1171

Delta R.T. -0.000 min

Lab File: VX048256.D




Acq: 21 Oct 2025 10:33


Tgt Ion: 63 Resp: 97758

Ion Ratio Lower Upper

63 100

106 29.4 22.2 33.2





#59

2-Hexanone

Concen: 87.804 ug/l

RT: 9.415 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :





VX1021WBSD01

Tgt Ion: 43 Resp: 104512

Ion Ratio Lower Upper

43 100

58 54.2 26.3 78.8



#60

Dibromochloromethane

Concen: 18.328 ug/l

RT: 9.500 min Scan# 1380

Delta R.T. -0.007 min

Lab File: VX048256.D




Acq: 21 Oct 2025 10:33


Tgt Ion: 129 Resp: 36743

Ion Ratio Lower Upper

129 100

127 77.5 38.6 115.7









#61
1,2-Dibromoethane
Concen: 18.555 ug/l
RT: 9.591 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion:107 Resp: 29504
Ion Ratio Lower Upper
107 100
109 95.5 77.0 115.4



#62
4-Bromofluorobenzene
Concen: 47.283 ug/l
RT: 11.061 min Scan# 1636
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 95 Resp: 96462
Ion Ratio Lower Upper
95 100
174 75.1 0.0 147.8
176 73.9 0.0 142.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.036 min Scan# 1468
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion:117 Resp: 203364
Ion Ratio Lower Upper
117 100
82 56.5 46.5 69.7
119 32.7 25.9 38.9



#64
Tetrachloroethene
Concen: 17.492 ug/l
RT: 9.256 min Scan# 1340
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion:164 Resp: 26555
Ion Ratio Lower Upper
164 100
166 135.3 104.5 156.7
129 98.9 81.8 122.6
131 97.9 77.0 115.6









#65
Chlorobenzene
Concen: 18.875 ug/l
RT: 10.061 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion:112 Resp: 86348
Ion Ratio Lower Upper
112 100
114 32.1 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 18.258 ug/l
RT: 10.146 min Scan# 1486
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion:131 Resp: 30678
Ion Ratio Lower Upper
131 100
133 94.6 47.6 142.8
119 63.3 32.5 97.5





#67

Ethyl Benzene

Concen: 18.320 ug/l

RT: 10.177 min Scan# 1491

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :





VX1021WBSD01

Tgt Ion: 91 Resp: 142694

Ion Ratio Lower Upper

91 100

106 31.4 23.8 35.6



#68

m/p-Xylenes

Concen: 37.497 ug/l

RT: 10.286 min Scan# 1509

Delta R.T. -0.000 min

Lab File: VX048256.D




Acq: 21 Oct 2025 10:33


Tgt Ion:106 Resp: 108512

Ion Ratio Lower Upper

106 100

91 209.5 167.9 251.9









#69
o-Xylene
Concen: 18.468 ug/l
RT: 10.622 min Scan# 1564
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion:106 Resp: 51598
Ion Ratio Lower Upper
106 100
91 219.4 111.3 333.9



#70
Styrene
Concen: 19.185 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion:104 Resp: 91270
Ion Ratio Lower Upper
104 100
78 51.0 42.2 63.4
103 53.4 43.0 64.4









#71
Bromoform
Concen: 17.711 ug/l
RT: 10.780 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion:173 Resp: 23785
Ion Ratio Lower Upper
173 100
175 47.7 24.0 72.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion:152 Resp: 101316
Ion Ratio Lower Upper
152 100
115 70.4 43.1 129.4
150 163.7 0.0 353.0





#73

Isopropylbenzene

Concen: 18.464 ug/l

RT: 10.945 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :





VX1021WBSD01

Tgt Ion:105 Resp: 136638

Ion Ratio Lower Upper

105 100

120 25.9 13.0 38.9



#74

N-ethyl acetate

Concen: 17.261 ug/l

RT: 10.829 min Scan# 1598

Delta R.T. 0.006 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Tgt Ion: 43 Resp: 45310




Ion Ratio Lower Upper


43 100

70 45.0 32.2 48.2

55 24.9 19.1 28.7

61 25.4 18.6 28.0









#75
1,1,2-Tetrachloroethane
Concen: 18.401 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 83 Resp: 37185
Ion Ratio Lower Upper
83 100
131 10.7 4.8 14.4
85 64.0 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 23.143 ug/l
RT: 11.219 min Scan# 1662
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 75 Resp: 38024
Ion Ratio Lower Upper
75 100
77 33.3 22.4 67.2









#77
Bromobenzene
Concen: 18.615 ug/l
RT: 11.183 min Scan# 1656
Delta R.T. 0.006 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion:156 Resp: 34541
Ion Ratio Lower Upper
156 100
77 161.0 87.0 261.0
158 96.2 47.8 143.4



#78
n-propylbenzene
Concen: 18.753 ug/l
RT: 11.286 min Scan# 1673
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 91 Resp: 160675
Ion Ratio Lower Upper
91 100
120 23.0 11.1 33.1









#79
2-Chlorotoluene
Concen: 18.767 ug/l
RT: 11.347 min Scan# 1683
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 91 Resp: 97491
Ion Ratio Lower Upper
91 100
126 33.9 16.4 49.4



#80
1,3,5-Trimethylbenzene
Concen: 18.700 ug/l
RT: 11.432 min Scan# 1697
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion:105 Resp: 114521
Ion Ratio Lower Upper
105 100
120 47.5 24.0 72.0









#81
trans-1,4-Dichloro-2-butene
Concen: 17.134 ug/l
RT: 11.000 min Scan# 1626
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion: 75 Resp: 11571
Ion Ratio Lower Upper
75 100
53 98.9 79.5 119.3
89 46.7 36.8 55.2



#82
4-Chlorotoluene
Concen: 18.713 ug/l
RT: 11.439 min Scan# 1698
Delta R.T. 0.006 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion: 91 Resp: 115393
Ion Ratio Lower Upper
91 100
126 29.6 14.1 42.4





#83

tert-Butylbenzene

Concen: 18.651 ug/l

RT: 11.695 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :

VX1021WBSD01





Tgt Ion:119 Resp: 117357

Ion Ratio Lower Upper

119 100

91 55.7 30.3 91.0

134 21.8 11.4 34.2



#84

1,2,4-Trimethylbenzene

Concen: 18.745 ug/l

RT: 11.731 min Scan# 1746

Delta R.T. -0.000 min

Lab File: VX048256.D




Acq: 21 Oct 2025 10:33


Tgt Ion:105 Resp: 114561

Ion Ratio Lower Upper

105 100

120 44.0 22.1 66.5









#85
sec-Butylbenzene
Concen: 18.364 ug/l
RT: 11.871 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion:105 Resp: 139017
Ion Ratio Lower Upper
105 100
134 20.1 10.1 30.1



#86
p-Isopropyltoluene
Concen: 18.135 ug/l
RT: 11.993 min Scan# 1789
Delta R.T. 0.006 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion:119 Resp: 118663
Ion Ratio Lower Upper
119 100
134 26.1 13.0 39.0
91 24.2 12.1 36.3









#87
1,3-Dichlorobenzene
Concen: 18.219 ug/l
RT: 11.951 min Scan# 1782
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01


Tgt Ion:146 Resp: 63730
Ion Ratio Lower Upper
146 100
111 41.6 20.4 61.2
148 64.5 31.9 95.9



#88
1,4-Dichlorobenzene
Concen: 18.203 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion:146 Resp: 65963
Ion Ratio Lower Upper
146 100
111 41.5 20.1 60.2
148 63.9 31.8 95.3





#89

n-Butylbenzene

Concen: 17.834 ug/l

RT: 12.316 min Scan# 1842

Delta R.T. 0.006 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :

VX1021WBSD01





Tgt Ion: 91 Resp: 106210

Ion Ratio Lower Upper

91 100

92 52.4 26.9 80.6

134 24.9 12.7 38.1



#90

Hexachloroethane

Concen: 17.435 ug/l

RT: 12.518 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VX048256.D




Acq: 21 Oct 2025 10:33


Tgt Ion: 117 Resp: 22528

Ion Ratio Lower Upper

117 100

201 71.4 34.1 102.2









#91
 1,2-Dichlorobenzene
 Concen: 18.612 ug/l
 RT: 12.316 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX048256.D
 Acq: 21 Oct 2025 10:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1021WBSD01


Tgt Ion:146 Resp: 61124
 Ion Ratio Lower Upper
 146 100
 111 41.7 21.9 65.5
 148 62.2 31.9 95.5



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 16.899 ug/l
 RT: 12.920 min Scan# 1941
 Delta R.T. -0.000 min
 Lab File: VX048256.D
 Acq: 21 Oct 2025 10:33

Tgt Ion: 75 Resp: 6932
 Ion Ratio Lower Upper
 75 100
 155 83.9 42.8 128.4
 157 104.3 54.2 162.6





#93

1,2,4-Trichlorobenzene

Concen: 16.758 ug/l

RT: 13.566 min Scan# 2047

Delta R.T. -0.000 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :

VX1021WBSD01





Tgt Ion:180 Resp: 37218

Ion Ratio Lower Upper

180 100

182 96.7 47.3 142.0

145 32.7 16.3 48.9



#94

Hexachlorobutadiene

Concen: 17.706 ug/l

RT: 13.706 min Scan# 2070

Delta R.T. 0.005 min

Lab File: VX048256.D

Acq: 21 Oct 2025 10:33




Tgt Ion:225 Resp: 14700


Ion Ratio Lower Upper

225 100

223 67.0 32.6 97.7

227 65.7 32.5 97.4









#95
Naphthalene
Concen: 17.005 ug/l
RT: 13.755 min Scan# 2078
Delta R.T. -0.000 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Instrument :
MSVOA_X
ClientSampleId :
VX1021WBSD01

Tgt Ion:	128	Resp:	100059
Ion	Ratio	Lower	Upper
128	100		
127	12.5	10.2	15.4
129	10.8	8.6	12.8



#96
1,2,3-Trichlorobenzene
Concen: 17.277 ug/l
RT: 13.944 min Scan# 2109
Delta R.T. 0.006 min
Lab File: VX048256.D
Acq: 21 Oct 2025 10:33

Tgt Ion:	180	Resp:	35212
Ion	Ratio	Lower	Upper
180	100		
182	96.8	46.6	139.7
145	34.4	16.9	50.7

