

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3376</u>
Lab File ID:	<u>VW032378.D</u>	Date Analyzed:	<u>10/20/2025</u>
Instrument ID:	<u>MSVOA_W</u>	Time Analyzed:	<u>10:20</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	234766	13.556				
UPPER LIMIT	469532	14.056				
LOWER LIMIT	117383	13.056				
EPA SAMPLE NO.						
PR132-S05-057082-20251016	193724	13.56				
PR132-COMP2-20251016	236278	13.56				
PR132-S12-047072-20251016	262504	13.56				
PR132-S12-020047-20251016	228417	13.56				
VW1020SBL01	188560	13.56				
VW1020SBS01	240746	13.56				
VW1020SBSD01	239754	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.