

Initial and Continuing Calibration Verification

Client: ALS Environmental

SDG No.: Q3380

Project: ALS Middletown

RunNo.: LB137617

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.97	1	97	90-110	10/22/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.97	1	97	90-110	10/22/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.96	1	96	90-110	10/22/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.96	1	96	90-110	10/22/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.96	1	96	90-110	10/22/2025
Sample ID: CCV5 Ammonia as N	mg/L	0.98	1	98	90-110	10/22/2025

Initial and Continuing Calibration Verification

Client: ALS Environmental

SDG No.: Q3380

Project: ALS Middletown

RunNo.: LB137617

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------