

Laboratory Control Sample Summary

Client: ALS Environmental

SDG No.: Q3380

Project: ALS Middletown

Run No.: LB137617

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170196BS							
Ammonia as N	mg/Kg	50	49.1		98	1	90-110	10/22/2025