

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q3385

Project: PVSC Monthly 2025

Run No.: LB137577

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137577BS							
BOD5	mg/L	198	186		94	1	84.6-115.4	10/17/2025

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q3385

Project: PVSC Monthly 2025

Run No.: LB137611

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137611BS							
TSS	mg/L	550	532		97	1	90-110	10/22/2025

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q3385

Project: PVSC Monthly 2025

Run No.: LB137604

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170155BS							
Cyanide	mg/L	0.1	0.094		94	1	85-115	10/21/2025