



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q3385

Contract: ARDM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB88	Mercury	0.027	+/-0.2	U	0.20	CV	10/27/2025	13:57	LB137660

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q3385

Contract: ARDM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB91	Mercury	0.027	+/-0.2	U	0.20	CV	10/27/2025	14:01	LB137660
CCB92	Mercury	0.027	+/-0.2	U	0.20	CV	10/27/2025	14:37	LB137660
CCB93	Mercury	0.027	+/-0.2	U	0.20	CV	10/27/2025	14:54	LB137660

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q3385

Contract: ARDM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Lead	2.30	+/-6	U	12.0	P	10/31/2025	11:50	LB137727
	Zinc	16.7	+/-20	U	40.0	P	10/31/2025	11:50	LB137727

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q3385

Contract: ARDM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Lead	2.30	+/-6	U	12.0	P	10/31/2025	12:19	LB137727
	Zinc	16.7	+/-20	U	40.0	P	10/31/2025	12:19	LB137727
CCB02	Lead	2.30	+/-6	U	12.0	P	10/31/2025	13:21	LB137727
	Zinc	16.7	+/-20	U	40.0	P	10/31/2025	13:21	LB137727
CCB03	Lead	2.30	+/-6	U	12.0	P	10/31/2025	14:13	LB137727
	Zinc	16.7	+/-20	U	40.0	P	10/31/2025	14:13	LB137727
CCB04	Lead	2.30	+/-6	U	12.0	P	10/31/2025	15:12	LB137727
	Zinc	16.7	+/-20	U	40.0	P	10/31/2025	15:12	LB137727
CCB05	Lead	2.30	+/-6	U	12.0	P	10/31/2025	16:02	LB137727
	Zinc	16.7	+/-20	U	40.0	P	10/31/2025	16:02	LB137727
CCB06	Lead	2.30	+/-6	U	12.0	P	10/31/2025	16:52	LB137727
	Zinc	16.7	+/-20	U	40.0	P	10/31/2025	16:52	LB137727
CCB07	Lead	2.30	+/-6	U	12.0	P	10/31/2025	17:43	LB137727
	Zinc	16.7	+/-20	U	40.0	P	10/31/2025	17:43	LB137727
CCB08	Lead	2.30	+/-6	U	12.0	P	10/31/2025	18:32	LB137727
	Zinc	16.7	+/-20	U	40.0	P	10/31/2025	18:32	LB137727
CCB09	Lead	2.30	+/-6	U	12.0	P	10/31/2025	18:49	LB137727
	Zinc	16.7	+/-20	U	40.0	P	10/31/2025	18:49	LB137727

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q3385

Contract: ARDM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------