

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Ardmore Chemical</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3385</u>
Contract:	<u>ARDM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3385-04</u>	Client ID:	<u>EFF-WWDUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3385-04DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Lead	ug/L	20	2.55	J	2.05	J	22		P
Mercury	ug/L	20	0.20	U	0.20	U			CV
Zinc	ug/L	20	175		178		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Ardmore Chemical</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3385</u>
Contract:	<u>ARDM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3385-04MS</u>	Client ID:	<u>EFF-WWMSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3385-04MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.03		3.02		0		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Ardmore Chemical</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3385</u>
Contract:	<u>ARDM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3385-05</u>	Client ID:	<u>EFF-WWMSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3385-06	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Lead	ug/L	20	461		452		2		P
Zinc	ug/L	20	277		275		1		P