

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3385

Client ID : SSTDCCC040

Date Analyzed: 10/27/2025

Lab File ID: BF144074.D

Time Analyzed: 10:27

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	54596	6.881	204208	8.16	113226	9.92
UPPER LIMIT	109192	7.381	408416	8.663	226452	10.422
LOWER LIMIT	27298	6.381	102104	7.663	56613	9.422
EPA SAMPLE NO.						
01 PB170210BS	69633	6.88	270699	8.16	145913	9.92
02 PB170210BL	69834	6.88	269261	8.16	147528	9.92
03 EFF-WW	47582	6.88	172761	8.16	88391	9.92
04 EFF-WWMS	48975	6.88	173973	8.16	86535	9.92
05 EFF-WWMSD	49404	6.88	184320	8.16	89189	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.