

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3385</u>	Analytical Method:	<u>SW624.1</u>
Client:	<u>Ardmore Chemical</u>	Datafile :	<u>VN088073.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1020WBS01	Chloromethane	20	19.0	ug/L	95			1	205	
	Vinyl Chloride	20	20.2	ug/L	101			5	195	
	Bromomethane	20	22.4	ug/L	112			15	185	
	Chloroethane	20	21.9	ug/L	110			40	160	
	Trichlorofluoromethane	20	22.7	ug/L	114			50	150	
	1,1-Dichloroethene	20	20.8	ug/L	104			50	150	
	Acrolein	100	72.0	ug/L	72			60	140	
	Acrylonitrile	100	99.1	ug/L	99			60	140	
	Methylene Chloride	20	19.2	ug/L	96			60	140	
	trans-1,2-Dichloroethene	20	19.7	ug/L	99			70	130	
	1,1-Dichloroethane	20	20.2	ug/L	101			70	130	
	Carbon Tetrachloride	20	21.5	ug/L	108			70	130	
	Chloroform	20	19.9	ug/L	100			70	135	
	1,1,1-Trichloroethane	20	21.2	ug/L	106			70	130	
	Benzene	20	19.8	ug/L	99			65	135	
	1,2-Dichloroethane	20	21.2	ug/L	106			70	130	
	Trichloroethene	20	19.6	ug/L	98			65	135	
	1,2-Dichloropropane	20	20.2	ug/L	101			35	165	
	Bromodichloromethane	20	20.4	ug/L	102			65	135	
	Toluene	20	20.0	ug/L	100			70	130	
	t-1,3-Dichloropropene	20	19.7	ug/L	99			50	150	
	cis-1,3-Dichloropropene	20	19.2	ug/L	96			25	175	
	1,1,2-Trichloroethane	20	20.0	ug/L	100			70	130	
	2-Chloroethyl vinyl ether	100	99.4	ug/L	99			1	225	
	Dibromochloromethane	20	19.4	ug/L	97			70	135	
	Tetrachloroethene	20	19.4	ug/L	97			70	130	
	Chlorobenzene	20	19.6	ug/L	98			65	135	
	Ethyl Benzene	20	20.3	ug/L	102			60	140	
	m/p-Xylenes	40	39.9	ug/L	100			87	111	
	o-Xylene	20	19.4	ug/L	97			87	111	
	Bromoform	20	18.1	ug/L	91			70	130	
	1,1,2,2-Tetrachloroethane	20	20.2	ug/L	101			60	140	
	1,3-Dichlorobenzene	20	19.3	ug/L	97			70	130	
	1,4-Dichlorobenzene	20	19.6	ug/L	98			65	135	
	1,2-Dichlorobenzene	20	19.1	ug/L	96			65	135	