

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ARDM01
 Lab Code: ACE SDG NO.: Q3385
 Lab File ID: VN088055.D Date Analyzed: 10/17/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:29
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	68741	7.79	379209	9.08	339769	11.84
UPPER LIMIT	137482	8.294	758418	9.576	679538	12.341
LOWER LIMIT	34370.5	7.294	189605	8.576	169885	11.341
EPA SAMPLE NO.						
EFF-WWMS	53200	7.80	297248	9.08	260365	11.84
EFF-WWMSD	62360	7.80	360102	9.08	317295	11.84
VN1017WBL01	56868	7.79	316193	9.08	274787	11.84
VN1017WBS01	61010	7.79	345518	9.08	303706	11.84

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.