

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ARDM01
 Lab Code: ACE SDG NO.: Q3385
 Lab File ID: VN088072.D Date Analyzed: 10/20/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:52
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	63726	7.79	362466	9.08	325295	11.85
UPPER LIMIT	127452	8.294	724932	9.582	650590	12.347
LOWER LIMIT	31863	7.294	181233	8.582	162648	11.347
EPA SAMPLE NO.						
EFF-WW	63544	7.80	356820	9.08	313448	11.84
VN1020WBL01	56116	7.80	302542	9.08	267258	11.85
VN1020WBS01	62285	7.80	343376	9.08	304684	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.