

DATA PACKAGE

METALS

PROJECT NAME : EDISON YARD

YANNUZZI GROUP, INC.

135 Kinnelon Rd Suite #102

Kinnelon, NJ - 07405

Phone No: 908-218-0880

ORDER ID : Q3391

ATTENTION : Alyssa Yannuzzi



Laboratory Certification ID # 20012



Table Of Contents for Q3391

1) Signature Page	3
2) Case Narrative	5
3) Qualifier Page	6
4) QA Checklist	7
5) Metals-SPLP Data	8
6) Shipping Document	38
6.1) CHAIN OF CUSTODY	39
6.2) Lab Certificate	40

1
2
3
4
5
6

DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

1

Laboratory Name : Alliance Technical Group LLC Client : Yannuzzi Group, Inc.
 Project Location : Edison, NJ Project Number : 2025-1
 Laboratory Sample ID(s) : Q3391 Sampling Date(s) : 9/18/2025
 List DKQP Methods Used (e.g., 8260,8270, et Cetra) **1312,6010D,SOP**

1	For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
1A	Were the method specified handling, preservation, and holding time requirements met?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
3	Were samples received at an appropriate temperature (4±2° C)?	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A
4	Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
5	a)Were reporting limits specified or referenced on the chain-of-custody or communicated to the laboratory prior to sample receipt? b)Were these reporting limits met?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
6	For each analytical method referenced in this laboratory report package, were results reported for all constituents identified in the method-specific analyte lists presented in the DKQP documents and/or site-specific QAPP?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
7	Are project-specific matrix spikes and/or laboratory duplicates included in this data set?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."

Cover Page

Order ID : Q3391

Project ID : Edison Yard

Client : Yannuzzi Group, Inc.

Lab Sample Number

Q3391-01

Client Sample Number

RCA

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 8:17 am, Nov 04, 2025

Date: 11/4/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Yannuzzi Group, Inc.

Project Name: Edison Yard

Project # N/A

Order ID # Q3391

Test Name: SPLP MetalGroup2

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 09/29/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: SPLP MetalGroup2. This data package contains results for SPLP MetalGroup2.

C. Analytical Techniques:

The analysis of SPLP MetalGroup2 was based on method 6010D and digestion based on method 3010 (water).

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike analysis met criteria for all compounds.

The Matrix Spike Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

The temperature of the samples at the time of receipt was 20.5°C.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

APPROVED

Signature_____

By Nimisha Pandya, QA/QC Supervisor at 8:17 am, Nov 04, 2025

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3391

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 11/04/2025

LAB CHRONICLE

OrderID:	Q3391	OrderDate:	10/17/2025 2:30:00 PM
Client:	Yannuzzi Group, Inc.	Project:	Edison Yard
Contact:	Alyssa Yannuzzi	Location:	D31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3391-01	RCA	Water	SPLP MetalGroup2	6010D	09/18/25	10/29/25	10/31/25	09/29/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3391

Order ID: Q3391

Client: Yannuzzi Group, Inc.

Project ID: Edison Yard

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : Q3391-01	RCA RCA	Water	Cobalt	6.57	J	1.13	15.0	ug/L



SAMPLE DATA

Report of Analysis

Client: Yannuzzi Group, Inc.
Project: Edison Yard
Client Sample ID: RCA
Lab Sample ID: Q3391-01
Level (low/med): low

Date Collected: 09/18/25
Date Received: 09/29/25
SDG No.: Q3391
Matrix: Water
% Solid: 0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-48-4	Cobalt	6.57	J	1	1.13	15.0	ug/L	10/29/25 12:32	10/31/25 13:33	6010D	

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cobalt	1920	2000	96	90 - 110	P	10/31/2025	11:34	LB137727

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cobalt	30.1	30.0	100	80 - 120	P	10/31/2025	11:45	LB137727

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cobalt	2470	2500	99	90 - 110	P	10/31/2025	12:15	LB137727
CCV02	Cobalt	2500	2500	100	90 - 110	P	10/31/2025	13:17	LB137727
CCV03	Cobalt	2440	2500	98	90 - 110	P	10/31/2025	14:09	LB137727
CCV04	Cobalt	2370	2500	95	90 - 110	P	10/31/2025	15:08	LB137727
CCV05	Cobalt	2470	2500	99	90 - 110	P	10/31/2025	15:58	LB137727
CCV06	Cobalt	2480	2500	99	90 - 110	P	10/31/2025	16:48	LB137727
CCV07	Cobalt	2440	2500	98	90 - 110	P	10/31/2025	17:39	LB137727
CCV08	Cobalt	2400	2500	96	90 - 110	P	10/31/2025	18:28	LB137727
CCV09	Cobalt	2410	2500	96	90 - 110	P	10/31/2025	18:45	LB137727



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cobalt	30.3	30.0	101	65 - 135	P	10/31/2025	11:54	LB137727



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Yannuzzi Group, Inc.</u>	SDG No.:	<u>Q3391</u>
Contract:	<u>YANN01</u>	Lab Code:	<u>ACE</u>

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cobalt	2.26	+/-15	U	30.0	P	10/31/2025	11:50	LB137727

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Yannuzzi Group, Inc. **SDG No.:** Q3391
Contract: YANN01 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cobalt	2.26	+/-15	U	30.0	P	10/31/2025	12:19	LB137727
CCB02	Cobalt	2.26	+/-15	U	30.0	P	10/31/2025	13:21	LB137727
CCB03	Cobalt	2.26	+/-15	U	30.0	P	10/31/2025	14:13	LB137727
CCB04	Cobalt	2.26	+/-15	U	30.0	P	10/31/2025	15:12	LB137727
CCB05	Cobalt	2.26	+/-15	U	30.0	P	10/31/2025	16:02	LB137727
CCB06	Cobalt	2.26	+/-15	U	30.0	P	10/31/2025	16:52	LB137727
CCB07	Cobalt	2.26	+/-15	U	30.0	P	10/31/2025	17:43	LB137727
CCB08	Cobalt	2.26	+/-15	U	30.0	P	10/31/2025	18:32	LB137727
CCB09	Cobalt	2.26	+/-15	U	30.0	P	10/31/2025	18:49	LB137727

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170310BL		WATER		Batch Number:	PB170310		Prep Date:	10/29/2025	
	Cobalt	1.13	<7.5	U	15.0	P	10/31/2025	14:45	LB137727
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170310TB		WATER		Batch Number:	PB170310		Prep Date:	10/29/2025	
	Cobalt	1.13	<7.5	U	15.0	P	10/31/2025	16:57	LB137727

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Yannuzzi Group, Inc.</u>	SDG No.: <u>Q3391</u>
Contract: <u>YANN01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cobalt	2.13			-30	30	10/31/2025	11:58	LB137727
ICSAB01	Cobalt	484	476	102	404	548	10/31/2025	12:02	LB137727



METAL QC DATA

A

B

C

D

E

F

G

H

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Yannuzzi Group, Inc. level: low sdg no.: Q3391
contract: YANN01 lab code: ACE
matrix: Water sample id: Q3391-01 client id: RCAMS
Percent Solids for Sample: NA Spiked ID: Q3391-01MS Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cobalt	ug/L	75 - 125	108		6.57	J	100	102		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Yannuzzi Group, Inc.</u>	level:	<u>low</u>	sdg no.:	<u>Q3391</u>
contract:	<u>YANN01</u>			lab code:	<u>ACE</u>
matrix:	<u>Water</u>	sample id:	<u>Q3391-01</u>	client id:	<u>RCAMSD</u>
Percent Solids for Sample:	NA	Spiked ID:	Q3391-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cobalt	ug/L	75 - 125	103		6.57	J	100	96		P

Metals
- 5b -

Client:

Yannuzzi Group, Inc.

Contract:

YANN01

Matrix:

Sample ID:

SDG No.:

Q3391

Lab Code:

ACE

Client ID:

Level:

LOW

Spiked ID:

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

.....

A

B

C

D

E

F

G

H

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Yannuzzi Group, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3391</u>
Contract:	<u>YANN01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3391-01</u>	Client ID:	<u>RCADUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3391-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cobalt	ug/L	20	6.57	J	6.56	J	0		P

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Yannuzzi Group, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3391</u>
Contract:	<u>YANN01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3391-01MS</u>	Client ID:	<u>RCAMSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3391-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cobalt	ug/L	20	108		103		5		P

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170310BS Cobalt	ug/L	100	99.9		100	80 - 120	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

RCAL

Lab Name: Alliance **Contract:** YANN01
Lab Code: ACE **Lb No.:** lb137727 **Lab Sample ID :** Q3391-01L **SDG No.:** Q3391
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I) C	Serial Dilution Result (S) C	% Differ- ence	Q	M
Cobalt	6.57 J	6.76 J	3		P



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMEN CORRECTION FACTORS

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Instrument ID:

Date:

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		As	Ba	Be	Cd	Co
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	-0.0007860

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Cobalt	228.616	0.0000000	0.0018800	0.0000000	0.0000000	0.0000000



METAL PREPARATION & ANALYICAL SUMMARY

Metals
- 13 -

SAMPLE PREPARATION SUMMARY

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Method: _____

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB170310							
PB170310BL	PB170310BL	MB	WATER	10/29/2025	50.0	25.0	
PB170310BS	PB170310BS	LCS	WATER	10/29/2025	50.0	25.0	
PB170310TB	PB170310TB	MB	WATER	10/29/2025	50.0	25.0	
Q3391-01	RCA	SAM	WATER	10/29/2025	50.0	25.0	
Q3391-01DUP	RCADUP	DUP	WATER	10/29/2025	50.0	25.0	
Q3391-01MS	RCAMS	MS	WATER	10/29/2025	50.0	25.0	
Q3391-01MSD	RCAMSD	MSD	WATER	10/29/2025	50.0	25.0	

metals
- 14 -
ANALYSIS RUN LOG

Client: Yannuzzi Group, Inc.

Contract: YANN01

Lab code: ACE

Sdg no.: Q3391

Instrument id number: _____

Method: _____

Run number: LB137727

Start date: 10/31/2025

End date: 10/31/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1109	Co
S1	S1	1	1113	Co
S2	S2	1	1117	Co
S3	S3	1	1121	Co
S4	S4	1	1126	Co
S5	S5	1	1130	Co
ICV01	ICV01	1	1134	Co
LLICV01	LLICV01	1	1145	Co
ICB01	ICB01	1	1150	Co
CRI01	CRI01	1	1154	Co
ICSA01	ICSA01	1	1158	Co
ICSAB01	ICSAB01	1	1202	Co
CCV01	CCV01	1	1215	Co
CCB01	CCB01	1	1219	Co
CCV02	CCV02	1	1317	Co
CCB02	CCB02	1	1321	Co
Q3391-01	RCA	1	1333	Co
CCV03	CCV03	1	1409	Co
CCB03	CCB03	1	1413	Co
Q3391-01DUP	RCADUP	1	1423	Co
Q3391-01L	RCAL	5	1428	Co
Q3391-01MS	RCAMS	1	1432	Co
Q3391-01MSD	RCAMSD	1	1436	Co
PB170310BL	PB170310BL	1	1445	Co
PB170310BS	PB170310BS	1	1449	Co
CCV04	CCV04	1	1508	Co
CCB04	CCB04	1	1512	Co
CCV05	CCV05	1	1558	Co
CCB05	CCB05	1	1602	Co
CCV06	CCV06	1	1648	Co
CCB06	CCB06	1	1652	Co
PB170310TB	PB170310TB	1	1657	Co
CCV07	CCV07	1	1739	Co
CCB07	CCB07	1	1743	Co
CCV08	CCV08	1	1828	Co
CCB08	CCB08	1	1832	Co
CCV09	CCV09	1	1845	Co
CCB09	CCB09	1	1849	Co



SHIPPING DOCUMENTS

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312